

DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

J	Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).
U	Indicates the analyte was analyzed for, but not detected.
ND	Indicates the analyte was analyzed for, but not detected
E	Indicates the reported value is estimated because of the presence of interference
M	Indicates Duplicate injection precision not met.
N	Indicates the spiked sample recovery is not within control limits.
S	Indicates the reported value was determined by the Method of Standard Addition (MSA).
*	Indicates that the duplicate analysis is not within control limits.
+	Indicates the correlation coefficient for the MSA is less than 0.995.
D	Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
M	Method qualifiers “P” for ICP instrument “PM” for ICP when Microwave Digestion is used “CV” for Manual Cold Vapor AA “AV” for automated Cold Vapor AA “CA” for MIDI-Distillation Spectrophotometric “AS” for Semi -Automated Spectrophotometric “C” for Manual Spectrophotometric “T” for Titrimetric “NR” for analyte not required to be analyzed
OR	Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis.
Q	Indicates the LCS did not meet the control limits requirements
H	Sample Analysis Out Of Hold Time

LAB CHRONICLE

OrderID: Q1150	OrderDate: 1/22/2025 12:07:00 PM
Client: Aramark Uniforms	Project: Monthly 2025
Contact: Jose Liceaga	Location: E11

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1150-01	GRAB	WATER			01/22/25 11:15			01/22/25
			TPH	1664A			01/23/25 10:00	
Q1150-02	COMP	WATER			01/22/25 11:18			01/22/25
			BOD5	SM5210 B			01/22/25 17:10	
			TSS	SM2540 D			01/22/25 16:40	



SAMPLE DATA

Report of Analysis

Client:	Aramark Uniforms	Date Collected:	01/22/25 11:15
Project:	Monthly 2025	Date Received:	01/22/25
Client Sample ID:	GRAB	SDG No.:	Q1150
Lab Sample ID:	Q1150-01	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
TPH	64.5		1	0.40	5.00	mg/L		01/23/25 10:00	1664A

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	Aramark Uniforms	Date Collected:	01/22/25 11:18
Project:	Monthly 2025	Date Received:	01/22/25
Client Sample ID:	COMP	SDG No.:	Q1150
Lab Sample ID:	Q1150-02	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
BOD5	621		1	0.17	2.00	mg/L		01/22/25 17:10	SM 5210 B-16
TSS	467		1	1.00	4.00	mg/L		01/22/25 16:40	SM 2540 D-15

Comments: _____

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits



QC RESULT SUMMARY

Preparation Blank Summary

Client: Aramark Uniforms

SDG No.: Q1150

Project: Monthly 2025

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: TSS	LB134370BL mg/L	1	2.0000	J	1	4	01/22/2025
Sample ID: TPH	LB134373BL mg/L	< 2.5000	2.5000	U	0.4	5.0	01/23/2025
Sample ID: BOD5	LB134390BL mg/L	< 0.2000	0.2000	U	0.17	2.0	01/22/2025

Duplicate Sample Summary

Client:	Aramark Uniforms	SDG No.:	Q1150
Project:	Monthly 2025	Sample ID:	LB134373BS
Client ID:	LB134373BSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TPH	mg/L	+/-18	16.8		17.1		1	1.77		01/23/2025

Duplicate Sample Summary

Client: Aramark Uniforms	SDG No.: Q1150
Project: Monthly 2025	Sample ID: Q1140-01
Client ID: FRAC-TANK-F06078DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	68.2		68.3		1	0.15		01/22/2025
BOD5	mg/L	+/-20	10.4		10.5		1	0.77		01/22/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1150

Project: Monthly 2025

Run No.: LB134370

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134370BS							
TSS	mg/L	550	532		97	1	90-110	01/22/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1150

Project: Monthly 2025

Run No.: LB134373

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134373BS							
TPH	mg/L	20.0	16.8		84	1	78-114	01/23/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1150

Project: Monthly 2025

Run No.: LB134373

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134373BSD							
TPH	mg/L	20.0	17.1		86	1	78-114	01/23/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1150

Project: Monthly 2025

Run No.: LB134390

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134390BS							
BOD5	mg/L	198	209		105	1	84.6-115.4	01/22/2025



RAW DATA

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: Iwona

ANALYST: jignesh

Date: 01/22/2025

Run Number: LB134370

BalanceID: WC SC-6

OvenID: WC OVEN-1

FilterID: 17416528

ThermometerID: WET OVEN#1

TEMP1 IN: 103 °C 01/22/2025 13:00 TEMP1 OUT: 103 °C 01/22/2025 14:00
 TEMP2 IN: 104 °C 01/22/2025 14:30 TEMP2 OUT: 103 °C 01/22/2025 15:30
 TEMP3 IN: 104 °C 01/22/2025 16:40 TEMP3 OUT: 103 °C 01/22/2025 18:00
 TEMP4 IN: 104 °C 01/22/2025 18:30 TEMP4 OUT: 103 °C 01/22/2025 19:45

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB134370BL	LB134370BL	1.4896	1.4896	100	1.4897	1.4897	1.4897	0.0001	1
2	LB134370BS	LB134370BS	1.3605	1.3605	100	1.4137	1.4137	1.4137	0.0532	532
3	Q1140-01	FRAC-TANK-F06078	1.3676	1.3676	1000	1.4358	1.4358	1.4358	0.0682	68.2
4	Q1140-01DUP	FRAC-TANK-F06078DUP	1.4837	1.4837	1000	1.5520	1.5520	1.5520	0.0683	68.3
5	Q1145-02	002-35TH AVE (DEC)	1.4893	1.4893	500	1.5050	1.5050	1.5050	0.0157	31.4
6	Q1150-02	COMP	1.3725	1.3725	100	1.4192	1.4192	1.4192	0.0467	467

A = Sample Volume (ml)
 B = Final Empty Dish Weight (g)
 C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)
 D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

134373

WorkList Name : tph q1140

WorkList ID : 187076

Department : Wet-Chemistry

Date : 01-23-2025 08:06:28

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1140-01	FRAC-TANK-F06078	Water	TPH	Conc H2SO4 to pH < 2	PSEG03	E11	01/21/2025	1664A
Q1150-01	GRAB	Water	TPH	Conc H2SO4 to pH < 2	ARAM01	E11	01/22/2025	1664A

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

01/23/25 08:15

TPH (WOC)

JD (SM)

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

01/23/25

JD (SM)

TPH (WOC)

11:30

Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB134373
Analysis Date: 01/23/2025
BalanceID: WC SC-6
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: Iwona
Extraction Date: 01/23/2025
Extraction IN Time: 08:25
Extraction OUT Time: 09:15
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB134373BL	LB134373BL	WATER	1.3	1000	100	2.7463	2.7463	3.02	2.7464	2.7464	0.0001	0.1
2	LB134373BS	LB134373BS	WATER	1.3	1000	100	3.0652	3.0652	3.01	3.0820	3.0820	0.0168	16.8
3	LB134373BSD	LB134373BSD	WATER	1.3	1000	100	3.1781	3.1781	3.03	3.1952	3.1952	0.0171	17.1
4	Q1140-01	FRAC-TANK-F06078	WATER	1.6	1000	100	3.0606	3.0606	3.04	3.1016	3.1016	0.0410	41
5	Q1150-01	GRAB	WATER	1.6	1000	100	3.0815	3.0815	3.05	3.1460	3.1460	0.0645	64.5

QC Batch# LB134373

Test: TPH

Analysis Date: 01/23/2025

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W3153
pH Paper 0-14	M6069
Sodium Sulfate	EP2580
1:1 HCL	WP110826
Silica Gel	W3079
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP100827
LCSWD	5.00 ML	WP100828
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-6

Before Analysis

0.0020 gram Balance: 0.0019 (0.0018-0.0022) In OVEN TEMP1 : 70 °C Dessicator Time In1 : 10:51

1.0000 gram Balance: 1.0004 (0.9950-1.0050) In Time1: 10:00

Bal Check Time: 08:37 Out OVEN TEMP1: 70 °C Dessicator Time Out1: 11:30

Out Time1: 10:50

After Analysis

0.0020 gram Balance: 0.0021 (0.0018-0.0022) In OVEN TEMP2 : 71 °C Dessicator Time In2 : 13:20

1.0000 gram Balance: 1.0005 (0.9950-1.0050) In Time2: 12:05

Bal Check Time: 13:30 Out OVEN TEMP2: 71 °C Dessicator Time Out2: 13:21

Out Time2: 12:37

WORKLIST(Hardcopy Internal Chain)

134373

WorkList Name : tph q1140

WorkList ID : 187076

Department : Wet-Chemistry

Date : 01-23-2025 08:06:28

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1140-01	FRAC-TANK-F06078	Water	TPH	Conc H2SO4 to pH < 2	PSEG03	E11	01/21/2025	1664A
Q1150-01	GRAB	Water	TPH	Conc H2SO4 to pH < 2	ARAM01	E11	01/22/2025	1664A

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

01/23/25 08:15

78 WCC

JD CSM

Date/Time 01/23/25

Raw Sample Received by:

Raw Sample Relinquished by:

11:30

JD CSM

78 WCC

BOD5 LOG

ANALYST: rubin

SUPERVISOR: Iwona

Analysis Date: 01/22/2025

MANGANOUS SULFATE SOLUTION: W3103

Alkaline Iodide Azide: W3109

Sodium Thiosulfate, 0.025N: W3105

NaOH, 1N: WP111323

IncubatorID: INCUBATOR #3

GuageID: 0511062

Zero DO: WP111324

QC BATCH ID: LB134390

BOD Water: WP111545

Starch: W3149

Sulfuric acid, 1N: WP110386

POLYSEED: WP111547

GGA: WP111546

Chlorine Strips: W3155

pH Strips: W3140

Lab SampleID	Client ID	Bottle No.	VOL. ML	Initial Reading (ML)	Final Reading (ML)	Difference	Average
WINKLER 1	WINKLER 1	1	300	0.0	9.7	9.7	9.7
WINKLER 2	WINKLER 2	2	300	9.8	19.5	9.7	9.7

Meter Calibration1: 9.80 Zero DO Reading1: 0.15 mg/L (<=0.2 Criteria)

Barometric Pressure1: 765 mmHg DO Meter BOD fluid reading for winkler comparison: 9.72

After Incubation

Meter Calibration2: 8.87 Zero DO Reading2: 0.10 mg/L (<=0.2 Criteria)

Barometric Pressure2: 760 mmHg

QC BATCH ID: LB134390

INCUBATOR TEMP IN(C): 20.2

INCUBATOR TEMP OUT(C): 19.6

TIME IN: 17:10

TIME OUT: 14:15

DATE IN: 01/22/2025

DATE OUT: 01/27/2025

Lab SampleID	Bottle No.	Check CL	Initial PH	Final PH	Temp °C	Sam Vol. (mL)	D.O.1 Initial	D.O.2 Final	Depletion	BOD Result (mg/L)	Avg Result (mg/L)	Comment
LB134390BL	1	No	6.51	N/A	20.40	300	9.71	9.69	0.02	0.02	0.02	
POLYSEED	1					10	9.67	7.06	2.61	0.52	0.62	
POLYSEED	2					15	9.64	4.64	5	0.67		
POLYSEED	3					20	9.61	2.80	6.81	0.68		
GGA	1					6	9.65	4.99	4.66	202	208.67	
GGA	2					6	9.64	4.91	4.73	205.5		
GGA	3					6	9.63	4.64	4.99	218.5		
Q1140-01	1	No	8.26	7.24	20.50	5	9.69	9.08	-	0	10.4	pH Adjusted
Q1140-01	2					20	9.66	8.60	-	0		
Q1140-01	3					50	9.62	8.48	-	0		
Q1140-01	4					150	9.51	3.69	5.82	10.4		
Q1140-01DUP	1	No	8.26	7.24	20.50	5	9.67	9.00	-	0	10.48	pH Adjusted
Q1140-01DUP	2					20	9.66	8.74	-	0		
Q1140-01DUP	3					50	9.61	8.31	-	0		
Q1140-01DUP	4					150	9.50	3.64	5.86	10.48		
Q1150-02	1	No	6.51	N/A	20.40	0.5	9.50	8.02	-	0	621	
Q1150-02	2					1	9.50	6.81	2.69	621		
Q1150-02	3					2	9.46	0.39	-	0		
Q1150-02	4					3	9.40	0.11	-	0		

NOTE: 2ml POLYSEED added to GGA and all the Samples, but not in Blank.

NOTE (For, CBOD5): 0.16 g Nitrification Inhibitor added to GGA and all the Samples, but not in Blank.

LB134390

WORKLIST(Hardcopy Internal Chain)

WorkList Name : BOD5-01222025

WorkList ID : 187069

Department : Wet-Chemistry

Date : 01-22-2025 16:06:55

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1150-02	COMP	Water	BOD5	Cool 4 deg C	ARAM01	E11	01/22/2025	SM5210 B

Date/Time 01.22.2025 16:40
Raw Sample Received by: NF(wc)
Raw Sample Relinquished by: NF(wc)

Date/Time 01.22.2025
Raw Sample Received by: NF(wc)
Raw Sample Relinquished by: NF(wc)

LB134390

WORKLIST(Hardcopy Internal Chain)

WorkList Name : BOD5- 01222025

WorkList ID : 137186

Department : Wet-Chemistry

Date : 01-22-2025 13:26:46

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1140-01	FRAC-TANK-F06078	Water	BOD5	Cool 4 deg C	PSEG03	E11	01/21/2025	SM5210 B

Date/Time 01.22.2025 15.25
Raw Sample Received by: NFCW
Raw Sample Relinquished by: NFCW

Date/Time 01.22.2025 17.15
Raw Sample Received by: Jol Cede
Raw Sample Relinquished by: NF (wcl)

Instrument ID: WC SC-3

Daily Analysis Runlog For Sequence/QC Batch ID # LB134370

Review By	jignesh	Review On	1/23/2025 11:10:17 AM
Supervise By	Iwona	Supervise On	1/23/2025 11:18:08 AM
SubDirectory	LB134370	Test	TSS
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	N/A		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	LB134370BL	LB134370BL	MB	01/22/25 16:40		jignesh	OK
2	LB134370BS	LB134370BS	LCS	01/22/25 16:40		jignesh	OK
3	Q1140-01	FRAC-TANK-F06078	SAM	01/22/25 16:40		jignesh	OK
4	Q1140-01DUP	FRAC-TANK-F06078	DUP	01/22/25 16:40		jignesh	OK
5	Q1145-02	002 35TH AVE (DEC)	SAM	01/22/25 16:40		jignesh	OK
6	Q1150-02	COMP	SAM	01/22/25 16:40		jignesh	OK

Instrument ID: WC SC-3

Daily Analysis Runlog For Sequence/QC Batch ID # LB134373

Review By	jignesh	Review On	1/23/2025 8:31:34 AM
Supervise By	Iwona	Supervise On	1/23/2025 9:23:19 AM
SubDirectory	LB134373	Test	TPH
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	W3153,M6069,EP2580,WP110826,W3079,NA,WP100827,WP100828,NA		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	LB134373BL	LB134373BL	MB	01/23/25 10:00		jignesh	OK
2	LB134373BS	LB134373BS	LCS	01/23/25 10:00		jignesh	OK
3	LB134373BSD	LB134373BSD	LCSD	01/23/25 10:00		jignesh	OK
4	Q1140-01	FRAC-TANK-F06078	SAM	01/23/25 10:00		jignesh	OK
5	Q1150-01	GRAB	SAM	01/23/25 10:00		jignesh	OK

Instrument ID: DO METER

Daily Analysis Runlog For Sequence/QC Batch ID # LB134390

Review By	rubina	Review On	1/27/2025 3:18:17 PM
Supervise By	Iwona	Supervise On	1/27/2025 3:41:29 PM
SubDirectory	LB134390	Test	BOD5
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	WP111545,W3149,WP110386,W3103,W3109,W3105,WP111547,WP111546,WP111323		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	LB134390BL	LB134390BL	MB	01/22/25 17:10		Niha	OK
2	LB134390BS	LB134390BS	LCS	01/22/25 17:10		Niha	OK
3	Q1140-01	FRAC-TANK-F06078	SAM	01/22/25 17:10		Niha	OK
4	Q1140-01DUP	FRAC-TANK-F06078	DUP	01/22/25 17:10		Niha	OK
5	Q1150-02	COMP	SAM	01/22/25 17:10		Niha	OK



SHIPPING DOCUMENTS

CHEMTECH

CHAIN OF CUSTODY RECORD

284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 • Fax (908) 789-8922
www.chemtech.net

CHEMTECH PROJECT NO. Q1150
QUOTE NO.
COC Number 2044249

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: Aramark Uniforms
ADDRESS: 740 Frelinghuysen Ave
CITY Newark STATE: NJ ZIP: 07114
ATTENTION: Jarrod Mills
PHONE: 973-824-1101 FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: Monthly
PROJECT NO.: LOCATION:
PROJECT MANAGER:
e-mail:
PHONE: FAX:

CLIENT BILLING INFORMATION

BILL TO: PO#:
ADDRESS:
CITY STATE: ZIP:
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) DAYS*
HARDCOPY (DATA PACKAGE): DAYS*
EDD: DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data) ☐ Other
☐ EDD FORMAT

TPH
BOD5
TSS

CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS	
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	A-HCl B-HNO3 C-H2SO4	D-NaOH E-ICE F-OTHER
1.	Grab Comp	W		✓	1-22-25	1115	1	✓										
2.		W	✓		1-22-25	1118	2		✓	✓								
3.																		
4.																		
5.																		
6.																		
7.																		
8.																		
9.																		
10.																		

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. <u>[Signature]</u>	DATE/TIME: <u>1-22-25 1120</u>	RECEIVED BY: 1. <u>[Signature]</u>	DATE/TIME: <u>1-22-25 1120</u>	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP <u>2.7C</u> °C
RELINQUISHED BY SAMPLER: 2. <u>[Signature]</u>	DATE/TIME: <u>1-22-25 1620</u>	RECEIVED BY: 2. <u>[Signature]</u>	DATE/TIME: <u>1-22-25 1620</u>	Comments:
RELINQUISHED BY SAMPLER: 3. <u>[Signature]</u>	DATE/TIME: <u>1-22-25 1620</u>	RECEIVED BY: 3. <u>[Signature]</u>	DATE/TIME: <u>1-22-25 1620</u>	

Page ____ of ____

CLIENT: ☐ Hand Delivered ☐ Other
CHEMTECH: ☐ Picked Up ☐ Field Sampling

Shipment Complete
☐ YES ☐ NO



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488