

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51973 MA No.: _____ SDG No.: MC0AZ0
 Instrument ID: P4 Start Date: 01/27/2025
 Analytical Method: ICP-AES Run Batch: LB134438
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	0	0.050	0	200	180	-10	38750	36000	-7
Antimony	0	0.010	0	60	56	-7	1250	1200	-4
Arsenic	0	0.0	0	10	10	0	1250	1200	-4
Barium	0	0.030	0	200	190	-5	2500	2400	-4
Beryllium	0	-0	0	5	5.3	6	125	130	4
Cadmium	0	-0	0	5	5.1	2	625	640	2
Calcium	0	0.14	0	5000	5000	0	62500	61000	-2
Chromium	0	-0	0	10	11	10	3875	4200	8
Cobalt	0	-0	0	50	50	0	625	640	2
Copper	0	-0	0	25	29	16	3750	3700	-1
Iron	0	-0.010	0	100	110	10	37500	38000	1
Lead	0	-0	0	10	11	10	6250	6400	2
Magnesium	0	0.99	0	5000	4400	-12	62500	59000	-6
Manganese	0	-0	0	15	15	0	3750	3700	-1
Nickel	0	-0	0	40	41	3	625	650	4
Potassium	0	0.86	0	5000	4600	-8	37500	35000	-7
Selenium	0	0.010	0	35	31	-11	1250	1200	-4
Silver	0	0.0	0	10	9.6	-4	312.5	300	-4
Sodium	0	0.66	0	5000	4900	-2	37500	36000	-4
Thallium	0	-0.010	0	25	28	12	1250	1300	4
Vanadium	0	-0.010	0	50	54	8	625	640	2
Zinc	0	0.0	0	60	59	-2	3750	3600	-4

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51973 MA No.: _____ SDG No.: MC0AZ0
 Instrument ID: P4 Start Date: 01/27/2025
 Analytical Method: ICP-AES Run Batch: LB134438
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	77500	75000	-3	155000	150000	-3	310000	310000	0
Antimony	2500	2500	0	5000	4800	-4	10000	10000	0
Arsenic	2500	2500	0	5000	4800	-4	10000	10000	0
Barium	5000	4900	-2	10000	9500	-5	20000	21000	5
Beryllium	250	260	4	500	500	0	1000	990	-1
Cadmium	1250	1300	4	2500	2500	0	5000	4900	-2
Calcium	125000	130000	4	250000	240000	-4	500000	500000	0
Chromium	7750	8300	7	15500	16000	3	31000	30000	-3
Cobalt	1250	1300	4	2500	2500	0	5000	5000	0
Copper	7500	7500	0	15000	14000	-7	30000	31000	3
Iron	75000	79000	5	150000	150000	0	300000	300000	0
Lead	12500	13000	4	25000	25000	0	50000	49000	-2
Magnesium	125000	120000	-4	250000	240000	-4	500000	510000	2
Manganese	7500	7700	3	15000	15000	0	30000	30000	0
Nickel	1250	1300	4	2500	2500	0	5000	4900	-2
Potassium	75000	74000	-1	150000	150000	0	300000	310000	3
Selenium	2500	2500	0	5000	4800	-4	10000	10000	0
Silver	625	620	-1	1250	1200	-4	2500	2600	4
Sodium	75000	74000	-1	150000	140000	-7	300000	300000	0
Thallium	2500	2700	8	5000	5000	0	10000	9700	-3
Vanadium	1250	1300	4	2500	2500	0	5000	5000	0
Zinc	7500	7600	1	15000	14000	-7	30000	31000	3

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51973 MA No.: _____ SDG No.: MC0AZ0
 Instrument ID: P4 Start Date: 01/27/2025
 Analytical Method: ICP-AES Run Batch: LB134438
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	800000	810000	1						
Antimony	0	0.0	0						
Arsenic	0	0.0	0						
Barium	0	0.0	0						
Beryllium	0	0.0	0						
Cadmium	0	0.0	0						
Calcium	800000	800000	0						
Chromium	0	0.0	0						
Cobalt	0	0.0	0						
Copper	0	0.0	0						
Iron	800000	790000	-1						
Lead	0	0.0	0						
Magnesium	800000	810000	1						
Manganese	0	0.0	0						
Nickel	0	0.0	0						
Potassium	0	0.0	0						
Selenium	0	0.0	0						
Silver	0	0.0	0						
Sodium	800000	810000	1						
Thallium	0	0.0	0						
Vanadium	0	0.0	0						
Zinc	0	0.0	0						

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51973 MA No.: _____ SDG No.: MC0AZ0
 Instrument ID: P4 Start Date: 01/28/2025
 Analytical Method: ICP-AES Run Batch: LB134460
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	0	0.080	0	200	170	-15	38750	35000	-10
Antimony	0	0.0	0	60	59	-2	1250	1200	-4
Arsenic	0	0.0	0	10	9.1	-9	1250	1200	-4
Barium	0	0.030	0	200	180	-10	2500	2400	-4
Beryllium	0	-0	0	5	5.4	8	125	130	4
Cadmium	0	-0	0	5	5.0	0	625	630	1
Calcium	0	0.72	0	5000	4600	-8	62500	60000	-4
Chromium	0	-0	0	10	11	10	3875	4100	6
Cobalt	0	-0	0	50	49	-2	625	630	1
Copper	0	-0	0	25	28	12	3750	3700	-1
Iron	0	0.0	0	100	99	-1	37500	38000	1
Lead	0	-0	0	10	10	0	6250	6200	-1
Magnesium	0	1.6	0	5000	4100	-18	62500	57000	-9
Manganese	0	0.0	0	15	15	0	3750	3700	-1
Nickel	0	-0	0	40	40	0	625	630	1
Potassium	0	1.1	0	5000	4300	-14	37500	35000	-7
Selenium	0	0.0	0	35	34	-3	1250	1200	-4
Silver	0	0.0	0	10	9.8	-2	312.5	300	-4
Sodium	0	1.4	0	5000	4500	-10	37500	35000	-7
Thallium	0	-0	0	25	26	4	1250	1300	4
Vanadium	0	-0	0	50	51	2	625	640	2
Zinc	0	0.010	0	60	55	-8	3750	3600	-4

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51973 MA No.: _____ SDG No.: MC0AZ0
 Instrument ID: P4 Start Date: 01/28/2025
 Analytical Method: ICP-AES Run Batch: LB134460
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	77500	72000	-7	155000	150000	-3	310000	300000	-3
Antimony	2500	2400	-4	5000	5000	0	10000	10000	0
Arsenic	2500	2500	0	5000	5000	0	10000	10000	0
Barium	5000	4900	-2	10000	10000	0	20000	20000	0
Beryllium	250	250	0	500	510	2	1000	980	-2
Cadmium	1250	1300	4	2500	2600	4	5000	4900	-2
Calcium	125000	120000	-4	250000	250000	0	500000	490000	-2
Chromium	7750	8100	5	15500	16000	3	31000	30000	-3
Cobalt	1250	1300	4	2500	2600	4	5000	4900	-2
Copper	7500	7300	-3	15000	15000	0	30000	30000	0
Iron	75000	76000	1	150000	150000	0	300000	290000	-3
Lead	12500	13000	4	25000	26000	4	50000	49000	-2
Magnesium	125000	120000	-4	250000	250000	0	500000	490000	-2
Manganese	7500	7500	0	15000	15000	0	30000	30000	0
Nickel	1250	1300	4	2500	2600	4	5000	4900	-2
Potassium	75000	73000	-3	150000	150000	0	300000	300000	0
Selenium	2500	2500	0	5000	5000	0	10000	10000	0
Silver	625	600	-4	1250	1300	4	2500	2500	0
Sodium	75000	70000	-7	150000	140000	-7	300000	290000	-3
Thallium	2500	2600	4	5000	5200	4	10000	9700	-3
Vanadium	1250	1300	4	2500	2600	4	5000	4900	-2
Zinc	7500	7400	-1	15000	15000	0	30000	30000	0

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51973 MA No.: _____ SDG No.: MC0AZ0
 Instrument ID: P4 Start Date: 01/28/2025
 Analytical Method: ICP-AES Run Batch: LB134460
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	800000	830000	4						
Antimony	0	0.0	0						
Arsenic	0	0.0	0						
Barium	0	0.0	0						
Beryllium	0	0.0	0						
Cadmium	0	0.0	0						
Calcium	800000	820000	3						
Chromium	0	0.0	0						
Cobalt	0	0.0	0						
Copper	0	0.0	0						
Iron	800000	800000	0						
Lead	0	0.0	0						
Magnesium	800000	830000	4						
Manganese	0	0.0	0						
Nickel	0	0.0	0						
Potassium	0	0.0	0						
Selenium	0	0.0	0						
Silver	0	0.0	0						
Sodium	800000	830000	4						
Thallium	0	0.0	0						
Vanadium	0	0.0	0						
Zinc	0	0.0	0						

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51973 MA No.: _____ SDG No.: MC0AZ0
 Instrument ID: P4 Start Date: 01/29/2025
 Analytical Method: ICP-AES Run Batch: LB134475
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	0	0.070	0	200	180	-10	38750	35000	-10
Antimony	0	0.010	0	60	57	-5	1250	1200	-4
Arsenic	0	0.0	0	10	10	0	1250	1200	-4
Barium	0	0.020	0	200	180	-10	2500	2500	0
Beryllium	0	-0	0	5	5.4	8	125	130	4
Cadmium	0	-0	0	5	5.1	2	625	630	1
Calcium	0	0.66	0	5000	4700	-6	62500	61000	-2
Chromium	0	-0	0	10	11	10	3875	4100	6
Cobalt	0	-0	0	50	50	0	625	620	-1
Copper	0	-0	0	25	29	16	3750	3700	-1
Iron	0	0.010	0	100	95	-5	37500	38000	1
Lead	0	-0	0	10	11	10	6250	6200	-1
Magnesium	0	1.6	0	5000	4200	-16	62500	58000	-7
Manganese	0	-0	0	15	15	0	3750	3800	1
Nickel	0	-0	0	40	41	3	625	630	1
Potassium	0	1.0	0	5000	4400	-12	37500	35000	-7
Selenium	0	-0	0	35	37	6	1250	1200	-4
Silver	0	0.0	0	10	9.6	-4	312.5	300	-4
Sodium	0	1.0	0	5000	4700	-6	37500	35000	-7
Thallium	0	-0	0	25	27	8	1250	1300	4
Vanadium	0	-0.010	0	50	55	10	625	640	2
Zinc	0	0.0	0	60	60	0	3750	3700	-1

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51973 MA No.: _____ SDG No.: MC0AZ0
 Instrument ID: P4 Start Date: 01/29/2025
 Analytical Method: ICP-AES Run Batch: LB134475
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	77500	70000	-10	155000	150000	-3	310000	290000	-6
Antimony	2500	2400	-4	5000	4900	-2	10000	10000	0
Arsenic	2500	2400	-4	5000	4900	-2	10000	10000	0
Barium	5000	4900	-2	10000	10000	0	20000	20000	0
Beryllium	250	260	4	500	500	0	1000	990	-1
Cadmium	1250	1300	4	2500	2500	0	5000	4900	-2
Calcium	125000	120000	-4	250000	250000	0	500000	480000	-4
Chromium	7750	8100	5	15500	16000	3	31000	30000	-3
Cobalt	1250	1300	4	2500	2500	0	5000	5000	0
Copper	7500	7400	-1	15000	15000	0	30000	30000	0
Iron	75000	74000	-1	150000	150000	0	300000	290000	-3
Lead	12500	13000	4	25000	25000	0	50000	50000	0
Magnesium	125000	120000	-4	250000	240000	-4	500000	480000	-4
Manganese	7500	7500	0	15000	16000	7	30000	29000	-3
Nickel	1250	1300	4	2500	2500	0	5000	4900	-2
Potassium	75000	72000	-4	150000	150000	0	300000	310000	3
Selenium	2500	2400	-4	5000	5000	0	10000	10000	0
Silver	625	610	-2	1250	1200	-4	2500	2500	0
Sodium	75000	71000	-5	150000	150000	0	300000	280000	-7
Thallium	2500	2600	4	5000	5100	2	10000	9700	-3
Vanadium	1250	1300	4	2500	2600	4	5000	4900	-2
Zinc	7500	7300	-3	15000	15000	0	30000	30000	0

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51973 MA No.: _____ SDG No.: MC0AZ0
 Instrument ID: P4 Start Date: 01/29/2025
 Analytical Method: ICP-AES Run Batch: LB134475
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	800000	840000	5						
Antimony	0	0.0	0						
Arsenic	0	0.0	0						
Barium	0	0.0	0						
Beryllium	0	0.0	0						
Cadmium	0	0.0	0						
Calcium	800000	830000	4						
Chromium	0	0.0	0						
Cobalt	0	0.0	0						
Copper	0	0.0	0						
Iron	800000	820000	3						
Lead	0	0.0	0						
Magnesium	800000	840000	5						
Manganese	0	0.0	0						
Nickel	0	0.0	0						
Potassium	0	0.0	0						
Selenium	0	0.0	0						
Silver	0	0.0	0						
Sodium	800000	830000	4						
Thallium	0	0.0	0						
Vanadium	0	0.0	0						
Zinc	0	0.0	0						

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51973 MA No.: _____ SDG No.: MC0AZ0
 Instrument ID: P4 Start Date: 01/31/2025
 Analytical Method: ICP-AES Run Batch: LB134517
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	0	0.070	0	200	180	-10	38750	34000	-12
Antimony	0	0.020	0	60	54	-10	1250	1100	-12
Arsenic	0	0.0	0	10	7.9	-21	1250	1100	-12
Barium	0	0.030	0	200	190	-5	2500	2400	-4
Beryllium	0	-0	0	5	5.2	4	125	130	4
Cadmium	0	-0	0	5	5.1	2	625	630	1
Calcium	0	0.82	0	5000	4800	-4	62500	60000	-4
Chromium	0	-0	0	10	11	10	3875	4000	3
Cobalt	0	-0	0	50	50	0	625	620	-1
Copper	0	0.0	0	25	24	-4	3750	3700	-1
Iron	0	0.010	0	100	97	-3	37500	37000	-1
Lead	0	0.0	0	10	9.1	-9	6250	6200	-1
Magnesium	0	1.6	0	5000	4400	-12	62500	58000	-7
Manganese	0	-0	0	15	16	7	3750	3700	-1
Nickel	0	-0	0	40	41	3	625	630	1
Potassium	0	1.3	0	5000	4500	-10	37500	34000	-9
Selenium	0	0.010	0	35	29	-17	1250	1100	-12
Silver	0	0.0	0	10	9.4	-6	312.5	300	-4
Sodium	0	0.87	0	5000	5100	2	37500	36000	-4
Thallium	0	-0	0	25	28	12	1250	1300	4
Vanadium	0	-0	0	50	53	6	625	630	1
Zinc	0	0.010	0	60	58	-3	3750	3500	-7

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51973 MA No.: _____ SDG No.: MC0AZ0
 Instrument ID: P4 Start Date: 01/31/2025
 Analytical Method: ICP-AES Run Batch: LB134517
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	77500	69000	-11	155000	140000	-10	310000	320000	3
Antimony	2500	2300	-8	5000	4900	-2	10000	10000	0
Arsenic	2500	2300	-8	5000	4900	-2	10000	10000	0
Barium	5000	4700	-6	10000	9100	-9	20000	21000	5
Beryllium	250	260	4	500	540	8	1000	950	-5
Cadmium	1250	1300	4	2500	2600	4	5000	4900	-2
Calcium	125000	120000	-4	250000	230000	-8	500000	510000	2
Chromium	7750	8000	3	15500	16000	3	31000	30000	-3
Cobalt	1250	1200	-4	2500	2500	0	5000	5000	0
Copper	7500	7300	-3	15000	14000	-7	30000	31000	3
Iron	75000	73000	-3	150000	140000	-7	300000	310000	3
Lead	12500	12000	-4	25000	26000	4	50000	50000	0
Magnesium	125000	110000	-12	250000	230000	-8	500000	510000	2
Manganese	7500	7300	-3	15000	14000	-7	30000	31000	3
Nickel	1250	1300	4	2500	2600	4	5000	4900	-2
Potassium	75000	69000	-8	150000	140000	-7	300000	320000	7
Selenium	2500	2300	-8	5000	4900	-2	10000	10000	0
Silver	625	600	-4	1250	1200	-4	2500	2500	0
Sodium	75000	70000	-7	150000	140000	-7	300000	290000	-3
Thallium	2500	2600	4	5000	5100	2	10000	9800	-2
Vanadium	1250	1300	4	2500	2500	0	5000	5000	0
Zinc	7500	7000	-7	15000	14000	-7	30000	32000	7

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51973 MA No.: _____ SDG No.: MC0AZ0
 Instrument ID: P4 Start Date: 01/31/2025
 Analytical Method: ICP-AES Run Batch: LB134517
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	800000	820000	3						
Antimony	0	0.0	0						
Arsenic	0	0.0	0						
Barium	0	0.0	0						
Beryllium	0	0.0	0						
Cadmium	0	0.0	0						
Calcium	800000	820000	3						
Chromium	0	0.0	0						
Cobalt	0	0.0	0						
Copper	0	0.0	0						
Iron	800000	800000	0						
Lead	0	0.0	0						
Magnesium	800000	830000	4						
Manganese	0	0.0	0						
Nickel	0	0.0	0						
Potassium	0	0.0	0						
Selenium	0	0.0	0						
Silver	0	0.0	0						
Sodium	800000	830000	4						
Thallium	0	0.0	0						
Vanadium	0	0.0	0						
Zinc	0	0.0	0						

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51973 MA No.: _____ SDG No.: MC0AZ0
 Instrument ID: P5 Start Date: 02/06/2025
 Analytical Method: ICP-AES Run Batch: LB134617
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	0	-0.020	0	200	220	10	38750	38000	-2
Antimony	0	0.010	0	60	55	-8	1250	1200	-4
Arsenic	0	-0	0	10	12	20	1250	1300	4
Barium	0	-0.010	0	200	210	5	2500	2500	0
Beryllium	0	-0	0	5	5.4	8	125	130	4
Cadmium	0	0.0	0	5	4.9	-2	625	610	-2
Calcium	0	-0.90	0	5000	5400	8	62500	65000	4
Chromium	0	-0	0	10	11	10	3875	3900	1
Cobalt	0	-0	0	50	52	4	625	630	1
Copper	0	-0	0	25	28	12	3750	3800	1
Iron	0	-0.030	0	100	120	20	37500	39000	4
Lead	0	-0	0	10	10	0	6250	6100	-2
Magnesium	0	-0.74	0	5000	5300	6	62500	65000	4
Manganese	0	-0	0	15	18	20	3750	3900	4
Nickel	0	0.0	0	40	39	-3	625	620	-1
Potassium	0	-1.3	0	5000	6200	24	37500	38000	1
Selenium	0	0.010	0	35	32	-9	1250	1200	-4
Silver	0	-0	0	10	11	10	312.5	300	-4
Sodium	0	0.040	0	5000	5000	0	37500	37000	-1
Thallium	0	-0	0	25	26	4	1250	1300	4
Vanadium	0	-0.010	0	50	55	10	625	640	2
Zinc	0	0.0	0	60	59	-2	3750	3600	-4

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51973 MA No.: _____ SDG No.: MC0AZ0
 Instrument ID: P5 Start Date: 02/06/2025
 Analytical Method: ICP-AES Run Batch: LB134617
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	77500	78000	1	155000	160000	3	310000	310000	0
Antimony	2500	2400	-4	5000	5000	0	10000	10000	0
Arsenic	2500	2600	4	5000	5100	2	10000	9800	-2
Barium	5000	5100	2	10000	10000	0	20000	20000	0
Beryllium	250	260	4	500	510	2	1000	980	-2
Cadmium	1250	1300	4	2500	2500	0	5000	5000	0
Calcium	125000	130000	4	250000	260000	4	500000	490000	-2
Chromium	7750	8000	3	15500	16000	3	31000	30000	-3
Cobalt	1250	1300	4	2500	2500	0	5000	4900	-2
Copper	7500	7700	3	15000	15000	0	30000	30000	0
Iron	75000	79000	5	150000	150000	0	300000	300000	0
Lead	12500	13000	4	25000	25000	0	50000	49000	-2
Magnesium	125000	130000	4	250000	250000	0	500000	490000	-2
Manganese	7500	7800	4	15000	15000	0	30000	29000	-3
Nickel	1250	1300	4	2500	2500	0	5000	5000	0
Potassium	75000	77000	3	150000	150000	0	300000	300000	0
Selenium	2500	2400	-4	5000	5000	0	10000	10000	0
Silver	625	620	-1	1250	1300	4	2500	2500	0
Sodium	75000	76000	1	150000	150000	0	300000	290000	-3
Thallium	2500	2600	4	5000	5100	2	10000	9700	-3
Vanadium	1250	1300	4	2500	2500	0	5000	4900	-2
Zinc	7500	7400	-1	15000	15000	0	30000	31000	3

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51973 MA No.: SDG No.: MC0AZ0
Instrument ID: P5 Start Date: 02/06/2025
Analytical Method: ICP-AES Run Batch: LB134617
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	800000	800000	0						
Calcium	800000	790000	-1						
Iron	800000	790000	-1						
Magnesium	800000	800000	0						
Sodium	800000	810000	1						