

MC0AZ7A

FORM 5B - IN

POST-DIGESTION/DISTILLATION SPIKE
SAMPLE RECOVERY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011

Lab Code: ACE Case No.: 51973 MA No. : SDG No.: MC0AZ0

Matrix : Soil Analytical Method: ICP-AES

% Solids: 41.4

Concentration Units (µg/L, mg/L or mg/kg dry weight): mg/kg

Analyte	Control Limit %R	Spiked Sample Result (SSR) Q	Sample Result (SR) Q	Spike Added (SA)	%R	Q
Arsenic	75 - 125	37	17	33.0	61	
Beryllium	75 - 125	12	3.9	8.8	92	
Copper	75 - 125	130	53	110	70	
Manganese	75 - 125	210	84	170	74	
Selenium	75 - 125	15	5.6 J	15	63	
Zinc	75 - 125	270	120	220	68	