



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tanagro Corp. **SDG No.:** Q1152
Contract: TANA01 **Lab Code:** CHEM **Case No.:** Q1152 **SAS No.:** Q1152
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Cadmium	5.56	6.0	93	40 - 160	P	01/30/2025	12:05	LB134493
	Chromium	9.97	10.0	100	40 - 160	P	01/30/2025	12:05	LB134493
	Copper	21.8	20.0	109	40 - 160	P	01/30/2025	12:05	LB134493
	Lead	10.9	12.0	91	40 - 160	P	01/30/2025	12:05	LB134493
	Molybdenum	205	200	103	40 - 160	P	01/30/2025	12:05	LB134493
	Nickel	38.1	40.0	95	40 - 160	P	01/30/2025	12:05	LB134493
	Silver	11.1	10.0	111	40 - 160	P	01/30/2025	12:05	LB134493
	Zinc	39.3	40.0	98	40 - 160	P	01/30/2025	12:05	LB134493
CRA	Mercury	0.050	0.05	101	40 - 160	CV	02/05/2025	17:30	LB134588