

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Tanagro Corp. **Level:** LOW **SDG No.:** Q1152
Contract: TANA01 **Lab Code:** CHEM **Case No.:** Q1152 **SAS No.:** Q1152
Matrix: Water **Sample ID:** Q1152-01 **Client ID:** EFFLUENT-DAY-1-METALSDUP
Percent Solids for Sample: NA **Duplicate ID** Q1152-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	0.40	J	0.38	J	5		P
Chromium	ug/L	20	2.64	J	2.79	J	6		P
Copper	ug/L	20	239		241		1		P
Lead	ug/L	20	84.8		85.2		0		P
Mercury	ug/L	20	0.25		0.20		22	*	CV
Molybdenum	ug/L	20	100	U	100	U			P
Nickel	ug/L	20	5.86	J	6.01	J	3		P
Silver	ug/L	20	48.9		50.2		3		P
Zinc	ug/L	20	1030		1050		2		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Tanagro Corp. **Level:** LOW **SDG No.:** Q1152
Contract: TANA01 **Lab Code:** CHEM **Case No.:** Q1152 **SAS No.:** Q1152
Matrix: Water **Sample ID:** Q1152-01MS **Client ID:** EFFLUENT-DAY-1-METALSMDS
Percent Solids for Sample: NA **Duplicate ID** Q1152-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	91.2		91.2		0		P
Chromium	ug/L	20	192		193		1		P
Copper	ug/L	20	375		377		1		P
Lead	ug/L	20	517		518		0		P
Mercury	ug/L	20	3.10		2.54		20		CV
Molybdenum	ug/L	20	200		201		0		P
Nickel	ug/L	20	236		236		0		P
Silver	ug/L	20	82.1		82.6		1		P
Zinc	ug/L	20	992		995		0		P