

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tanagro Corp. **SDG No.:** Q1152
Contract: TANA01 **Lab Code:** CHEM **Case No.:** Q1152 **SAS No.:** Q1152

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166190BS							
Cadmium	ug/L	100	94.1		94	85 - 115	P
Chromium	ug/L	200	202		101	85 - 115	P
Copper	ug/L	150	152		101	85 - 115	P
Lead	ug/L	500	476		95	85 - 115	P
Molybdenum	ug/L	200	203		102	85 - 115	P
Nickel	ug/L	250	239		96	85 - 115	P
Silver	ug/L	37.5	36.7		98	85 - 115	P
Zinc	ug/L	100	101		101	85 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tanagro Corp.

SDG No.: Q1152

Contract: TANA01

Lab Code: CHEM

Case No.: Q1152

SAS No.: Q1152

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166583BS Mercury	ug/L	4.0	4.26		106	85 - 115	CV