



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: PSEG SDG No.: Q1154
Contract: PSEG03 Lab Code: CHEM Case No.: Q1154 SAS No.: Q1154
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.17	0.2	83	40 - 160	CV	01/23/2025	15:28	LB134389
CRI01	Aluminum	88.0	100	88	40 - 160	P	01/27/2025	11:29	LB134427
	Antimony	52.5	50.0	105	40 - 160	P	01/27/2025	11:29	LB134427
	Arsenic	21.5	20.0	107	40 - 160	P	01/27/2025	11:29	LB134427
	Barium	93.3	100	93	40 - 160	P	01/27/2025	11:29	LB134427
	Beryllium	5.69	6.0	95	40 - 160	P	01/27/2025	11:29	LB134427
	Cadmium	6.00	6.0	100	40 - 160	P	01/27/2025	11:29	LB134427
	Calcium	1960	2000	98	40 - 160	P	01/27/2025	11:29	LB134427
	Chromium	10.2	10.0	102	40 - 160	P	01/27/2025	11:29	LB134427
	Cobalt	29.4	30.0	98	40 - 160	P	01/27/2025	11:29	LB134427
	Copper	22.1	20.0	110	40 - 160	P	01/27/2025	11:29	LB134427
	Iron	96.9	100	97	40 - 160	P	01/27/2025	11:29	LB134427
	Lead	11.4	12.0	95	40 - 160	P	01/27/2025	11:29	LB134427
	Magnesium	2040	2000	102	40 - 160	P	01/27/2025	11:29	LB134427
	Manganese	19.2	20.0	96	40 - 160	P	01/27/2025	11:29	LB134427
	Nickel	39.7	40.0	99	40 - 160	P	01/27/2025	11:29	LB134427
	Potassium	2150	2000	107	40 - 160	P	01/27/2025	11:29	LB134427
	Selenium	21.6	20.0	108	40 - 160	P	01/27/2025	11:29	LB134427
	Silver	10.5	10.0	105	40 - 160	P	01/27/2025	11:29	LB134427
	Sodium	1970	2000	98	40 - 160	P	01/27/2025	11:29	LB134427
	Thallium	40.0	40.0	100	40 - 160	P	01/27/2025	11:29	LB134427
	Vanadium	38.5	40.0	96	40 - 160	P	01/27/2025	11:29	LB134427
	Zinc	45.0	40.0	112	40 - 160	P	01/27/2025	11:29	LB134427