



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: CHA Companies, Inc.

SDG No.: Q1157

Project: 355320 Kingsboro PC Bldgs. 2 & 28 Interior Reno

RunNo.: LB134371

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Residual Chlorine	mg/L	0.378	0.4	94	90-110	01/23/2025
Sample ID: CCV1 Residual Chlorine	mg/L	0.407	0.4	102	90-110	01/23/2025
Sample ID: CCV2 Residual Chlorine	mg/L	0.397	0.4	99	90-110	01/23/2025
Sample ID: CCV3 Residual Chlorine	mg/L	0.388	0.4	97	90-110	01/23/2025

Initial and Continuing Calibration Verification

Client: CHA Companies, Inc.

SDG No.: Q1157

Project: 355320 Kingsboro PC Bldgs. 2 & 28 Interior Reno

RunNo.: LB134384

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Turbidity	ICV	NTU	9.920	10	99	90-110	01/23/2025
Sample ID: Turbidity	CCV1	NTU	10.193	10	102	90-110	01/23/2025
Sample ID: Turbidity	CCV2	NTU	10.046	10	100	90-110	01/23/2025
Sample ID: Turbidity	CCV3	NTU	9.384	10	94	90-110	01/23/2025

Initial and Continuing Calibration Verification

Client: CHA Companies, Inc.

SDG No.: Q1157

Project: 355320 Kingsboro PC Bldgs. 2 & 28 Interior Reno

RunNo.: LB134384

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------