

Laboratory Control Sample Summary

Client: CHA Companies, Inc.

SDG No.: Q1157

Project: 355320 Kingsboro PC Bldgs. 2 & 28 Interior Reno

Run No.: LB134371

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134371BS							
Residual Chlorine	mg/L	0.4	0.38		94	1	90-110	01/23/2025