

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51955 MA No. : SDG No.: YE8C9
 Initial Calibration Verification Source : MP84213
 Continuing Calibration Verification Source : MP84216
 Run Batch: LB134475 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV005				ID: CCV028				ID: CCV029		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	2500	2400	94	1	400000	410000	101	1	400000	100	2
Calcium	10000	10000	101	1	400000	410000	103	1	410000	102	2
Iron	10000	9300	93	1	400000	410000	102	0	410000	102	2
Magnesium	6000	5800	96	0	400000	410000	102	0	400000	101	2
Potassium	9900	9500	96	1	150000	160000	109	1	160000	110	2
Sodium	10000	10000	103	2	400000	410000	103	0	390000	97	0