

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134475

Review By	Kareem Khairalla	Review On	1/30/2025 3:36:15 PM
Supervise By	Mohan Bera	Supervise On	1/31/2025 7:33:47 AM
STD. NAME	STD REF.#		
ICAL Standard	MP84204,MP84212,MP84210,MP84209,MP84208,MP84207,MP84206		
ICV Standard	MP84213		
CCV Standard	MP84216		
ICSA Standard	MP84214,MP84215		
CRI Standard			
LCS Standard			
Chk Standard	MP84218,MP84219		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	01/29/25 14:56		Kareem	OK
2	S1	S01	CAL2	01/29/25 15:00		Kareem	OK
3	S2	S02	CAL3	01/29/25 15:04		Kareem	OK
4	S3	S03	CAL4	01/29/25 15:08		Kareem	OK
5	S4	S04	CAL5	01/29/25 15:12		Kareem	OK
6	S5	S05	CAL6	01/29/25 15:16		Kareem	OK
7	S6	S06	CAL7	01/29/25 15:21		Kareem	OK
8	ICV005	ICV005	ICV	01/29/25 15:31	Inst. paused for calibration review	Kareem	OK
9	ICB005	ICB005	ICB	01/29/25 15:35		Kareem	OK
10	ICSA005	ICSA005	ICSA	01/29/25 15:41	Inst. paused for review	Kareem	OK
11	ICSAB005	ICSAB005	ICSAB	01/29/25 15:45		Kareem	OK
12	CCV028	CCV028	CCV	01/29/25 15:49		Kareem	OK
13	CCB028	CCB028	CCB	01/29/25 15:53		Kareem	OK
14	Q1136-02	YE8F7	SAM	01/29/25 15:58		Kareem	OK
15	Q1136-03	YE8F7D	DUP	01/29/25 16:01		Kareem	OK
16	Q1136-02L	YE8F7L	SD	01/29/25 16:05		Kareem	OK
17	Q1136-04	YE8F7S	MS	01/29/25 16:09	MS fail for Cd,Co,Pb,Ni,Ag,Tl	Kareem	OK
18	Q1151-01DL	MC0AZ0	SAM	01/29/25 16:13	20x for Cr,Co,Ni, Still Co,Ni high	Kareem	Dilution

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134475

Review By	Kareem Khairalla	Review On	1/30/2025 3:36:15 PM
Supervise By	Mohan Bera	Supervise On	1/31/2025 7:33:47 AM
STD. NAME	STD REF.#		
ICAL Standard	MP84204,MP84212,MP84210,MP84209,MP84208,MP84207,MP84206		
ICV Standard	MP84213		
CCV Standard	MP84216		
ICSA Standard	MP84214,MP84215		
CRI Standard			
LCS Standard			
Chk Standard	MP84218,MP84219		

19	Q1151-02DL	MC0AZ1	SAM	01/29/25 16:18	5x for Co,Ni	Kareem	Confirms
20	Q1159-01	YE8C9	SAM	01/29/25 16:21		Kareem	OK
21	Q1151-01DL2	MC0AZ0	SAM	01/29/25 16:25	40x for Co,Ni	Kareem	Confirms
22	Q1159-02	YE8C9D	DUP	01/29/25 16:29		Kareem	OK
23	Q1159-01L	YE8C9L	SD	01/29/25 16:33		Kareem	OK
24	Q1159-03	YE8C9S	MS	01/29/25 16:38		Kareem	OK
25	Q1159-04	YE8D1	SAM	01/29/25 16:42		Kareem	OK
26	Q1159-05	YE8D3	SAM	01/29/25 16:46		Kareem	OK
27	Q1159-06	YE8E2	SAM	01/29/25 16:50		Kareem	OK
28	PB166301BL	PBS301	MB	01/29/25 16:54		Kareem	OK
29	PB166301BS	LCS301	LCS	01/29/25 16:58		Kareem	OK
30	CCV029	CCV029	CCV	01/29/25 17:02		Kareem	OK
31	CCB029	CCB029	CCB	01/29/25 17:06		Kareem	OK
32	Q1136-02A	YE8F7A	PS	01/29/25 17:49	PS for Cd,Co,Pb,Ni,Tl(0.01 ML MP84292, 0.02 ML MP84296, 0.02 ML MP84294, 0.002 ML M5961, 0.005 ML MP84293)	Kareem	OK
33	CCV030	CCV030	CCV	01/29/25 18:01		Kareem	OK
34	CCB030	CCB030	CCB	01/29/25 18:06		Kareem	OK