


Batch Folder: D:\Agilent\ICPMH\1\DATA\P8021125 MS.b\
Analysis File: P8021125 MS.batch.bin
DA Date-Time: 2025-02-11 16:41:16
Calibration Title:
Calibration Method: External Calibration
VIS Interpolation Fit:

Level	Standard Data File	Sample Name	Acq. Date-Time
1	004CALB.d	S00	2025-02-11 12:05:13
2	005CALS.d	S02	2025-02-11 12:08:35
3	006CALB.d	S03	2025-02-11 12:11:58
4	007CALS.d	S04	2025-02-11 12:15:02
5	008CALS.d	S05	2025-02-11 12:18:01
6	009CALS.d	S06	2025-02-11 12:20:51
7	010CALS.d	S07	2025-02-11 12:23:38
8	011CALS.d	S08	2025-02-11 12:26:23

9 Be [No Gas] ISTD: 6 Li [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	361.63	0.0000	P	1.2
2	☐	1.000	1.074	6278.90	0.0008	P	1.1
3	☐	50.000	51.377	280845.21	0.0379	P	2.5
4	☐	125.000	125.718	687140.12	0.0927	P	2.1
5	☐	250.000	253.365	1368162.02	0.1867	A	2.4
6	☐	500.000	505.733	2695154.28	0.3726	A	0.9
7	☐	1000.000	996.134	5226624.33	0.7338	A	0.3
8	☐			1942.39	0.0003	P	10.4

$$y = 7.3661E-004 * x + 4.8717E-005$$

R = 1.0000


DL = 0.00238

BEC = 0.06614

Weight: <None>


Min Conc: 0

12 C [No Gas] No available calibration points




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐			3360557.45		A	1.2
2	☐			3369094.92		A	0.1
3	☐			3438303.32		A	0.4
4	☐			3592506.75		A	0.6
5	☐			3714614.04		A	1.3
6	☐			3972299.70		A	1.7
7	☐			4509665.94		A	3.1
8	☐			4069612.27		A	2.3

12 C [He] No available calibration points



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			22745.42		P	1.3
2	☐			22698.71		P	1.5
3	☐			23193.93		P	0.5
4	☐			24332.42		P	1.1
5	☐			26180.01		P	1.6
6	☐			28867.23		P	1.0
7	☐			34647.49		P	2.1
8	☐			38443.49		P	0.4

23 Na [He] ISTD:45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	56834.43	0.3075	P	1.5
2	☐	500.000	541.926	296212.48	1.5827	P	0.9
3	☐	5000.000	5395.509	2384490.40	13.0036	A	1.7
4	☐	12500.000	13008.194	5714264.08	30.9169	A	0.1
5	☐	25000.000	26266.120	11455199.41	62.1139	A	2.8
6	☐	50000.000	52959.770	22737010.22	124.9263	A	0.9
7	☐	100000.00	103433.66	44802141.00	243.6956	A	0.1
8	☐	500000.00	498937.28	205456994.7	1,174.348	A	1.2

$$y = 0.0024 * x + 0.3075$$


$$R = 1.0000$$

$$DL = 5.923$$

$$BEC = 130.7$$

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	4133.96	0.0224	P	4.3
2	☐	500.000	535.768	126672.02	0.6768	P	0.5
3	☐	5000.000	5346.022	1201561.62	6.5525	A	1.4
4	☐	12500.000	12958.730	2929735.09	15.8514	A	0.7
5	☐	25000.000	25946.260	5849317.27	31.7157	A	2.8
6	☐	50000.000	51175.425	11381502.05	62.5332	A	0.4
7	☐	100000.00	100144.98	22493654.67	122.3495	A	0.6
8	☐	500000.00	499791.18	106819078.4	610.5175	A	0.3

$$y = 0.0012 * x + 0.0224$$


$$R = 1.0000$$

$$DL = 2.376$$

$$BEC = 18.3$$

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	113.33	0.0006	P	11.1
2	☐	20.000	18.890	1317.85	0.0070	P	4.6
3	☐	1000.000	1002.990	62685.38	0.3418	P	0.2
4	☐	2500.000	2460.172	154800.30	0.8376	P	0.8
5	☐	5000.000	4942.658	310252.35	1.6821	P	2.1
6	☐	10000.000	9829.508	608743.34	3.3446	P	0.5
7	☐	20000.000	19944.046	1247461.68	6.7856	A	1.6
8	☐	100000.00	100032.07	5954172.07	34.0317	A	0.9

$$y = 3.4020E-004 * x + 6.1331E-004$$

$$R = 1.0000$$


$$DL = 0.5992$$

$$BEC = 1.803$$

Weight: <None>

Min Conc: 0

28 Si [No Gas] ISTD :45 Sc [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	513265.19	0.0490	P	0.6
2	☐	10.000	6.376	590083.28	0.0567	P	0.4
3	☐	50.000	51.081	1168455.04	0.1105	A	0.8
4	☐	125.000	128.798	2134505.19	0.2041	A	1.2
5	☐	250.000	253.013	3706115.15	0.3536	A	2.6
6	☐	500.000	512.571	6836667.26	0.6661	A	2.9
7	☐	1000.000	992.469	12718038.98	1.2438	A	1.5
8	☐			670331.67	0.0697	P	2.2

$$y = 0.0012 * x + 0.0490$$

$$R = 0.9999$$


$$DL = 0.7172$$

$$BEC = 40.75$$

Weight: <None>

Min Conc: 0

31 P [He] ISTD :45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	-11.458	86.67	0.0005	P	16.3
2	☐	0.000	11.458	154.45	0.0008	P	18.3
3	☐	1000.000	982.833	2924.77	0.0160	P	5.7
4	☐	2500.000	2513.659	7354.15	0.0398	P	0.4
5	☐	5000.000	4977.904	14412.81	0.0782	P	3.8
6	☐	10000.000	10324.689	29379.74	0.1614	P	0.3
7	☐	20000.000	19842.330	56922.18	0.3096	P	0.9
8	☐			88.89	0.0005	P	18.3

$$y = 1.5572E-005 * x + 6.4699E-004$$

$$R = 0.9998$$


$$DL = 21.95$$

$$BEC = 41.55$$

Weight: <None>

Min Conc: 0

34 S [No Gas] ISTD:45 Sc [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	-122.073	746544.49	0.0713	P	0.2
2	☐	0.000	122.073	758825.73	0.0729	P	0.2
3	☐	1000.000	1186.137	844998.56	0.0799	P	0.6
4	☐	2500.000	2786.306	946100.82	0.0905	A	1.5
5	☐	5000.000	5051.511	1104109.44	0.1053	A	2.0
6	☐	10000.000	10299.215	1435460.71	0.1398	A	1.9
7	☐	20000.000	19792.420	2067919.90	0.2022	A	1.3
8	☐			682747.53	0.0710	P	1.3

$$y = 6.5727E-006 * x + 0.0721$$

R = 0.9997


DL = 52.29

BEC = 1.098E+04

Weight: <None>


Min Conc: 0

35 Cl [No Gas] No available calibration points




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			127507.58		P	0.4
2	☐			127300.34		P	1.0
3	☐			126941.70		P	0.1
4	☐			128789.58		P	0.7
5	☐			132751.66		P	1.2
6	☐			143977.44		P	1.3
7	☐			165127.62		P	1.8
8	☐			125792.96		P	1.6

35 Cl [He] No available calibration points



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			496.68		P	4.4
2	☐			426.68		P	6.8
3	☐			451.12		P	10.6
4	☐			468.90		P	15.1
5	☐			474.46		P	5.4
6	☐			556.68		P	2.2
7	☐			630.02		P	7.4
8	☐			472.23		P	12.0

39 K [He] ISTD :45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	9404.26	0.0509	P	2.0
2	☐	500.000	511.948	72701.71	0.3885	P	0.9
3	☐	2500.000	2515.987	313557.84	1.7099	P	0.8
4	☐	6250.000	6130.526	756536.93	4.0933	P	0.6
5	☐	12500.000	12535.253	1533826.82	8.3166	A	2.5
6	☐	25000.000	25173.078	3030279.68	16.6499	A	1.2
7	☐	50000.000	48434.432	5880779.98	31.9883	A	0.9
8	☐	250000.00	250296.84	28881381.23	165.0952	A	2.5

$$y = 6.5939E-004 * x + 0.0509$$

$$R = 1.0000$$


$$DL = 4.646$$

$$BEC = 77.16$$

Weight: <None>

Min Conc: 0

43 Ca [No Gas] ISTD:45 Sc [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	418.90	0.0000	P	8.2
2	☐	500.000	557.274	28606.29	0.0027	P	1.0
3	☐	5000.000	5237.620	269627.71	0.0255	P	1.6
4	☐	12500.000	12948.521	658945.87	0.0630	P	1.0
5	☐	25000.000	25865.046	1318677.40	0.1258	A	2.3
6	☐	50000.000	52812.055	2636408.17	0.2569	A	2.6
7	☐	100000.00	102909.14	5117125.24	0.5005	A	1.9
8	☐	500000.00	499080.06	23340811.32	2.4270	A	1.8

$$y = 4.8628E-006 * x + 4.0045E-005$$

R = 1.0000


DL = 2.029

BEC = 8.235

Weight: <None>

Min Conc: 0

44 Ca [No Gas] ISTD:45 Sc [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	29255.37	0.0028	P	1.1
2	☐	500.000	550.279	482897.91	0.0464	P	0.9
3	☐	5000.000	5237.700	4418528.03	0.4180	A	1.7
4	☐	12500.000	12884.254	10712373.38	1.0243	A	0.1
5	☐	25000.000	25348.293	21091365.24	2.0124	A	2.7
6	☐	50000.000	52048.012	42391909.09	4.1292	A	1.5
7	☐	100000.00	102351.35	83002634.32	8.1172	A	1.1
8	☐	500000.00	499295.47	380710749.8	39.5869	A	1.9

$$y = 7.9280E-005 * x + 0.0028$$

R = 1.0000


DL = 1.132

BEC = 35.26

Weight: <None>

Min Conc: 0

47 Ti [No Gas] ISTD:45 Sc [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	640.08	0.0001	P	15.8
2	☐	5.000	5.289	13608.75	0.0013	P	0.9
3	☐	500.000	512.437	1277689.90	0.1209	A	1.8
4	☐	1250.000	1266.661	3124153.88	0.2987	A	0.9
5	☐	2500.000	2461.336	6083835.88	0.5804	A	1.1
6	☐	5000.000	5058.425	12245010.10	1.1927	A	1.2
7	☐	10000.000	9977.749	24056862.42	2.3526	A	1.3
8	☐			4429.62	0.0005	P	3.8

$$y = 2.3578E-004 * x + 6.1210E-005$$

$$R = 1.0000$$


$$DL = 0.1232$$

$$BEC = 0.2596$$

Weight: <None>

Min Conc: 0

51 V [He] ISTD:45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	15.55	0.0001	P	45.5
2	☐	5.000	5.287	13518.69	0.0722	P	1.4
3	☐	50.000	50.766	127068.63	0.6929	P	0.7
4	☐	125.000	124.016	312836.60	1.6926	P	0.2
5	☐	250.000	246.737	621038.67	3.3674	P	2.7
6	☐	500.000	497.839	1236563.13	6.7943	P	1.2
7	☐	1000.000	1001.980	2513928.26	13.6746	A	1.4
8	☐			721.14	0.0041	P	5.7

$$y = 0.0136 * x + 8.4396E-005$$

$$R = 1.0000$$


$$DL = 0.008434$$

$$BEC = 0.006184$$

Weight: <None>

Min Conc: 0

52 Cr [He] ISTD:45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	746.69	0.0040	P	5.8
2	☐	2.000	2.074	7658.78	0.0409	P	1.3
3	☐	50.000	50.527	165473.33	0.9024	P	0.7
4	☐	125.000	123.544	406715.60	2.2006	P	0.4
5	☐	250.000	243.673	799782.35	4.3364	P	2.2
6	☐	500.000	506.313	1639015.11	9.0059	A	2.1
7	☐	1000.000	998.581	3264720.71	17.7580	A	0.5
8	☐			10331.59	0.0590	P	1.9

$$y = 0.0178 * x + 0.0040$$

$$R = 0.9999$$


$$DL = 0.03972$$

$$BEC = 0.2271$$

Weight: <None>

Min Conc: 0

54 Fe [He] ISTD:45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	363.34	0.0020	P	15.0
2	☐	50.000	56.743	7796.63	0.0417	P	1.1
3	☐	2500.000	2696.478	346196.43	1.8879	P	0.6
4	☐	6250.000	6564.854	848945.15	4.5934	P	1.1
5	☐	12500.000	13288.875	1714697.87	9.2961	A	1.8
6	☐	25000.000	26612.619	3387818.60	18.6147	A	1.5
7	☐	50000.000	51759.812	6655597.40	36.2025	A	0.4
8	☐	250000.00	249437.49	30521899.54	174.4570	A	1.2

$$y = 6.9939E-004 * x + 0.0020$$


$$R = 1.0000$$

$$DL = 1.267$$

$$BEC = 2.814$$

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	164.45	0.0009	P	13.9
2	☐	1.000	1.039	1540.10	0.0082	P	3.3
3	☐	500.000	514.385	666753.06	3.6359	P	0.4
4	☐	1250.000	1288.953	1683574.99	9.1096	A	1.8
5	☐	2500.000	2554.575	3330009.78	18.0535	A	1.5
6	☐	5000.000	5051.624	6497271.29	35.6995	A	1.6
7	☐	10000.000	9954.956	12933611.89	70.3502	A	0.3
8	☐			5867.92	0.0335	P	2.6

$$y = 0.0071 * x + 8.8853E-004$$


$$R = 1.0000$$

$$DL = 0.05231$$

$$BEC = 0.1257$$

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	5986.86	0.0324	P	1.5
2	☐	50.000	55.418	139578.56	0.7458	P	0.7
3	☐	2500.000	2687.743	6350587.89	34.6310	A	0.8
4	☐	6250.000	6542.381	15571277.27	84.2508	A	0.8
5	☐	12500.000	12986.197	30836203.98	167.2004	A	2.6
6	☐	25000.000	25941.773	60784408.54	333.9742	A	1.1
7	☐	50000.000	50686.796	119961205.9	652.5106	A	0.3
8	☐	250000.00	249734.96	562514476.0	3,214.807	A	0.7

$$y = 0.0129 * x + 0.0324$$


$$R = 1.0000$$

$$DL = 0.1131$$

$$BEC = 2.516$$

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	137.78	0.0007	P	10.4
2	☐	50.000	54.915	3446.01	0.0184	P	2.5
3	☐	2500.000	2653.277	156656.71	0.8543	P	0.5
4	☐	6250.000	6481.522	385499.25	2.0858	P	0.7
5	☐	12500.000	12791.724	759108.12	4.1157	P	2.0
6	☐	25000.000	26686.958	1562671.56	8.5857	A	0.8
7	☐	50000.000	51971.331	3073859.16	16.7194	A	1.2
8	☐	250000.00	249415.13	14037437.98	80.2350	A	1.6

$$y = 3.2169\text{E-}004 * x + 7.4458\text{E-}004$$


R = 1.0000

DL = 0.722

BEC = 2.315

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	174.45	0.0009	P	11.1
2	☐	1.000	1.137	7277.47	0.0389	P	2.1
3	☐	50.000	52.741	322956.36	1.7612	P	0.8
4	☐	125.000	127.362	785770.42	4.2516	P	1.0
5	☐	250.000	257.217	1583511.76	8.5854	A	2.1
6	☐	500.000	515.545	3131815.65	17.2069	A	0.6
7	☐	1000.000	989.991	6074573.17	33.0413	A	0.8
8	☐			12716.88	0.0727	P	1.8

$$y = 0.0334 * x + 9.4281\text{E-}004$$

R = 0.9998


DL = 0.009399

BEC = 0.02825

Weight: <None>

Min Conc: 0

60 Ni [He] ISTD:45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	911.15	0.0049	P	3.1
2	☐	1.000	1.088	2815.86	0.0150	P	1.5
3	☐	50.000	52.412	90244.72	0.4921	P	0.8
4	☐	125.000	127.271	219562.39	1.1880	P	0.9
5	☐	250.000	250.554	430456.00	2.3340	P	2.4
6	☐	500.000	494.517	837554.47	4.6017	P	0.4
7	☐	1000.000	1002.199	1713577.08	9.3209	A	1.1
8	☐			6671.63	0.0381	P	3.2

$$y = 0.0093 * x + 0.0049$$

$$R = 1.0000$$


$$DL = 0.04913$$

$$BEC = 0.5304$$

Weight: <None>

Min Conc: 0

63 Cu [He] ISTD:45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	3408.22	0.0184	P	4.3
2	☐	2.000	2.254	14341.79	0.0766	P	1.1
3	☐	500.000	549.997	2607152.39	14.2175	A	1.0
4	☐	1250.000	1309.061	6249460.25	33.8139	A	1.1
5	☐	2500.000	2606.155	12411753.15	67.3005	A	2.8
6	☐	5000.000	5119.625	24058980.75	132.1897	A	0.8
7	☐	10000.000	9903.766	47009129.30	255.6999	A	0.8
8	☐			11176.73	0.0639	P	2.1

$$y = 0.0258 * x + 0.0184$$


$$R = 0.9998$$

$$DL = 0.09155$$

$$BEC = 0.7144$$

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	195.56	0.0011	P	19.7
2	☐	2.000	5.719	3142.60	0.0168	P	3.5
3	☐	500.000	529.429	267299.64	1.4576	P	0.9
4	☐	1250.000	1280.881	651513.04	3.5251	P	0.8
5	☐	2500.000	2534.639	1286339.64	6.9745	P	2.2
6	☐	5000.000	5189.570	2598898.36	14.2788	A	0.1
7	☐	10000.000	9891.223	5003111.38	27.2142	A	0.9
8	☐			2770.30	0.0158	P	5.2

$$y = 0.0028 * x + 0.0011$$


$$R = 0.9997$$

$$DL = 0.228$$

$$BEC = 0.3851$$

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	11.11	0.0000	P	86.6
2	☐	1.000	1.035	321.12	0.0002	P	6.4
3	☐	50.000	51.271	15063.70	0.0088	P	1.3
4	☐	125.000	125.336	36952.74	0.0215	P	0.1
5	☐	250.000	254.368	74181.62	0.0437	P	4.4
6	☐	500.000	510.113	146407.79	0.0877	P	1.7
7	☐	1000.000	993.746	289459.19	0.1708	P	1.3
8	☐			161.11	0.0001	P	11.4

$$y = 1.7185E-004 * x + 6.4940E-006$$

$$R = 0.9999$$


$$DL = 0.09818$$

$$BEC = 0.03779$$

Weight: <None>

Min Conc: 0

77 Se [He] ISTD:89 Y [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	1.11	0.0000	P	173.2
2	☐	5.000	4.860	18.89	0.0000	P	26.6
3	☐	50.000	53.081	191.11	0.0001	P	8.7
4	☐	125.000	130.536	470.01	0.0003	P	2.6
5	☐	250.000	264.367	938.93	0.0006	P	12.2
6	☐	500.000	513.088	1795.69	0.0011	P	0.6
7	☐	1000.000	989.019	3512.70	0.0021	P	2.4
8	☐			5.56	0.0000	P	91.3

$$y = 2.0943E-006 * x + 6.4547E-007$$

R = 0.9998


DL = 1.601

BEC = 0.3082

Weight: <None>

Min Conc: 0

78 Se [He] ISTD:89 Y [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	3.33	0.0000	P	99.6
2	☐	5.000	5.251	68.89	0.0000	P	19.6
3	☐	50.000	53.807	661.13	0.0004	P	6.4
4	☐	125.000	130.981	1610.11	0.0009	P	3.7
5	☐	250.000	261.912	3183.72	0.0019	P	1.4
6	☐	500.000	503.494	6016.91	0.0036	P	2.2
7	☐	1000.000	994.336	12056.36	0.0071	P	1.1
8	☐			15.56	0.0000	P	52.3

$$y = 7.1519E-006 * x + 1.9431E-006$$

R = 0.9999


DL = 0.8116

BEC = 0.2717

Weight: <None>


Min Conc: 0

81 Br [No Gas] No available calibration points



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			3632.73		P	4.5
2	☐			3846.13		P	1.9
3	☐			3836.12		P	1.4
4	☐			3917.26		P	1.7
5	☐			3969.49		P	2.4
6	☐			3911.70		P	1.8
7	☐			4038.40		P	3.3
8	☐			4264.02		P	2.2

82 Se [No Gas] ISTD: 89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	-35.56	0.0000	P	-108.
2	☐	5.000	5.690	1360.11	0.0001	P	6.7
3	☐	50.000	51.353	12673.84	0.0005	P	3.7
4	☐	125.000	125.574	30977.02	0.0011	P	0.3
5	☐	250.000	255.037	62439.44	0.0023	P	3.5
6	☐	500.000	500.836	122328.15	0.0046	P	1.9
7	☐	1000.000	998.180	240507.81	0.0091	P	1.5
8	☐			-31.12	0.0000	P	-86.4

$$y = 9.1072E-006 * x - 1.3157E-006$$

$$R = 1.0000$$


$$DL = 0.4707$$

$$BEC = -0.1445$$

Weight: <None>


Min Conc: 0

83 Kr [No Gas] No available calibration points



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			285.56		P	17.3
2	☐			291.12		P	15.6
3	☐			262.23		P	15.9
4	☐			277.78		P	5.9
5	☐			264.45		P	9.5
6	☐			291.12		P	0.7
7	☐			317.79		P	4.2
8	☐			357.79		P	6.9

86 Sr [No Gas] ISTD : 89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	781.14	0.0000	P	6.5
2	☐	1.000	1.039	6325.94	0.0002	P	2.4
3	☐	50.000	50.116	270837.62	0.0100	P	1.2
4	☐	125.000	123.384	664050.36	0.0245	P	1.2
5	☐	250.000	245.451	1309660.60	0.0487	P	2.7
6	☐	500.000	504.909	2685975.69	0.1001	A	1.5
7	☐	1000.000	998.879	5240817.01	0.1980	A	0.2
8	☐			15965.88	0.0006	P	1.7

$$y = 1.9824E-004 * x + 2.8879E-005$$

$$R = 1.0000$$


$$DL = 0.02836$$

$$BEC = 0.1457$$

Weight: <None>

Min Conc: 0

88 Sr [No Gas] ISTD: 89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	2825.87	0.0001	P	5.6
2	☐	1.000	1.051	51967.28	0.0019	P	0.6
3	☐	50.000	51.020	2410345.68	0.0887	A	1.4
4	☐	125.000	125.158	5894713.10	0.2174	A	1.2
5	☐	250.000	250.752	11713536.63	0.4354	A	2.0
6	☐	500.000	498.039	23198508.54	0.8647	A	1.0
7	☐	1000.000	1000.722	45973874.87	1.7374	A	0.6
8	☐			136274.86	0.0055	P	0.8

$$y = 0.0017 * x + 1.0448E-004$$

R = 1.0000


DL = 0.01018

BEC = 0.06018

Weight: <None>

Min Conc: 0

90 Zr [No Gas] ISTD: 89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	913.37	0.0000	P	0.9
2	☐	1.000	1.028	30819.64	0.0011	P	0.7
3	☐	50.000	51.000	1497319.02	0.0551	A	2.0
4	☐	125.000	125.544	3675817.06	0.1356	A	1.3
5	☐	250.000	250.667	7280292.88	0.2706	A	2.2
6	☐	500.000	506.355	14665439.92	0.5466	A	1.7
7	☐	1000.000	996.538	28467033.47	1.0758	A	0.8
8	☐			46238.69	0.0019	P	0.2

$$y = 0.0011 * x + 3.3766E-005$$

R = 1.0000


DL = 0.0008525

BEC = 0.03128

Weight: <None>

Min Conc: 0

91 Zr [No Gas] ISTD : 89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	174.45	0.0000	P	10.5
2	☐	1.000	1.011	6780.59	0.0003	P	2.2
3	☐	50.000	49.304	325406.46	0.0120	P	2.6
4	☐	125.000	121.120	797307.97	0.0294	P	1.0
5	☐	250.000	250.640	1636637.43	0.0608	A	2.2
6	☐	500.000	504.843	3287332.35	0.1225	A	2.3
7	☐	1000.000	997.938	6409197.75	0.2422	A	1.4
8	☐			10289.43	0.0004	P	0.9

$$y = 2.4270E-004 * x + 6.4489E-006$$

R = 1.0000


DL = 0.008347

BEC = 0.02657

Weight: <None>

Min Conc: 0

94 Mo [No Gas] ISTD : 89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	391.12	0.0000	P	11.5
2	☐	5.000	6.230	36412.05	0.0014	P	1.0
3	☐	500.000	504.362	2942752.00	0.1083	A	2.7
4	☐	1250.000	1234.020	7183080.38	0.2649	A	0.1
5	☐	2500.000	2481.328	14327663.40	0.5326	A	2.3
6	☐	5000.000	5027.630	28951671.24	1.0791	A	0.8
7	☐	10000.000	9992.632	56753204.71	2.1448	A	1.4
8	☐			40232.46	0.0016	P	0.9

$$y = 2.1464E-004 * x + 1.4461E-005$$

R = 1.0000


DL = 0.02315

BEC = 0.06737

Weight: <None>

Min Conc: 0

95 Mo [No Gas] ISTD:89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	152.23	0.0000	P	12.5
2	☐	5.000	5.167	43726.33	0.0016	P	0.6
3	☐	500.000	502.824	4278614.90	0.1574	A	1.4
4	☐	1250.000	1248.540	10597854.98	0.3908	A	1.0
5	☐	2500.000	2487.536	20947410.10	0.7786	A	2.1
6	☐	5000.000	5012.688	42095429.37	1.5691	A	1.1
7	☐	10000.000	9996.813	82800687.10	3.1292	A	1.4
8	☐			41343.55	0.0017	P	1.8

$$y = 3.1302\text{E-}004 * x + 5.6281\text{E-}006$$

$$R = 1.0000$$


$$DL = 0.006769$$

$$BEC = 0.01798$$

Weight: <None>

Min Conc: 0

96 Mo [No Gas] ISTD:89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	245.56	0.0000	P	6.5
2	☐	5.000	5.321	49197.85	0.0018	P	1.4
3	☐	500.000	506.172	4699074.41	0.1729	A	1.3
4	☐	1250.000	1246.948	11548227.88	0.4259	A	0.6
5	☐	2500.000	2496.006	22931215.21	0.8524	A	2.5
6	☐	5000.000	5008.971	45893952.65	1.7106	A	1.3
7	☐	10000.000	9996.586	90334759.76	3.4140	A	2.0
8	☐			47771.86	0.0019	P	1.7

$$y = 3.4151\text{E-}004 * x + 9.0788\text{E-}006$$

$$R = 1.0000$$


$$DL = 0.005151$$

$$BEC = 0.02658$$

Weight: <None>

Min Conc: 0

97 Mo [No Gas] ISTD:89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	122.22	0.0000	P	11.2
2	☐	5.000	5.281	27557.40	0.0010	P	0.6
3	☐	500.000	510.788	2677493.29	0.0985	A	1.7
4	☐	1250.000	1246.636	6519029.76	0.2404	A	1.1
5	☐	2500.000	2490.455	12919576.75	0.4803	A	2.4
6	☐	5000.000	4979.357	25761610.17	0.9602	A	0.8
7	☐	10000.000	10012.589	51090981.46	1.9308	A	1.7
8	☐			25497.94	0.0010	P	1.0

$$y = 1.9284E-004 * x + 4.5188E-006$$

R = 1.0000


DL = 0.007839

BEC = 0.02343

Weight: <None>

Min Conc: 0

98 Mo [No Gas] ISTD:89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	196.67	0.0000	P	9.6
2	☐	5.000	5.179	70761.38	0.0026	P	0.6
3	☐	500.000	502.440	6906871.50	0.2541	A	2.0
4	☐	1250.000	1247.750	17112852.80	0.6310	A	0.6
5	☐	2500.000	2500.990	34024028.66	1.2649	A	2.8
6	☐	5000.000	5027.166	68212822.32	2.5425	A	0.8
7	☐	10000.000	9986.329	133643052.4	5.0505	A	1.1
8	☐			65447.77	0.0026	P	0.6

$$y = 5.0574E-004 * x + 7.2715E-006$$

R = 1.0000


DL = 0.004143

BEC = 0.01438

Weight: <None>

Min Conc: 0

106 Cd [No Gas] ISTD : 159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	8269.21	0.0004	P	0.7
2	☐	1.000	0.475	8583.87	0.0004	P	4.0
3	☐	50.000	53.290	33574.09	0.0016	P	1.4
4	☐	125.000	130.198	69305.01	0.0033	P	1.1
5	☐	250.000	255.105	127219.42	0.0061	P	2.4
6	☐	500.000	503.967	244847.64	0.0116	P	0.5
7	☐	1000.000	995.927	467069.89	0.0225	P	0.6
8	☐			7897.89	0.0004	P	2.8

$$y = 2.2169\text{E-}005 * x + 4.0012\text{E-}004$$

R = 1.0000


DL = 0.361

BEC = 18.05

Weight: <None>

Min Conc: 0

107 Ag [No Gas] ISTD : 159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	232.23	0.0000	P	7.7
2	☐	1.000	1.052	21977.59	0.0011	P	1.9
3	☐	50.000	52.011	1091906.39	0.0514	P	1.4
4	☐	125.000	132.119	2755006.56	0.1306	A	0.4
5	☐	250.000	260.734	5416344.64	0.2578	A	2.1
6	☐	500.000	506.633	10598469.63	0.5009	A	1.0
7	☐	1000.000	993.010	20400324.14	0.9818	A	0.7
8	☐			3576.06	0.0002	P	2.7

$$y = 9.8873\text{E-}004 * x + 1.1244\text{E-}005$$

R = 0.9999


DL = 0.002643

BEC = 0.01137

Weight: <None>

Min Conc: 0

108 Cd [No Gas] ISTD : 159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	20.00	0.0000	P	43.1
2	☐	1.000	1.069	374.45	0.0000	P	5.8
3	☐	50.000	52.508	17692.57	0.0008	P	2.3
4	☐	125.000	131.307	43924.10	0.0021	P	0.7
5	☐	250.000	257.135	85667.25	0.0041	P	2.9
6	☐	500.000	504.141	169137.44	0.0080	P	0.8
7	☐	1000.000	995.232	327895.52	0.0158	P	0.6
8	☐			264.45	0.0000	P	22.2

$$y = 1.5855E-005 * x + 9.6494E-007$$

$$R = 0.9999$$


$$DL = 0.07876$$

$$BEC = 0.06086$$

Weight: <None>

Min Conc: 0

109 Ag [No Gas] ISTD : 159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	66.66	0.0000	P	17.9
2	☐	1.000	1.067	21104.05	0.0010	P	1.9
3	☐	50.000	52.147	1043710.47	0.0492	P	1.8
4	☐	125.000	132.450	2633330.61	0.1249	A	1.1
5	☐	250.000	258.777	5125797.43	0.2440	A	2.1
6	☐	500.000	505.377	10080838.12	0.4765	A	1.2
7	☐	1000.000	994.079	19473775.13	0.9372	A	0.8
8	☐			3310.43	0.0002	P	8.7

$$y = 9.4279E-004 * x + 3.2280E-006$$

$$R = 0.9999$$


$$DL = 0.001837$$

$$BEC = 0.003424$$

Weight: <None>

Min Conc: 0

111 Cd [No Gas] ISTD : 159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	5794.45	0.0003	P	0.8
2	☐	1.000	1.092	10342.96	0.0005	P	2.3
3	☐	50.000	52.063	222924.08	0.0105	P	1.6
4	☐	125.000	128.359	537297.32	0.0255	P	1.7
5	☐	250.000	252.065	1045505.93	0.0498	P	2.6
6	☐	500.000	507.835	2115317.97	0.1000	A	2.4
7	☐	1000.000	995.043	4064870.65	0.1956	A	0.1
8	☐			6083.44	0.0003	P	3.6

$$y = 1.9632E-004 * x + 2.8038E-004$$

R = 0.9999


DL = 0.03577

BEC = 1.428

Weight: <None>

Min Conc: 0

118 Sn [No Gas] ISTD : 159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	3199.33	0.0002	P	1.4
2	☐	5.000	5.406	75531.26	0.0036	P	1.0
3	☐	50.000	51.155	697930.70	0.0329	P	1.7
4	☐	125.000	128.995	1743462.38	0.0827	A	0.2
5	☐	250.000	255.483	3437226.72	0.1636	A	1.4
6	☐	500.000	499.030	6757281.01	0.3194	A	0.6
7	☐	1000.000	998.555	13275721.05	0.6389	A	0.7
8	☐			4106.22	0.0002	P	7.3

$$y = 6.3969E-004 * x + 1.5482E-004$$


R = 1.0000

DL = 0.01045

BEC = 0.242

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	52.22	0.0000	P	20.8
2	☐	2.000	2.170	36398.34	0.0017	P	1.1
3	☐	50.000	50.695	862187.81	0.0406	P	1.5
4	☐	125.000	128.994	2179274.22	0.1033	A	0.3
5	☐	250.000	254.133	4277443.27	0.2036	A	1.9
6	☐	500.000	503.651	8536571.33	0.4035	A	0.8
7	☐	1000.000	996.607	16589281.14	0.7984	A	0.0
8	☐			13604.64	0.0007	P	3.0

$$y = 8.0110E-004 * x + 2.5284E-006$$


$$R = 1.0000$$

$$DL = 0.001972$$

$$BEC = 0.003156$$

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	258.90	0.0000	P	3.4
2	☐	10.000	10.674	43544.16	0.0021	P	1.5
3	☐	250.000	252.123	1038393.20	0.0489	P	1.1
4	☐	625.000	650.192	2659638.67	0.1261	A	1.5
5	☐	1250.000	1270.549	5177905.86	0.2464	A	1.9
6	☐	2500.000	2516.167	10325885.82	0.4880	A	0.2
7	☐	5000.000	4983.523	20084869.70	0.9666	A	0.5
8	☐			7615.55	0.0004	P	2.4

$$y = 1.9396E-004 * x + 1.2526E-005$$

$$R = 1.0000$$


$$DL = 0.006518$$

$$BEC = 0.06458$$

Weight: <None>

Min Conc: 0

137 Ba [No Gas] ISTD :159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	338.90	0.0000	P	6.9
2	☐	10.000	10.756	75685.51	0.0036	P	1.4
3	☐	250.000	257.171	1829544.00	0.0862	A	1.1
4	☐	625.000	640.558	4525920.42	0.2146	A	2.3
5	☐	1250.000	1263.268	8892373.48	0.4233	A	2.6
6	☐	2500.000	2501.767	17734934.73	0.8382	A	0.6
7	☐	5000.000	4993.495	34762336.70	1.6731	A	1.2
8	☐			13198.77	0.0007	P	2.6

$$y = 3.3505E-004 * x + 1.6387E-005$$

$$R = 1.0000$$


$$DL = 0.01019$$

$$BEC = 0.04891$$


Weight: <None>

Min Conc: 0


150 Nd [No Gas] No available calibration points



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			10.00		P	57.7
2	☐			13.33		P	50.0
3	☐			92.22		P	32.4
4	☐			184.45		P	25.6
5	☐			342.23		P	7.2
6	☐			650.02		P	7.2
7	☐			1277.85		P	6.8
8	☐			373.35		P	4.1

150 Nd [He] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			1.11		P	173.2
2	☐			3.33		P	100.1
3	☐			3.33		P	100.1
4	☐			3.33		P	173.2
5	☐			12.22		P	56.8
6	☐			16.67		P	20.0
7	☐			41.11		P	32.8
8	☐			70.00		P	26.5

156 Dy [No Gas] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			136.67		P	21.7
2	☐			150.00		P	9.7
3	☐			203.34		P	16.6
4	☐			267.78		P	12.0
5	☐			314.45		P	3.4
6	☐			354.45		P	6.1
7	☐			524.46		P	5.8
8	☐			696.69		P	7.8

156 Dy [He] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			3.33		P	100.1
2	☐			12.22		P	31.5
3	☐			11.11		P	96.4
4	☐			8.89		P	94.4
5	☐			23.33		P	49.5
6	☐			27.78		P	18.3
7	☐			44.44		P	18.9
8	☐			196.67		P	10.6

160 Gd [No Gas] Noavailable calibration poin...

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			135.56		P	14.8
2	☐			131.12		P	8.2
3	☐			123.34		P	27.0
4	☐			145.56		P	16.6
5	☐			165.56		P	15.1
6	☐			203.34		P	12.8
7	☐			245.56		P	5.7
8	☐			626.69		P	4.5

160 Gd [He] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			25.56		P	49.4
2	☐			18.89		P	27.0
3	☐			16.67		P	20.0
4	☐			28.89		P	37.1
5	☐			37.78		P	35.7
6	☐			54.44		P	9.4
7	☐			70.00		P	40.7
8	☐			233.34		P	8.0

164 Er [No Gas] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			122.22		P	20.7
2	☐			86.67		P	24.0
3	☐			143.34		P	15.2
4	☐			140.00		P	8.6
5	☐			138.89		P	6.9
6	☐			182.22		P	11.2
7	☐			271.12		P	6.2
8	☐			733.36		P	6.1

164 Er [He] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			32.22		P	21.5
2	☐			26.67		P	37.5
3	☐			41.11		P	18.7
4	☐			46.67		P	28.6
5	☐			54.45		P	19.7
6	☐			55.55		P	12.5
7	☐			97.78		P	20.5
8	☐			225.56		P	14.6

172 Yb [No Gas] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			125.56		P	8.1
2	☐			128.89		P	9.1
3	☐			103.33		P	8.5
4	☐			144.45		P	25.5
5	☐			148.89		P	7.2
6	☐			230.01		P	12.9
7	☐			313.34		P	15.1
8	☐			862.26		P	7.6


172 Yb [He] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			34.44		P	47.7
2	☐			40.00		P	30.0
3	☐			47.78		P	14.5
4	☐			54.44		P	9.4
5	☐			75.55		P	22.2
6	☐			58.89		P	18.2
7	☐			127.78		P	6.6
8	☐			332.23		P	11.2

176 Yb [No Gas] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			2189.10		P	15.6
2	☐			2255.77		P	4.4
3	☐			6351.59		P	1.6
4	☐			12213.51		P	1.8
5	☐			22188.75		P	1.4
6	☐			41934.12		P	2.2
7	☐			81353.87		P	0.6
8	☐			2636.96		P	2.1

176 Yb [He] No available calibration points



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			372.23		P	8.5
2	☐			364.45		P	9.8
3	☐			1597.89		P	4.4
4	☐			3502.72		P	0.7
5	☐			6598.38		P	1.5
6	☐			13139.95		P	1.5
7	☐			25828.58		P	0.3
8	☐			527.79		P	6.0

203 Tl [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	378.90	0.0000	P	9.1
2	☐	1.000	1.022	18387.57	0.0014	P	2.3
3	☐	50.000	49.097	870915.41	0.0652	P	1.3
4	☐	125.000	125.576	2198868.73	0.1667	A	1.2
5	☐	250.000	253.695	4390439.31	0.3368	A	2.2
6	☐	500.000	505.342	8583471.26	0.6708	A	1.7
7	☐	1000.000	996.378	16498236.97	1.3225	A	0.8
8	☐			1481.21	0.0001	P	6.1

$$y = 0.0013 * x + 2.8773E-005$$

$$R = 1.0000$$


$$DL = 0.005924$$

$$BEC = 0.02168$$

Weight: <None>

Min Conc: 0

205 Tl [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	924.49	0.0001	P	6.8
2	☐	1.000	1.029	44425.90	0.0033	P	0.8
3	☐	50.000	50.454	2146047.26	0.1607	A	1.4
4	☐	125.000	125.413	5266100.86	0.3992	A	0.2
5	☐	250.000	246.593	10233455.05	0.7849	A	2.0
6	☐	500.000	504.412	20543266.50	1.6055	A	2.2
7	☐	1000.000	998.572	39649768.30	3.1782	A	0.7
8	☐			3630.55	0.0003	P	12.2

$$y = 0.0032 * x + 7.0190E-005$$

$$R = 1.0000$$


$$DL = 0.004501$$

$$BEC = 0.02205$$

Weight: <None>

Min Conc: 0

206 [Pb] [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	2149.09	0.0002	P	4.1
2	☐	1.000	0.992	16577.36	0.0012	P	1.5
3	☐	250.000	256.478	3751403.03	0.2808	A	1.3
4	☐	625.000	619.413	8942510.35	0.6780	A	1.1
5	☐	1250.000	1241.290	17711257.79	1.3585	A	2.0
6	☐	2500.000	2531.068	35444687.25	2.7698	A	1.3
7	☐	5000.000	4987.018	68080247.88	5.4573	A	0.6
8	☐			14608.32	0.0013	P	2.0

$$y = 0.0011 * x + 1.6316E-004$$

$$R = 1.0000$$


$$DL = 0.01856$$

$$BEC = 0.1491$$

Weight: <None>

Min Conc: 0

207 [Pb] [No Gas] ISTD : 209 Bi [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	1734.58	0.0001	P	4.7
2	☐	1.000	0.989	14113.35	0.0011	P	0.5
3	☐	250.000	253.715	3193850.86	0.2391	A	1.1
4	☐	625.000	622.848	7738943.15	0.5867	A	1.2
5	☐	1250.000	1229.328	15095505.61	1.1579	A	2.4
6	☐	2500.000	2514.433	30306275.94	2.3682	A	0.9
7	☐	5000.000	4998.034	58722289.68	4.7072	A	0.8
8	☐			12411.59	0.0011	P	4.1

$$y = 9.4178E-004 * x + 1.3168E-004$$

R = 1.0000


DL = 0.01967

BEC = 0.1398

Weight: <None>

Min Conc: 0

208 Pb [No Gas] ISTD : 209 Bi [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	8146.64	0.0006	P	2.5
2	☐	1.000	0.998	65729.19	0.0050	P	0.4
3	☐	250.000	252.313	14643378.98	1.0962	A	1.0
4	☐	625.000	623.682	35727693.22	2.7086	A	0.9
5	☐	1250.000	1233.854	69852444.24	5.3580	A	2.4
6	☐	2500.000	2515.575	139786392.0	10.9232	A	0.9
7	☐	5000.000	4996.298	270640306.5	21.6944	A	0.6
8	☐			57768.68	0.0053	P	3.9

$$y = 0.0043 * x + 6.1846E-004$$

R = 1.0000


DL = 0.01061

BEC = 0.1424

Weight: <None>

Min Conc: 0

238 U [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	164.45	0.0000	P	24.8
2	☐	1.000	0.965	56322.17	0.0042	P	0.7
3	☐	50.000	49.554	2901633.08	0.2172	A	0.7
4	☐	125.000	122.784	7098609.54	0.5382	A	1.2
5	☐	250.000	242.499	13858141.04	1.0629	A	2.0
6	☐	500.000	495.160	27770835.42	2.1703	A	2.0
7	☐	1000.000	1004.594	54930364.18	4.4031	A	0.7
8	☐			2419.15	0.0002	P	9.0

$$y = 0.0044 * x + 1.2491E-005$$

$$R = 1.0000$$


$$DL = 0.002124$$

$$BEC = 0.00285$$


Weight: <None>

Min Conc: 0


6 Li (ISTD) [No Gas]




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		7423441.25		A	1.3
2	☐	1.000		7477615.36		A	0.8
3	☐	1.000		7413901.76		A	2.1
4	☐	1.000		7417953.29		A	1.7
5	☐	1.000		7331348.65		A	2.1
6	☐	1.000		7234025.45		A	1.0
7	☐	1.000		7122742.22		A	1.2
8	☐	1.000		6504285.31		A	1.5

45 Sc (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		10464420.81		A	0.7
2	☐	1.000		10402583.87		A	0.4
3	☐	1.000		10570873.94		A	1.1
4	☐	1.000		10458657.14		A	0.4
5	☐	1.000		10482539.64		A	1.2
6	☐	1.000		10267841.31		A	2.0
7	☐	1.000		10226560.54		A	1.4
8	☐	1.000		9618838.75		A	1.5

45 Sc (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		184856.14		P	1.5
2	☐	1.000		187166.28		P	0.8
3	☐	1.000		183381.76		P	0.5
4	☐	1.000		184827.42		P	0.7
5	☐	1.000		184491.19		P	2.0
6	☐	1.000		182010.14		P	0.7
7	☐	1.000		183845.21		P	0.3
8	☐	1.000		174968.61		P	1.1

89 Y (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		27049771.54		A	0.2
2	☐	1.000		26941753.77		A	0.9
3	☐	1.000		27185464.87		A	1.4
4	☐	1.000		27118289.32		A	0.8
5	☐	1.000		26909883.49		A	2.1
6	☐	1.000		26830286.27		A	1.1
7	☐	1.000		26462815.44		A	0.9
8	☐	1.000		24838338.52		A	1.1

89 Y (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		1714175.79		A	0.8
2	☐	1.000		1742951.64		A	1.3
3	☐	1.000		1708394.14		A	1.4
4	☐	1.000		1715072.35		A	0.6
5	☐	1.000		1698412.36		A	3.4
6	☐	1.000		1670128.03		A	0.9
7	☐	1.000		1695056.34		A	1.4
8	☐	1.000		1624114.92		A	2.4

103 Rh (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		14734637.00		A	1.2
2	☐	1.000		14971739.50		A	0.3
3	☐	1.000		14808382.14		A	0.8
4	☐	1.000		14683971.45		A	1.1
5	☐	1.000		14422062.98		A	2.0
6	☐	1.000		14146720.90		A	2.0
7	☐	1.000		13695039.24		A	0.9
8	☐	1.000		12196135.93		A	1.7

103 Rh (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		5414031.86		A	0.9
2	☐	1.000		5440344.36		A	0.7
3	☐	1.000		5350975.03		A	0.6
4	☐	1.000		5313886.52		A	0.6
5	☐	1.000		5250853.19		A	2.2
6	☐	1.000		5075515.90		A	1.1
7	☐	1.000		4980942.39		A	1.3
8	☐	1.000		4445461.63		A	1.1

115 In (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		16037263.07		A	0.3
2	☐	1.000		16147299.06		A	0.7
3	☐	1.000		16138883.73		A	0.9
4	☐	1.000		15937920.51		A	0.7
5	☐	1.000		15889066.42		A	2.1
6	☐	1.000		15514528.60		A	1.8
7	☐	1.000		15001307.48		A	0.6
8	☐	1.000		13929933.79		A	1.6

115 In (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		1386412.82		P	0.3
2	☐	1.000		1398261.54		P	0.7
3	☐	1.000		1375348.97		P	0.6
4	☐	1.000		1365568.58		P	0.2
5	☐	1.000		1349042.93		P	2.7
6	☐	1.000		1306108.62		P	0.3
7	☐	1.000		1276745.53		P	0.4
8	☐	1.000		1227208.67		P	1.8

159 Tb (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		20667778.17		A	1.4
2	☐	1.000		20907060.52		A	0.7
3	☐	1.000		21230676.63		A	1.1
4	☐	1.000		21088796.21		A	0.8
5	☐	1.000		21014332.47		A	1.6
6	☐	1.000		21157596.76		A	0.1
7	☐	1.000		20778610.11		A	0.7
8	☐	1.000		19114230.83		A	0.1

159 Tb (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		6324544.28		A	0.3
2	☐	1.000		6272043.17		A	0.5
3	☐	1.000		6237170.18		A	0.4
4	☐	1.000		6241130.19		A	1.0
5	☐	1.000		6254895.67		A	2.9
6	☐	1.000		6191560.88		A	1.3
7	☐	1.000		6151420.25		A	1.3
8	☐	1.000		5858378.73		A	1.8

165 Ho (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		20556050.94		A	0.9
2	☐	1.000		20696992.19		A	0.3
3	☐	1.000		20675908.02		A	0.6
4	☐	1.000		20629429.42		A	0.3
5	☐	1.000		20735622.33		A	2.2
6	☐	1.000		20600766.22		A	0.7
7	☐	1.000		20305066.09		A	0.9
8	☐	1.000		19042269.30		A	1.0

165 Ho (ISTD) [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		6514911.64		A	1.0
2	☐	1.000		6552088.93		A	0.9
3	☐	1.000		6532102.61		A	0.7
4	☐	1.000		6576114.35		A	0.9
5	☐	1.000		6475222.33		A	2.1
6	☐	1.000		6457489.56		A	0.4
7	☐	1.000		6449639.21		A	0.6
8	☐	1.000		6103709.08		A	0.4

209 Bi (ISTD) [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		13173314.94		A	0.7
2	☐	1.000		13274347.72		A	0.5
3	☐	1.000		13358852.16		A	0.3
4	☐	1.000		13190892.16		A	0.7
5	☐	1.000		13041023.84		A	1.9
6	☐	1.000		12798423.28		A	1.5
7	☐	1.000		12475585.37		A	0.9
8	☐	1.000		10847304.42		A	0.4

209 Bi (ISTD) [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		4950087.91		A	0.8
2	☐	1.000		4968419.44		A	1.2
3	☐	1.000		4901413.89		A	0.1
4	☐	1.000		4844027.50		A	0.5
5	☐	1.000		4848751.70		A	2.1
6	☐	1.000		4674437.26		A	1.0
7	☐	1.000		4658149.07		A	0.4
8	☐	1.000		4090868.55		A	1.0

US EPA Tune Check Report

Operator Name Jaswal
Acq/Data Batch D:\Agilent\ICPMH\1\DATA\IP8021125 MS.b
Acq. Date-Time 2025-02-11 11:22:52
Report Comment ---
Instrument Name G8403A SG19224459

[No Gas]

Sensitivity

Mass	Conc. [ug/l]	Count	CPS	Resp (Required) [cps/ug/l]	Resp (Flag)	RSD%	RSD% (Required)
9		33681	336808.90			0.576	5.000
24		317868	3178681.60			0.536	5.000
25		42868	428682.00			0.422	5.000
26		47889	478886.16			0.369	5.000
59		185789	1857886.03			0.885	5.000
113		22196	221964.29			0.425	5.000
115		292044	2920437.87			0.944	5.000
206		71039	710393.19			0.629	5.000
207		61658	616576.70			0.147	5.000
208		151245	1512454.00			1.139	5.000
220		1	5.60			27.808	

Mass	RSD% (Flag)
9	
24	
25	
26	
59	
113	
115	
206	
207	
208	
220	


Mass	Rep#1 Count	Rep#2 Count	Rep#3 Count	Rep#4 Count	Rep#5 Count
9	33985	33739	33573	33481	33626
24	319455	319239	318551	315819	316277
25	43132	42889	42923	42709	42689
26	48036	47949	48043	47636	47779
59	186794	183813	187922	184624	185790
113	22279	22315	22152	22100	22137
115	296208	290284	292469	292333	288926
206	71568	71184	71292	70628	70524
207	61722	61735	61715	61554	61563
208	152775	151737	150449	148645	152622

US EPA Tune Check Report

Mass	Rep#1 Count	Rep#2 Count	Rep#3 Count	Rep#4 Count	Rep#5 Count
220	1	1	0	1	0

Integration Time [sec] 0.1

Resolution/Axis



Mass	Peak Height	Axis	Axis (Required)	Axis (Flag)
9	56451.76	9.10	8.90 - 9.10	
24	540315.86	24.00	23.90 - 24.10	
25	72844.30	25.00	24.90 - 25.10	
26	82942.97	26.00	25.90 - 26.10	
59	334772.90	59.00	58.90 - 59.10	
113	43756.33	113.05	112.90 - 113.10	
115	573082.19	115.05	114.90 - 115.10	
206	141412.16	206.00	205.90 - 206.10	
207	122171.00	206.95	206.90 - 207.10	
208	302103.53	207.95	207.90 - 208.10	
220	0.80	220.50	-	

Mass	W-50%	W-5%	W-5% (Required)	W-5% (Flag)
9	0.63	0.777	0.900	
24	0.63	0.787	0.900	
25	0.63	0.792	0.900	
26	0.62	0.788	0.900	
59	0.59	0.779	0.900	
113	0.52	0.733	0.900	
115	0.53	0.734	0.900	
206	0.50	0.778	0.900	
207	0.50	0.784	0.900	
208	0.50	0.791	0.900	
220	0.27	0.388		

Integration Time [sec] 0.1

Acquisition Time [sec] 256.770000000002

Y Axis Linear

Tune Parameters

Plasma Parameters

Plasma Mode --- Nebulizer Gas 0.72 L/min Dilution Gas 0.40 L/min

US EPA Tune Check Report

RF Power	1600 W	Option Gas	---	Auxiliary Gas	0.90 L/min
RF Matching	1.80 V	Nebulizer Pump	0.10 rps	Plasma Gas	15.0 L/min
Sample Depth	9.0 mm	S/C Temp	2 °C		

Lens Parameters

Extract 1	-6.0 V	Omega Lens	8.6 V	Deflect	14.2 V
Extract 2	-210.0 V	Cell Entrance	-30 V	Plate Bias	-50 V
Omega Bias	-95 V	Cell Exit	-50 V		

Cell Parameters

Use Gas	No	3rd Gas Flow	---	Energy Discrimination	4.0 V
He Flow	0.0 mL/min	OctP Bias	-8.0 V		
H2 Flow	---	OctP RF	170 V		

QP Parameters

Mass Gain	103	Axis Gain	0.9959	QP Bias	-4.0 V
Mass Offset	125	Axis Offset	0.05		

Hardware Settings

Torch

Torch H	-0.2 mm	Torch V	1.5 mm
---------	---------	---------	--------

EM

Discriminator	4.0 mV	Analog HV	2179 V	Pulse HV	1089 V
---------------	--------	-----------	--------	----------	--------

[He]

Sensitivity

Mass	Conc. [ug/l]	Count	CPS	Resp (Required) [cps/ug/l]	Resp (Flag)	RSD%	RSD% (Required)
59		32890	328904.85			0.370	
89		31748	317475.54			0.361	
205		36794	367944.69			0.422	

Mass	RSD% (Flag)
59	
89	
205	

Mass	Rep#1 Count	Rep#2 Count	Rep#3 Count	Rep#4 Count	Rep#5 Count
59	32961	32954	32822	33005	32711
89	31847	31724	31819	31788	31559
205	36762	36981	36899	36755	36575

Integration Time [sec] 0.1

Tune Parameters

Plasma Parameters

Plasma Mode	---	Nebulizer Gas	0.72 L/min	Dilution Gas	0.40 L/min
RF Power	1600 W	Option Gas	---	Auxiliary Gas	0.90 L/min
RF Matching	1.80 V	Nebulizer Pump	0.10 rps	Plasma Gas	15.0 L/min
Sample Depth	9.0 mm	S/C Temp	2 °C		

Lens Parameters

Extract 1	-6.0 V	Omega Lens	8.6 V	Deflect	3.4 V
-----------	--------	------------	-------	---------	-------

US EPA Tune Check Report

Extract 2	-210.0 V	Cell Entrance	-50 V	Plate Bias	-60 V
Omega Bias	-95 V	Cell Exit	-70 V		
Cell Parameters					
Use Gas	Yes	3rd Gas Flow	---	Energy Discrimination	5.0 V
He Flow	4.5 mL/min	OctP Bias	-18.0 V		
H2 Flow	---	OctP RF	200 V		
QP Parameters					
Mass Gain	103	Axis Gain	0.9959	QP Bias	-13.0 V
Mass Offset	125	Axis Offset	0.05		
Hardware Settings					
Torch					
Torch H	-0.2 mm	Torch V	1.5 mm		
EM					
Discriminator	4.0 mV	Analog HV	2179 V	Pulse HV	1089 V

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S0 Instrumnet Name : P8
Client Sample ID : S0 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:05:13 DataFile Name : 004CALB.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	-0.14	0.23	-0.08	0.00	N/A	ppb
Antimony	121-1	0.00	0.00	0.00	0.00	N/A	ppb
Arsenic	75-2	0.02	-0.04	0.02	0.00	N/A	ppb
Barium	135-1	0.00	0.00	0.00	0.00	N/A	ppb
Barium	137-1	0.00	0.00	0.00	0.00	N/A	ppb
Beryllium	9-1	0.00	0.00	0.00	0.00	N/A	ppb
Bismuth	209-1				100		%
Bismuth	209-2				100		%
Bromine	81-1						cps
Cadmium	108-1	-0.02	-0.01	0.03	0.00	N/A	ppb
Cadmium	106-1	0.12	-0.12	0.00	0.00	N/A	ppb
Cadmium	111-1	0.01	-0.01	0.00	0.00	N/A	ppb
Calcium	43-1	0.00	-0.68	0.67	0.00	N/A	ppb
Calcium	44-1	0.40	-0.35	-0.06	0.00	N/A	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	-0.01	-0.01	0.02	0.00	N/A	ppb
Cobalt	59-2	0.00	0.00	0.00	0.00	N/A	ppb
Copper	63-2	0.03	0.00	-0.03	0.00	N/A	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				100		%
Holmium	165-2				100		%
Indium	115-1				100		%
Indium	115-2				100		%
Iron	54-2	0.18	0.30	-0.48	0.00	N/A	ppb
Iron	56-2	-0.04	0.03	0.02	0.00	N/A	ppb
Iron	57-2	-0.11	-0.17	0.28	0.00	N/A	ppb
Krypton	83-1						cps
Lead	206-1	0.00	0.00	0.01	0.00	N/A	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S0 Instrumnet Name : P8
Client Sample ID : S0 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:05:13 DataFile Name : 004CALB.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	0.00	0.01	0.00	0.00	N/A	ppb
Lead	208-1	0.00	0.00	0.00	0.00	N/A	ppb
Lithium	6-1				100		%
Magnesium	24-2	0.11	-0.84	0.73	0.00	N/A	ppb
Manganese	55-2	-0.02	0.00	0.02	0.00	N/A	ppb
Molybdenum	94-1	0.01	0.00	-0.01	0.00	N/A	ppb
Molybdenum	95-1	0.00	0.00	0.00	0.00	N/A	ppb
Molybdenum	96-1	0.00	0.00	0.00	0.00	N/A	ppb
Molybdenum	97-1	0.00	0.00	0.00	0.00	N/A	ppb
Molybdenum	98-1	0.00	0.00	0.00	0.00	N/A	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.00	0.02	-0.02	0.00	N/A	ppb
Phosphorus	31-2	-17.01	-7.69	-9.67	-11.46		ppb
Potassium	39-2	1.78	-0.99	-0.80	0.00	N/A	ppb
Rhodium	103-1				100		%
Rhodium	103-2				100		%
Scandium	45-1				100		%
Scandium	45-2				100		%
Selenium	82-1	-0.17	0.02	0.14	0.00	N/A	ppb
Selenium	77-2	-0.31	0.62	-0.31	0.00	N/A	ppb
Selenium	78-2	0.00	-0.27	0.27	0.00	N/A	ppb
Silicon	28-1	-0.05	-0.21	0.26	0.00	N/A	ppb
Silver	107-1	0.00	0.00	0.00	0.00	N/A	ppb
Silver	109-1	0.00	0.00	0.00	0.00	N/A	ppb
Sodium	23-2	1.79	0.33	-2.12	0.00	N/A	ppb
Strontium	86-1	-0.01	0.01	0.00	0.00	N/A	ppb
Strontium	88-1	0.00	0.00	0.00	0.00	N/A	ppb
Sulfur	34-1	-105.34	-141.26	-119.61	-122.07		ppb
Terbium	159-1				100		%
Terbium	159-2				100		%
Thallium	203-1	0.00	0.00	0.00	0.00	N/A	ppb
Thallium	205-1	0.00	0.00	0.00	0.00	N/A	ppb
Tin	118-1	0.00	0.00	0.00	0.00	N/A	ppb
Titanium	47-1	0.00	-0.04	0.04	0.00	N/A	ppb
Uranium	238-1	0.00	0.00	0.00	0.00	N/A	ppb

3

LB Number :	LB134674	Operator :	Jaswal
Lab Sample ID :	S02	Instrumnet Name :	P8
Client Sample ID :	S02	Dilution Factor :	1
Date & Time Acquired :	2025-02-11 12:08:35	DataFile Name :	005CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	18.64	19.93	18.09	18.89	5.00	ppb
Antimony	121-1	2.17	2.14	2.19	2.17	1.14	ppb
Arsenic	75-2	0.98	1.11	1.02	1.03	6.60	ppb
Barium	135-1	10.54	10.63	10.85	10.67	1.48	ppb
Barium	137-1	10.83	10.59	10.85	10.76	1.37	ppb
Beryllium	9-1	1.09	1.06	1.07	1.07	1.22	ppb
Bismuth	209-1				101		%
Bismuth	209-2				100		%
Bromine	81-1						cps
Cadmium	108-1	1.01	1.14	1.06	1.07	6.11	ppb
Cadmium	106-1	0.96	-0.37	0.84	0.47	154.54	ppb
Cadmium	111-1	1.12	1.02	1.13	1.09	5.37	ppb
Calcium	43-1	551.26	561.62	558.94	557.27	0.96	ppb
Calcium	44-1	546.24	556.45	548.15	550.28	0.99	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2.08	2.10	2.04	2.07	1.49	ppb
Cobalt	59-2	1.12	1.12	1.16	1.14	2.10	ppb
Copper	63-2	2.23	2.24	2.29	2.25	1.45	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				101		%
Holmium	165-2				101		%
Indium	115-1				101		%
Indium	115-2				101		%
Iron	54-2	56.02	57.30	56.90	56.74	1.16	ppb
Iron	56-2	55.85	55.21	55.19	55.42	0.68	ppb
Iron	57-2	55.31	56.12	53.32	54.91	2.62	ppb
Krypton	83-1						cps
Lead	206-1	0.97	1.00	1.00	0.99	1.70	ppb

LB Number :
Lab Sample ID :
Client Sample ID :
Date & Time Acquired :

LB134674
S02
S02
2025-02-11 12:08:35

Operator :
Instrumnet Name :
Dilution Factor :
DataFile Name :

Jaswal
P8
1
005CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	0.98	0.99	0.99	0.99	0.55	ppb
Lead	208-1	0.99	1.00	1.00	1.00	0.42	ppb
Lithium	6-1				101		%
Magnesium	24-2	534.80	533.80	538.70	535.77	0.48	ppb
Manganese	55-2	1.07	1.00	1.05	1.04	3.70	ppb
Molybdenum	94-1	6.24	6.28	6.16	6.23	0.96	ppb
Molybdenum	95-1	5.20	5.16	5.14	5.17	0.64	ppb
Molybdenum	96-1	5.31	5.40	5.25	5.32	1.39	ppb
Molybdenum	97-1	5.31	5.25	5.27	5.28	0.59	ppb
Molybdenum	98-1	5.16	5.21	5.16	5.18	0.57	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	1.06	1.10	1.10	1.09	2.20	ppb
Phosphorus	31-2	5.65	6.05	22.68	11.46	84.83	ppb
Potassium	39-2	515.69	505.55	514.61	511.95	1.09	ppb
Rhodium	103-1				102		%
Rhodium	103-2				100		%
Scandium	45-1				99		%
Scandium	45-2				101		%
Selenium	82-1	5.95	5.86	5.26	5.69	6.55	ppb
Selenium	77-2	6.10	3.39	5.09	4.86	28.26	ppb
Selenium	78-2	4.02	5.68	6.05	5.25	20.61	ppb
Silicon	28-1	6.57	6.39	6.16	6.38	3.21	ppb
Silver	107-1	1.03	1.05	1.07	1.05	1.92	ppb
Silver	109-1	1.06	1.05	1.09	1.07	1.91	ppb
Sodium	23-2	545.49	534.97	545.32	541.93	1.11	ppb
Strontium	86-1	1.07	1.03	1.02	1.04	2.69	ppb
Strontium	88-1	1.06	1.05	1.05	1.05	0.59	ppb
Sulfur	34-1	110.19	141.26	114.77	122.07	13.74	ppb
Terbium	159-1				101		%
Terbium	159-2				99		%
Thallium	203-1	1.02	1.04	1.00	1.02	2.36	ppb
Thallium	205-1	1.02	1.04	1.03	1.03	0.78	ppb
Tin	118-1	5.37	5.38	5.47	5.41	1.03	ppb
Titanium	47-1	5.32	5.23	5.32	5.29	0.91	ppb
Uranium	238-1	0.96	0.96	0.97	0.97	0.71	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S02 Instrumnet Name : P8
Client Sample ID : S02 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:08:35 DataFile Name : 005CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Vanadium	51-2	5.34	5.20	5.32	5.29	1.41	ppb
Ytterbium	172-1						cps
Ytterbium	172-2						cps
Ytterbium	176-1						cps
Ytterbium	176-2						cps
Yttrium	89-1				100		%
Yttrium	89-2				102		%
Zinc	66-2	5.79	5.48	5.89	5.72	3.76	ppb
Zirconium	90-1	1.03	1.02	1.03	1.03	0.70	ppb
Zirconium	91-1	1.02	1.02	0.98	1.01	2.24	ppb

LB Number :	LB134674	Operator :	Jaswal
Lab Sample ID :	S03	Instrumnet Name :	P8
Client Sample ID :	S03	Dilution Factor :	1
Date & Time Acquired :	2025-02-11 12:11:58	DataFile Name :	006CALB.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	1000.11	1004.42	1004.44	1002.99	0.25	ppb
Antimony	121-1	50.05	50.52	51.51	50.70	1.47	ppb
Arsenic	75-2	50.55	51.41	51.85	51.27	1.28	ppb
Barium	135-1	249.33	251.94	255.10	252.12	1.15	ppb
Barium	137-1	253.97	258.84	258.70	257.17	1.08	ppb
Beryllium	9-1	50.34	52.78	51.01	51.38	2.46	ppb
Bismuth	209-1				101		%
Bismuth	209-2				99		%
Bromine	81-1						cps
Cadmium	108-1	51.10	53.11	53.32	52.51	2.34	ppb
Cadmium	106-1	52.75	52.71	54.41	53.29	1.81	ppb
Cadmium	111-1	51.10	52.31	52.78	52.06	1.67	ppb
Calcium	43-1	5157.25	5231.17	5324.44	5237.62	1.60	ppb
Calcium	44-1	5175.53	5200.27	5337.31	5237.70	1.66	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	50.15	50.53	50.90	50.53	0.74	ppb
Cobalt	59-2	52.32	53.14	52.76	52.74	0.78	ppb
Copper	63-2	543.79	553.81	552.39	550.00	0.99	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				101		%
Holmium	165-2				100		%
Indium	115-1				101		%
Indium	115-2				99		%
Iron	54-2	2686.28	2687.32	2715.83	2696.48	0.62	ppb
Iron	56-2	2668.67	2710.60	2683.97	2687.74	0.79	ppb
Iron	57-2	2640.37	2665.37	2654.10	2653.28	0.47	ppb
Krypton	83-1						cps
Lead	206-1	252.95	259.35	257.13	256.48	1.27	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S03 Instrumnet Name : P8
Client Sample ID : S03 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:11:58 DataFile Name : 006CALB.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	255.35	255.37	250.43	253.71	1.12	ppb
Lead	208-1	249.78	254.64	252.51	252.31	0.96	ppb
Lithium	6-1				100		%
Magnesium	24-2	5263.02	5363.97	5411.07	5346.02	1.41	ppb
Manganese	55-2	511.92	514.75	516.49	514.38	0.45	ppb
Molybdenum	94-1	490.82	518.46	503.81	504.36	2.74	ppb
Molybdenum	95-1	497.38	510.59	500.51	502.82	1.37	ppb
Molybdenum	96-1	498.69	508.80	511.03	506.17	1.30	ppb
Molybdenum	97-1	500.99	517.68	513.70	510.79	1.71	ppb
Molybdenum	98-1	490.85	507.33	509.14	502.44	2.01	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	51.99	52.78	52.47	52.41	0.76	ppb
Phosphorus	31-2	924.57	1041.83	982.10	982.83	5.97	ppb
Potassium	39-2	2492.98	2521.59	2533.38	2515.99	0.83	ppb
Rhodium	103-1				101		%
Rhodium	103-2				99		%
Scandium	45-1				101		%
Scandium	45-2				99		%
Selenium	82-1	49.15	52.40	52.51	51.35	3.72	ppb
Selenium	77-2	48.28	57.51	53.46	53.08	8.71	ppb
Selenium	78-2	52.26	57.78	51.38	53.81	6.44	ppb
Silicon	28-1	50.54	50.81	51.89	51.08	1.40	ppb
Silver	107-1	51.25	52.03	52.75	52.01	1.44	ppb
Silver	109-1	51.07	52.72	52.65	52.15	1.79	ppb
Sodium	23-2	5285.78	5437.97	5462.77	5395.51	1.78	ppb
Strontium	86-1	49.44	50.44	50.47	50.12	1.17	ppb
Strontium	88-1	50.19	51.44	51.42	51.02	1.40	ppb
Sulfur	34-1	1170.00	1116.80	1271.62	1186.14	6.63	ppb
Terbium	159-1				103		%
Terbium	159-2				99		%
Thallium	203-1	48.36	49.42	49.51	49.10	1.31	ppb
Thallium	205-1	50.09	50.02	51.25	50.45	1.37	ppb
Tin	118-1	50.13	51.51	51.82	51.15	1.76	ppb
Titanium	47-1	505.16	509.07	523.08	512.44	1.84	ppb
Uranium	238-1	49.90	49.24	49.53	49.55	0.66	ppb

LB Number :

LB134674

Operator :

Jaswal

Lab Sample ID :

S03

Instrumnet Name :

P8

Client Sample ID :

S03

Dilution Factor :

1

Date & Time Acquired :

2025-02-11 12:11:58

DataFile Name :

006CALB.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Vanadium	51-2	50.89	51.04	50.36	50.77	0.70	ppb
Ytterbium	172-1						cps
Ytterbium	172-2						cps
Ytterbium	176-1						cps
Ytterbium	176-2						cps
Yttrium	89-1				101		%
Yttrium	89-2				100		%
Zinc	66-2	525.46	528.11	534.71	529.43	0.90	ppb
Zirconium	90-1	49.85	51.84	51.31	51.00	2.02	ppb
Zirconium	91-1	47.83	50.25	49.83	49.30	2.63	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S04 Instrumnet Name : P8
Client Sample ID : S04 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:15:02 DataFile Name : 007CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	2452.01	2482.76	2445.74	2460.17	0.81	ppb
Antimony	121-1	128.58	129.34	129.06	128.99	0.30	ppb
Arsenic	75-2	125.33	125.43	125.25	125.34	0.07	ppb
Barium	135-1	641.19	649.16	660.22	650.19	1.47	ppb
Barium	137-1	626.80	656.04	638.83	640.56	2.29	ppb
Beryllium	9-1	123.17	125.60	128.38	125.72	2.07	ppb
Bismuth	209-1				100		%
Bismuth	209-2				98		%
Bromine	81-1						cps
Cadmium	108-1	130.89	132.33	130.70	131.31	0.68	ppb
Cadmium	106-1	128.74	129.94	131.91	130.20	1.23	ppb
Cadmium	111-1	125.87	129.90	129.30	128.36	1.69	ppb
Calcium	43-1	12871.82	12875.29	13098.45	12948.52	1.00	ppb
Calcium	44-1	12892.34	12882.67	12877.75	12884.25	0.06	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	123.46	124.13	123.04	123.54	0.45	ppb
Cobalt	59-2	127.17	128.71	126.21	127.36	0.99	ppb
Copper	63-2	1313.74	1320.59	1292.85	1309.06	1.10	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				100		%
Holmium	165-2				101		%
Indium	115-1				99		%
Indium	115-2				98		%
Iron	54-2	6528.84	6646.90	6518.82	6564.85	1.09	ppb
Iron	56-2	6543.12	6597.46	6486.57	6542.38	0.85	ppb
Iron	57-2	6446.64	6535.56	6462.37	6481.52	0.73	ppb
Krypton	83-1						cps
Lead	206-1	613.17	618.22	626.85	619.41	1.12	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S04 Instrumnet Name : P8
Client Sample ID : S04 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:15:02 DataFile Name : 007CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	619.82	617.05	631.68	622.85	1.25	ppb
Lead	208-1	619.77	621.16	630.11	623.68	0.90	ppb
Lithium	6-1				100		%
Magnesium	24-2	12857.97	13045.94	12972.29	12958.73	0.73	ppb
Manganese	55-2	1291.88	1310.64	1264.34	1288.95	1.81	ppb
Molybdenum	94-1	1232.34	1235.58	1234.14	1234.02	0.13	ppb
Molybdenum	95-1	1243.85	1262.15	1239.62	1248.54	0.96	ppb
Molybdenum	96-1	1246.51	1254.70	1239.64	1246.95	0.60	ppb
Molybdenum	97-1	1243.74	1262.09	1234.07	1246.64	1.14	ppb
Molybdenum	98-1	1239.97	1249.14	1254.15	1247.75	0.58	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	126.80	128.56	126.45	127.27	0.89	ppb
Phosphorus	31-2	2503.53	2512.45	2525.00	2513.66	0.43	ppb
Potassium	39-2	6140.98	6162.98	6087.62	6130.53	0.63	ppb
Rhodium	103-1				100		%
Rhodium	103-2				98		%
Scandium	45-1				100		%
Scandium	45-2				100		%
Selenium	82-1	125.09	125.91	125.72	125.57	0.34	ppb
Selenium	77-2	129.50	127.73	134.38	130.54	2.64	ppb
Selenium	78-2	125.80	135.58	131.56	130.98	3.75	ppb
Silicon	28-1	130.02	126.49	129.88	128.80	1.55	ppb
Silver	107-1	131.64	132.04	132.67	132.12	0.40	ppb
Silver	109-1	130.81	133.38	133.16	132.45	1.07	ppb
Sodium	23-2	13002.08	13021.11	13001.39	13008.19	0.09	ppb
Strontium	86-1	121.73	124.39	124.03	123.38	1.17	ppb
Strontium	88-1	123.43	126.02	126.02	125.16	1.20	ppb
Sulfur	34-1	2686.04	3018.98	2653.91	2786.31	7.25	ppb
Terbium	159-1				102		%
Terbium	159-2				99		%
Thallium	203-1	123.82	126.09	126.81	125.58	1.24	ppb
Thallium	205-1	125.69	125.07	125.48	125.41	0.25	ppb
Tin	118-1	129.34	128.92	128.73	128.99	0.24	ppb
Titanium	47-1	1258.34	1262.23	1279.41	1266.66	0.88	ppb
Uranium	238-1	121.27	122.79	124.29	122.78	1.23	ppb

3

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S05 Instrumnet Name : P8
Client Sample ID : S05 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:18:01 DataFile Name : 008CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	4976.89	5023.54	4827.55	4942.66	2.07	ppb
Antimony	121-1	248.56	257.69	256.15	254.13	1.92	ppb
Arsenic	75-2	253.94	265.65	243.51	254.37	4.35	ppb
Barium	135-1	1243.70	1289.50	1278.45	1270.55	1.88	ppb
Barium	137-1	1227.65	1290.54	1271.61	1263.27	2.55	ppb
Beryllium	9-1	246.82	254.51	258.76	253.37	2.39	ppb
Bismuth	209-1				99		%
Bismuth	209-2				98		%
Bromine	81-1						cps
Cadmium	108-1	250.61	265.10	255.70	257.13	2.86	ppb
Cadmium	106-1	248.19	261.43	255.69	255.10	2.60	ppb
Cadmium	111-1	244.72	257.77	253.71	252.06	2.65	ppb
Calcium	43-1	25182.11	26272.37	26140.66	25865.05	2.30	ppb
Calcium	44-1	24569.66	25761.95	25713.26	25348.29	2.66	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	243.84	248.96	238.22	243.67	2.20	ppb
Cobalt	59-2	259.44	261.13	251.08	257.22	2.09	ppb
Copper	63-2	2591.07	2685.56	2541.84	2606.16	2.80	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				101		%
Holmium	165-2				99		%
Indium	115-1				99		%
Indium	115-2				97		%
Iron	54-2	13441.49	13415.10	13010.04	13288.87	1.82	ppb
Iron	56-2	12987.00	13323.70	12647.90	12986.20	2.60	ppb
Iron	57-2	12736.75	13075.28	12563.14	12791.72	2.04	ppb
Krypton	83-1						cps
Lead	206-1	1214.61	1246.17	1263.09	1241.29	1.98	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S05 Instrumnet Name : P8
Client Sample ID : S05 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:18:01 DataFile Name : 008CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	1195.71	1241.10	1251.17	1229.33	2.40	ppb
Lead	208-1	1200.16	1246.56	1254.84	1233.85	2.39	ppb
Lithium	6-1				99		%
Magnesium	24-2	26315.73	26400.75	25122.30	25946.26	2.76	ppb
Manganese	55-2	2554.89	2593.99	2514.85	2554.58	1.55	ppb
Molybdenum	94-1	2415.32	2515.21	2513.45	2481.33	2.30	ppb
Molybdenum	95-1	2427.42	2527.43	2507.75	2487.54	2.13	ppb
Molybdenum	96-1	2426.05	2540.19	2521.78	2496.01	2.46	ppb
Molybdenum	97-1	2423.84	2506.70	2540.83	2490.46	2.42	ppb
Molybdenum	98-1	2419.32	2535.62	2548.04	2500.99	2.84	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	250.16	256.81	244.69	250.55	2.42	ppb
Phosphorus	31-2	4992.78	5158.69	4782.24	4977.90	3.79	ppb
Potassium	39-2	12508.10	12857.28	12240.38	12535.25	2.47	ppb
Rhodium	103-1				98		%
Rhodium	103-2				97		%
Scandium	45-1				100		%
Scandium	45-2				100		%
Selenium	82-1	245.18	262.84	257.09	255.04	3.53	ppb
Selenium	77-2	278.17	287.32	227.61	264.37	12.16	ppb
Selenium	78-2	262.02	265.48	258.24	261.91	1.38	ppb
Silicon	28-1	245.83	252.27	260.94	253.01	3.00	ppb
Silver	107-1	255.07	266.16	260.96	260.73	2.13	ppb
Silver	109-1	252.96	263.78	259.59	258.78	2.11	ppb
Sodium	23-2	26488.44	26881.63	25428.29	26266.12	2.86	ppb
Strontium	86-1	238.03	247.47	250.86	245.45	2.71	ppb
Strontium	88-1	244.86	253.91	253.49	250.75	2.04	ppb
Sulfur	34-1	4706.98	5102.59	5344.96	5051.51	6.38	ppb
Terbium	159-1				102		%
Terbium	159-2				99		%
Thallium	203-1	247.15	257.06	256.88	253.69	2.24	ppb
Thallium	205-1	240.95	249.27	249.56	246.59	1.98	ppb
Tin	118-1	251.39	256.84	258.22	255.48	1.41	ppb
Titanium	47-1	2434.28	2489.18	2460.55	2461.34	1.12	ppb
Uranium	238-1	237.15	246.29	244.06	242.50	1.96	ppb

3

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S06 Instrumnet Name : P8
Client Sample ID : S06 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:20:51 DataFile Name : 009CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	9883.79	9794.63	9810.11	9829.51	0.48	ppb
Antimony	121-1	499.07	505.10	506.78	503.65	0.81	ppb
Arsenic	75-2	500.30	517.61	512.43	510.11	1.74	ppb
Barium	135-1	2513.58	2523.15	2511.77	2516.17	0.24	ppb
Barium	137-1	2516.93	2489.70	2498.67	2501.77	0.55	ppb
Beryllium	9-1	507.98	500.56	508.66	505.73	0.89	ppb
Bismuth	209-1				97		%
Bismuth	209-2				94		%
Bromine	81-1						cps
Cadmium	108-1	503.90	500.35	508.18	504.14	0.78	ppb
Cadmium	106-1	503.93	506.41	501.56	503.97	0.48	ppb
Cadmium	111-1	500.58	522.01	500.92	507.83	2.42	ppb
Calcium	43-1	51417.80	54154.09	52864.28	52812.05	2.59	ppb
Calcium	44-1	51209.24	52209.30	52725.49	52048.01	1.48	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	517.83	503.85	497.25	506.31	2.08	ppb
Cobalt	59-2	516.27	512.29	518.08	515.54	0.57	ppb
Copper	63-2	5166.74	5101.16	5090.98	5119.63	0.80	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				100		%
Holmium	165-2				99		%
Indium	115-1				97		%
Indium	115-2				94		%
Iron	54-2	27081.10	26348.20	26408.56	26612.62	1.53	ppb
Iron	56-2	26194.12	26000.63	25630.57	25941.77	1.10	ppb
Iron	57-2	26721.47	26878.62	26460.78	26686.96	0.79	ppb
Krypton	83-1						cps
Lead	206-1	2503.78	2568.80	2520.62	2531.07	1.33	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S06 Instrumnet Name : P8
Client Sample ID : S06 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:20:51 DataFile Name : 009CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	2502.93	2541.56	2498.81	2514.43	0.94	ppb
Lead	208-1	2500.96	2542.70	2503.06	2515.58	0.93	ppb
Lithium	6-1				97		%
Magnesium	24-2	51355.01	50925.34	51245.92	51175.42	0.44	ppb
Manganese	55-2	5131.32	5058.12	4965.43	5051.62	1.65	ppb
Molybdenum	94-1	4985.02	5068.08	5029.78	5027.63	0.83	ppb
Molybdenum	95-1	4947.27	5041.60	5049.20	5012.69	1.13	ppb
Molybdenum	96-1	4937.70	5019.56	5069.65	5008.97	1.33	ppb
Molybdenum	97-1	4938.52	5016.56	4982.99	4979.36	0.79	ppb
Molybdenum	98-1	5018.96	5069.36	4993.18	5027.17	0.77	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	495.33	495.81	492.42	494.52	0.37	ppb
Phosphorus	31-2	10353.82	10324.76	10295.48	10324.69	0.28	ppb
Potassium	39-2	25511.32	24922.23	25085.69	25173.08	1.21	ppb
Rhodium	103-1				96		%
Rhodium	103-2				94		%
Scandium	45-1				98		%
Scandium	45-2				98		%
Selenium	82-1	494.54	511.98	495.99	500.84	1.93	ppb
Selenium	77-2	509.92	516.16	513.18	513.09	0.61	ppb
Selenium	78-2	499.41	495.24	515.83	503.49	2.16	ppb
Silicon	28-1	494.84	525.63	517.24	512.57	3.11	ppb
Silver	107-1	501.21	507.30	511.39	506.63	1.01	ppb
Silver	109-1	498.69	509.69	507.75	505.38	1.16	ppb
Sodium	23-2	53437.00	52517.32	52924.99	52959.77	0.87	ppb
Strontium	86-1	497.27	512.31	505.14	504.91	1.49	ppb
Strontium	88-1	492.20	501.31	500.62	498.04	1.02	ppb
Sulfur	34-1	9939.85	10748.57	10209.22	10299.22	4.00	ppb
Terbium	159-1				102		%
Terbium	159-2				98		%
Thallium	203-1	495.69	512.45	507.89	505.34	1.72	ppb
Thallium	205-1	493.47	515.22	504.54	504.41	2.16	ppb
Tin	118-1	496.15	501.81	499.13	499.03	0.57	ppb
Titanium	47-1	4998.97	5118.22	5058.09	5058.43	1.18	ppb
Uranium	238-1	488.89	506.71	489.88	495.16	2.02	ppb

3

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S07 Instrumnet Name : P8
Client Sample ID : S07 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:23:38 DataFile Name : 010CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	20040.16	20199.80	19592.17	19944.05	1.58	ppb
Antimony	121-1	996.63	997.03	996.16	996.61	0.04	ppb
Arsenic	75-2	981.78	1007.52	991.94	993.75	1.30	ppb
Barium	135-1	4996.89	4957.10	4996.59	4983.52	0.46	ppb
Barium	137-1	5059.63	4975.99	4944.87	4993.49	1.19	ppb
Beryllium	9-1	998.40	992.81	997.19	996.13	0.29	ppb
Bismuth	209-1				95		%
Bismuth	209-2				94		%
Bromine	81-1						cps
Cadmium	108-1	992.14	1002.25	991.31	995.23	0.61	ppb
Cadmium	106-1	996.24	1002.23	989.31	995.93	0.65	ppb
Cadmium	111-1	995.79	994.87	994.47	995.04	0.07	ppb
Calcium	43-1	102320.07	105137.98	101269.39	102909.15	1.94	ppb
Calcium	44-1	102047.33	103629.06	101377.68	102351.36	1.13	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1004.68	996.88	994.19	998.58	0.55	ppb
Cobalt	59-2	994.35	980.75	994.88	989.99	0.81	ppb
Copper	63-2	9997.46	9861.91	9851.93	9903.77	0.82	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				99		%
Holmium	165-2				99		%
Indium	115-1				94		%
Indium	115-2				92		%
Iron	54-2	51723.88	51963.26	51592.29	51759.81	0.36	ppb
Iron	56-2	50852.48	50551.90	50656.01	50686.80	0.30	ppb
Iron	57-2	51801.28	51429.18	52683.53	51971.33	1.24	ppb
Krypton	83-1						cps
Lead	206-1	4993.66	4951.94	5015.45	4987.02	0.65	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S07 Instrumnet Name : P8
Client Sample ID : S07 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:23:38 DataFile Name : 010CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	5034.46	4955.81	5003.84	4998.03	0.79	ppb
Lead	208-1	5012.52	4963.14	5013.24	4996.30	0.57	ppb
Lithium	6-1				96		%
Magnesium	24-2	100216.27	99523.94	100694.75	100144.98	0.59	ppb
Manganese	55-2	9980.64	9916.80	9967.43	9954.96	0.34	ppb
Molybdenum	94-1	9832.46	10076.92	10068.52	9992.63	1.39	ppb
Molybdenum	95-1	9832.32	10057.16	10100.96	9996.81	1.44	ppb
Molybdenum	96-1	9778.17	10042.93	10168.66	9996.59	1.99	ppb
Molybdenum	97-1	9824.77	10061.13	10151.87	10012.59	1.69	ppb
Molybdenum	98-1	9864.05	10064.97	10029.97	9986.33	1.07	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	1013.46	1002.59	990.55	1002.20	1.14	ppb
Phosphorus	31-2	19938.05	19957.51	19631.43	19842.33	0.92	ppb
Potassium	39-2	48504.25	48818.05	47980.99	48434.43	0.87	ppb
Rhodium	103-1				93		%
Rhodium	103-2				92		%
Scandium	45-1				98		%
Scandium	45-2				99		%
Selenium	82-1	981.40	1007.40	1005.74	998.18	1.46	ppb
Selenium	77-2	1008.55	962.83	995.68	989.02	2.38	ppb
Selenium	78-2	982.07	996.32	1004.62	994.34	1.15	ppb
Silicon	28-1	984.63	1009.92	982.86	992.47	1.53	ppb
Silver	107-1	1000.56	990.85	987.62	993.01	0.68	ppb
Silver	109-1	989.94	1003.64	988.66	994.08	0.84	ppb
Sodium	23-2	103396.66	103579.29	103325.04	103433.66	0.13	ppb
Strontium	86-1	999.26	997.18	1000.20	998.88	0.15	ppb
Strontium	88-1	995.67	1007.02	999.48	1000.72	0.58	ppb
Sulfur	34-1	19857.95	20140.36	19378.96	19792.42	1.94	ppb
Terbium	159-1				101		%
Terbium	159-2				97		%
Thallium	203-1	999.10	987.27	1002.77	996.38	0.81	ppb
Thallium	205-1	993.21	996.06	1006.44	998.57	0.70	ppb
Tin	118-1	996.36	1006.89	992.42	998.56	0.75	ppb
Titanium	47-1	10020.64	10077.57	9835.04	9977.75	1.27	ppb
Uranium	238-1	1000.74	999.94	1013.11	1004.59	0.74	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S07 Instrumnet Name : P8
Client Sample ID : S07 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:23:38 DataFile Name : 010CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Vanadium	51-2	1008.13	1012.17	985.64	1001.98	1.43	ppb
Ytterbium	172-1						cps
Ytterbium	172-2						cps
Ytterbium	176-1						cps
Ytterbium	176-2						cps
Yttrium	89-1				98		%
Yttrium	89-2				99		%
Zinc	66-2	9930.79	9952.97	9789.91	9891.22	0.89	ppb
Zirconium	90-1	987.30	1000.81	1001.50	996.54	0.80	ppb
Zirconium	91-1	984.69	996.58	1012.55	997.94	1.40	ppb

LB Number :	LB134674	Operator :	Jaswal
Lab Sample ID :	S08	Instrumnet Name :	P8
Client Sample ID :	S08	Dilution Factor :	1
Date & Time Acquired :	2025-02-11 12:26:23	DataFile Name :	011CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	100001.44	99145.65	100949.12	100032.07	0.90	ppb
Antimony	121-1	0.92	0.88	0.86	0.89	3.02	ppb
Arsenic	75-2	0.47	0.57	0.59	0.54	12.16	ppb
Barium	135-1	2.03	2.01	1.93	1.99	2.45	ppb
Barium	137-1	2.07	1.98	1.98	2.01	2.69	ppb
Beryllium	9-1	0.39	0.32	0.31	0.34	12.43	ppb
Bismuth	209-1				82		%
Bismuth	209-2				83		%
Bromine	81-1						cps
Cadmium	108-1	0.92	0.93	0.59	0.81	23.82	ppb
Cadmium	106-1	1.13	0.53	0.11	0.59	87.38	ppb
Cadmium	111-1	0.26	0.17	0.15	0.19	30.09	ppb
Calcium	43-1	496458.64	491689.72	509091.84	499080.07	1.80	ppb
Calcium	44-1	500231.92	489395.68	508258.84	499295.48	1.90	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	3.15	3.02	3.11	3.09	2.07	ppb
Cobalt	59-2	2.11	2.16	2.18	2.15	1.80	ppb
Copper	63-2	1.74	1.82	1.72	1.76	3.02	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				93		%
Holmium	165-2				94		%
Indium	115-1				87		%
Indium	115-2				89		%
Iron	54-2	247511.87	247946.59	252854.03	249437.49	1.19	ppb
Iron	56-2	249835.80	251502.37	247866.72	249734.97	0.73	ppb
Iron	57-2	249954.84	245133.83	253156.72	249415.13	1.62	ppb
Krypton	83-1						cps
Lead	206-1	1.10	1.09	1.05	1.08	2.31	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S08 Instrumnet Name : P8
Client Sample ID : S08 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:26:23 DataFile Name : 011CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	1.11	1.10	1.02	1.08	4.63	ppb
Lead	208-1	1.13	1.10	1.03	1.08	4.43	ppb
Lithium	6-1				88		%
Magnesium	24-2	498044.63	499922.23	501406.70	499791.18	0.34	ppb
Manganese	55-2	4.48	4.67	4.71	4.62	2.69	ppb
Molybdenum	94-1	7.41	7.49	7.54	7.48	0.87	ppb
Molybdenum	95-1	5.32	5.20	5.39	5.30	1.83	ppb
Molybdenum	96-1	5.49	5.67	5.66	5.61	1.73	ppb
Molybdenum	97-1	5.24	5.35	5.31	5.30	1.05	ppb
Molybdenum	98-1	5.18	5.18	5.23	5.20	0.56	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	3.45	3.55	3.71	3.57	3.66	ppb
Phosphorus	31-2	-3.91	-7.46	-15.54	-8.97		ppb
Potassium	39-2	246285.44	247204.60	257400.49	250296.85	2.46	ppb
Rhodium	103-1				83		%
Rhodium	103-2				82		%
Scandium	45-1				92		%
Scandium	45-2				95		%
Selenium	82-1	0.03	-0.12	0.11	0.01	1565.73	ppb
Selenium	77-2	2.57	1.64	-0.31	1.30	112.87	ppb
Selenium	78-2	1.70	1.16	0.32	1.06	65.72	ppb
Silicon	28-1	18.54	16.06	16.86	17.15	7.37	ppb
Silver	107-1	0.18	0.18	0.17	0.18	2.86	ppb
Silver	109-1	0.20	0.18	0.16	0.18	8.87	ppb
Sodium	23-2	493861.44	497536.67	505413.75	498937.28	1.18	ppb
Strontium	86-1	3.15	3.04	3.11	3.10	1.79	ppb
Strontium	88-1	3.09	3.08	3.13	3.10	0.78	ppb
Sulfur	34-1	-151.56	-325.66	-49.80	-175.67		ppb
Terbium	159-1				92		%
Terbium	159-2				93		%
Thallium	203-1	0.08	0.09	0.07	0.08	7.75	ppb
Thallium	205-1	0.10	0.08	0.07	0.08	15.49	ppb
Tin	118-1	0.11	0.10	0.07	0.09	26.00	ppb
Titanium	47-1	1.61	1.72	1.75	1.69	4.42	ppb
Uranium	238-1	0.05	0.05	0.04	0.05	9.57	ppb

3

LB Number :

LB134674

Operator :

Jaswal

Lab Sample ID :

ICV005

Instrumnet Name :

P8

Client Sample ID :

ICV005

Dilution Factor :

1

Date & Time Acquired :

2025-02-11 13:29:04

DataFile Name :

012ICV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	474.38	474.26	479.22	475.96	0.59	ppb
Antimony	121-1	203.39	205.37	208.33	205.70	1.21	ppb
Arsenic	75-2	204.81	205.88	210.12	206.93	1.36	ppb
Barium	135-1	99.44	100.04	101.23	100.24	0.91	ppb
Barium	137-1	99.75	101.13	102.64	101.17	1.43	ppb
Beryllium	9-1	100.05	103.59	101.43	101.69	1.76	ppb
Bismuth	209-1				102		%
Bismuth	209-2				101		%
Bromine	81-1						cps
Cadmium	108-1	130.63	135.70	136.19	134.17	2.29	ppb
Cadmium	106-1	79.71	80.20	82.71	80.87	1.99	ppb
Cadmium	111-1	103.51	105.21	105.35	104.69	0.98	ppb
Calcium	43-1	2228.40	2264.92	2255.92	2249.74	0.85	ppb
Calcium	44-1	2182.53	2171.05	2225.64	2193.08	1.31	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	101.29	101.03	101.25	101.19	0.14	ppb
Cobalt	59-2	105.02	106.03	105.30	105.45	0.49	ppb
Copper	63-2	106.08	107.80	107.31	107.06	0.83	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				100		%
Indium	115-1				101		%
Indium	115-2				100		%
Iron	54-2	2115.83	2152.87	2133.01	2133.90	0.87	ppb
Iron	56-2	2084.94	2069.91	2056.00	2070.28	0.70	ppb
Iron	57-2	2071.76	2042.39	2063.45	2059.20	0.74	ppb
Krypton	83-1						cps
Lead	206-1	207.14	205.46	203.00	205.20	1.01	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : ICV005 Instrumnet Name : P8
Client Sample ID : ICV005 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:29:04 DataFile Name : 012ICV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	192.93	197.39	195.74	195.35	1.15	ppb
Lead	208-1	197.10	199.05	197.91	198.02	0.50	ppb
Lithium	6-1				102		%
Magnesium	24-2	1145.83	1144.18	1153.22	1147.74	0.42	ppb
Manganese	55-2	98.59	101.58	99.41	99.86	1.55	ppb
Molybdenum	94-1	3842.15	3894.58	3854.33	3863.69	0.71	ppb
Molybdenum	95-1	4712.46	4804.91	4737.07	4751.48	1.01	ppb
Molybdenum	96-1	4640.67	4669.45	4688.96	4666.36	0.52	ppb
Molybdenum	97-1	4648.16	4742.05	4780.20	4723.47	1.44	ppb
Molybdenum	98-1	4655.45	4808.26	4725.23	4729.64	1.62	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	104.55	106.57	104.77	105.29	1.05	ppb
Phosphorus	31-2	-9.47	-16.25	-11.58	-12.44		ppb
Potassium	39-2	1902.58	1911.28	1905.26	1906.37	0.23	ppb
Rhodium	103-1				101		%
Rhodium	103-2				100		%
Scandium	45-1				103		%
Scandium	45-2				101		%
Selenium	82-1	206.07	211.41	213.32	210.27	1.79	ppb
Selenium	77-2	202.75	241.20	235.75	226.57	9.18	ppb
Selenium	78-2	219.67	213.51	224.26	219.15	2.46	ppb
Silicon	28-1	1.21	1.33	1.70	1.41	18.19	ppb
Silver	107-1	47.78	49.27	49.02	48.69	1.64	ppb
Silver	109-1	48.15	49.05	49.75	48.98	1.63	ppb
Sodium	23-2	2042.96	2019.09	2032.96	2031.67	0.59	ppb
Strontium	86-1	522.12	526.86	536.76	528.58	1.41	ppb
Strontium	88-1	514.76	523.90	526.72	521.79	1.20	ppb
Sulfur	34-1	-30.06	2.37	20.86	-2.28		ppb
Terbium	159-1				103		%
Terbium	159-2				99		%
Thallium	203-1	205.48	201.73	200.40	202.54	1.30	ppb
Thallium	205-1	200.43	199.20	199.02	199.55	0.38	ppb
Tin	118-1	0.04	0.06	0.06	0.06	21.39	ppb
Titanium	47-1	0.44	0.44	0.46	0.45	3.13	ppb
Uranium	238-1	0.01	0.01	0.01	0.01	9.50	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : ICV005 Instrumnet Name : P8
Client Sample ID : ICV005 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:29:04 DataFile Name : 012ICV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Vanadium	51-2	98.19	99.33	98.44	98.65	0.61	ppb
Ytterbium	172-1						cps
Ytterbium	172-2						cps
Ytterbium	176-1						cps
Ytterbium	176-2						cps
Yttrium	89-1				102		%
Yttrium	89-2				100		%
Zinc	66-2	199.47	204.25	200.58	201.43	1.24	ppb
Zirconium	90-1	0.02	0.02	0.02	0.02	16.21	ppb
Zirconium	91-1	0.12	0.11	0.14	0.12	11.20	ppb

LB Number : LB134674

Operator : Jaswal

Lab Sample ID : ICB005

Instrumnet Name : P8

Client Sample ID : ICB005

Dilution Factor : 1

Date & Time Acquired : 2025-02-11 13:33:28

DataFile Name : 013CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	-0.97	-0.51	-0.04	-0.51		ppb
Antimony	121-1	0.04	0.03	0.03	0.03	17.21	ppb
Arsenic	75-2	-0.02	-0.03	-0.03	-0.02		ppb
Barium	135-1	0.01	0.01	0.00	0.01	87.14	ppb
Barium	137-1	0.01	0.01	0.01	0.01	14.33	ppb
Beryllium	9-1	0.01	0.01	0.01	0.01	30.13	ppb
Bismuth	209-1				102		%
Bismuth	209-2				102		%
Bromine	81-1						cps
Cadmium	108-1	0.02	0.06	0.06	0.04	50.68	ppb
Cadmium	106-1	-0.24	-0.65	-1.28	-0.72		ppb
Cadmium	111-1	-0.01	-0.05	-0.10	-0.05		ppb
Calcium	43-1	0.47	-1.32	0.19	-0.22		ppb
Calcium	44-1	-4.99	-5.50	-6.09	-5.53		ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	0.00	0.01	0.00	0.00	313.17	ppb
Cobalt	59-2	-0.01	-0.01	-0.01	-0.01		ppb
Copper	63-2	-0.11	-0.16	-0.11	-0.13		ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				101		%
Holmium	165-2				101		%
Indium	115-1				102		%
Indium	115-2				100		%
Iron	54-2	0.35	0.31	0.20	0.29	26.23	ppb
Iron	56-2	0.03	-0.16	-0.05	-0.06		ppb
Iron	57-2	-0.44	-0.07	0.04	-0.16		ppb
Krypton	83-1						cps
Lead	206-1	0.01	0.02	0.02	0.02	32.46	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : ICB005 Instrumnet Name : P8
Client Sample ID : ICB005 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:33:28 DataFile Name : 013CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	0.03	0.01	0.03	0.02	40.64	ppb
Lead	208-1	0.02	0.02	0.02	0.02	15.86	ppb
Lithium	6-1				103		%
Magnesium	24-2	-7.24	-6.36	-7.31	-6.97		ppb
Manganese	55-2	0.00	0.02	0.00	0.01	162.70	ppb
Molybdenum	94-1	0.09	0.05	0.07	0.07	23.72	ppb
Molybdenum	95-1	0.10	0.09	0.07	0.09	17.36	ppb
Molybdenum	96-1	0.09	0.08	0.08	0.08	11.10	ppb
Molybdenum	97-1	0.09	0.08	0.08	0.08	6.97	ppb
Molybdenum	98-1	0.10	0.10	0.08	0.09	12.23	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	-0.08	-0.12	-0.12	-0.10		ppb
Phosphorus	31-2	-14.26	-21.16	-6.53	-13.99		ppb
Potassium	39-2	-1.16	-0.70	-1.00	-0.95		ppb
Rhodium	103-1				102		%
Rhodium	103-2				101		%
Scandium	45-1				102		%
Scandium	45-2				102		%
Selenium	82-1	0.20	-0.06	0.21	0.12	127.21	ppb
Selenium	77-2	-0.31	-0.31	-0.31	-0.31		ppb
Selenium	78-2	0.00	0.00	0.00	0.00		ppb
Silicon	28-1	-0.44	-0.47	-0.70	-0.53		ppb
Silver	107-1	0.01	0.01	0.01	0.01	31.91	ppb
Silver	109-1	0.01	0.01	0.01	0.01	8.40	ppb
Sodium	23-2	-5.25	-4.47	-6.40	-5.37		ppb
Strontium	86-1	-0.02	-0.01	-0.03	-0.02		ppb
Strontium	88-1	-0.01	-0.01	-0.01	-0.01		ppb
Sulfur	34-1	-69.67	-118.01	-132.30	-106.66		ppb
Terbium	159-1				103		%
Terbium	159-2				100		%
Thallium	203-1	0.01	0.01	0.00	0.00	55.64	ppb
Thallium	205-1	0.00	0.00	0.00	0.00	2.06	ppb
Tin	118-1	0.00	0.00	-0.01	0.00		ppb
Titanium	47-1	-0.03	-0.07	-0.05	-0.05		ppb
Uranium	238-1	0.00	0.00	0.00	0.00		ppb

LB Number :

LB134674

Operator :

Jaswal

Lab Sample ID :

ICB005

Instrumnet Name :

P8

Client Sample ID :

ICB005

Dilution Factor :

1

Date & Time Acquired :

2025-02-11 13:33:28

DataFile Name :

013CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Vanadium	51-2	0.00	0.00	0.00	0.00		ppb
Ytterbium	172-1						cps
Ytterbium	172-2						cps
Ytterbium	176-1						cps
Ytterbium	176-2						cps
Yttrium	89-1				102		%
Yttrium	89-2				101		%
Zinc	66-2	0.03	0.15	0.25	0.14	78.78	ppb
Zirconium	90-1	0.00	0.00	0.00	0.00		ppb
Zirconium	91-1	0.00	0.01	0.01	0.01	83.82	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : ICSA005 Instrumnet Name : P8
Client Sample ID : ICSA005 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:40:53 DataFile Name : 015ICSA.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	93753.99	95541.18	93981.70	94425.62	1.03	ppb
Antimony	121-1	1.19	1.15	1.17	1.17	1.75	ppb
Arsenic	75-2	0.34	0.22	0.21	0.26	28.27	ppb
Barium	135-1	1.40	1.35	1.39	1.38	1.87	ppb
Barium	137-1	1.44	1.44	1.47	1.45	1.23	ppb
Beryllium	9-1	0.29	0.29	0.30	0.30	2.31	ppb
Bismuth	209-1				92		%
Bismuth	209-2				92		%
Bromine	81-1						cps
Cadmium	108-1	19.33	18.14	18.36	18.61	3.38	ppb
Cadmium	106-1	-2.58	-1.61	-2.62	-2.27		ppb
Cadmium	111-1	0.44	0.56	0.44	0.48	14.44	ppb
Calcium	43-1	107230.28	106378.29	108303.47	107304.01	0.90	ppb
Calcium	44-1	104007.83	104537.69	106853.62	105133.05	1.44	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	19.67	20.30	20.28	20.09	1.78	ppb
Cobalt	59-2	1.16	1.20	1.18	1.18	1.52	ppb
Copper	63-2	8.01	8.20	8.16	8.13	1.24	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				100		%
Holmium	165-2				100		%
Indium	115-1				96		%
Indium	115-2				95		%
Iron	54-2	103355.70	105396.89	104119.06	104290.55	0.99	ppb
Iron	56-2	103556.33	105216.25	104118.86	104297.15	0.81	ppb
Iron	57-2	103931.65	107096.75	104442.56	105156.98	1.62	ppb
Krypton	83-1						cps
Lead	206-1	4.86	4.80	4.83	4.83	0.71	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : ICSA005 Instrumnet Name : P8
Client Sample ID : ICSA005 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:40:53 DataFile Name : 015ICSA.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	4.45	4.39	4.48	4.44	1.13	ppb
Lead	208-1	4.55	4.48	4.51	4.51	0.80	ppb
Lithium	6-1				96		%
Magnesium	24-2	96157.04	98164.13	98421.79	97580.99	1.27	ppb
Manganese	55-2	7.68	7.67	7.98	7.78	2.24	ppb
Molybdenum	94-1	1602.24	1609.96	1640.90	1617.70	1.26	ppb
Molybdenum	95-1	1956.97	1975.72	2029.65	1987.45	1.90	ppb
Molybdenum	96-1	1947.66	1964.03	2020.11	1977.27	1.92	ppb
Molybdenum	97-1	2001.49	1984.19	2025.85	2003.84	1.04	ppb
Molybdenum	98-1	1983.79	1987.71	2032.34	2001.28	1.35	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	5.00	5.15	5.21	5.12	2.11	ppb
Phosphorus	31-2	103511.66	104900.69	104989.08	104467.14	0.79	ppb
Potassium	39-2	99496.96	100608.07	100327.40	100144.14	0.58	ppb
Rhodium	103-1				94		%
Rhodium	103-2				94		%
Scandium	45-1				99		%
Scandium	45-2				101		%
Selenium	82-1	0.13	0.58	0.23	0.31	75.27	ppb
Selenium	77-2	-0.31	0.62	-0.31	0.00	20462.25	ppb
Selenium	78-2	0.26	0.82	0.80	0.63	50.24	ppb
Silicon	28-1	25.53	26.12	28.05	26.57	4.96	ppb
Silver	107-1	0.02	0.02	0.02	0.02	20.11	ppb
Silver	109-1	0.02	0.02	0.03	0.02	30.13	ppb
Sodium	23-2	105227.50	109112.01	106874.08	107071.20	1.82	ppb
Strontium	86-1	33.93	34.72	34.79	34.48	1.39	ppb
Strontium	88-1	35.29	35.74	35.56	35.53	0.63	ppb
Sulfur	34-1	98761.46	99159.30	99908.90	99276.55	0.59	ppb
Terbium	159-1				101		%
Terbium	159-2				98		%
Thallium	203-1	0.06	0.05	0.07	0.06	13.35	ppb
Thallium	205-1	0.05	0.05	0.06	0.05	13.41	ppb
Tin	118-1	0.12	0.10	0.10	0.11	9.36	ppb
Titanium	47-1	2081.60	2090.38	2152.89	2108.29	1.84	ppb
Uranium	238-1	0.01	0.01	0.02	0.01	9.36	ppb

LB Number :

LB134674

Operator :

Jaswal

Lab Sample ID :

ICSA005

Instrumnet Name :

P8

Client Sample ID :

ICSA005

Dilution Factor :

1

Date & Time Acquired :

2025-02-11 13:40:53

DataFile Name :

015ICSA.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Vanadium	51-2	0.13	0.16	0.14	0.14	9.13	ppb
Ytterbium	172-1						cps
Ytterbium	172-2						cps
Ytterbium	176-1						cps
Ytterbium	176-2						cps
Yttrium	89-1				100		%
Yttrium	89-2				101		%
Zinc	66-2	10.82	11.31	11.08	11.07	2.21	ppb
Zirconium	90-1	0.01	0.01	0.01	0.01	17.68	ppb
Zirconium	91-1	0.01	0.02	0.01	0.01	30.62	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : ICSAB005 Instrumnet Name : P8
Client Sample ID : ICSAB005 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:44:07 DataFile Name : 016ICSB.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	92838.93	94469.11	93063.81	93457.28	0.95	ppb
Antimony	121-1	22.07	22.24	22.10	22.14	0.40	ppb
Arsenic	75-2	21.18	22.22	21.86	21.75	2.44	ppb
Barium	135-1	22.31	21.93	22.27	22.17	0.94	ppb
Barium	137-1	22.44	22.72	22.35	22.50	0.87	ppb
Beryllium	9-1	21.42	20.87	21.72	21.34	2.02	ppb
Bismuth	209-1				94		%
Bismuth	209-2				92		%
Bromine	81-1						cps
Cadmium	108-1	34.81	34.99	34.94	34.91	0.26	ppb
Cadmium	106-1	13.22	14.10	13.41	13.58	3.38	ppb
Cadmium	111-1	20.78	20.64	20.52	20.65	0.62	ppb
Calcium	43-1	105683.57	106579.58	108123.02	106795.39	1.16	ppb
Calcium	44-1	104794.82	105729.87	105617.30	105380.66	0.48	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	39.71	39.76	39.90	39.79	0.24	ppb
Cobalt	59-2	21.38	21.89	21.63	21.63	1.18	ppb
Copper	63-2	28.41	28.36	28.53	28.43	0.32	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				100		%
Holmium	165-2				100		%
Indium	115-1				97		%
Indium	115-2				96		%
Iron	54-2	104816.97	104548.16	103730.56	104365.23	0.54	ppb
Iron	56-2	102910.05	104311.05	103267.01	103496.04	0.70	ppb
Iron	57-2	104934.36	105261.70	103525.25	104573.77	0.88	ppb
Krypton	83-1						cps
Lead	206-1	25.97	24.96	25.75	25.56	2.07	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : ICSAB005 Instrumnet Name : P8
Client Sample ID : ICSAB005 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:44:07 DataFile Name : 016ICSB.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	24.71	24.23	24.66	24.53	1.08	ppb
Lead	208-1	25.02	24.47	25.01	24.83	1.27	ppb
Lithium	6-1				97		%
Magnesium	24-2	96682.97	97260.00	97640.71	97194.56	0.50	ppb
Manganese	55-2	27.46	27.92	27.30	27.56	1.17	ppb
Molybdenum	94-1	1619.40	1608.19	1644.86	1624.15	1.16	ppb
Molybdenum	95-1	2019.46	1990.68	2005.22	2005.12	0.72	ppb
Molybdenum	96-1	1982.97	1946.65	1956.62	1962.08	0.96	ppb
Molybdenum	97-1	2014.48	1996.99	1995.50	2002.32	0.53	ppb
Molybdenum	98-1	2022.90	1966.70	2004.50	1998.03	1.43	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	24.81	25.20	25.15	25.05	0.85	ppb
Phosphorus	31-2	103987.60	104893.04	104661.69	104514.11	0.45	ppb
Potassium	39-2	99869.44	100245.41	99341.59	99818.81	0.45	ppb
Rhodium	103-1				94		%
Rhodium	103-2				93		%
Scandium	45-1				101		%
Scandium	45-2				102		%
Selenium	82-1	21.09	21.15	20.39	20.88	2.03	ppb
Selenium	77-2	21.47	26.60	24.55	24.21	10.67	ppb
Selenium	78-2	23.64	29.89	26.15	26.56	11.83	ppb
Silicon	28-1	28.23	29.29	30.26	29.26	3.48	ppb
Silver	107-1	19.41	19.55	19.61	19.52	0.52	ppb
Silver	109-1	19.58	19.49	19.65	19.57	0.42	ppb
Sodium	23-2	105675.64	107130.02	107019.12	106608.26	0.76	ppb
Strontium	86-1	34.39	34.23	34.42	34.34	0.29	ppb
Strontium	88-1	35.44	35.27	35.66	35.46	0.55	ppb
Sulfur	34-1	96935.89	97847.31	99160.36	97981.19	1.14	ppb
Terbium	159-1				102		%
Terbium	159-2				98		%
Thallium	203-1	21.19	21.13	21.08	21.13	0.25	ppb
Thallium	205-1	21.36	20.71	21.35	21.14	1.76	ppb
Tin	118-1	0.16	0.15	0.15	0.15	1.26	ppb
Titanium	47-1	2118.93	2117.62	2144.11	2126.89	0.70	ppb
Uranium	238-1	0.02	0.01	0.01	0.01	8.87	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : ICSAB005 Instrumnet Name : P8
Client Sample ID : ICSAB005 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:44:07 DataFile Name : 016ICSB.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Vanadium	51-2	20.07	20.23	20.18	20.16	0.40	ppb
Ytterbium	172-1						cps
Ytterbium	172-2						cps
Ytterbium	176-1						cps
Ytterbium	176-2						cps
Yttrium	89-1				102		%
Yttrium	89-2				101		%
Zinc	66-2	31.09	31.28	31.66	31.34	0.92	ppb
Zirconium	90-1	0.01	0.01	0.01	0.01	4.08	ppb
Zirconium	91-1	0.02	0.01	0.03	0.02	38.41	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : CCV021 Instrumnet Name : P8
Client Sample ID : CCV021 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:47:19 DataFile Name : 017CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	49746.73	49962.40	49673.39	49794.17	0.30	ppb
Antimony	121-1	492.27	507.99	502.80	501.02	1.60	ppb
Arsenic	75-2	489.17	502.13	498.52	496.60	1.35	ppb
Barium	135-1	2560.98	2595.31	2559.67	2571.98	0.79	ppb
Barium	137-1	2549.68	2577.32	2530.49	2552.49	0.92	ppb
Beryllium	9-1	500.37	509.58	508.55	506.17	1.00	ppb
Bismuth	209-1				90		%
Bismuth	209-2				88		%
Bromine	81-1						cps
Cadmium	108-1	492.93	502.30	489.47	494.90	1.34	ppb
Cadmium	106-1	484.39	496.86	489.92	490.39	1.27	ppb
Cadmium	111-1	497.56	514.74	499.46	503.92	1.87	ppb
Calcium	43-1	261716.52	264345.47	260408.05	262156.68	0.76	ppb
Calcium	44-1	259926.82	258880.09	261089.02	259965.31	0.43	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	493.18	507.34	496.51	499.01	1.48	ppb
Cobalt	59-2	490.85	492.97	496.95	493.59	0.63	ppb
Copper	63-2	4796.25	4769.19	4751.78	4772.41	0.47	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				98		%
Holmium	165-2				99		%
Indium	115-1				94		%
Indium	115-2				93		%
Iron	54-2	125145.17	126385.42	125291.99	125607.53	0.54	ppb
Iron	56-2	124763.32	126027.15	123663.34	124817.94	0.95	ppb
Iron	57-2	126361.39	126526.43	125710.97	126199.60	0.34	ppb
Krypton	83-1						cps
Lead	206-1	2535.08	2592.49	2547.84	2558.47	1.18	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : CCV021 Instrumnet Name : P8
Client Sample ID : CCV021 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:47:19 DataFile Name : 017CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	2507.54	2587.80	2491.55	2528.96	2.04	ppb
Lead	208-1	2527.18	2594.66	2509.28	2543.71	1.77	ppb
Lithium	6-1				94		%
Magnesium	24-2	248815.27	248704.63	246424.06	247981.32	0.54	ppb
Manganese	55-2	4868.97	4970.08	4908.91	4915.99	1.04	ppb
Molybdenum	94-1	5019.61	5102.96	5111.91	5078.16	1.00	ppb
Molybdenum	95-1	4986.78	5107.91	5010.17	5034.95	1.28	ppb
Molybdenum	96-1	4953.94	5045.19	5056.66	5018.60	1.12	ppb
Molybdenum	97-1	4977.78	5008.55	5064.46	5016.93	0.88	ppb
Molybdenum	98-1	5011.44	5050.31	5029.28	5030.34	0.39	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	465.89	462.76	459.05	462.57	0.74	ppb
Phosphorus	31-2	10304.56	10225.77	10004.64	10178.32	1.53	ppb
Potassium	39-2	125522.40	124299.99	123606.73	124476.37	0.78	ppb
Rhodium	103-1				90		%
Rhodium	103-2				91		%
Scandium	45-1				98		%
Scandium	45-2				101		%
Selenium	82-1	498.03	495.27	497.46	496.92	0.29	ppb
Selenium	77-2	501.74	496.66	503.50	500.63	0.71	ppb
Selenium	78-2	482.50	496.57	493.30	490.79	1.50	ppb
Silicon	28-1	506.51	512.50	522.84	513.95	1.61	ppb
Silver	107-1	492.30	496.81	491.40	493.50	0.59	ppb
Silver	109-1	484.14	487.15	484.71	485.33	0.33	ppb
Sodium	23-2	259278.15	260931.82	253667.77	257959.25	1.48	ppb
Strontium	86-1	509.64	506.46	508.21	508.10	0.31	ppb
Strontium	88-1	495.64	503.77	504.26	501.22	0.97	ppb
Sulfur	34-1	9637.46	9749.82	9754.52	9713.93	0.68	ppb
Terbium	159-1				99		%
Terbium	159-2				97		%
Thallium	203-1	505.41	518.79	509.41	511.20	1.34	ppb
Thallium	205-1	505.47	516.84	515.98	512.76	1.23	ppb
Tin	118-1	501.13	501.98	503.51	502.21	0.24	ppb
Titanium	47-1	5053.52	5005.06	5030.24	5029.61	0.48	ppb
Uranium	238-1	513.31	521.68	506.84	513.94	1.45	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : CCV021 Instrumnet Name : P8
Client Sample ID : CCV021 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:47:19 DataFile Name : 017CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Vanadium	51-2	494.98	498.21	503.46	498.88	0.86	ppb
Ytterbium	172-1						cps
Ytterbium	172-2						cps
Ytterbium	176-1						cps
Ytterbium	176-2						cps
Yttrium	89-1				97		%
Yttrium	89-2				100		%
Zinc	66-2	4926.43	4849.19	4750.30	4841.97	1.82	ppb
Zirconium	90-1	503.35	513.17	511.35	509.29	1.03	ppb
Zirconium	91-1	509.81	520.64	501.30	510.58	1.90	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : CCB021 Instrumnet Name : P8
Client Sample ID : CCB021 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:50:25 DataFile Name : 018CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	1.57	0.55	0.74	0.95	56.70	ppb
Antimony	121-1	0.11	0.11	0.11	0.11	3.97	ppb
Arsenic	75-2	-0.02	0.03	-0.01	0.00	1376.36	ppb
Barium	135-1	0.05	0.07	0.02	0.05	53.82	ppb
Barium	137-1	0.07	0.06	0.06	0.07	13.62	ppb
Beryllium	9-1	0.14	0.12	0.11	0.12	12.79	ppb
Bismuth	209-1				103		%
Bismuth	209-2				102		%
Bromine	81-1						cps
Cadmium	108-1	0.09	0.09	0.00	0.06	89.63	ppb
Cadmium	106-1	-0.82	-0.90	-1.03	-0.92		ppb
Cadmium	111-1	-0.03	-0.05	-0.06	-0.05		ppb
Calcium	43-1	6.86	5.14	4.39	5.46	23.15	ppb
Calcium	44-1	2.75	2.16	1.42	2.11	31.72	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	-0.01	-0.02	-0.02	-0.01		ppb
Cobalt	59-2	-0.01	0.00	-0.01	0.00		ppb
Copper	63-2	-0.02	-0.07	-0.03	-0.04		ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				103		%
Indium	115-1				105		%
Indium	115-2				103		%
Iron	54-2	3.10	3.80	3.58	3.49	10.18	ppb
Iron	56-2	3.39	3.51	2.66	3.18	14.38	ppb
Iron	57-2	4.06	3.76	2.92	3.58	16.64	ppb
Krypton	83-1						cps
Lead	206-1	0.19	0.17	0.18	0.18	5.38	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : CCB021 Instrumnet Name : P8
Client Sample ID : CCB021 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:50:25 DataFile Name : 018CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	0.23	0.19	0.18	0.20	13.37	ppb
Lead	208-1	0.20	0.18	0.17	0.19	7.74	ppb
Lithium	6-1				106		%
Magnesium	24-2	-7.45	-7.79	-7.77	-7.67		ppb
Manganese	55-2	0.12	0.08	0.08	0.10	23.51	ppb
Molybdenum	94-1	0.19	0.14	0.16	0.17	14.97	ppb
Molybdenum	95-1	0.20	0.16	0.14	0.17	17.58	ppb
Molybdenum	96-1	0.18	0.15	0.15	0.16	12.28	ppb
Molybdenum	97-1	0.20	0.17	0.12	0.16	25.60	ppb
Molybdenum	98-1	0.21	0.16	0.15	0.17	18.77	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	-0.04	-0.01	-0.06	-0.04		ppb
Phosphorus	31-2	-14.10	-13.15	-9.90	-12.38		ppb
Potassium	39-2	23.30	23.49	17.96	21.59	14.55	ppb
Rhodium	103-1				106		%
Rhodium	103-2				105		%
Scandium	45-1				106		%
Scandium	45-2				106		%
Selenium	82-1	0.13	0.03	0.38	0.18	99.36	ppb
Selenium	77-2	1.46	-0.31	-0.31	0.28	363.03	ppb
Selenium	78-2	-0.27	-0.02	-0.27	-0.19		ppb
Silicon	28-1	0.51	0.53	0.28	0.44	31.16	ppb
Silver	107-1	0.05	0.05	0.04	0.05	6.99	ppb
Silver	109-1	0.05	0.05	0.05	0.05	8.31	ppb
Sodium	23-2	41.53	37.40	37.65	38.86	5.95	ppb
Strontium	86-1	-0.03	-0.03	-0.02	-0.03		ppb
Strontium	88-1	-0.03	-0.03	-0.03	-0.03		ppb
Sulfur	34-1	-895.39	-901.33	-837.69	-878.14		ppb
Terbium	159-1				103		%
Terbium	159-2				102		%
Thallium	203-1	0.07	0.06	0.05	0.06	14.39	ppb
Thallium	205-1	0.06	0.06	0.06	0.06	6.55	ppb
Tin	118-1	0.03	0.03	0.03	0.03	10.03	ppb
Titanium	47-1	0.14	0.08	0.05	0.09	51.00	ppb
Uranium	238-1	0.01	0.01	0.00	0.01	30.82	ppb

LB Number :

LB134674

Operator :

Jaswal

Lab Sample ID :

CCB021

Instrumnet Name :

P8

Client Sample ID :

CCB021

Dilution Factor :

1

Date & Time Acquired :

2025-02-11 13:50:25

DataFile Name :

018CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Vanadium	51-2	0.01	0.00	0.00	0.00	177.05	ppb
Ytterbium	172-1						cps
Ytterbium	172-2						cps
Ytterbium	176-1						cps
Ytterbium	176-2						cps
Yttrium	89-1				106		%
Yttrium	89-2				105		%
Zinc	66-2	0.24	0.28	0.14	0.22	32.21	ppb
Zirconium	90-1	0.02	0.02	0.02	0.02	4.02	ppb
Zirconium	91-1	0.02	0.02	0.02	0.02	6.96	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-01DLX2 Instrumnet Name : P8
Client Sample ID : YE8C9 Dilution Factor : 2
Date & Time Acquired : 2025-02-11 13:53:45 DataFile Name : 019AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	13023.36	14394.10	11685.31	13034.26	10.39	ppb
Antimony	121-1	0.30	0.31	0.32	0.31	2.84	ppb
Arsenic	75-2	30.52	32.85	28.60	30.66	6.93	ppb
Barium	135-1	122.54	121.85	124.77	123.05	1.24	ppb
Barium	137-1	124.03	122.86	126.67	124.52	1.57	ppb
Beryllium	9-1	0.54	0.56	0.58	0.56	3.14	ppb
Bismuth	209-1				102		%
Bismuth	209-2				92		%
Bromine	81-1						cps
Cadmium	108-1	0.24	0.32	0.35	0.30	17.89	ppb
Cadmium	106-1	-0.74	-2.10	-0.88	-1.24		ppb
Cadmium	111-1	0.03	-0.07	0.04	0.00		ppb
Calcium	43-1	3008.15	3068.24	3074.69	3050.36	1.20	ppb
Calcium	44-1	3071.52	3128.00	3114.64	3104.72	0.95	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	30.35	33.66	27.15	30.39	10.71	ppb
Cobalt	59-2	17.41	19.05	15.35	17.27	10.74	ppb
Copper	63-2	57.85	63.83	52.54	58.07	9.73	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				93		%
Indium	115-1				104		%
Indium	115-2				92		%
Iron	54-2	37361.70	40875.82	33179.24	37138.92	10.37	ppb
Iron	56-2	37468.64	40947.72	33117.73	37178.03	10.55	ppb
Iron	57-2	38292.27	42114.28	33816.90	38074.48	10.91	ppb
Krypton	83-1						cps
Lead	206-1	12.44	13.27	13.07	12.92	3.36	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-01DLX2 Instrumnet Name : P8
Client Sample ID : YE8C9 Dilution Factor : 2
Date & Time Acquired : 2025-02-11 13:53:45 DataFile Name : 019AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	11.66	12.26	12.29	12.07	2.95	ppb
Lead	208-1	11.98	12.69	12.60	12.43	3.12	ppb
Lithium	6-1				106		%
Magnesium	24-2	6468.64	7253.99	5886.28	6536.31	10.50	ppb
Manganese	55-2	281.46	307.38	248.21	279.02	10.63	ppb
Molybdenum	94-1	1.04	1.10	1.12	1.08	3.90	ppb
Molybdenum	95-1	0.43	0.44	0.46	0.44	3.32	ppb
Molybdenum	96-1	0.52	0.53	0.55	0.54	3.07	ppb
Molybdenum	97-1	0.44	0.43	0.47	0.45	5.50	ppb
Molybdenum	98-1	0.45	0.44	0.45	0.45	1.03	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	44.36	48.92	39.56	44.28	10.58	ppb
Phosphorus	31-2	363.48	392.51	331.15	362.38	8.47	ppb
Potassium	39-2	381.04	467.73	354.22	401.00	14.80	ppb
Rhodium	103-1				103		%
Rhodium	103-2				93		%
Scandium	45-1				106		%
Scandium	45-2				97		%
Selenium	82-1	0.28	0.69	0.44	0.47	43.97	ppb
Selenium	77-2	10.28	18.52	13.72	14.17	29.19	ppb
Selenium	78-2	3.67	3.10	2.81	3.19	13.78	ppb
Silicon	28-1	4646.41	4621.42	4555.06	4607.63	1.02	ppb
Silver	107-1	0.22	0.20	0.20	0.20	6.79	ppb
Silver	109-1	0.22	0.21	0.20	0.21	5.34	ppb
Sodium	23-2	95.23	120.94	74.20	96.79	24.19	ppb
Strontium	86-1	25.73	26.66	26.77	26.38	2.16	ppb
Strontium	88-1	25.76	26.41	27.11	26.42	2.56	ppb
Sulfur	34-1	-1264.00	-1287.51	-1350.87	-1300.79		ppb
Terbium	159-1				104		%
Terbium	159-2				91		%
Thallium	203-1	0.15	0.15	0.15	0.15	1.48	ppb
Thallium	205-1	0.15	0.16	0.15	0.15	1.92	ppb
Tin	118-1	0.06	0.04	0.04	0.05	30.26	ppb
Titanium	47-1	7.72	17.21	8.95	11.30	45.69	ppb
Uranium	238-1	0.22	0.23	0.24	0.23	4.52	ppb

LB Number :

LB134674

Operator :

Jaswal

Lab Sample ID :

Q1159-01DLX2

Instrumnet Name :

P8

Client Sample ID :

YE8C9

Dilution Factor :

2

Date & Time Acquired :

2025-02-11 13:53:45

DataFile Name :

019AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Vanadium	51-2	36.91	41.04	32.92	36.96	11.00	ppb
Ytterbium	172-1						cps
Ytterbium	172-2						cps
Ytterbium	176-1						cps
Ytterbium	176-2						cps
Yttrium	89-1				106		%
Yttrium	89-2				97		%
Zinc	66-2	79.68	87.61	71.45	79.58	10.15	ppb
Zirconium	90-1	0.42	0.52	0.40	0.45	14.05	ppb
Zirconium	91-1	0.45	0.43	0.44	0.44	2.86	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-02DLX2 Instrumnet Name : P8
Client Sample ID : YE8C9D Dilution Factor : 2
Date & Time Acquired : 2025-02-11 13:57:03 DataFile Name : 020AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	11561.18	11886.15	11611.28	11686.20	1.50	ppb
Antimony	121-1	0.29	0.31	0.29	0.29	4.87	ppb
Arsenic	75-2	27.26	27.31	27.30	27.29	0.09	ppb
Barium	135-1	122.65	123.90	123.83	123.46	0.57	ppb
Barium	137-1	124.22	125.41	126.74	125.46	1.00	ppb
Beryllium	9-1	0.53	0.54	0.53	0.53	0.86	ppb
Bismuth	209-1				101		%
Bismuth	209-2				102		%
Bromine	81-1						cps
Cadmium	108-1	0.32	0.27	0.31	0.30	8.46	ppb
Cadmium	106-1	-0.95	-1.07	-0.33	-0.78		ppb
Cadmium	111-1	0.02	-0.02	0.07	0.02	170.11	ppb
Calcium	43-1	2972.83	2977.04	3038.29	2996.06	1.22	ppb
Calcium	44-1	3064.28	3107.29	3060.81	3077.46	0.84	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	26.72	27.54	26.92	27.06	1.58	ppb
Cobalt	59-2	15.27	15.65	15.16	15.36	1.67	ppb
Copper	63-2	51.93	53.12	51.96	52.34	1.30	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				103		%
Indium	115-1				104		%
Indium	115-2				103		%
Iron	54-2	33969.75	34104.99	33496.20	33856.98	0.94	ppb
Iron	56-2	33247.33	33763.27	33265.81	33425.47	0.88	ppb
Iron	57-2	33419.82	34680.30	33739.07	33946.40	1.93	ppb
Krypton	83-1						cps
Lead	206-1	12.93	13.00	12.83	12.92	0.67	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-02DLX2 Instrumnet Name : P8
Client Sample ID : YE8C9D Dilution Factor : 2
Date & Time Acquired : 2025-02-11 13:57:03 DataFile Name : 020AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	12.21	12.32	12.00	12.18	1.36	ppb
Lead	208-1	12.50	12.61	12.33	12.48	1.12	ppb
Lithium	6-1				107		%
Magnesium	24-2	5759.46	5886.64	5713.79	5786.63	1.55	ppb
Manganese	55-2	247.83	252.15	247.37	249.11	1.06	ppb
Molybdenum	94-1	1.08	1.08	1.07	1.08	0.73	ppb
Molybdenum	95-1	0.44	0.43	0.41	0.43	3.08	ppb
Molybdenum	96-1	0.60	0.51	0.53	0.55	7.97	ppb
Molybdenum	97-1	0.43	0.40	0.43	0.42	4.68	ppb
Molybdenum	98-1	0.43	0.42	0.45	0.43	3.51	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	39.02	40.07	39.13	39.41	1.46	ppb
Phosphorus	31-2	284.44	294.45	281.87	286.92	2.32	ppb
Potassium	39-2	342.05	334.95	326.08	334.36	2.39	ppb
Rhodium	103-1				106		%
Rhodium	103-2				104		%
Scandium	45-1				108		%
Scandium	45-2				108		%
Selenium	82-1	0.65	0.57	0.57	0.60	7.54	ppb
Selenium	77-2	21.14	16.38	13.33	16.95	23.22	ppb
Selenium	78-2	5.00	3.59	2.72	3.77	30.53	ppb
Silicon	28-1	4436.20	4555.18	4722.87	4571.42	3.15	ppb
Silver	107-1	0.17	0.18	0.19	0.18	5.40	ppb
Silver	109-1	0.16	0.18	0.19	0.17	8.56	ppb
Sodium	23-2	57.58	63.40	58.45	59.81	5.24	ppb
Strontium	86-1	26.33	26.30	26.34	26.32	0.08	ppb
Strontium	88-1	26.22	26.10	26.30	26.20	0.38	ppb
Sulfur	34-1	-1856.16	-1819.99	-1776.82	-1817.66		ppb
Terbium	159-1				104		%
Terbium	159-2				102		%
Thallium	203-1	0.13	0.14	0.15	0.14	8.79	ppb
Thallium	205-1	0.12	0.13	0.15	0.14	10.27	ppb
Tin	118-1	0.04	0.03	0.05	0.04	20.09	ppb
Titanium	47-1	10.97	9.88	10.81	10.55	5.54	ppb
Uranium	238-1	0.23	0.24	0.23	0.23	2.32	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-02DLX2 Instrumnet Name : P8
Client Sample ID : YE8C9D Dilution Factor : 2
Date & Time Acquired : 2025-02-11 13:57:03 DataFile Name : 020AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Vanadium	51-2	33.13	33.71	32.46	33.10	1.89	ppb
Ytterbium	172-1						cps
Ytterbium	172-2						cps
Ytterbium	176-1						cps
Ytterbium	176-2						cps
Yttrium	89-1				108		%
Yttrium	89-2				108		%
Zinc	66-2	69.55	72.57	70.08	70.73	2.28	ppb
Zirconium	90-1	0.36	0.39	0.39	0.38	5.30	ppb
Zirconium	91-1	2.01	0.39	0.40	0.93	99.83	ppb

LB Number :	LB134674	Operator :	Jaswal
Lab Sample ID :	Q1159-01LDLX10	Instrumnet Name :	P8
Client Sample ID :	YE8C9L	Dilution Factor :	10
Date & Time Acquired :	2025-02-11 14:00:19	DataFile Name :	021AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	2400.28	2390.35	2349.19	2379.94	1.14	ppb
Antimony	121-1	0.06	0.07	0.06	0.06	10.99	ppb
Arsenic	75-2	5.55	5.50	5.42	5.49	1.15	ppb
Barium	135-1	25.07	25.01	25.12	25.07	0.22	ppb
Barium	137-1	25.29	25.32	25.25	25.28	0.15	ppb
Beryllium	9-1	0.11	0.14	0.12	0.12	10.04	ppb
Bismuth	209-1				101		%
Bismuth	209-2				101		%
Bromine	81-1						cps
Cadmium	108-1	0.03	0.11	0.06	0.06	61.76	ppb
Cadmium	106-1	-1.25	-2.51	-1.39	-1.72		ppb
Cadmium	111-1	-0.06	-0.18	-0.09	-0.11		ppb
Calcium	43-1	603.43	613.59	614.48	610.50	1.01	ppb
Calcium	44-1	598.40	604.87	602.29	601.85	0.54	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	5.43	5.50	5.53	5.49	0.90	ppb
Cobalt	59-2	3.18	3.16	3.08	3.14	1.74	ppb
Copper	63-2	10.63	10.64	10.45	10.57	1.01	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				103		%
Indium	115-1				104		%
Indium	115-2				103		%
Iron	54-2	6818.56	6734.57	6797.43	6783.52	0.64	ppb
Iron	56-2	6943.74	6870.90	6739.86	6851.50	1.51	ppb
Iron	57-2	6833.96	6719.21	6704.60	6752.59	1.05	ppb
Krypton	83-1						cps
Lead	206-1	2.68	2.69	2.71	2.69	0.53	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-01LDLX10 Instrumnet Name : P8
Client Sample ID : YE8C9L Dilution Factor : 10
Date & Time Acquired : 2025-02-11 14:00:19 DataFile Name : 021AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	2.60	2.55	2.61	2.59	1.13	ppb
Lead	208-1	2.64	2.60	2.65	2.63	0.92	ppb
Lithium	6-1				106		%
Magnesium	24-2	1155.73	1165.48	1147.05	1156.09	0.80	ppb
Manganese	55-2	51.37	52.55	50.56	51.49	1.94	ppb
Molybdenum	94-1	0.21	0.24	0.26	0.24	10.05	ppb
Molybdenum	95-1	0.09	0.09	0.09	0.09	2.58	ppb
Molybdenum	96-1	0.12	0.10	0.10	0.10	9.12	ppb
Molybdenum	97-1	0.11	0.08	0.09	0.09	13.91	ppb
Molybdenum	98-1	0.10	0.09	0.10	0.09	6.03	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	8.03	7.98	8.00	8.00	0.36	ppb
Phosphorus	31-2	52.51	55.53	52.81	53.61	3.11	ppb
Potassium	39-2	73.45	70.78	73.02	72.42	1.98	ppb
Rhodium	103-1				106		%
Rhodium	103-2				104		%
Scandium	45-1				107		%
Scandium	45-2				105		%
Selenium	82-1	0.13	0.14	0.16	0.14	8.81	ppb
Selenium	77-2	1.45	1.43	1.42	1.43	1.05	ppb
Selenium	78-2	1.27	0.24	0.99	0.84	64.18	ppb
Silicon	28-1	1072.16	925.27	921.76	973.06	8.82	ppb
Silver	107-1	0.03	0.04	0.03	0.03	9.23	ppb
Silver	109-1	0.04	0.04	0.04	0.04	4.44	ppb
Sodium	23-2	19.72	18.98	15.41	18.04	12.78	ppb
Strontium	86-1	5.26	5.37	5.32	5.32	1.03	ppb
Strontium	88-1	5.26	5.33	5.35	5.31	0.88	ppb
Sulfur	34-1	-2065.35	-2018.08	-2105.45	-2062.96		ppb
Terbium	159-1				103		%
Terbium	159-2				101		%
Thallium	203-1	0.04	0.04	0.04	0.04	7.85	ppb
Thallium	205-1	0.04	0.04	0.04	0.04	8.44	ppb
Tin	118-1	-0.03	-0.05	-0.05	-0.04		ppb
Titanium	47-1	1.31	1.62	1.92	1.62	18.81	ppb
Uranium	238-1	0.04	0.04	0.04	0.04	2.83	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-01LDLX10 Instrumnet Name : P8
Client Sample ID : YE8C9L Dilution Factor : 10
Date & Time Acquired : 2025-02-11 14:00:19 DataFile Name : 021AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Vanadium	51-2	6.67	6.66	6.91	6.75	2.12	ppb
Ytterbium	172-1						cps
Ytterbium	172-2						cps
Ytterbium	176-1						cps
Ytterbium	176-2						cps
Yttrium	89-1				106		%
Yttrium	89-2				107		%
Zinc	66-2	14.79	15.05	14.50	14.78	1.86	ppb
Zirconium	90-1	0.07	0.07	0.08	0.07	5.33	ppb
Zirconium	91-1	0.16	0.09	0.08	0.11	35.29	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-03DLX2 Instrumnet Name : P8
Client Sample ID : YE8C9S Dilution Factor : 2
Date & Time Acquired : 2025-02-11 14:03:40 DataFile Name : 022AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	11817.93	11598.56	11695.75	11704.08	0.94	ppb
Antimony	121-1	10.63	10.66	10.82	10.71	0.95	ppb
Arsenic	75-2	31.28	29.60	30.06	30.32	2.87	ppb
Barium	135-1	319.70	316.23	324.64	320.19	1.32	ppb
Barium	137-1	333.76	325.61	340.15	333.17	2.19	ppb
Beryllium	9-1	5.24	5.52	5.38	5.38	2.53	ppb
Bismuth	209-1				101		%
Bismuth	209-2				102		%
Bromine	81-1						cps
Cadmium	108-1	4.80	5.06	4.75	4.87	3.47	ppb
Cadmium	106-1	2.74	2.31	2.32	2.46	10.00	ppb
Cadmium	111-1	5.21	5.01	5.18	5.13	2.07	ppb
Calcium	43-1	3469.65	3419.18	3466.52	3451.78	0.82	ppb
Calcium	44-1	3388.72	3342.93	3337.58	3356.41	0.84	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	46.23	45.55	46.60	46.13	1.15	ppb
Cobalt	59-2	65.59	63.55	65.21	64.78	1.68	ppb
Copper	63-2	76.84	75.67	75.73	76.08	0.87	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				104		%
Indium	115-1				103		%
Indium	115-2				104		%
Iron	54-2	33676.35	32324.28	32977.96	32992.86	2.05	ppb
Iron	56-2	32908.36	32418.20	32906.12	32744.23	0.86	ppb
Iron	57-2	33510.95	32992.49	32868.36	33123.93	1.03	ppb
Krypton	83-1						cps
Lead	206-1	14.74	14.72	14.85	14.77	0.49	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-03DLX2 Instrumnet Name : P8
Client Sample ID : YE8C9S Dilution Factor : 2
Date & Time Acquired : 2025-02-11 14:03:40 DataFile Name : 022AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	14.07	13.94	14.07	14.03	0.52	ppb
Lead	208-1	14.32	14.26	14.32	14.30	0.26	ppb
Lithium	6-1				107		%
Magnesium	24-2	5828.44	5634.15	5669.03	5710.54	1.81	ppb
Manganese	55-2	295.45	288.16	294.65	292.75	1.37	ppb
Molybdenum	94-1	42.43	42.64	43.55	42.87	1.39	ppb
Molybdenum	95-1	50.52	50.99	51.69	51.06	1.16	ppb
Molybdenum	96-1	49.31	50.27	50.70	50.09	1.42	ppb
Molybdenum	97-1	50.99	51.28	52.13	51.46	1.15	ppb
Molybdenum	98-1	50.43	51.31	51.35	51.03	1.02	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	89.04	87.19	88.84	88.36	1.15	ppb
Phosphorus	31-2	307.28	321.91	335.74	321.64	4.42	ppb
Potassium	39-2	328.75	320.98	317.06	322.26	1.85	ppb
Rhodium	103-1				106		%
Rhodium	103-2				104		%
Scandium	45-1				108		%
Scandium	45-2				108		%
Selenium	82-1	2.50	2.28	2.44	2.40	4.63	ppb
Selenium	77-2	16.51	16.57	18.74	17.27	7.35	ppb
Selenium	78-2	4.65	3.68	5.05	4.46	15.81	ppb
Silicon	28-1	4814.84	4428.17	4447.94	4563.65	4.77	ppb
Silver	107-1	5.09	5.12	5.22	5.14	1.27	ppb
Silver	109-1	5.10	5.08	5.27	5.15	2.06	ppb
Sodium	23-2	51.34	48.19	52.67	50.73	4.54	ppb
Strontium	86-1	1013.38	1013.92	1008.09	1011.80	0.32	ppb
Strontium	88-1	997.55	1000.36	1009.06	1002.32	0.60	ppb
Sulfur	34-1	-2288.41	-2352.42	-2366.66	-2335.83		ppb
Terbium	159-1				104		%
Terbium	159-2				102		%
Thallium	203-1	4.90	4.98	5.01	4.96	1.12	ppb
Thallium	205-1	4.98	5.01	5.02	5.00	0.35	ppb
Tin	118-1	0.02	0.00	0.01	0.01	80.18	ppb
Titanium	47-1	7.87	14.07	18.01	13.32	38.38	ppb
Uranium	238-1	0.23	0.24	0.23	0.23	1.30	ppb

LB Number :

LB134674

Operator :

Jaswal

Lab Sample ID :

Q1159-03DLX2

Instrumnet Name :

P8

Client Sample ID :

YE8C9S

Dilution Factor :

2

Date & Time Acquired :

2025-02-11 14:03:40

DataFile Name :

022AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Vanadium	51-2	80.97	79.34	81.19	80.50	1.26	ppb
Ytterbium	172-1						cps
Ytterbium	172-2						cps
Ytterbium	176-1						cps
Ytterbium	176-2						cps
Yttrium	89-1				107		%
Yttrium	89-2				107		%
Zinc	66-2	120.38	116.92	119.63	118.98	1.53	ppb
Zirconium	90-1	0.39	0.38	0.41	0.40	3.64	ppb
Zirconium	91-1	0.41	0.41	0.64	0.49	27.50	ppb

LB Number :LB134674Operator :Jaswal

Lab Sample ID :PB166570BLInstrumnet Name :P8

Client Sample ID :PBW570Dilution Factor :1

Date & Time Acquired :2025-02-11 14:15:28DataFile Name :023CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	-0.44	-0.06	-0.57	-0.36		ppb
Antimony	121-1	0.01	0.00	0.00	0.01	27.52	ppb
Arsenic	75-2	-0.01	0.02	-0.02	0.00		ppb
Barium	135-1	-0.01	-0.01	-0.02	-0.02		ppb
Barium	137-1	-0.01	-0.01	-0.01	-0.01		ppb
Beryllium	9-1	0.00	0.00	0.00	0.00	177.07	ppb
Bismuth	209-1				103		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Cadmium	108-1	0.04	0.04	0.04	0.04	3.00	ppb
Cadmium	106-1	-1.04	-2.27	-0.58	-1.30		ppb
Cadmium	111-1	-0.09	-0.18	-0.05	-0.10		ppb
Calcium	43-1	-1.33	0.21	-0.47	-0.53		ppb
Calcium	44-1	-5.67	-5.48	-6.31	-5.82		ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	-0.02	0.00	-0.02	-0.01		ppb
Cobalt	59-2	0.00	-0.01	0.00	0.00		ppb
Copper	63-2	-0.17	-0.19	-0.20	-0.19		ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				104		%
Indium	115-1				105		%
Indium	115-2				104		%
Iron	54-2	0.95	0.99	0.80	0.91	10.67	ppb
Iron	56-2	0.70	0.67	0.59	0.65	8.86	ppb
Iron	57-2	0.29	1.74	1.54	1.19	65.91	ppb
Krypton	83-1						cps
Lead	206-1	0.04	0.03	0.04	0.04	11.08	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : PB166570BL Instrumnet Name : P8
Client Sample ID : PBW570 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:15:28 DataFile Name : 023CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	0.04	0.04	0.03	0.03	17.37	ppb
Lead	208-1	0.04	0.04	0.04	0.04	7.15	ppb
Lithium	6-1				107		%
Magnesium	24-2	-6.62	-7.13	-6.81	-6.85		ppb
Manganese	55-2	-0.01	0.00	0.00	-0.01		ppb
Molybdenum	94-1	0.00	0.01	0.01	0.01	28.98	ppb
Molybdenum	95-1	0.00	0.01	0.00	0.00	75.43	ppb
Molybdenum	96-1	0.00	0.01	-0.01	0.00		ppb
Molybdenum	97-1	0.00	-0.01	0.00	0.00		ppb
Molybdenum	98-1	0.00	0.00	0.00	0.00	49.42	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	-0.05	-0.06	-0.06	-0.06		ppb
Phosphorus	31-2	-7.49	-19.80	-24.33	-17.21		ppb
Potassium	39-2	3.85	1.80	3.12	2.92	35.55	ppb
Rhodium	103-1				105		%
Rhodium	103-2				105		%
Scandium	45-1				107		%
Scandium	45-2				107		%
Selenium	82-1	0.17	0.00	0.39	0.19	104.02	ppb
Selenium	77-2	-0.31	-0.31	-0.31	-0.31		ppb
Selenium	78-2	0.98	-0.27	0.99	0.57	128.24	ppb
Silicon	28-1	-0.46	-1.52	-1.63	-1.20		ppb
Silver	107-1	0.00	0.00	0.00	0.00	747.58	ppb
Silver	109-1	0.00	0.00	0.00	0.00	33.98	ppb
Sodium	23-2	3.59	-0.52	-0.45	0.87	268.83	ppb
Strontium	86-1	-0.04	-0.05	-0.04	-0.04		ppb
Strontium	88-1	-0.04	-0.04	-0.03	-0.03		ppb
Sulfur	34-1	-1377.16	-1655.54	-1520.32	-1517.67		ppb
Terbium	159-1				104		%
Terbium	159-2				102		%
Thallium	203-1	0.02	0.02	0.02	0.02	8.36	ppb
Thallium	205-1	0.02	0.02	0.02	0.02	2.03	ppb
Tin	118-1	-0.01	-0.02	-0.01	-0.02		ppb
Titanium	47-1	-0.06	-0.04	-0.08	-0.06		ppb
Uranium	238-1	0.00	0.00	0.00	0.00		ppb

LB Number :

LB134674

Operator :

Jaswal

Lab Sample ID :

PB166570BL

Instrumnet Name :

P8

Client Sample ID :

PBW570

Dilution Factor :

1

Date & Time Acquired :

2025-02-11 14:15:28

DataFile Name :

023CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Vanadium	51-2	0.00	0.00	0.00	0.00		ppb
Ytterbium	172-1						cps
Ytterbium	172-2						cps
Ytterbium	176-1						cps
Ytterbium	176-2						cps
Yttrium	89-1				104		%
Yttrium	89-2				108		%
Zinc	66-2	0.13	0.00	-0.02	0.04	230.14	ppb
Zirconium	90-1	0.00	0.00	0.00	0.00	210.07	ppb
Zirconium	91-1	0.01	0.00	0.00	0.01	97.99	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : PB166570BS Instrumnet Name : P8
Client Sample ID : LCS570 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:18:49 DataFile Name : 024LCSE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	41.27	41.94	40.43	41.22	1.83	ppb
Antimony	121-1	4.33	4.44	4.32	4.36	1.52	ppb
Arsenic	75-2	2.29	2.31	2.31	2.30	0.50	ppb
Barium	135-1	20.75	21.44	21.01	21.07	1.64	ppb
Barium	137-1	20.69	21.35	21.13	21.06	1.59	ppb
Beryllium	9-1	2.15	2.18	2.13	2.15	1.08	ppb
Bismuth	209-1				102		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Cadmium	108-1	2.65	2.30	2.43	2.46	7.11	ppb
Cadmium	106-1	1.04	1.23	2.02	1.43	36.25	ppb
Cadmium	111-1	2.09	2.17	2.23	2.17	3.38	ppb
Calcium	43-1	1074.02	1108.22	1084.79	1089.01	1.61	ppb
Calcium	44-1	1048.57	1074.68	1066.03	1063.09	1.25	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	3.95	4.11	4.07	4.04	1.99	ppb
Cobalt	59-2	2.11	2.18	2.16	2.15	1.65	ppb
Copper	63-2	4.52	4.58	4.59	4.56	0.90	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				104		%
Indium	115-1				105		%
Indium	115-2				105		%
Iron	54-2	426.73	438.04	442.64	435.80	1.88	ppb
Iron	56-2	426.95	434.63	438.82	433.47	1.39	ppb
Iron	57-2	435.76	431.17	440.14	435.69	1.03	ppb
Krypton	83-1						cps
Lead	206-1	1.98	2.09	2.07	2.05	2.93	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : PB166570BS Instrumnet Name : P8
Client Sample ID : LCS570 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:18:49 DataFile Name : 024LCSE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	2.04	2.06	2.16	2.09	3.06	ppb
Lead	208-1	2.01	2.09	2.09	2.07	2.21	ppb
Lithium	6-1				107		%
Magnesium	24-2	988.05	1005.11	1021.90	1005.02	1.68	ppb
Manganese	55-2	2.04	2.07	2.11	2.07	1.68	ppb
Molybdenum	94-1	21.14	21.64	21.79	21.52	1.57	ppb
Molybdenum	95-1	21.27	21.91	21.84	21.67	1.62	ppb
Molybdenum	96-1	21.10	21.74	21.67	21.50	1.62	ppb
Molybdenum	97-1	21.43	22.10	22.02	21.85	1.68	ppb
Molybdenum	98-1	20.97	21.66	21.72	21.45	1.93	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	2.03	2.00	2.16	2.06	4.24	ppb
Phosphorus	31-2	37.48	48.04	29.93	38.48	23.65	ppb
Potassium	39-2	1007.03	1024.04	1046.03	1025.70	1.91	ppb
Rhodium	103-1				106		%
Rhodium	103-2				105		%
Scandium	45-1				107		%
Scandium	45-2				108		%
Selenium	82-1	10.37	11.04	10.92	10.78	3.34	ppb
Selenium	77-2	10.76	6.73	7.66	8.38	25.20	ppb
Selenium	78-2	12.20	13.12	11.40	12.24	7.01	ppb
Silicon	28-1	22.33	23.73	23.24	23.10	3.08	ppb
Silver	107-1	2.11	2.20	2.18	2.16	2.22	ppb
Silver	109-1	2.13	2.19	2.16	2.16	1.50	ppb
Sodium	23-2	1047.44	1061.84	1096.05	1068.44	2.34	ppb
Strontium	86-1	3.55	3.74	3.73	3.68	2.86	ppb
Strontium	88-1	3.61	3.72	3.74	3.69	1.93	ppb
Sulfur	34-1	-1344.14	-1162.81	-1173.22	-1226.72		ppb
Terbium	159-1				104		%
Terbium	159-2				103		%
Thallium	203-1	2.00	2.04	2.08	2.04	1.92	ppb
Thallium	205-1	2.08	2.06	2.11	2.08	1.23	ppb
Tin	118-1	10.53	10.92	10.83	10.76	1.87	ppb
Titanium	47-1	2.49	2.48	2.49	2.48	0.22	ppb
Uranium	238-1	1.85	1.91	1.90	1.89	1.82	ppb

LB Number :

LB134674

Operator :

Jaswal

Lab Sample ID :

PB166570BS

Instrumnet Name :

P8

Client Sample ID :

LCS570

Dilution Factor :

1

Date & Time Acquired :

2025-02-11 14:18:49

DataFile Name :

024LCSE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Vanadium	51-2	10.50	10.55	10.69	10.58	0.92	ppb
Ytterbium	172-1						cps
Ytterbium	172-2						cps
Ytterbium	176-1						cps
Ytterbium	176-2						cps
Yttrium	89-1				106		%
Yttrium	89-2				106		%
Zinc	66-2	10.84	10.69	10.44	10.66	1.87	ppb
Zirconium	90-1	1.98	2.08	2.02	2.03	2.36	ppb
Zirconium	91-1	1.95	2.03	2.00	1.99	2.05	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1135-09 Instrumnet Name : P8
Client Sample ID : YE8H7 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:22:18 DataFile Name : 025AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	0.92	0.71	0.43	0.69	35.58	ppb
Antimony	121-1	0.01	0.01	0.01	0.01	9.62	ppb
Arsenic	75-2	-0.02	-0.01	0.02	0.00		ppb
Barium	135-1	0.01	0.03	0.03	0.02	59.93	ppb
Barium	137-1	0.02	0.01	0.02	0.02	36.15	ppb
Beryllium	9-1	0.00	0.00	0.00	0.00		ppb
Bismuth	209-1				106		%
Bismuth	209-2				106		%
Bromine	81-1						cps
Cadmium	108-1	0.04	-0.01	0.02	0.02	158.61	ppb
Cadmium	106-1	-1.57	-1.07	-0.66	-1.10		ppb
Cadmium	111-1	-0.12	-0.08	-0.05	-0.08		ppb
Calcium	43-1	5.89	6.03	4.72	5.55	12.95	ppb
Calcium	44-1	-0.70	-0.97	-0.61	-0.76		ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	0.20	0.21	0.20	0.20	3.59	ppb
Cobalt	59-2	-0.02	-0.01	-0.01	-0.01		ppb
Copper	63-2	-0.18	-0.18	-0.18	-0.18		ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				105		%
Holmium	165-2				105		%
Indium	115-1				108		%
Indium	115-2				107		%
Iron	54-2	1.60	1.62	2.71	1.98	32.12	ppb
Iron	56-2	1.46	1.26	1.42	1.38	7.95	ppb
Iron	57-2	1.69	1.05	0.71	1.15	43.11	ppb
Krypton	83-1						cps
Lead	206-1	0.15	0.11	0.09	0.11	26.33	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1135-09 Instrumnet Name : P8
Client Sample ID : YE8H7 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:22:18 DataFile Name : 025AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	0.15	0.11	0.09	0.12	25.40	ppb
Lead	208-1	0.15	0.10	0.09	0.11	25.54	ppb
Lithium	6-1				110		%
Magnesium	24-2	-6.21	-6.11	-4.80	-5.71		ppb
Manganese	55-2	0.36	0.32	0.44	0.37	17.13	ppb
Molybdenum	94-1	-0.02	0.00	0.00	-0.01		ppb
Molybdenum	95-1	0.00	0.00	0.00	0.00		ppb
Molybdenum	96-1	0.00	0.00	0.00	0.00		ppb
Molybdenum	97-1	0.00	-0.01	0.00	0.00		ppb
Molybdenum	98-1	0.00	0.00	0.00	0.00		ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	-0.08	-0.08	-0.05	-0.07		ppb
Phosphorus	31-2	-5.50	-8.36	-5.50	-6.45		ppb
Potassium	39-2	-0.07	3.15	1.86	1.64	98.63	ppb
Rhodium	103-1				109		%
Rhodium	103-2				106		%
Scandium	45-1				109		%
Scandium	45-2				110		%
Selenium	82-1	0.11	-0.19	0.36	0.09	296.96	ppb
Selenium	77-2	0.55	-0.31	-0.31	-0.02		ppb
Selenium	78-2	-0.02	0.23	-0.27	-0.02		ppb
Silicon	28-1	9.11	9.60	10.48	9.73	7.17	ppb
Silver	107-1	0.01	0.01	0.01	0.01	22.27	ppb
Silver	109-1	0.01	0.01	0.01	0.01	20.21	ppb
Sodium	23-2	16.24	13.37	16.33	15.31	11.01	ppb
Strontium	86-1	-0.02	0.01	0.03	0.01	302.91	ppb
Strontium	88-1	0.03	0.02	0.02	0.02	6.22	ppb
Sulfur	34-1	-1364.05	-1398.69	-1313.04	-1358.59		ppb
Terbium	159-1				108		%
Terbium	159-2				105		%
Thallium	203-1	0.03	0.02	0.01	0.02	29.65	ppb
Thallium	205-1	0.03	0.02	0.01	0.02	31.39	ppb
Tin	118-1	-0.02	-0.01	-0.01	-0.01		ppb
Titanium	47-1	-0.13	-0.14	-0.15	-0.14		ppb
Uranium	238-1	0.00	0.00	0.00	0.00		ppb

LB Number :

LB134674

Operator :

Jaswal

Lab Sample ID :

Q1135-09

Instrumnet Name :

P8

Client Sample ID :

YE8H7

Dilution Factor :

1

Date & Time Acquired :

2025-02-11 14:22:18

DataFile Name :

025AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Vanadium	51-2	0.00	0.00	0.00	0.00		ppb
Ytterbium	172-1						cps
Ytterbium	172-2						cps
Ytterbium	176-1						cps
Ytterbium	176-2						cps
Yttrium	89-1				110		%
Yttrium	89-2				107		%
Zinc	66-2	0.06	0.04	0.07	0.06	23.92	ppb
Zirconium	90-1	0.00	0.00	0.00	0.00	455.02	ppb
Zirconium	91-1	0.00	0.00	0.00	0.00	685.88	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : PB166317BL Instrumnet Name : P8
Client Sample ID : PBW317 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:25:40 DataFile Name : 026CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	-0.55	-0.39	-0.10	-0.35		ppb
Antimony	121-1	0.00	0.00	0.00	0.00	27.86	ppb
Arsenic	75-2	0.01	-0.03	-0.04	-0.02		ppb
Barium	135-1	-0.03	0.00	-0.02	-0.02		ppb
Barium	137-1	0.00	-0.01	-0.01	-0.01		ppb
Beryllium	9-1	-0.01	-0.01	0.00	-0.01		ppb
Bismuth	209-1				102		%
Bismuth	209-2				102		%
Bromine	81-1						cps
Cadmium	108-1	0.06	0.03	0.00	0.03	106.61	ppb
Cadmium	106-1	-0.32	-1.94	-1.65	-1.31		ppb
Cadmium	111-1	-0.03	-0.15	-0.13	-0.10		ppb
Calcium	43-1	0.71	-0.96	-0.85	-0.37		ppb
Calcium	44-1	-6.00	-6.37	-6.05	-6.14		ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	0.00	-0.02	-0.03	-0.02		ppb
Cobalt	59-2	0.00	-0.01	-0.01	-0.01		ppb
Copper	63-2	-0.19	-0.18	-0.16	-0.18		ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				101		%
Holmium	165-2				102		%
Indium	115-1				104		%
Indium	115-2				103		%
Iron	54-2	0.62	0.60	0.74	0.66	11.65	ppb
Iron	56-2	0.26	0.16	0.33	0.25	34.51	ppb
Iron	57-2	-0.46	0.47	0.71	0.24	259.86	ppb
Krypton	83-1						cps
Lead	206-1	0.00	0.00	-0.01	0.00		ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : PB166317BL Instrumnet Name : P8
Client Sample ID : PBW317 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:25:40 DataFile Name : 026CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	0.00	0.00	0.00	0.00	431.31	ppb
Lead	208-1	-0.01	0.00	0.00	0.00		ppb
Lithium	6-1				107		%
Magnesium	24-2	-6.23	-4.83	-5.74	-5.60		ppb
Manganese	55-2	-0.02	-0.04	-0.05	-0.04		ppb
Molybdenum	94-1	0.01	0.00	0.00	0.00	108.41	ppb
Molybdenum	95-1	0.00	0.00	0.00	0.00	123.60	ppb
Molybdenum	96-1	0.00	0.00	0.00	0.00		ppb
Molybdenum	97-1	0.00	0.00	0.00	0.00		ppb
Molybdenum	98-1	0.00	0.01	0.00	0.00	40.48	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	-0.10	-0.08	-0.13	-0.10		ppb
Phosphorus	31-2	-12.00	-15.04	-17.47	-14.84		ppb
Potassium	39-2	-2.26	4.77	1.38	1.29	271.38	ppb
Rhodium	103-1				105		%
Rhodium	103-2				104		%
Scandium	45-1				105		%
Scandium	45-2				105		%
Selenium	82-1	-0.12	0.11	-0.11	-0.04		ppb
Selenium	77-2	-0.31	1.43	-0.31	0.27	369.72	ppb
Selenium	78-2	0.76	0.75	0.25	0.59	49.33	ppb
Silicon	28-1	0.81	0.53	0.90	0.75	25.56	ppb
Silver	107-1	0.00	0.00	0.00	0.00	80.06	ppb
Silver	109-1	0.00	0.00	0.00	0.00	18.23	ppb
Sodium	23-2	-2.97	-0.92	-2.43	-2.11		ppb
Strontium	86-1	-0.04	-0.04	-0.03	-0.04		ppb
Strontium	88-1	-0.03	-0.03	-0.03	-0.03		ppb
Sulfur	34-1	-784.35	-839.37	-769.86	-797.86		ppb
Terbium	159-1				103		%
Terbium	159-2				102		%
Thallium	203-1	0.01	0.01	0.01	0.01	37.44	ppb
Thallium	205-1	0.01	0.00	0.00	0.00	36.05	ppb
Tin	118-1	-0.01	-0.01	-0.02	-0.01		ppb
Titanium	47-1	-0.04	-0.06	-0.06	-0.05		ppb
Uranium	238-1	0.00	0.00	0.00	0.00		ppb

LB Number :

LB134674

Operator :

Jaswal

Lab Sample ID :

PB166317BL

Instrumnet Name :

P8

Client Sample ID :

PBW317

Dilution Factor :

1

Date & Time Acquired :

2025-02-11 14:25:40

DataFile Name :

026CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Vanadium	51-2	0.00	0.00	0.00	0.00	388.11	ppb
Ytterbium	172-1						cps
Ytterbium	172-2						cps
Ytterbium	176-1						cps
Ytterbium	176-2						cps
Yttrium	89-1				105		%
Yttrium	89-2				105		%
Zinc	66-2	0.11	0.08	-0.02	0.06	120.33	ppb
Zirconium	90-1	0.00	0.00	0.00	0.00	37.00	ppb
Zirconium	91-1	0.00	0.00	0.00	0.00		ppb

LB Number :	LB134674	Operator :	Jaswal
Lab Sample ID :	PB166317BS	Instrumnet Name :	P8
Client Sample ID :	LCS317	Dilution Factor :	1
Date & Time Acquired :	2025-02-11 14:29:05	DataFile Name :	027LCSE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	42.31	42.79	44.10	43.07	2.15	ppb
Antimony	121-1	4.26	4.25	4.41	4.30	2.03	ppb
Arsenic	75-2	2.23	2.25	1.98	2.16	7.12	ppb
Barium	135-1	20.53	20.66	21.20	20.80	1.72	ppb
Barium	137-1	21.16	20.98	21.35	21.17	0.86	ppb
Beryllium	9-1	2.13	2.13	2.13	2.13	0.04	ppb
Bismuth	209-1				102		%
Bismuth	209-2				102		%
Bromine	81-1						cps
Cadmium	108-1	2.38	2.52	2.55	2.48	3.60	ppb
Cadmium	106-1	0.75	-0.04	0.50	0.41	99.68	ppb
Cadmium	111-1	2.10	2.06	2.11	2.09	1.21	ppb
Calcium	43-1	1088.25	1085.67	1107.36	1093.76	1.08	ppb
Calcium	44-1	1052.90	1079.72	1069.48	1067.37	1.27	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	3.96	4.01	3.96	3.98	0.68	ppb
Cobalt	59-2	2.20	2.19	2.18	2.19	0.59	ppb
Copper	63-2	4.52	4.52	4.49	4.51	0.43	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				101		%
Holmium	165-2				103		%
Indium	115-1				103		%
Indium	115-2				104		%
Iron	54-2	434.99	440.43	442.39	439.27	0.87	ppb
Iron	56-2	431.43	431.64	432.54	431.87	0.14	ppb
Iron	57-2	424.58	435.20	425.04	428.27	1.40	ppb
Krypton	83-1						cps
Lead	206-1	2.00	1.99	2.00	2.00	0.16	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : PB166317BS Instrumnet Name : P8
Client Sample ID : LCS317 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:29:05 DataFile Name : 027LCSE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	2.07	2.02	2.01	2.04	1.75	ppb
Lead	208-1	2.03	2.01	1.99	2.01	0.82	ppb
Lithium	6-1				107		%
Magnesium	24-2	1006.20	1006.22	998.56	1003.66	0.44	ppb
Manganese	55-2	2.01	2.13	2.05	2.06	2.86	ppb
Molybdenum	94-1	11.84	11.98	12.10	11.97	1.07	ppb
Molybdenum	95-1	9.86	9.90	10.30	10.02	2.40	ppb
Molybdenum	96-1	10.20	10.35	10.47	10.34	1.29	ppb
Molybdenum	97-1	9.92	10.23	10.21	10.12	1.67	ppb
Molybdenum	98-1	9.81	10.04	10.34	10.06	2.61	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	2.08	2.01	2.06	2.05	1.62	ppb
Phosphorus	31-2	27.21	33.22	20.52	26.98	23.54	ppb
Potassium	39-2	1036.66	1027.66	1018.88	1027.73	0.87	ppb
Rhodium	103-1				106		%
Rhodium	103-2				104		%
Scandium	45-1				105		%
Scandium	45-2				108		%
Selenium	82-1	10.58	10.13	11.63	10.78	7.10	ppb
Selenium	77-2	10.30	6.79	11.19	9.43	24.71	ppb
Selenium	78-2	11.63	12.45	14.24	12.77	10.43	ppb
Silicon	28-1	23.48	24.72	24.85	24.35	3.10	ppb
Silver	107-1	2.08	2.16	2.19	2.15	2.67	ppb
Silver	109-1	2.13	2.16	2.18	2.15	1.21	ppb
Sodium	23-2	1069.85	1074.81	1064.19	1069.62	0.50	ppb
Strontium	86-1	1.91	2.10	2.03	2.01	4.72	ppb
Strontium	88-1	1.97	1.99	2.07	2.01	2.47	ppb
Sulfur	34-1	-1022.63	-878.23	-794.16	-898.34		ppb
Terbium	159-1				104		%
Terbium	159-2				102		%
Thallium	203-1	2.07	2.02	2.03	2.04	1.07	ppb
Thallium	205-1	2.04	2.03	2.03	2.03	0.37	ppb
Tin	118-1	10.58	10.68	10.81	10.69	1.09	ppb
Titanium	47-1	2.54	2.44	2.55	2.51	2.45	ppb
Uranium	238-1	1.87	1.88	1.87	1.87	0.31	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : PB166317BS Instrumnet Name : P8
Client Sample ID : LCS317 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:29:05 DataFile Name : 027LCSE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Vanadium	51-2	10.75	10.54	10.39	10.56	1.70	ppb
Ytterbium	172-1						cps
Ytterbium	172-2						cps
Ytterbium	176-1						cps
Ytterbium	176-2						cps
Yttrium	89-1				105		%
Yttrium	89-2				105		%
Zinc	66-2	10.33	10.49	10.33	10.38	0.87	ppb
Zirconium	90-1	1.98	2.03	2.05	2.02	1.81	ppb
Zirconium	91-1	1.97	1.99	1.98	1.98	0.42	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1186-01 Instrumnet Name : P8
Client Sample ID : ME2948 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:32:25 DataFile Name : 028AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	7.88	8.23	7.75	7.95	3.10	ppb
Antimony	121-1	0.05	0.05	0.05	0.05	2.94	ppb
Arsenic	75-2	0.38	0.42	0.29	0.36	18.40	ppb
Barium	135-1	18.49	18.85	18.32	18.55	1.45	ppb
Barium	137-1	18.38	19.41	18.88	18.89	2.73	ppb
Beryllium	9-1	-0.01	0.00	-0.01	-0.01		ppb
Bismuth	209-1				93		%
Bismuth	209-2				93		%
Bromine	81-1						cps
Cadmium	108-1	0.05	0.02	0.04	0.04	38.81	ppb
Cadmium	106-1	-1.26	-0.50	-0.46	-0.74		ppb
Cadmium	111-1	-0.09	-0.04	-0.03	-0.06		ppb
Calcium	43-1	374945.38	382182.11	378143.76	378423.75	0.96	ppb
Calcium	44-1	369342.36	378084.91	371452.94	372960.07	1.22	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	0.46	0.48	0.37	0.44	13.37	ppb
Cobalt	59-2	0.02	0.02	0.03	0.02	11.96	ppb
Copper	63-2	-0.07	-0.07	-0.08	-0.07		ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				100		%
Holmium	165-2				100		%
Indium	115-1				99		%
Indium	115-2				97		%
Iron	54-2	832.33	822.93	845.38	833.55	1.35	ppb
Iron	56-2	861.23	856.04	854.50	857.26	0.41	ppb
Iron	57-2	828.22	818.88	849.52	832.21	1.89	ppb
Krypton	83-1						cps
Lead	206-1	0.08	0.08	0.07	0.08	9.53	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1186-01 Instrumnet Name : P8
Client Sample ID : ME2948 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:32:25 DataFile Name : 028AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	0.08	0.09	0.08	0.08	8.32	ppb
Lead	208-1	0.09	0.09	0.08	0.08	5.00	ppb
Lithium	6-1				102		%
Magnesium	24-2	115372.05	114712.95	117417.01	115834.00	1.22	ppb
Manganese	55-2	29.45	28.85	30.14	29.48	2.19	ppb
Molybdenum	94-1	0.47	0.44	0.43	0.45	4.07	ppb
Molybdenum	95-1	0.35	0.38	0.37	0.37	3.43	ppb
Molybdenum	96-1	0.36	0.35	0.36	0.36	1.97	ppb
Molybdenum	97-1	0.38	0.35	0.39	0.37	5.75	ppb
Molybdenum	98-1	0.35	0.36	0.37	0.36	2.35	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.14	0.11	0.24	0.16	38.58	ppb
Phosphorus	31-2	-2.59	-7.24	-16.73	-8.85		ppb
Potassium	39-2	3592.20	3548.37	3600.33	3580.30	0.78	ppb
Rhodium	103-1				96		%
Rhodium	103-2				95		%
Scandium	45-1				101		%
Scandium	45-2				104		%
Selenium	82-1	0.48	0.75	1.15	0.80	41.92	ppb
Selenium	77-2	-0.31	0.60	-0.31	0.00		ppb
Selenium	78-2	0.26	0.00	1.35	0.54	134.40	ppb
Silicon	28-1	5575.50	5692.76	5550.32	5606.19	1.36	ppb
Silver	107-1	0.02	0.02	0.02	0.02	4.85	ppb
Silver	109-1	0.01	0.01	0.01	0.01	14.79	ppb
Sodium	23-2	31661.82	31845.38	32335.28	31947.49	1.09	ppb
Strontium	86-1	13988.20	13770.49	13947.12	13901.94	0.83	ppb
Strontium	88-1	14047.29	13802.44	13960.21	13936.65	0.89	ppb
Sulfur	34-1	405453.53	410350.89	405708.20	407170.87	0.68	ppb
Terbium	159-1				101		%
Terbium	159-2				99		%
Thallium	203-1	0.01	0.01	0.01	0.01	20.51	ppb
Thallium	205-1	0.01	0.01	0.01	0.01	21.90	ppb
Tin	118-1	0.95	0.94	0.92	0.94	1.80	ppb
Titanium	47-1	0.44	0.50	0.52	0.49	9.27	ppb
Uranium	238-1	0.21	0.21	0.20	0.21	1.57	ppb

LB Number :

LB134674

Operator :

Jaswal

Lab Sample ID :

Q1186-01

Instrumnet Name :

P8

Client Sample ID :

ME2948

Dilution Factor :

1

Date & Time Acquired :

2025-02-11 14:32:25

DataFile Name :

028AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Vanadium	51-2	0.09	0.11	0.11	0.11	11.23	ppb
Ytterbium	172-1						cps
Ytterbium	172-2						cps
Ytterbium	176-1						cps
Ytterbium	176-2						cps
Yttrium	89-1				101		%
Yttrium	89-2				102		%
Zinc	66-2	1.93	1.84	1.66	1.81	7.41	ppb
Zirconium	90-1	0.08	0.08	0.08	0.08	5.74	ppb
Zirconium	91-1	0.09	0.08	0.08	0.08	4.69	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-01DL2X4 Instrumnet Name : P8
Client Sample ID : YE8C9 Dilution Factor : 4
Date & Time Acquired : 2025-02-11 14:35:40 DataFile Name : 029AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	6120.53	6733.47	6024.12	6292.71	6.11	ppb
Antimony	121-1	0.13	0.15	0.13	0.14	5.82	ppb
Arsenic	75-2	13.92	15.11	13.64	14.22	5.47	ppb
Barium	135-1	61.94	62.33	63.67	62.65	1.45	ppb
Barium	137-1	62.01	62.76	63.86	62.88	1.48	ppb
Beryllium	9-1	0.27	0.25	0.24	0.25	4.96	ppb
Bismuth	209-1				103		%
Bismuth	209-2				96		%
Bromine	81-1						cps
Cadmium	108-1	0.13	0.18	0.11	0.14	26.71	ppb
Cadmium	106-1	-0.63	-0.53	-0.51	-0.56		ppb
Cadmium	111-1	-0.01	-0.01	0.01	0.00		ppb
Calcium	43-1	1553.81	1560.82	1605.62	1573.42	1.79	ppb
Calcium	44-1	1606.90	1581.39	1631.06	1606.45	1.55	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	14.31	15.63	14.08	14.67	5.69	ppb
Cobalt	59-2	7.95	8.73	7.82	8.17	6.04	ppb
Copper	63-2	27.02	30.02	26.35	27.80	7.03	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				104		%
Holmium	165-2				99		%
Indium	115-1				104		%
Indium	115-2				97		%
Iron	54-2	17690.27	19473.78	17272.69	18145.58	6.44	ppb
Iron	56-2	17495.32	19344.66	16867.81	17902.59	7.19	ppb
Iron	57-2	17252.54	19015.37	16941.64	17736.52	6.31	ppb
Krypton	83-1						cps
Lead	206-1	6.72	6.71	6.90	6.78	1.55	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-01DL2X4 Instrumnet Name : P8
Client Sample ID : YE8C9 Dilution Factor : 4
Date & Time Acquired : 2025-02-11 14:35:40 DataFile Name : 029AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	6.39	6.34	6.45	6.39	0.81	ppb
Lead	208-1	6.50	6.44	6.63	6.52	1.51	ppb
Lithium	6-1				108		%
Magnesium	24-2	2970.90	3304.07	2915.53	3063.50	6.86	ppb
Manganese	55-2	131.55	143.90	127.23	134.23	6.45	ppb
Molybdenum	94-1	0.53	0.82	2.08	1.14	72.07	ppb
Molybdenum	95-1	0.20	0.22	0.21	0.21	4.27	ppb
Molybdenum	96-1	0.27	0.25	0.27	0.26	4.03	ppb
Molybdenum	97-1	0.21	0.21	0.23	0.22	6.08	ppb
Molybdenum	98-1	0.22	0.20	0.21	0.21	3.09	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	20.62	22.58	20.09	21.09	6.22	ppb
Phosphorus	31-2	184.17	167.87	147.01	166.35	11.20	ppb
Potassium	39-2	177.62	205.22	202.41	195.09	7.79	ppb
Rhodium	103-1				105		%
Rhodium	103-2				98		%
Scandium	45-1				106		%
Scandium	45-2				101		%
Selenium	82-1	-0.01	0.07	0.21	0.09	122.26	ppb
Selenium	77-2	9.64	3.72	6.72	6.69	44.24	ppb
Selenium	78-2	2.11	0.61	1.27	1.33	56.40	ppb
Silicon	28-1	2264.31	2262.80	2400.82	2309.31	3.43	ppb
Silver	107-1	0.10	0.09	0.10	0.10	5.14	ppb
Silver	109-1	0.10	0.10	0.10	0.10	3.02	ppb
Sodium	23-2	38.00	55.69	35.00	42.90	26.06	ppb
Strontium	86-1	13.93	13.78	14.17	13.96	1.42	ppb
Strontium	88-1	13.82	13.83	14.31	13.99	2.00	ppb
Sulfur	34-1	-472.09	-658.16	-433.60	-521.28		ppb
Terbium	159-1				105		%
Terbium	159-2				98		%
Thallium	203-1	0.06	0.05	0.05	0.05	9.18	ppb
Thallium	205-1	0.05	0.05	0.05	0.05	5.32	ppb
Tin	118-1	0.01	-0.02	-0.02	-0.01		ppb
Titanium	47-1	4.39	6.02	6.02	5.48	17.21	ppb
Uranium	238-1	0.11	0.11	0.11	0.11	0.72	ppb

LB Number :

LB134674

Operator :

Jaswal

Lab Sample ID :

Q1159-01DL2X4

Instrumnet Name :

P8

Client Sample ID :

YE8C9

Dilution Factor :

4

Date & Time Acquired :

2025-02-11 14:35:40

DataFile Name :

029AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Vanadium	51-2	17.20	19.14	17.03	17.79	6.60	ppb
Ytterbium	172-1						cps
Ytterbium	172-2						cps
Ytterbium	176-1						cps
Ytterbium	176-2						cps
Yttrium	89-1				106		%
Yttrium	89-2				100		%
Zinc	66-2	37.30	41.40	34.96	37.89	8.60	ppb
Zirconium	90-1	0.22	0.19	0.24	0.22	10.97	ppb
Zirconium	91-1	0.18	0.21	0.23	0.21	10.23	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-02DL2X4 Instrumnet Name : P8
Client Sample ID : YE8C9D Dilution Factor : 4
Date & Time Acquired : 2025-02-11 14:38:57 DataFile Name : 030AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	5782.66	5937.85	6258.60	5993.04	4.05	ppb
Antimony	121-1	0.13	0.13	0.13	0.13	1.45	ppb
Arsenic	75-2	13.13	13.62	13.23	13.33	1.95	ppb
Barium	135-1	59.91	61.64	61.84	61.13	1.74	ppb
Barium	137-1	59.87	62.10	62.30	61.42	2.20	ppb
Beryllium	9-1	0.26	0.24	0.25	0.25	3.35	ppb
Bismuth	209-1				105		%
Bismuth	209-2				100		%
Bromine	81-1						cps
Cadmium	108-1	0.14	0.09	0.13	0.12	21.58	ppb
Cadmium	106-1	-1.18	-0.49	-0.05	-0.57		ppb
Cadmium	111-1	-0.05	0.01	0.05	0.00	1589.24	ppb
Calcium	43-1	1497.54	1518.41	1499.20	1505.05	0.77	ppb
Calcium	44-1	1531.91	1587.68	1524.09	1547.89	2.24	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	13.41	13.71	14.25	13.79	3.11	ppb
Cobalt	59-2	7.60	7.77	8.14	7.84	3.53	ppb
Copper	63-2	25.87	26.34	27.68	26.63	3.52	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				106		%
Holmium	165-2				102		%
Indium	115-1				107		%
Indium	115-2				102		%
Iron	54-2	17128.44	17166.35	17876.01	17390.27	2.42	ppb
Iron	56-2	16652.95	16717.50	17607.67	16992.71	3.14	ppb
Iron	57-2	16193.05	16623.95	17514.74	16777.25	4.02	ppb
Krypton	83-1						cps
Lead	206-1	6.25	6.29	6.45	6.33	1.69	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-02DL2X4 Instrumnet Name : P8
Client Sample ID : YE8C9D Dilution Factor : 4
Date & Time Acquired : 2025-02-11 14:38:57 DataFile Name : 030AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	5.88	5.98	6.04	5.97	1.37	ppb
Lead	208-1	5.99	6.08	6.18	6.09	1.57	ppb
Lithium	6-1				107		%
Magnesium	24-2	2807.75	2860.61	3024.39	2897.58	3.90	ppb
Manganese	55-2	123.72	126.12	131.82	127.22	3.27	ppb
Molybdenum	94-1	0.48	0.50	0.49	0.49	2.69	ppb
Molybdenum	95-1	0.19	0.20	0.19	0.19	1.64	ppb
Molybdenum	96-1	0.39	0.24	0.24	0.29	29.64	ppb
Molybdenum	97-1	0.19	0.18	0.20	0.19	4.94	ppb
Molybdenum	98-1	0.18	0.21	0.19	0.19	7.55	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	19.42	19.45	21.18	20.01	5.03	ppb
Phosphorus	31-2	127.15	164.37	172.30	154.61	15.59	ppb
Potassium	39-2	159.33	164.65	181.12	168.37	6.75	ppb
Rhodium	103-1				107		%
Rhodium	103-2				104		%
Scandium	45-1				109		%
Scandium	45-2				106		%
Selenium	82-1	0.71	0.14	0.67	0.51	62.04	ppb
Selenium	77-2	9.29	4.95	4.27	6.17	44.18	ppb
Selenium	78-2	2.28	1.52	3.21	2.34	36.11	ppb
Silicon	28-1	2412.98	2281.59	2251.33	2315.30	3.71	ppb
Silver	107-1	0.07	0.08	0.09	0.08	10.45	ppb
Silver	109-1	0.08	0.08	0.08	0.08	1.56	ppb
Sodium	23-2	26.02	27.70	37.00	30.24	19.56	ppb
Strontium	86-1	12.73	13.29	13.23	13.08	2.36	ppb
Strontium	88-1	12.92	13.42	13.19	13.18	1.89	ppb
Sulfur	34-1	-1081.76	-1034.96	-1116.36	-1077.69		ppb
Terbium	159-1				106		%
Terbium	159-2				101		%
Thallium	203-1	0.04	0.04	0.05	0.04	10.09	ppb
Thallium	205-1	0.04	0.04	0.04	0.04	8.37	ppb
Tin	118-1	-0.02	-0.03	-0.03	-0.03		ppb
Titanium	47-1	3.65	5.44	3.86	4.32	22.70	ppb
Uranium	238-1	0.10	0.11	0.30	0.17	65.55	ppb

LB Number :

LB134674

Operator :

Jaswal

Lab Sample ID :

Q1159-02DL2X4

Instrumnet Name :

P8

Client Sample ID :

YE8C9D

Dilution Factor :

4

Date & Time Acquired :

2025-02-11 14:38:57

DataFile Name :

030AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Vanadium	51-2	16.53	16.81	17.83	17.06	4.03	ppb
Ytterbium	172-1						cps
Ytterbium	172-2						cps
Ytterbium	176-1						cps
Ytterbium	176-2						cps
Yttrium	89-1				109		%
Yttrium	89-2				105		%
Zinc	66-2	35.42	35.57	37.40	36.13	3.05	ppb
Zirconium	90-1	0.52	0.18	0.20	0.30	63.48	ppb
Zirconium	91-1	0.21	0.19	0.17	0.19	10.60	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-01LDL2X20 Instrumnet Name : P8
Client Sample ID : YE8C9L Dilution Factor : 20
Date & Time Acquired : 2025-02-11 14:42:16 DataFile Name : 031AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	1185.86	1170.09	1181.89	1179.28	0.70	ppb
Antimony	121-1	0.03	0.03	0.03	0.03	3.62	ppb
Arsenic	75-2	2.72	2.86	2.88	2.82	3.14	ppb
Barium	135-1	12.27	12.56	12.20	12.34	1.54	ppb
Barium	137-1	12.72	12.45	12.14	12.43	2.35	ppb
Beryllium	9-1	0.05	0.05	0.04	0.05	12.60	ppb
Bismuth	209-1				103		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Cadmium	108-1	0.03	0.04	0.05	0.04	24.26	ppb
Cadmium	106-1	-0.57	-1.17	-1.07	-0.94		ppb
Cadmium	111-1	-0.04	-0.08	-0.08	-0.07		ppb
Calcium	43-1	309.95	313.56	312.82	312.11	0.61	ppb
Calcium	44-1	300.54	308.19	301.31	303.34	1.39	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2.76	2.74	2.81	2.77	1.32	ppb
Cobalt	59-2	1.52	1.47	1.54	1.51	2.51	ppb
Copper	63-2	5.15	4.97	5.19	5.11	2.25	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				104		%
Indium	115-1				106		%
Indium	115-2				103		%
Iron	54-2	3396.72	3295.62	3391.22	3361.19	1.69	ppb
Iron	56-2	3430.21	3340.32	3394.33	3388.29	1.34	ppb
Iron	57-2	3362.70	3314.57	3327.66	3334.98	0.75	ppb
Krypton	83-1						cps
Lead	206-1	1.25	1.28	1.27	1.27	1.06	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-01LDL2X20 Instrumnet Name : P8
Client Sample ID : YE8C9L Dilution Factor : 20
Date & Time Acquired : 2025-02-11 14:42:16 DataFile Name : 031AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	1.21	1.20	1.21	1.20	0.64	ppb
Lead	208-1	1.23	1.23	1.24	1.24	0.47	ppb
Lithium	6-1				106		%
Magnesium	24-2	574.27	562.38	569.19	568.62	1.05	ppb
Manganese	55-2	25.08	24.98	25.37	25.15	0.81	ppb
Molybdenum	94-1	0.13	0.09	0.08	0.10	26.45	ppb
Molybdenum	95-1	0.04	0.04	0.04	0.04	4.71	ppb
Molybdenum	96-1	0.05	0.05	0.05	0.05	0.98	ppb
Molybdenum	97-1	0.03	0.03	0.04	0.03	13.73	ppb
Molybdenum	98-1	0.04	0.04	0.04	0.04	5.36	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	4.04	3.83	3.94	3.94	2.62	ppb
Phosphorus	31-2	31.46	9.79	12.80	18.01	65.18	ppb
Potassium	39-2	34.81	34.61	35.99	35.14	2.12	ppb
Rhodium	103-1				107		%
Rhodium	103-2				104		%
Scandium	45-1				107		%
Scandium	45-2				107		%
Selenium	82-1	-0.10	0.21	0.05	0.06	273.95	ppb
Selenium	77-2	1.43	1.46	-0.31	0.86	117.73	ppb
Selenium	78-2	0.74	-0.01	0.24	0.32	118.89	ppb
Silicon	28-1	456.61	463.09	460.45	460.05	0.71	ppb
Silver	107-1	0.02	0.02	0.01	0.02	19.40	ppb
Silver	109-1	0.01	0.02	0.02	0.02	8.13	ppb
Sodium	23-2	8.58	5.59	6.63	6.93	21.92	ppb
Strontium	86-1	2.62	2.58	2.72	2.64	2.67	ppb
Strontium	88-1	2.71	2.66	2.69	2.68	0.85	ppb
Sulfur	34-1	-861.10	-781.07	-978.68	-873.61		ppb
Terbium	159-1				104		%
Terbium	159-2				102		%
Thallium	203-1	0.01	0.01	0.01	0.01	31.18	ppb
Thallium	205-1	0.01	0.01	0.01	0.01	16.17	ppb
Tin	118-1	-0.07	-0.06	-0.06	-0.06		ppb
Titanium	47-1	0.67	1.38	0.76	0.94	41.10	ppb
Uranium	238-1	0.02	0.02	0.02	0.02	5.48	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-01LDL2X20 Instrumnet Name : P8
Client Sample ID : YE8C9L Dilution Factor : 20
Date & Time Acquired : 2025-02-11 14:42:16 DataFile Name : 031AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Vanadium	51-2	3.37	3.30	3.24	3.30	1.95	ppb
Ytterbium	172-1						cps
Ytterbium	172-2						cps
Ytterbium	176-1						cps
Ytterbium	176-2						cps
Yttrium	89-1				107		%
Yttrium	89-2				106		%
Zinc	66-2	7.26	7.44	7.64	7.45	2.56	ppb
Zirconium	90-1	0.03	0.04	0.05	0.04	17.19	ppb
Zirconium	91-1	0.04	0.03	0.05	0.04	24.86	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-03DL2X4 Instrumnet Name : P8
Client Sample ID : YE8C9S Dilution Factor : 4
Date & Time Acquired : 2025-02-11 14:45:35 DataFile Name : 032AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	5861.59	5939.44	5911.28	5904.10	0.67	ppb
Antimony	121-1	5.33	5.38	5.40	5.37	0.67	ppb
Arsenic	75-2	15.75	15.82	16.65	16.08	3.10	ppb
Barium	135-1	157.52	161.28	161.89	160.23	1.48	ppb
Barium	137-1	162.49	162.13	160.43	161.68	0.68	ppb
Beryllium	9-1	2.72	2.68	2.68	2.69	0.75	ppb
Bismuth	209-1				102		%
Bismuth	209-2				102		%
Bromine	81-1						cps
Cadmium	108-1	2.58	2.82	2.56	2.65	5.44	ppb
Cadmium	106-1	1.83	1.61	1.65	1.70	7.04	ppb
Cadmium	111-1	2.69	2.64	2.54	2.62	2.89	ppb
Calcium	43-1	1735.13	1761.32	1756.07	1750.84	0.79	ppb
Calcium	44-1	1681.63	1743.82	1689.34	1704.93	1.99	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	23.10	23.66	23.50	23.42	1.22	ppb
Cobalt	59-2	32.49	32.92	33.12	32.84	0.99	ppb
Copper	63-2	38.13	38.68	38.72	38.51	0.85	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				104		%
Holmium	165-2				104		%
Indium	115-1				105		%
Indium	115-2				103		%
Iron	54-2	17011.13	17097.42	17037.73	17048.76	0.26	ppb
Iron	56-2	16430.19	16967.92	16656.89	16685.00	1.62	ppb
Iron	57-2	16160.89	16417.63	16404.64	16327.72	0.89	ppb
Krypton	83-1						cps
Lead	206-1	7.31	7.43	7.31	7.35	0.95	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-03DL2X4 Instrumnet Name : P8
Client Sample ID : YE8C9S Dilution Factor : 4
Date & Time Acquired : 2025-02-11 14:45:35 DataFile Name : 032AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	6.92	6.93	6.95	6.93	0.21	ppb
Lead	208-1	7.04	7.12	7.10	7.09	0.59	ppb
Lithium	6-1				108		%
Magnesium	24-2	2796.90	2853.05	2845.39	2831.78	1.08	ppb
Manganese	55-2	146.05	149.86	147.88	147.93	1.29	ppb
Molybdenum	94-1	21.19	21.60	21.79	21.53	1.43	ppb
Molybdenum	95-1	25.00	25.52	25.84	25.45	1.66	ppb
Molybdenum	96-1	24.67	25.60	25.07	25.11	1.87	ppb
Molybdenum	97-1	25.61	25.99	25.99	25.86	0.85	ppb
Molybdenum	98-1	25.04	25.68	25.66	25.46	1.42	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	44.57	45.10	45.63	45.10	1.17	ppb
Phosphorus	31-2	145.62	150.27	131.23	142.37	6.97	ppb
Potassium	39-2	159.20	163.31	162.03	161.51	1.30	ppb
Rhodium	103-1				106		%
Rhodium	103-2				103		%
Scandium	45-1				108		%
Scandium	45-2				107		%
Selenium	82-1	0.98	1.25	1.08	1.11	12.34	ppb
Selenium	77-2	8.18	8.50	7.64	8.10	5.37	ppb
Selenium	78-2	4.45	1.28	2.83	2.85	55.68	ppb
Silicon	28-1	2289.32	2274.59	2256.08	2273.33	0.73	ppb
Silver	107-1	2.54	2.56	2.60	2.57	1.18	ppb
Silver	109-1	2.53	2.56	2.59	2.56	1.25	ppb
Sodium	23-2	23.02	25.35	24.77	24.38	4.99	ppb
Strontium	86-1	505.30	516.54	514.22	512.02	1.16	ppb
Strontium	88-1	498.56	503.13	503.41	501.70	0.54	ppb
Sulfur	34-1	-944.73	-1030.80	-1124.20	-1033.24		ppb
Terbium	159-1				104		%
Terbium	159-2				102		%
Thallium	203-1	2.44	2.46	2.43	2.45	0.73	ppb
Thallium	205-1	2.45	2.47	2.48	2.47	0.64	ppb
Tin	118-1	-0.02	-0.02	-0.03	-0.03		ppb
Titanium	47-1	4.66	4.19	3.91	4.25	8.81	ppb
Uranium	238-1	0.11	0.11	0.11	0.11	0.98	ppb

LB Number :

LB134674

Operator :

Jaswal

Lab Sample ID :

Q1159-03DL2X4

Instrumnet Name :

P8

Client Sample ID :

YE8C9S

Dilution Factor :

4

Date & Time Acquired :

2025-02-11 14:45:35

DataFile Name :

032AREF.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Vanadium	51-2	40.38	41.14	41.46	40.99	1.35	ppb
Ytterbium	172-1						cps
Ytterbium	172-2						cps
Ytterbium	176-1						cps
Ytterbium	176-2						cps
Yttrium	89-1				108		%
Yttrium	89-2				107		%
Zinc	66-2	59.73	60.48	60.43	60.21	0.69	ppb
Zirconium	90-1	0.27	0.20	0.17	0.21	23.51	ppb
Zirconium	91-1	0.31	0.21	0.19	0.24	27.93	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : CCV022 Instrumnet Name : P8
Client Sample ID : CCV022 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:50:23 DataFile Name : 033CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	49473.08	50852.87	49571.92	49965.96	1.54	ppb
Antimony	121-1	500.02	494.86	504.80	499.89	0.99	ppb
Arsenic	75-2	487.40	491.14	504.57	494.37	1.83	ppb
Barium	135-1	2496.16	2493.42	2549.64	2513.07	1.26	ppb
Barium	137-1	2501.67	2474.06	2558.13	2511.29	1.71	ppb
Beryllium	9-1	500.52	507.65	513.65	507.27	1.30	ppb
Bismuth	209-1				89		%
Bismuth	209-2				88		%
Bromine	81-1						cps
Cadmium	108-1	490.12	489.13	497.07	492.11	0.88	ppb
Cadmium	106-1	485.95	486.92	498.25	490.38	1.40	ppb
Cadmium	111-1	494.44	489.39	489.82	491.22	0.57	ppb
Calcium	43-1	251891.84	256836.15	257808.16	255512.05	1.24	ppb
Calcium	44-1	251129.51	254691.67	258252.44	254691.21	1.40	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	494.07	509.62	501.30	501.66	1.55	ppb
Cobalt	59-2	483.63	500.88	494.21	492.91	1.76	ppb
Copper	63-2	4675.94	4785.34	4839.99	4767.09	1.75	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				99		%
Holmium	165-2				98		%
Indium	115-1				94		%
Indium	115-2				92		%
Iron	54-2	126998.97	126869.85	126346.19	126738.34	0.27	ppb
Iron	56-2	124234.33	124820.26	126837.43	125297.34	1.09	ppb
Iron	57-2	123184.09	125502.80	128336.49	125674.46	2.05	ppb
Krypton	83-1						cps
Lead	206-1	2522.20	2536.34	2554.74	2537.76	0.64	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : CCV022 Instrumnet Name : P8
Client Sample ID : CCV022 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:50:23 DataFile Name : 033CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	2504.87	2554.04	2554.47	2537.79	1.12	ppb
Lead	208-1	2510.03	2528.91	2567.73	2535.56	1.16	ppb
Lithium	6-1				95		%
Magnesium	24-2	248488.64	254072.12	251785.19	251448.65	1.12	ppb
Manganese	55-2	4907.71	4910.09	4980.29	4932.70	0.84	ppb
Molybdenum	94-1	4970.75	5022.10	5000.65	4997.83	0.52	ppb
Molybdenum	95-1	4889.76	5034.01	5021.77	4981.85	1.61	ppb
Molybdenum	96-1	5018.54	4996.00	5009.63	5008.06	0.23	ppb
Molybdenum	97-1	5003.00	4979.73	5046.74	5009.82	0.68	ppb
Molybdenum	98-1	4948.35	4974.32	5009.33	4977.33	0.61	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	454.13	461.18	465.59	460.30	1.26	ppb
Phosphorus	31-2	10071.28	10178.72	10008.37	10086.12	0.85	ppb
Potassium	39-2	124622.77	124504.20	124051.56	124392.84	0.24	ppb
Rhodium	103-1				92		%
Rhodium	103-2				91		%
Scandium	45-1				101		%
Scandium	45-2				101		%
Selenium	82-1	487.62	482.38	487.99	486.00	0.64	ppb
Selenium	77-2	488.48	506.20	518.79	504.49	3.02	ppb
Selenium	78-2	497.94	509.27	484.95	497.39	2.45	ppb
Silicon	28-1	495.28	503.58	512.53	503.80	1.71	ppb
Silver	107-1	479.98	485.27	491.56	485.60	1.19	ppb
Silver	109-1	482.97	480.11	492.81	485.30	1.37	ppb
Sodium	23-2	255912.08	259221.77	262756.68	259296.84	1.32	ppb
Strontium	86-1	504.63	506.68	504.67	505.33	0.23	ppb
Strontium	88-1	495.14	501.11	501.71	499.32	0.73	ppb
Sulfur	34-1	9714.57	9591.88	9677.16	9661.20	0.65	ppb
Terbium	159-1				99		%
Terbium	159-2				96		%
Thallium	203-1	513.72	509.84	519.45	514.34	0.94	ppb
Thallium	205-1	505.56	503.58	514.11	507.75	1.10	ppb
Tin	118-1	494.57	492.56	511.79	499.64	2.12	ppb
Titanium	47-1	4907.00	4955.37	5009.43	4957.26	1.03	ppb
Uranium	238-1	502.77	501.12	520.47	508.12	2.11	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : CCV022 Instrumnet Name : P8
Client Sample ID : CCV022 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:50:23 DataFile Name : 033CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Vanadium	51-2	490.40	493.86	496.20	493.49	0.59	ppb
Ytterbium	172-1						cps
Ytterbium	172-2						cps
Ytterbium	176-1						cps
Ytterbium	176-2						cps
Yttrium	89-1				99		%
Yttrium	89-2				100		%
Zinc	66-2	4799.76	4821.03	4892.93	4837.91	1.01	ppb
Zirconium	90-1	496.36	506.21	506.57	503.04	1.15	ppb
Zirconium	91-1	492.65	513.94	511.09	505.90	2.28	ppb

LB Number :

LB134674

Operator :

Jaswal

Lab Sample ID :

CCB022

Instrumnet Name :

P8

Client Sample ID :

CCB022

Dilution Factor :

1

Date & Time Acquired :

2025-02-11 14:53:49

DataFile Name :

034CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	0.21	0.35	0.53	0.37	43.41	ppb
Antimony	121-1	0.11	0.10	0.09	0.10	9.96	ppb
Arsenic	75-2	-0.02	-0.02	0.01	-0.01		ppb
Barium	135-1	0.06	0.04	0.03	0.04	40.70	ppb
Barium	137-1	0.09	0.05	0.05	0.07	31.20	ppb
Beryllium	9-1	0.10	0.09	0.09	0.09	9.24	ppb
Bismuth	209-1				103		%
Bismuth	209-2				100		%
Bromine	81-1						cps
Cadmium	108-1	0.07	0.03	0.11	0.07	62.70	ppb
Cadmium	106-1	-2.03	-0.94	-2.09	-1.69		ppb
Cadmium	111-1	-0.15	-0.05	-0.14	-0.11		ppb
Calcium	43-1	4.61	4.53	3.57	4.24	13.59	ppb
Calcium	44-1	1.47	-1.02	-1.53	-0.36		ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	-0.02	0.02	0.01	0.00	656.28	ppb
Cobalt	59-2	0.00	0.00	0.00	0.00		ppb
Copper	63-2	-0.08	-0.05	-0.04	-0.06		ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				102		%
Indium	115-1				104		%
Indium	115-2				102		%
Iron	54-2	2.58	2.55	2.44	2.52	2.88	ppb
Iron	56-2	2.08	2.16	1.88	2.04	6.95	ppb
Iron	57-2	1.85	2.24	2.46	2.18	14.30	ppb
Krypton	83-1						cps
Lead	206-1	0.13	0.11	0.12	0.12	8.23	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : CCB022 Instrumnet Name : P8
Client Sample ID : CCB022 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:53:49 DataFile Name : 034CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	207-1	0.14	0.11	0.11	0.12	11.70	ppb
Lead	208-1	0.14	0.12	0.11	0.12	13.58	ppb
Lithium	6-1				105		%
Magnesium	24-2	-7.64	-8.31	-7.42	-7.79		ppb
Manganese	55-2	0.09	0.09	0.12	0.10	13.07	ppb
Molybdenum	94-1	0.17	0.10	0.11	0.13	29.44	ppb
Molybdenum	95-1	0.17	0.11	0.12	0.13	23.74	ppb
Molybdenum	96-1	0.16	0.13	0.10	0.13	25.47	ppb
Molybdenum	97-1	0.15	0.12	0.09	0.12	23.51	ppb
Molybdenum	98-1	0.17	0.13	0.10	0.13	27.55	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.17	0.19	0.28	0.21	27.26	ppb
Phosphorus	31-2	-15.11	-16.37	-11.62	-14.36		ppb
Potassium	39-2	9.76	10.98	8.69	9.81	11.67	ppb
Rhodium	103-1				106		%
Rhodium	103-2				104		%
Scandium	45-1				105		%
Scandium	45-2				105		%
Selenium	82-1	0.26	0.08	0.44	0.26	69.22	ppb
Selenium	77-2	-0.31	0.57	-0.31	-0.02		ppb
Selenium	78-2	-0.27	0.50	1.56	0.60	154.39	ppb
Silicon	28-1	0.36	0.17	0.21	0.25	40.52	ppb
Silver	107-1	0.04	0.03	0.02	0.03	21.86	ppb
Silver	109-1	0.05	0.04	0.03	0.04	21.61	ppb
Sodium	23-2	20.93	20.89	19.72	20.51	3.38	ppb
Strontium	86-1	-0.03	-0.03	-0.04	-0.03		ppb
Strontium	88-1	-0.02	-0.02	-0.02	-0.02		ppb
Sulfur	34-1	-948.11	-935.85	-935.81	-939.92		ppb
Terbium	159-1				104		%
Terbium	159-2				100		%
Thallium	203-1	0.04	0.03	0.03	0.03	9.60	ppb
Thallium	205-1	0.03	0.03	0.03	0.03	14.66	ppb
Tin	118-1	0.01	0.01	0.01	0.01	28.03	ppb
Titanium	47-1	0.08	0.06	0.05	0.06	23.17	ppb
Uranium	238-1	0.01	0.01	0.00	0.01	49.94	ppb

LB Number :LB134674

Operator :Jaswal

Lab Sample ID :CCB022

Instrumnet Name :P8

Client Sample ID :CCB022

Dilution Factor :1

Date & Time Acquired :2025-02-11 14:53:49

DataFile Name :034CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Vanadium	51-2	0.00	0.01	0.00	0.00	176.56	ppb
Ytterbium	172-1						cps
Ytterbium	172-2						cps
Ytterbium	176-1						cps
Ytterbium	176-2						cps
Yttrium	89-1				105		%
Yttrium	89-2				105		%
Zinc	66-2	0.23	0.12	0.15	0.17	33.12	ppb
Zirconium	90-1	0.01	0.02	0.01	0.01	29.26	ppb
Zirconium	91-1	0.02	0.02	0.00	0.01	78.92	ppb

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S0 Instrumnet Name : P8
Client Sample ID : S0 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:05:13 DataFile Name : 004CALB.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	103	127	110	113	10.61	cps
Antimony	121-1	57	40	60	52	20.52	cps
Arsenic	75-2	17	0	17	11	86.60	cps
Barium	135-1	253	270	253	259	3.72	cps
Barium	137-1	307	357	353	339	8.25	cps
Beryllium	9-1	366	360	359	362	1.11	cps
Bismuth	209-1	13275325	13140858	13103762	13173315	0.69	cps
Bismuth	209-2	4988262	4906578	4955423	4950088	0.83	cps
Bromine	81-1	3557	3821	3520	3633	4.51	cps
Cadmium	108-1	13	17	30	20	44.10	cps
Cadmium	106-1	8193	8269	8346	8269	0.93	cps
Cadmium	111-1	5752	5787	5844	5794	0.81	cps
Calcium	43-1	420	387	450	419	7.56	cps
Calcium	44-1	29641	29133	28993	29255	1.17	cps
Carbon	12-1	3404220	3349154	3328299	3360557	1.17	cps
Carbon	12-2	23021	22427	22788	22745	1.32	cps
Chlorine	35-1	127939	127697	126887	127508	0.43	cps
Chlorine	35-2	477	520	493	497	4.40	cps
Chromium	52-2	717	713	810	747	7.35	cps
Cobalt	59-2	153	173	197	174	12.43	cps
Copper	63-2	3527	3374	3324	3408	3.11	cps
Dysprosium	156-1	147	103	160	137	21.68	cps
Dysprosium	156-2	3	0	7	3	100.05	cps
Erbium	164-1	133	93	140	122	20.65	cps
Erbium	164-2	40	27	30	32	21.53	cps
Gadolinium	160-1	133	157	117	136	14.82	cps
Gadolinium	160-2	17	20	40	26	49.38	cps
Holmium	165-1	20745007	20390432	20532713	20556051	0.87	cps
Holmium	165-2	6570019	6536123	6438593	6514912	1.05	cps
Indium	115-1	15990482	16095920	16025387	16037263	0.33	cps
Indium	115-2	1383269	1384341	1391629	1386413	0.33	cps
Iron	54-2	383	400	307	363	13.70	cps
Iron	56-2	5831	6001	6128	5987	2.49	cps
Iron	57-2	130	127	157	138	11.93	cps
Krypton	83-1	340	273	243	286	17.33	cps
Lead	206-1	2120	2087	2240	2149	3.75	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S0 Instrumnet Name : P8
Client Sample ID : S0 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:05:13 DataFile Name : 004CALB.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	1710	1823	1670	1735	4.59	cps
Lead	208-1	8051	8355	8034	8147	2.21	cps
Lithium	6-1	7490438	7316961	7462925	7423441	1.26	cps
Magnesium	24-2	4121	3911	4371	4134	5.57	cps
Manganese	55-2	140	163	190	164	15.21	cps
Molybdenum	94-1	440	380	353	391	11.35	cps
Molybdenum	95-1	173	147	137	152	12.45	cps
Molybdenum	96-1	237	263	237	246	6.27	cps
Molybdenum	97-1	110	137	120	122	11.02	cps
Molybdenum	98-1	213	200	177	197	9.44	cps
Neodymium	150-1	17	7	7	10	57.72	cps
Neodymium	150-2	0	3	0	1	173.21	cps
Nickel	60-2	907	930	897	911	1.88	cps
Phosphorus	31-2	70	97	93	87	16.77	cps
Potassium	39-2	9537	9210	9467	9404	1.83	cps
Rhodium	103-1	14912726	14564954	14726231	14734637	1.18	cps
Rhodium	103-2	5463921	5364977	5413198	5414032	0.91	cps
Scandium	45-1	10482385	10523577	10387300	10464421	0.67	cps
Scandium	45-2	183202	183361	188006	184856	1.48	cps
Selenium	82-1	-77	-30	0	-36	-108.66	cps
Selenium	77-2	0	3	0	1	173.21	cps
Selenium	78-2	3	0	7	3	100.05	cps
Silicon	28-1	513562	513497	512737	513265	0.09	cps
Silver	107-1	247	233	217	232	6.47	cps
Silver	109-1	73	53	73	67	17.32	cps
Sodium	23-2	57104	56526	56874	56834	0.51	cps
Strontium	86-1	730	830	783	781	6.41	cps
Strontium	88-1	2950	2874	2654	2826	5.45	cps
Sulfur	34-1	748981	749440	741213	746544	0.62	cps
Terbium	159-1	20338462	20804449	20860423	20667778	1.39	cps
Terbium	159-2	6329934	6305846	6337854	6324544	0.26	cps
Thallium	203-1	343	407	387	379	8.55	cps
Thallium	205-1	883	993	897	924	6.49	cps
Tin	118-1	3200	3187	3210	3199	0.37	cps
Titanium	47-1	633	547	740	640	15.14	cps
Uranium	238-1	133	210	150	164	24.52	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S0 Instrumnet Name : P8
Client Sample ID : S0 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:05:13 DataFile Name : 004CALB.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	13	23	10	16	44.60	cps
Ytterbium	172-1	117	123	137	126	8.11	cps
Ytterbium	172-2	27	53	23	34	47.73	cps
Ytterbium	176-1	2584	2007	1977	2189	15.62	cps
Ytterbium	176-2	337	397	383	372	8.46	cps
Yttrium	89-1	27036724	26993352	27119239	27049772	0.24	cps
Yttrium	89-2	1699120	1719689	1723718	1714176	0.77	cps
Zinc	66-2	237	183	167	196	18.70	cps
Zirconium	90-1	903	917	920	913	0.97	cps
Zirconium	91-1	193	157	173	174	10.52	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S02 Instrumnet Name : P8
Client Sample ID : S02 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:08:35 DataFile Name : 005CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1293	1397	1263	1318	5.31	cps
Antimony	121-1	36368	36245	36582	36398	0.47	cps
Arsenic	75-2	303	340	320	321	5.72	cps
Barium	135-1	42958	43680	43994	43544	1.22	cps
Barium	137-1	76106	75060	75891	75686	0.73	cps
Beryllium	9-1	6302	6256	6279	6279	0.37	cps
Bismuth	209-1	13305169	13319359	13198515	13274348	0.50	cps
Bismuth	209-2	4897531	5006140	5001587	4968419	1.24	cps
Bromine	81-1	3884	3764	3891	3846	1.85	cps
Cadmium	108-1	353	400	370	374	6.32	cps
Cadmium	106-1	8796	8256	8700	8584	3.36	cps
Cadmium	111-1	10444	10143	10441	10343	1.67	cps
Calcium	43-1	28308	28715	28796	28606	0.91	cps
Calcium	44-1	479668	486118	482908	482898	0.67	cps
Carbon	12-1	3363525	3372602	3371157	3369095	0.14	cps
Carbon	12-2	22377	23045	22674	22699	1.47	cps
Chlorine	35-1	126567	126605	128729	127300	0.97	cps
Chlorine	35-2	393	443	443	427	6.77	cps
Chromium	52-2	7622	7826	7529	7659	1.98	cps
Cobalt	59-2	7135	7269	7429	7277	2.02	cps
Copper	63-2	14134	14411	14481	14342	1.28	cps
Dysprosium	156-1	133	157	160	150	9.68	cps
Dysprosium	156-2	17	10	10	12	31.50	cps
Erbium	164-1	70	80	110	87	24.02	cps
Erbium	164-2	37	17	27	27	37.50	cps
Gadolinium	160-1	127	143	123	131	8.17	cps
Gadolinium	160-2	20	13	23	19	26.96	cps
Holmium	165-1	20748497	20702416	20640064	20696992	0.26	cps
Holmium	165-2	6496169	6612263	6547834	6552089	0.89	cps
Indium	115-1	16067441	16277713	16096744	16147299	0.71	cps
Indium	115-2	1390285	1408611	1395889	1398262	0.67	cps
Iron	54-2	7652	7942	7796	7797	1.86	cps
Iron	56-2	139733	140366	138636	139579	0.63	cps
Iron	57-2	3447	3550	3340	3446	3.05	cps
Krypton	83-1	270	343	260	291	15.63	cps
Lead	206-1	16337	16728	16667	16577	1.27	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S02 Instrumnet Name : P8
Client Sample ID : S02 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:08:35 DataFile Name : 005CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	14084	14235	14021	14113	0.78	cps
Lead	208-1	65676	65889	65622	65729	0.21	cps
Lithium	6-1	7417028	7533299	7482519	7477615	0.78	cps
Magnesium	24-2	125642	127387	126987	126672	0.72	cps
Manganese	55-2	1570	1497	1553	1540	2.50	cps
Molybdenum	94-1	36150	36735	36351	36412	0.82	cps
Molybdenum	95-1	43618	43702	43859	43726	0.28	cps
Molybdenum	96-1	48615	49963	49016	49198	1.41	cps
Molybdenum	97-1	27477	27430	27764	27557	0.66	cps
Molybdenum	98-1	69878	71265	71141	70761	1.08	cps
Neodymium	150-1	13	7	20	13	49.99	cps
Neodymium	150-2	0	7	3	3	100.05	cps
Nickel	60-2	2750	2860	2837	2816	2.06	cps
Phosphorus	31-2	137	140	187	154	18.10	cps
Potassium	39-2	72698	72578	72829	72702	0.17	cps
Rhodium	103-1	14933407	15023952	14957859	14971740	0.31	cps
Rhodium	103-2	5415561	5419026	5486447	5440344	0.73	cps
Scandium	45-1	10404605	10362599	10440548	10402584	0.38	cps
Scandium	45-2	185967	188890	186642	187166	0.82	cps
Selenium	82-1	1410	1403	1267	1360	5.95	cps
Selenium	77-2	23	13	20	19	26.96	cps
Selenium	78-2	53	73	80	69	20.15	cps
Silicon	28-1	592652	588031	589567	590083	0.40	cps
Silver	107-1	21587	22044	22301	21978	1.65	cps
Silver	109-1	20906	20973	21433	21104	1.36	cps
Sodium	23-2	295886	295862	296889	296212	0.20	cps
Strontium	86-1	6432	6291	6255	6326	1.47	cps
Strontium	88-1	51819	51832	52250	51967	0.47	cps
Sulfur	34-1	758163	757218	761096	758826	0.27	cps
Terbium	159-1	20876308	21064451	20780423	20907061	0.69	cps
Terbium	159-2	6254713	6310590	6250826	6272043	0.53	cps
Thallium	203-1	18463	18857	17842	18388	2.78	cps
Thallium	205-1	44195	44921	44162	44426	0.97	cps
Tin	118-1	74953	75711	75929	75531	0.68	cps
Titanium	47-1	13680	13420	13727	13609	1.22	cps
Uranium	238-1	56194	56314	56458	56322	0.24	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S02 Instrumnet Name : P8
Client Sample ID : S02 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:08:35 DataFile Name : 005CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	13557	13423	13577	13519	0.62	cps
Ytterbium	172-1	140	130	117	129	9.08	cps
Ytterbium	172-2	50	43	27	40	30.04	cps
Ytterbium	176-1	2177	2367	2224	2256	4.39	cps
Ytterbium	176-2	333	357	403	364	9.78	cps
Yttrium	89-1	26691519	26956347	27177395	26941754	0.90	cps
Yttrium	89-2	1737480	1722980	1768395	1742952	1.33	cps
Zinc	66-2	3157	3047	3224	3143	2.84	cps
Zirconium	90-1	30577	30607	31275	30820	1.28	cps
Zirconium	91-1	6805	6868	6668	6781	1.51	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S03 Instrumnet Name : P8
Client Sample ID : S03 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:11:58 DataFile Name : 006CALB.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	62842	62641	62574	62685	0.22	cps
Antimony	121-1	861745	857215	867603	862188	0.60	cps
Arsenic	75-2	14811	15335	15045	15064	1.74	cps
Barium	135-1	1039569	1035146	1040465	1038393	0.27	cps
Barium	137-1	1829036	1837006	1822590	1829544	0.39	cps
Beryllium	9-1	280527	282353	279656	280845	0.49	cps
Bismuth	209-1	13375628	13388117	13312811	13358852	0.30	cps
Bismuth	209-2	4898733	4904587	4900921	4901414	0.06	cps
Bromine	81-1	3841	3781	3887	3836	1.39	cps
Cadmium	108-1	17431	17855	17792	17693	1.29	cps
Cadmium	106-1	33731	33223	33768	33574	0.91	cps
Cadmium	111-1	221619	223413	223740	222924	0.51	cps
Calcium	43-1	267443	270725	270715	269628	0.70	cps
Calcium	44-1	4398491	4410493	4446600	4418528	0.57	cps
Carbon	12-1	3434325	3453165	3427420	3438303	0.39	cps
Carbon	12-2	23108	23148	23325	23194	0.50	cps
Chlorine	35-1	126987	126803	127035	126942	0.10	cps
Chlorine	35-2	440	410	503	451	10.56	cps
Chromium	52-2	165134	165134	166152	165473	0.36	cps
Cobalt	59-2	322127	324722	322021	322956	0.47	cps
Copper	63-2	2591688	2619645	2610124	2607152	0.55	cps
Dysprosium	156-1	167	210	233	203	16.64	cps
Dysprosium	156-2	3	7	23	11	96.43	cps
Erbium	164-1	123	167	140	143	15.25	cps
Erbium	164-2	37	50	37	41	18.72	cps
Gadolinium	160-1	123	157	90	123	27.03	cps
Gadolinium	160-2	13	20	17	17	20.01	cps
Holmium	165-1	20808719	20581035	20637970	20675908	0.57	cps
Holmium	165-2	6534923	6575449	6485936	6532103	0.69	cps
Indium	115-1	16309522	16038421	16068709	16138884	0.92	cps
Indium	115-2	1365235	1378947	1381864	1375349	0.65	cps
Iron	54-2	346744	344288	347558	346196	0.49	cps
Iron	56-2	6339515	6390928	6321320	6350588	0.57	cps
Iron	57-2	156734	157035	156201	156657	0.27	cps
Krypton	83-1	263	303	220	262	15.90	cps
Lead	206-1	3704497	3801680	3748032	3751403	1.30	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S03 Instrumnet Name : P8
Client Sample ID : S03 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:11:58 DataFile Name : 006CALB.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	3218431	3221582	3141539	3193851	1.42	cps
Lead	208-1	14514907	14810707	14604523	14643379	1.04	cps
Lithium	6-1	7556054	7252764	7432887	7413902	2.06	cps
Magnesium	24-2	1189372	1203052	1212261	1201562	0.96	cps
Manganese	55-2	667130	665803	667326	666753	0.12	cps
Molybdenum	94-1	2910849	3005718	2911690	2942752	1.85	cps
Molybdenum	95-1	4301342	4316458	4218045	4278615	1.24	cps
Molybdenum	96-1	4705375	4692972	4698877	4699074	0.13	cps
Molybdenum	97-1	2669178	2696189	2667114	2677493	0.61	cps
Molybdenum	98-1	6858357	6929595	6932663	6906871	0.61	cps
Neodymium	150-1	77	73	127	92	32.40	cps
Neodymium	150-2	3	0	7	3	100.05	cps
Nickel	60-2	90005	90677	90052	90245	0.42	cps
Phosphorus	31-2	2774	3087	2914	2925	5.37	cps
Potassium	39-2	312452	313571	314651	313558	0.35	cps
Rhodium	103-1	14903637	14669506	14852004	14808382	0.83	cps
Rhodium	103-2	5322813	5342154	5387958	5350975	0.63	cps
Scandium	45-1	10647223	10625834	10439565	10570874	1.08	cps
Scandium	45-2	184366	182989	182790	183382	0.47	cps
Selenium	82-1	12329	12853	12840	12674	2.36	cps
Selenium	77-2	173	210	190	191	9.60	cps
Selenium	78-2	640	720	623	661	7.82	cps
Silicon	28-1	1170052	1171124	1164189	1168455	0.32	cps
Silver	107-1	1089316	1089754	1096649	1091906	0.38	cps
Silver	109-1	1034800	1052768	1043563	1043710	0.86	cps
Sodium	23-2	2349817	2397796	2405859	2384490	1.27	cps
Strontium	86-1	271544	270835	270134	270838	0.26	cps
Strontium	88-1	2410179	2414751	2406107	2410346	0.18	cps
Sulfur	34-1	850008	844585	840403	844999	0.57	cps
Terbium	159-1	21490709	21178267	21023054	21230677	1.12	cps
Terbium	159-2	6226428	6262761	6222322	6237170	0.36	cps
Thallium	203-1	858876	878597	875274	870915	1.21	cps
Thallium	205-1	2133244	2132464	2172434	2146047	1.06	cps
Tin	118-1	692535	701098	700159	697931	0.67	cps
Titanium	47-1	1268812	1276069	1288188	1277690	0.77	cps
Uranium	238-1	2925307	2889464	2890129	2901633	0.71	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S03 Instrumnet Name : P8
Client Sample ID : S03 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:11:58 DataFile Name : 006CALB.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	128063	127488	125655	127069	0.99	cps
Ytterbium	172-1	93	107	110	103	8.54	cps
Ytterbium	172-2	40	50	53	48	14.52	cps
Ytterbium	176-1	6368	6442	6245	6352	1.57	cps
Ytterbium	176-2	1517	1640	1637	1598	4.40	cps
Yttrium	89-1	27626859	27006856	26922680	27185465	1.41	cps
Yttrium	89-2	1703513	1734378	1687292	1708394	1.40	cps
Zinc	66-2	266730	266070	269099	267300	0.60	cps
Zirconium	90-1	1487592	1512282	1492083	1497319	0.88	cps
Zirconium	91-1	320865	329555	325799	325406	1.34	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S04 Instrumnet Name : P8
Client Sample ID : S04 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:15:02 DataFile Name : 007CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	154440	155120	154841	154800	0.22	cps
Antimony	121-1	2191182	2173691	2172950	2179274	0.47	cps
Arsenic	75-2	36721	36988	37149	36953	0.58	cps
Barium	135-1	2645811	2641562	2691543	2659639	1.04	cps
Barium	137-1	4467736	4611294	4498732	4525920	1.67	cps
Beryllium	9-1	683968	688060	689392	687140	0.41	cps
Bismuth	209-1	13265025	13219506	13088146	13190892	0.70	cps
Bismuth	209-2	4846478	4819631	4865973	4844027	0.48	cps
Bromine	81-1	3901	3861	3991	3917	1.70	cps
Cadmium	108-1	44167	44033	43572	43924	0.71	cps
Cadmium	106-1	69226	68820	69869	69305	0.76	cps
Cadmium	111-1	531645	540851	539395	537297	0.92	cps
Calcium	43-1	653566	658458	664813	658946	0.86	cps
Calcium	44-1	10694724	10763806	10678590	10712373	0.42	cps
Carbon	12-1	3568904	3600141	3608475	3592507	0.58	cps
Carbon	12-2	24177	24183	24637	24332	1.09	cps
Chlorine	35-1	127730	129358	129281	128790	0.71	cps
Chlorine	35-2	550	437	420	469	15.09	cps
Chromium	52-2	406849	405764	407534	406716	0.22	cps
Cobalt	59-2	785348	788507	783457	785770	0.32	cps
Copper	63-2	6278072	6260123	6210186	6249460	0.56	cps
Dysprosium	156-1	297	273	233	268	11.96	cps
Dysprosium	156-2	0	17	10	9	94.38	cps
Erbium	164-1	127	150	143	140	8.58	cps
Erbium	164-2	33	60	47	47	28.58	cps
Gadolinium	160-1	130	173	133	146	16.57	cps
Gadolinium	160-2	33	37	17	29	37.08	cps
Holmium	165-1	20665574	20655454	20567261	20629429	0.26	cps
Holmium	165-2	6614905	6506883	6606556	6576114	0.91	cps
Indium	115-1	16051284	15834460	15928018	15937921	0.68	cps
Indium	115-2	1367194	1367196	1362316	1365569	0.21	cps
Iron	54-2	845134	853505	848197	848945	0.50	cps
Iron	56-2	15588463	15591745	15533624	15571277	0.21	cps
Iron	57-2	383801	385971	386725	385499	0.39	cps
Krypton	83-1	270	267	297	278	5.92	cps
Lead	206-1	8902627	8945081	8979823	8942510	0.43	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S04 Instrumnet Name : P8
Client Sample ID : S04 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:15:02 DataFile Name : 007CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	7744967	7683947	7787916	7738943	0.68	cps
Lead	208-1	35704779	35662100	35816200	35727693	0.22	cps
Lithium	6-1	7534394	7433388	7286078	7417953	1.68	cps
Magnesium	24-2	2909807	2928592	2950806	2929735	0.70	cps
Manganese	55-2	1689134	1699911	1661680	1683575	1.17	cps
Molybdenum	94-1	7208966	7126941	7213334	7183080	0.68	cps
Molybdenum	95-1	10610964	10616673	10565929	10597855	0.26	cps
Molybdenum	96-1	11601820	11514836	11528028	11548228	0.41	cps
Molybdenum	97-1	6536549	6540304	6480236	6519030	0.52	cps
Molybdenum	98-1	17090663	16976508	17271388	17112853	0.87	cps
Neodymium	150-1	130	210	213	184	25.58	cps
Neodymium	150-2	10	0	0	3	173.21	cps
Nickel	60-2	218976	220218	219494	219562	0.28	cps
Phosphorus	31-2	7332	7299	7432	7354	0.94	cps
Potassium	39-2	758554	755128	755929	756537	0.24	cps
Rhodium	103-1	14654109	14546289	14851517	14683971	1.05	cps
Rhodium	103-2	5339703	5322778	5279178	5313887	0.59	cps
Scandium	45-1	10434892	10510160	10430920	10458657	0.43	cps
Scandium	45-2	185004	183519	185959	184827	0.67	cps
Selenium	82-1	31012	30778	31142	30977	0.60	cps
Selenium	77-2	463	460	487	470	3.09	cps
Selenium	78-2	1537	1667	1627	1610	4.14	cps
Silicon	28-1	2145101	2115919	2142495	2134505	0.76	cps
Silver	107-1	2768892	2738971	2757157	2755007	0.55	cps
Silver	109-1	2623589	2637883	2638520	2633331	0.32	cps
Sodium	23-2	5717072	5679413	5746307	5714264	0.59	cps
Strontium	86-1	658443	663419	670289	664050	0.90	cps
Strontium	88-1	5842458	5881541	5960140	5894713	1.02	cps
Sulfur	34-1	937034	966793	934475	946101	1.90	cps
Terbium	159-1	21272376	20977451	21016563	21088796	0.76	cps
Terbium	159-2	6301849	6242359	6179182	6241130	0.98	cps
Thallium	203-1	2180431	2212830	2203345	2198869	0.76	cps
Thallium	205-1	5307194	5263150	5227958	5266101	0.75	cps
Tin	118-1	1763345	1733193	1733849	1743462	0.99	cps
Titanium	47-1	3096615	3128598	3147248	3124154	0.82	cps
Uranium	238-1	7051129	7114458	7130242	7098610	0.59	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S04 Instrumnet Name : P8
Client Sample ID : S04 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:15:02 DataFile Name : 007CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	312258	310936	315316	312837	0.72	cps
Ytterbium	172-1	110	183	140	144	25.53	cps
Ytterbium	172-2	53	60	50	54	9.35	cps
Ytterbium	176-1	12089	12473	12079	12214	1.84	cps
Ytterbium	176-2	3507	3524	3477	3503	0.68	cps
Yttrium	89-1	27252918	26872171	27229779	27118289	0.79	cps
Yttrium	89-2	1704380	1715493	1725344	1715072	0.61	cps
Zinc	66-2	648243	652368	653929	651513	0.45	cps
Zirconium	90-1	3661290	3697553	3668607	3675817	0.52	cps
Zirconium	91-1	794656	798387	798881	797308	0.29	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S05 Instrumnet Name : P8
Client Sample ID : S05 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:18:01 DataFile Name : 008CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	312364	309106	309287	310252	0.59	cps
Antimony	121-1	4252500	4266716	4313114	4277443	0.74	cps
Arsenic	75-2	73845	75042	73658	74182	1.01	cps
Barium	135-1	5151990	5169594	5212134	5177906	0.60	cps
Barium	137-1	8784702	8937141	8955277	8892373	1.05	cps
Beryllium	9-1	1361679	1373684	1369123	1368162	0.44	cps
Bismuth	209-1	13313163	12990345	12819564	13041024	1.92	cps
Bismuth	209-2	4764676	4822292	4959287	4848752	2.06	cps
Bromine	81-1	3901	3927	4081	3969	2.45	cps
Cadmium	108-1	84879	86892	85231	85667	1.25	cps
Cadmium	106-1	126049	128057	127552	127219	0.82	cps
Cadmium	111-1	1032022	1051711	1052785	1045506	1.12	cps
Calcium	43-1	1299438	1340927	1315668	1318677	1.59	cps
Calcium	44-1	20692978	21459617	21121501	21091365	1.82	cps
Carbon	12-1	3662095	3721340	3760407	3714614	1.33	cps
Carbon	12-2	26460	25709	26370	26180	1.57	cps
Chlorine	35-1	130946	133482	133828	132752	1.19	cps
Chlorine	35-2	447	480	497	474	5.37	cps
Chromium	52-2	800262	801017	798068	799782	0.19	cps
Cobalt	59-2	1597019	1575898	1577618	1583512	0.74	cps
Copper	63-2	12339749	12538734	12356776	12411753	0.89	cps
Dysprosium	156-1	310	327	307	314	3.41	cps
Dysprosium	156-2	30	30	10	23	49.49	cps
Erbium	164-1	133	133	150	139	6.93	cps
Erbium	164-2	67	47	50	54	19.68	cps
Gadolinium	160-1	167	140	190	166	15.11	cps
Gadolinium	160-2	50	23	40	38	35.67	cps
Holmium	165-1	21209266	20709873	20287729	20735622	2.22	cps
Holmium	165-2	6530785	6320486	6574396	6475222	2.10	cps
Indium	115-1	16271167	15630076	15765957	15889066	2.13	cps
Indium	115-2	1338590	1319639	1388900	1349043	2.65	cps
Iron	54-2	1734082	1696726	1713286	1714698	1.09	cps
Iron	56-2	30837090	31015804	30655718	30836204	0.58	cps
Iron	57-2	755760	760624	760940	759108	0.38	cps
Krypton	83-1	293	253	247	264	9.54	cps
Lead	206-1	17696720	17716294	17720760	17711258	0.07	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S05 Instrumnet Name : P8
Client Sample ID : S05 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:18:01 DataFile Name : 008CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	14993644	15185476	15107397	15095506	0.64	cps
Lead	208-1	69383994	70318296	69855043	69852444	0.67	cps
Lithium	6-1	7487475	7325423	7181148	7331349	2.09	cps
Magnesium	24-2	5932262	5834666	5781024	5849317	1.31	cps
Manganese	55-2	3329841	3314469	3345719	3330010	0.47	cps
Molybdenum	94-1	14283499	14390014	14309477	14327663	0.39	cps
Molybdenum	95-1	20934362	21087313	20820556	20947410	0.64	cps
Molybdenum	96-1	22827276	23123186	22843183	22931215	0.73	cps
Molybdenum	97-1	12877961	12884673	12996096	12919577	0.51	cps
Molybdenum	98-1	33710703	34181160	34180223	34024029	0.80	cps
Neodymium	150-1	333	370	323	342	7.18	cps
Neodymium	150-2	7	20	10	12	56.76	cps
Nickel	60-2	429757	432499	429113	430456	0.42	cps
Phosphorus	31-2	14457	14641	14140	14413	1.76	cps
Potassium	39-2	1530442	1542048	1528991	1533827	0.47	cps
Rhodium	103-1	14729377	14394934	14141878	14422063	2.04	cps
Rhodium	103-2	5212580	5156799	5383180	5250853	2.25	cps
Scandium	45-1	10608094	10492659	10346866	10482540	1.25	cps
Scandium	45-2	184421	180803	188250	184491	2.02	cps
Selenium	82-1	61482	63769	62068	62439	1.90	cps
Selenium	77-2	987	990	840	939	9.12	cps
Selenium	78-2	3174	3124	3254	3184	2.06	cps
Silicon	28-1	3659628	3701099	3757618	3706115	1.33	cps
Silver	107-1	5386278	5439302	5423455	5416345	0.50	cps
Silver	109-1	5093295	5139964	5144134	5125797	0.55	cps
Sodium	23-2	11551552	11492227	11321819	11455199	1.04	cps
Strontium	86-1	1300823	1308379	1319779	1309661	0.73	cps
Strontium	88-1	11714221	11751618	11674771	11713537	0.33	cps
Sulfur	34-1	1093495	1108879	1109954	1104109	0.83	cps
Terbium	159-1	21356366	20668220	21018412	21014332	1.64	cps
Terbium	159-2	6156864	6143251	6464572	6254896	2.91	cps
Thallium	203-1	4367558	4432607	4371154	4390439	0.83	cps
Thallium	205-1	10210487	10306705	10183173	10233455	0.63	cps
Tin	118-1	3437706	3398946	3475028	3437227	1.11	cps
Titanium	47-1	6089271	6158833	6003403	6083836	1.28	cps
Uranium	238-1	13838342	14022877	13713204	13858141	1.12	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S05 Instrumnet Name : P8
Client Sample ID : S05 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:18:01 DataFile Name : 008CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	624364	623205	615548	621039	0.77	cps
Ytterbium	172-1	157	153	137	149	7.20	cps
Ytterbium	172-2	93	60	73	76	22.20	cps
Ytterbium	176-1	22503	21892	22172	22189	1.38	cps
Ytterbium	176-2	6512	6575	6708	6598	1.52	cps
Yttrium	89-1	27551317	26654479	26523855	26909883	2.08	cps
Yttrium	89-2	1691862	1643530	1759845	1698412	3.44	cps
Zinc	66-2	1286808	1288963	1283248	1286340	0.22	cps
Zirconium	90-1	7264335	7319011	7257533	7280293	0.46	cps
Zirconium	91-1	1635766	1627491	1646655	1636637	0.59	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S06 Instrumnet Name : P8
Client Sample ID : S06 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:20:51 DataFile Name : 009CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	607203	609205	609822	608743	0.22	cps
Antimony	121-1	8463293	8562131	8584291	8536571	0.75	cps
Arsenic	75-2	145118	148051	146054	146408	1.02	cps
Barium	135-1	10320615	10355662	10301380	10325886	0.27	cps
Barium	137-1	17851688	17651255	17701861	17734935	0.59	cps
Beryllium	9-1	2727373	2678639	2679451	2695154	1.04	cps
Bismuth	209-1	12942676	12583387	12869207	12798423	1.48	cps
Bismuth	209-2	4657111	4725707	4640494	4674437	0.97	cps
Bromine	81-1	3991	3884	3861	3912	1.77	cps
Cadmium	108-1	169143	167884	170385	169137	0.74	cps
Cadmium	106-1	244958	246018	243566	244848	0.50	cps
Cadmium	111-1	2086245	2174443	2085266	2115318	2.42	cps
Calcium	43-1	2620377	2650772	2638075	2636408	0.58	cps
Calcium	44-1	42570386	41686527	42918814	42391909	1.50	cps
Carbon	12-1	3906252	3971130	4039518	3972300	1.68	cps
Carbon	12-2	28985	29075	28541	28867	0.99	cps
Chlorine	35-1	142735	143116	146082	143977	1.27	cps
Chlorine	35-2	553	570	547	557	2.16	cps
Chromium	52-2	1662985	1638207	1615854	1639015	1.44	cps
Cobalt	59-2	3111036	3125440	3158972	3131816	0.79	cps
Copper	63-2	24086260	24076182	24014500	24058981	0.16	cps
Dysprosium	156-1	353	377	333	354	6.12	cps
Dysprosium	156-2	27	23	33	28	18.33	cps
Erbium	164-1	200	160	187	182	11.18	cps
Erbium	164-2	53	63	50	56	12.49	cps
Gadolinium	160-1	220	217	173	203	12.80	cps
Gadolinium	160-2	60	53	50	54	9.35	cps
Holmium	165-1	20759480	20549427	20493392	20600766	0.68	cps
Holmium	165-2	6442398	6488947	6441124	6457490	0.42	cps
Indium	115-1	15762101	15568416	15213068	15514529	1.79	cps
Indium	115-2	1302560	1305982	1309783	1306109	0.28	cps
Iron	54-2	3420010	3368823	3374623	3387819	0.83	cps
Iron	56-2	60885091	61186556	60281579	60784409	0.76	cps
Iron	57-2	1552134	1580667	1555214	1562672	1.00	cps
Krypton	83-1	293	290	290	291	0.66	cps
Lead	206-1	35462491	35373334	35498237	35444687	0.18	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S06 Instrumnet Name : P8
Client Sample ID : S06 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:20:51 DataFile Name : 009CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	30510363	30121232	30287233	30306276	0.64	cps
Lead	208-1	140553358	138932536	139873282	139786392	0.58	cps
Lithium	6-1	7287956	7263823	7150297	7234025	1.02	cps
Magnesium	24-2	11329883	11374769	11439855	11381502	0.49	cps
Manganese	55-2	6547180	6534003	6410631	6497271	1.16	cps
Molybdenum	94-1	29031846	28887644	28935524	28951671	0.25	cps
Molybdenum	95-1	42017699	41907868	42360721	42095429	0.56	cps
Molybdenum	96-1	45754286	45523233	46404339	45893953	1.00	cps
Molybdenum	97-1	25840026	25689890	25754914	25761610	0.29	cps
Molybdenum	98-1	68871937	68083521	67683009	68212822	0.89	cps
Neodymium	150-1	650	697	603	650	7.18	cps
Neodymium	150-2	13	20	17	17	20.01	cps
Nickel	60-2	832191	843350	837122	837554	0.67	cps
Phosphorus	31-2	29226	29507	29407	29380	0.48	cps
Potassium	39-2	3046379	3013239	3031221	3030280	0.55	cps
Rhodium	103-1	14460019	14043339	13936805	14146721	1.95	cps
Rhodium	103-2	5084232	5127604	5014711	5075516	1.12	cps
Scandium	45-1	10478442	10064474	10260609	10267841	2.02	cps
Scandium	45-2	180549	182793	182689	182010	0.70	cps
Selenium	82-1	122166	123786	121032	122328	1.13	cps
Selenium	77-2	1803	1800	1783	1796	0.60	cps
Selenium	78-2	6031	5898	6121	6017	1.87	cps
Silicon	28-1	6755868	6861979	6892155	6836667	1.05	cps
Silver	107-1	10490453	10613701	10691255	10598470	0.96	cps
Silver	109-1	9952605	10168089	10121820	10080838	1.13	cps
Sodium	23-2	22758015	22645304	22807711	22737010	0.37	cps
Strontium	86-1	2675469	2697756	2684703	2685976	0.42	cps
Strontium	88-1	23186521	23113263	23295741	23198509	0.40	cps
Sulfur	34-1	1440525	1437112	1428745	1435461	0.42	cps
Terbium	159-1	21168585	21159863	21144342	21157597	0.06	cps
Terbium	159-2	6117703	6281247	6175733	6191561	1.34	cps
Thallium	203-1	8515575	8559202	8675637	8583471	0.96	cps
Thallium	205-1	20328353	20635199	20666248	20543267	0.91	cps
Tin	118-1	6721864	6795603	6754376	6757281	0.55	cps
Titanium	47-1	12351255	12146266	12237509	12245010	0.84	cps
Uranium	238-1	27733745	27946810	27631951	27770835	0.58	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S06 Instrumnet Name : P8
Client Sample ID : S06 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:20:51 DataFile Name : 009CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	1244265	1234892	1230532	1236563	0.57	cps
Ytterbium	172-1	197	253	240	230	12.88	cps
Ytterbium	172-2	47	67	63	59	18.19	cps
Ytterbium	176-1	42602	40863	42337	41934	2.23	cps
Ytterbium	176-2	12960	13117	13343	13140	1.47	cps
Yttrium	89-1	27132938	26555681	26802240	26830286	1.08	cps
Yttrium	89-2	1687718	1664263	1658404	1670128	0.93	cps
Zinc	66-2	2576205	2610806	2609684	2598898	0.76	cps
Zirconium	90-1	14580628	14517316	14898375	14665440	1.39	cps
Zirconium	91-1	3252433	3246713	3362852	3287332	1.99	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S07 Instrumnet Name : P8
Client Sample ID : S07 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:23:38 DataFile Name : 010CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1253403	1259425	1229557	1247462	1.27	cps
Antimony	121-1	16472252	16591189	16704403	16589281	0.70	cps
Arsenic	75-2	290723	290754	286900	289459	0.77	cps
Barium	135-1	19996131	19972217	20286261	20084870	0.87	cps
Barium	137-1	34975251	34631722	34680037	34762337	0.53	cps
Beryllium	9-1	5173412	5268338	5238123	5226624	0.93	cps
Bismuth	209-1	12406000	12609147	12411609	12475585	0.93	cps
Bismuth	209-2	4678991	4655771	4639686	4658149	0.42	cps
Bromine	81-1	3901	4044	4171	4038	3.35	cps
Cadmium	108-1	324564	330107	329015	327896	0.90	cps
Cadmium	106-1	463915	469838	467456	467070	0.64	cps
Cadmium	111-1	4039130	4062925	4092556	4064871	0.66	cps
Calcium	43-1	5099674	5152446	5099255	5117125	0.60	cps
Calcium	44-1	82942632	82818747	83246524	83002634	0.27	cps
Carbon	12-1	4366607	4520629	4641762	4509666	3.06	cps
Carbon	12-2	34110	34340	35493	34647	2.14	cps
Chlorine	35-1	161836	165790	167757	165128	1.83	cps
Chlorine	35-2	630	583	677	630	7.41	cps
Chromium	52-2	3284374	3248647	3261141	3264721	0.56	cps
Cobalt	59-2	6100678	5998346	6124696	6074573	1.10	cps
Copper	63-2	47449666	46659593	46918129	47009129	0.86	cps
Dysprosium	156-1	490	547	537	524	5.77	cps
Dysprosium	156-2	53	37	43	44	18.87	cps
Erbium	164-1	273	287	253	271	6.19	cps
Erbium	164-2	100	77	117	98	20.55	cps
Gadolinium	160-1	257	250	230	246	5.65	cps
Gadolinium	160-2	100	43	67	70	40.69	cps
Holmium	165-1	20252447	20507497	20155254	20305066	0.90	cps
Holmium	165-2	6470568	6469976	6408374	6449639	0.55	cps
Indium	115-1	14918787	15100736	14984400	15001307	0.61	cps
Indium	115-2	1276170	1281772	1272294	1276746	0.37	cps
Iron	54-2	6650436	6660262	6656094	6655597	0.07	cps
Iron	56-2	120342328	119255895	120285395	119961206	0.51	cps
Iron	57-2	3063447	3031906	3126224	3073859	1.56	cps
Krypton	83-1	333	310	310	318	4.24	cps
Lead	206-1	67793259	68327722	68119762	68080248	0.40	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S07 Instrumnet Name : P8
Client Sample ID : S07 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:23:38 DataFile Name : 010CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	58822957	58852269	58491642	58722290	0.34	cps
Lead	208-1	270014228	271731893	270174799	270640307	0.35	cps
Lithium	6-1	7034089	7203445	7130692	7122742	1.19	cps
Magnesium	24-2	22507369	22281817	22691778	22493655	0.91	cps
Manganese	55-2	12965757	12842416	12992662	12933612	0.62	cps
Molybdenum	94-1	56381386	56842971	57035257	56753205	0.59	cps
Molybdenum	95-1	82222499	82734274	83445289	82800687	0.74	cps
Molybdenum	96-1	89213772	90138465	91652042	90334760	1.36	cps
Molybdenum	97-1	50615808	50990074	51667063	51090981	1.04	cps
Molybdenum	98-1	133276021	133777898	133875238	133643052	0.24	cps
Neodymium	150-1	1267	1370	1197	1278	6.82	cps
Neodymium	150-2	27	43	53	41	32.76	cps
Nickel	60-2	1732698	1708735	1699299	1713577	1.00	cps
Phosphorus	31-2	57192	57068	56506	56922	0.64	cps
Potassium	39-2	5888830	5908282	5845228	5880780	0.55	cps
Rhodium	103-1	13592335	13824641	13668142	13695039	0.87	cps
Rhodium	103-2	5052150	4924117	4966561	4980942	1.31	cps
Scandium	45-1	10248553	10077105	10354024	10226561	1.37	cps
Scandium	45-2	183829	183252	184454	183845	0.33	cps
Selenium	82-1	238742	241081	241701	240508	0.65	cps
Selenium	77-2	3641	3387	3510	3513	3.61	cps
Selenium	78-2	12105	11969	12095	12056	0.63	cps
Silicon	28-1	12650303	12745443	12758371	12718039	0.46	cps
Silver	107-1	20410596	20350261	20440115	20400324	0.22	cps
Silver	109-1	19255426	19655148	19510751	19473775	1.04	cps
Sodium	23-2	44782258	44720581	44903584	44802141	0.21	cps
Strontium	86-1	5292860	5195921	5233670	5240817	0.93	cps
Strontium	88-1	46179984	45946558	45795083	45973875	0.42	cps
Sulfur	34-1	2077008	2060966	2065786	2067920	0.40	cps
Terbium	159-1	20631501	20772228	20932101	20778610	0.72	cps
Terbium	159-2	6240808	6136736	6076717	6151420	1.35	cps
Thallium	203-1	16451770	16523171	16519770	16498237	0.24	cps
Thallium	205-1	39217547	39974092	39757666	39649768	0.98	cps
Tin	118-1	13152923	13382500	13291741	13275721	0.87	cps
Titanium	47-1	24214779	23944945	24010863	24056862	0.58	cps
Uranium	238-1	54415436	55262309	55113348	54930364	0.82	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S07 Instrumnet Name : P8
Client Sample ID : S07 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:23:38 DataFile Name : 010CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	2529206	2531372	2481206	2513928	1.13	cps
Ytterbium	172-1	297	367	277	313	15.08	cps
Ytterbium	172-2	137	127	120	128	6.57	cps
Ytterbium	176-1	81700	81553	80808	81354	0.59	cps
Ytterbium	176-2	25899	25735	25852	25829	0.33	cps
Yttrium	89-1	26715645	26280949	26391853	26462815	0.85	cps
Yttrium	89-2	1723020	1679193	1682955	1695056	1.43	cps
Zinc	66-2	5022769	5018198	4968367	5003111	0.60	cps
Zirconium	90-1	28473997	28393913	28533190	28467033	0.25	cps
Zirconium	91-1	6384881	6356800	6485913	6409198	1.06	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S08 Instrumnet Name : P8
Client Sample ID : S08 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:26:23 DataFile Name : 011CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	5997963	5929553	5935001	5954172	0.64	cps
Antimony	121-1	14071	13460	13283	13605	3.04	cps
Arsenic	75-2	143	170	170	161	9.55	cps
Barium	135-1	7749	7692	7405	7616	2.42	cps
Barium	137-1	13600	12990	13006	13199	2.64	cps
Beryllium	9-1	2172	1882	1772	1942	10.64	cps
Bismuth	209-1	10846868	10800313	10894732	10847304	0.44	cps
Bismuth	209-2	4114555	4113727	4044323	4090869	0.99	cps
Bromine	81-1	4161	4291	4341	4264	2.18	cps
Cadmium	108-1	297	300	197	264	22.21	cps
Cadmium	106-1	8129	7876	7689	7898	2.80	cps
Cadmium	111-1	6329	6013	5908	6083	3.61	cps
Calcium	43-1	23143771	23365488	23513175	23340811	0.80	cps
Calcium	44-1	380211594	379181701	382738954	380710750	0.48	cps
Carbon	12-1	3965885	4095558	4147394	4069612	2.30	cps
Carbon	12-2	38413	38300	38617	38443	0.42	cps
Chlorine	35-1	128061	124644	124674	125793	1.56	cps
Chlorine	35-2	530	417	470	472	12.01	cps
Chromium	52-2	10574	10157	10264	10332	2.10	cps
Cobalt	59-2	12559	12856	12736	12717	1.17	cps
Copper	63-2	11171	11501	10858	11177	2.88	cps
Dysprosium	156-1	710	743	637	697	7.83	cps
Dysprosium	156-2	190	220	180	197	10.58	cps
Erbium	164-1	747	683	770	733	6.12	cps
Erbium	164-2	263	210	203	226	14.58	cps
Gadolinium	160-1	597	630	653	627	4.54	cps
Gadolinium	160-2	213	250	237	233	7.95	cps
Holmium	165-1	19232303	18871630	19022875	19042269	0.95	cps
Holmium	165-2	6129933	6094578	6086617	6103709	0.38	cps
Indium	115-1	14053919	14057400	13678483	13929934	1.56	cps
Indium	115-2	1231167	1247569	1202890	1227209	1.84	cps
Iron	54-2	30519369	30485153	30561177	30521900	0.12	cps
Iron	56-2	566999965	569142645	551400818	562514476	1.72	cps
Iron	57-2	14176053	13862700	14073561	14037438	1.14	cps
Krypton	83-1	367	377	330	358	6.87	cps
Lead	206-1	14862	14618	14345	14608	1.77	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S08 Instrumnet Name : P8
Client Sample ID : S08 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:26:23 DataFile Name : 011CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	12766	12586	11882	12412	3.76	cps
Lead	208-1	59731	58038	55537	57769	3.65	cps
Lithium	6-1	6501263	6605274	6406319	6504285	1.53	cps
Magnesium	24-2	107258228	107353318	105845688	106819078	0.79	cps
Manganese	55-2	5738	5955	5911	5868	1.95	cps
Molybdenum	94-1	40154	40505	40037	40232	0.61	cps
Molybdenum	95-1	41793	40756	41482	41344	1.29	cps
Molybdenum	96-1	47166	48548	47601	47772	1.48	cps
Molybdenum	97-1	25403	25891	25200	25498	1.39	cps
Molybdenum	98-1	65755	65557	65031	65448	0.57	cps
Neodymium	150-1	377	387	357	373	4.09	cps
Neodymium	150-2	90	53	67	70	26.51	cps
Nickel	60-2	6528	6672	6815	6672	2.15	cps
Phosphorus	31-2	103	93	70	89	19.24	cps
Potassium	39-2	28639990	28664270	29339885	28881381	1.38	cps
Rhodium	103-1	12204182	12400409	11983816	12196136	1.71	cps
Rhodium	103-2	4457509	4488489	4390386	4445462	1.13	cps
Scandium	45-1	9586493	9772201	9497822	9618839	1.46	cps
Scandium	45-2	176300	175794	172812	174969	1.08	cps
Selenium	82-1	-27	-60	-7	-31	-86.60	cps
Selenium	77-2	10	7	0	6	91.64	cps
Selenium	78-2	23	17	7	16	53.90	cps
Silicon	28-1	684150	668266	658579	670332	1.93	cps
Silver	107-1	3537	3687	3504	3576	2.73	cps
Silver	109-1	3584	3340	3007	3310	8.75	cps
Sodium	23-2	204932184	205863817	205574984	205456995	0.23	cps
Strontium	86-1	16333	15759	15806	15966	2.00	cps
Strontium	88-1	136805	136295	135725	136275	0.40	cps
Sulfur	34-1	682056	684087	682100	682748	0.17	cps
Terbium	159-1	19117023	19123370	19102300	19114231	0.06	cps
Terbium	159-2	5916969	5920016	5738151	5858379	1.78	cps
Thallium	203-1	1523	1537	1383	1481	5.73	cps
Thallium	205-1	4134	3450	3307	3631	12.17	cps
Tin	118-1	4324	4231	3764	4106	7.31	cps
Titanium	47-1	4224	4561	4504	4430	4.07	cps
Uranium	238-1	2630	2424	2204	2419	8.82	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : S08 Instrumnet Name : P8
Client Sample ID : S08 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 12:26:23 DataFile Name : 011CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	683	763	717	721	5.57	cps
Ytterbium	172-1	837	813	937	862	7.60	cps
Ytterbium	172-2	290	360	347	332	11.19	cps
Ytterbium	176-1	2664	2574	2674	2637	2.09	cps
Ytterbium	176-2	563	517	503	528	5.97	cps
Yttrium	89-1	25019185	24974695	24521135	24838339	1.11	cps
Yttrium	89-2	1658061	1631685	1582599	1624115	2.36	cps
Zinc	66-2	2640	2800	2870	2770	4.26	cps
Zirconium	90-1	46524	46591	45601	46239	1.20	cps
Zirconium	91-1	10447	10241	10180	10289	1.36	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : ICV005 Instrumnet Name : P8
Client Sample ID : ICV005 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:29:04 DataFile Name : 012ICV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	30272	30151	30395	30273	0.40	cps
Antimony	121-1	3516543	3504548	3549952	3523681	0.67	cps
Arsenic	75-2	60432	61571	60881	60961	0.94	cps
Barium	135-1	416525	413590	417893	416003	0.53	cps
Barium	137-1	721616	722114	731831	725187	0.79	cps
Beryllium	9-1	570422	571316	566929	569556	0.41	cps
Bismuth	209-1	13542598	13433284	13461624	13479169	0.42	cps
Bismuth	209-2	4961829	5012488	4997408	4990575	0.52	cps
Bromine	81-1	3794	3814	3871	3826	1.04	cps
Cadmium	108-1	44719	45849	45950	45506	1.50	cps
Cadmium	106-1	46772	46394	47512	46893	1.21	cps
Cadmium	111-1	444618	445941	445885	445481	0.17	cps
Calcium	43-1	118516	118563	116853	117977	0.83	cps
Calcium	44-1	1915951	1876150	1902369	1898157	1.07	cps
Carbon	12-1	4283972	4408284	4393500	4361919	1.56	cps
Carbon	12-2	30121	30161	29954	30079	0.37	cps
Chlorine	35-1	120728	119004	119471	119734	0.74	cps
Chlorine	35-2	433	533	413	460	13.98	cps
Chromium	52-2	337270	335155	335097	335841	0.37	cps
Cobalt	59-2	655095	658932	652897	655641	0.47	cps
Copper	63-2	515175	521522	517964	518220	0.61	cps
Dysprosium	156-1	100	110	93	101	8.30	cps
Dysprosium	156-2	3	13	7	8	65.47	cps
Erbium	164-1	127	103	140	123	15.04	cps
Erbium	164-2	30	50	37	39	26.18	cps
Gadolinium	160-1	123	100	153	126	21.30	cps
Gadolinium	160-2	17	20	13	17	20.01	cps
Holmium	165-1	21315476	20733988	20844497	20964653	1.47	cps
Holmium	165-2	6508822	6500566	6488707	6499365	0.16	cps
Indium	115-1	16337170	16282833	16102265	16240756	0.76	cps
Indium	115-2	1390167	1384283	1386422	1386958	0.21	cps
Iron	54-2	276886	280677	277450	278338	0.74	cps
Iron	56-2	5021218	4966507	4921785	4969837	1.00	cps
Iron	57-2	124676	122453	123428	123519	0.90	cps
Krypton	83-1	267	240	290	266	9.42	cps
Lead	206-1	3071860	3022364	2992500	3028908	1.32	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : ICV005 Instrumnet Name : P8
Client Sample ID : ICV005 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:29:04 DataFile Name : 012ICV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	2462501	2498942	2483373	2481605	0.74	cps
Lead	208-1	11597939	11618483	11576345	11597589	0.18	cps
Lithium	6-1	7734904	7482279	7583045	7600076	1.67	cps
Magnesium	24-2	265715	264350	265791	265286	0.31	cps
Manganese	55-2	130355	133803	130647	131602	1.45	cps
Molybdenum	94-1	23123643	22878052	22585228	22862308	1.18	cps
Molybdenum	95-1	41360599	41162430	40480217	41001082	1.13	cps
Molybdenum	96-1	44438539	43643673	43716906	43933039	1.00	cps
Molybdenum	97-1	25133275	25027085	25165682	25108681	0.29	cps
Molybdenum	98-1	66017892	66552564	65240604	65937020	1.00	cps
Neodymium	150-1	57	47	67	57	17.65	cps
Neodymium	150-2	3	7	7	6	34.70	cps
Nickel	60-2	182518	185330	181802	183217	1.02	cps
Phosphorus	31-2	93	73	87	84	12.06	cps
Potassium	39-2	243935	244095	242794	243608	0.29	cps
Rhodium	103-1	15167069	14817089	14778421	14920860	1.43	cps
Rhodium	103-2	5406739	5414773	5356488	5392667	0.59	cps
Scandium	45-1	10896802	10725958	10613273	10745344	1.33	cps
Scandium	45-2	186862	186166	185737	186255	0.30	cps
Selenium	82-1	52586	52656	53001	52747	0.42	cps
Selenium	77-2	730	880	833	814	9.43	cps
Selenium	78-2	2700	2660	2707	2689	0.94	cps
Silicon	28-1	550322	543271	542306	545300	0.80	cps
Silver	107-1	1019729	1037840	1031261	1029610	0.89	cps
Silver	109-1	979772	985028	997669	987490	0.93	cps
Sodium	23-2	955750	941739	945627	947705	0.76	cps
Strontium	86-1	2902978	2859200	2905685	2889287	0.90	cps
Strontium	88-1	25059375	24893650	24965774	24972933	0.33	cps
Sulfur	34-1	783983	773978	767137	775033	1.09	cps
Terbium	159-1	21581733	21301225	21270942	21384633	0.80	cps
Terbium	159-2	6304495	6245242	6273710	6274482	0.47	cps
Thallium	203-1	3693848	3597194	3581029	3624024	1.68	cps
Thallium	205-1	8639799	8517454	8527706	8561653	0.79	cps
Tin	118-1	3921	4097	4174	4064	3.20	cps
Titanium	47-1	1807	1757	1807	1790	1.61	cps
Uranium	238-1	747	820	713	760	7.18	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : ICV005 Instrumnet Name : P8
Client Sample ID : ICV005 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:29:04 DataFile Name : 012ICV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	250410	252391	249556	250786	0.58	cps
Ytterbium	172-1	120	127	123	123	2.70	cps
Ytterbium	172-2	37	27	47	37	27.27	cps
Ytterbium	176-1	1893	2110	2094	2032	5.93	cps
Ytterbium	176-2	383	457	417	419	8.77	cps
Yttrium	89-1	28039401	27368147	27300040	27569196	1.48	cps
Yttrium	89-2	1716657	1739921	1685710	1714096	1.59	cps
Zinc	66-2	102745	104814	102695	103418	1.17	cps
Zirconium	90-1	1407	1483	1543	1478	4.64	cps
Zirconium	91-1	970	917	1083	990	8.60	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : ICB005 Instrumnet Name : P8
Client Sample ID : ICB005 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:33:28 DataFile Name : 013CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	53	83	113	83	36.01	cps
Antimony	121-1	693	527	550	590	15.30	cps
Arsenic	75-2	7	3	3	4	43.40	cps
Barium	135-1	307	293	270	290	6.40	cps
Barium	137-1	423	440	453	439	3.42	cps
Beryllium	9-1	429	446	427	434	2.41	cps
Bismuth	209-1	13542910	13478139	13431154	13484068	0.42	cps
Bismuth	209-2	5023546	5052107	5098193	5057949	0.74	cps
Bromine	81-1	3781	3704	3741	3742	1.03	cps
Cadmium	108-1	27	40	40	36	21.64	cps
Cadmium	106-1	8379	8203	7976	8186	2.47	cps
Cadmium	111-1	5900	5764	5592	5752	2.69	cps
Calcium	43-1	450	357	437	414	12.18	cps
Calcium	44-1	25509	25025	24648	25061	1.72	cps
Carbon	12-1	3370678	3347989	3363624	3360764	0.35	cps
Carbon	12-2	23112	22654	22684	22817	1.12	cps
Chlorine	35-1	118402	118587	118399	118463	0.09	cps
Chlorine	35-2	503	477	367	449	16.14	cps
Chromium	52-2	763	790	753	769	2.47	cps
Cobalt	59-2	123	130	143	132	7.70	cps
Copper	63-2	2944	2717	2957	2873	4.70	cps
Dysprosium	156-1	67	77	40	61	31.02	cps
Dysprosium	156-2	0	0	3	1	173.21	cps
Erbium	164-1	67	80	87	78	13.09	cps
Erbium	164-2	47	23	23	31	43.32	cps
Gadolinium	160-1	130	113	120	121	6.92	cps
Gadolinium	160-2	13	17	7	12	41.65	cps
Holmium	165-1	20726861	20697862	20552567	20659097	0.45	cps
Holmium	165-2	6563018	6656614	6557052	6592228	0.85	cps
Indium	115-1	16348563	16375735	16276626	16333641	0.31	cps
Indium	115-2	1384810	1388247	1400582	1391213	0.60	cps
Iron	54-2	417	413	400	410	2.15	cps
Iron	56-2	6175	5735	6011	5974	3.73	cps
Iron	57-2	113	137	143	131	12.01	cps
Krypton	83-1	260	297	267	274	7.12	cps
Lead	206-1	2374	2457	2514	2448	2.88	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : ICB005 Instrumnet Name : P8
Client Sample ID : ICB005 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:33:28 DataFile Name : 013CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	2154	1927	2094	2058	5.71	cps
Lead	208-1	9258	9358	9508	9375	1.34	cps
Lithium	6-1	7831695	7526308	7659804	7672603	2.00	cps
Magnesium	24-2	2544	2757	2544	2615	4.71	cps
Manganese	55-2	167	190	170	176	7.19	cps
Molybdenum	94-1	907	710	840	819	12.21	cps
Molybdenum	95-1	1040	890	780	903	14.45	cps
Molybdenum	96-1	1140	1033	963	1046	8.51	cps
Molybdenum	97-1	600	550	543	564	5.49	cps
Molybdenum	98-1	1560	1533	1280	1458	10.60	cps
Neodymium	150-1	3	30	17	17	80.01	cps
Neodymium	150-2	0	3	3	2	86.60	cps
Nickel	60-2	797	727	727	750	5.39	cps
Phosphorus	31-2	80	60	103	81	26.74	cps
Potassium	39-2	9436	9530	9517	9494	0.53	cps
Rhodium	103-1	15177009	14880696	14964620	15007442	1.02	cps
Rhodium	103-2	5459983	5501569	5487777	5483110	0.39	cps
Scandium	45-1	10629244	10604165	10655321	10629577	0.24	cps
Scandium	45-2	188292	189013	189489	188932	0.32	cps
Selenium	82-1	13	-50	17	-7	-563.47	cps
Selenium	77-2	0	0	0	0	0.00	cps
Selenium	78-2	3	3	3	3	0.00	cps
Silicon	28-1	515777	514132	513703	514537	0.21	cps
Silver	107-1	487	463	370	440	14.03	cps
Silver	109-1	293	303	270	289	5.92	cps
Sodium	23-2	55572	56134	55415	55707	0.68	cps
Strontium	86-1	660	733	653	682	6.51	cps
Strontium	88-1	2427	2330	2230	2329	4.22	cps
Sulfur	34-1	761967	756800	759449	759405	0.34	cps
Terbium	159-1	21228205	21269873	21450672	21316250	0.55	cps
Terbium	159-2	6331759	6283236	6392654	6335883	0.87	cps
Thallium	203-1	497	503	417	472	10.21	cps
Thallium	205-1	1143	1137	1127	1136	0.74	cps
Tin	118-1	3250	3347	3244	3280	1.76	cps
Titanium	47-1	567	477	520	521	8.64	cps
Uranium	238-1	60	40	60	53	21.65	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : ICB005 Instrumnet Name : P8
Client Sample ID : ICB005 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:33:28 DataFile Name : 013CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	13	17	10	13	25.01	cps
Ytterbium	172-1	143	150	83	126	29.25	cps
Ytterbium	172-2	43	37	47	42	12.06	cps
Ytterbium	176-1	2124	1890	2087	2034	6.17	cps
Ytterbium	176-2	377	373	473	408	13.93	cps
Yttrium	89-1	27524396	27509838	27491057	27508430	0.06	cps
Yttrium	89-2	1726713	1728911	1737181	1730935	0.32	cps
Zinc	66-2	213	277	330	273	21.37	cps
Zirconium	90-1	943	860	950	918	5.46	cps
Zirconium	91-1	180	237	220	212	13.72	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : ICSA005 Instrumnet Name : P8
Client Sample ID : ICSA005 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:40:53 DataFile Name : 015ICSA.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	6006961	6017654	5975984	6000200	0.36	cps
Antimony	121-1	19811	19554	19634	19666	0.67	cps
Arsenic	75-2	113	77	73	88	25.28	cps
Barium	135-1	5885	5821	5911	5872	0.79	cps
Barium	137-1	10337	10554	10631	10508	1.45	cps
Beryllium	9-1	1885	1909	1940	1911	1.44	cps
Bismuth	209-1	12029610	12211727	12245480	12162272	0.95	cps
Bismuth	209-2	4573909	4578975	4543534	4565473	0.42	cps
Bromine	81-1	5815	5731	6145	5897	3.71	cps
Cadmium	108-1	6372	6115	6105	6197	2.44	cps
Cadmium	106-1	7109	7722	7145	7325	4.70	cps
Cadmium	111-1	7618	8290	7680	7863	4.73	cps
Calcium	43-1	5446983	5383944	5352264	5394397	0.89	cps
Calcium	44-1	86158472	86280525	86114075	86184358	0.10	cps
Carbon	12-1	42514864	42940103	43689409	43048125	1.38	cps
Carbon	12-2	298031	297683	295211	296975	0.52	cps
Chlorine	35-1	82030902	86693605	88162952	85629153	3.74	cps
Chlorine	35-2	345867	349596	352022	349162	0.89	cps
Chromium	52-2	66631	67569	68161	67454	1.14	cps
Cobalt	59-2	7482	7575	7519	7525	0.63	cps
Copper	63-2	42430	42630	42831	42630	0.47	cps
Dysprosium	156-1	143	150	120	138	11.43	cps
Dysprosium	156-2	27	33	17	26	32.81	cps
Erbium	164-1	157	143	197	166	16.76	cps
Erbium	164-2	53	37	20	37	45.45	cps
Gadolinium	160-1	187	150	170	169	10.87	cps
Gadolinium	160-2	37	50	37	41	18.72	cps
Holmium	165-1	20617565	20658430	20432159	20569385	0.59	cps
Holmium	165-2	6575024	6500270	6485179	6520158	0.74	cps
Indium	115-1	15318846	15519039	15363973	15400619	0.68	cps
Indium	115-2	1333767	1308138	1324311	1322072	0.98	cps
Iron	54-2	13614068	13647490	13610834	13624131	0.15	cps
Iron	56-2	251060403	250758690	250514036	250777710	0.11	cps
Iron	57-2	6296719	6378436	6279789	6318315	0.83	cps
Krypton	83-1	307	223	330	287	19.56	cps
Lead	206-1	65993	66077	66694	66255	0.58	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : ICSA005 Instrumnet Name : P8
Client Sample ID : ICSA005 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:40:53 DataFile Name : 015ICSA.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	52004	52047	53333	52461	1.44	cps
Lead	208-1	245054	244953	247366	245791	0.56	cps
Lithium	6-1	7116545	7220420	7096544	7144503	0.93	cps
Magnesium	24-2	22124691	22203375	22474322	22267462	0.82	cps
Manganese	55-2	10394	10204	10707	10435	2.44	cps
Molybdenum	94-1	9372882	9387651	9455733	9405422	0.47	cps
Molybdenum	95-1	16694674	16800269	17056196	16850380	1.10	cps
Molybdenum	96-1	18127939	18221306	18521591	18290279	1.12	cps
Molybdenum	97-1	10519083	10394462	10488139	10467228	0.62	cps
Molybdenum	98-1	27343355	27308917	27594248	27415507	0.57	cps
Neodymium	150-1	57	117	67	80	40.18	cps
Neodymium	150-2	10	27	23	20	44.10	cps
Nickel	60-2	9680	9773	9970	9808	1.51	cps
Phosphorus	31-2	303682	302537	305684	303968	0.52	cps
Potassium	39-2	12365498	12291436	12374273	12343736	0.37	cps
Rhodium	103-1	13889221	13856253	13797442	13847639	0.34	cps
Rhodium	103-2	5155276	5050942	5055983	5087400	1.16	cps
Scandium	45-1	10445305	10407103	10161973	10338127	1.49	cps
Scandium	45-2	188330	185136	186905	186790	0.86	cps
Selenium	82-1	-3	107	20	41	140.98	cps
Selenium	77-2	0	3	0	1	173.21	cps
Selenium	78-2	7	13	13	11	34.61	cps
Silicon	28-1	833414	837674	841634	837574	0.49	cps
Silver	107-1	570	580	707	619	12.31	cps
Silver	109-1	383	460	633	492	26.02	cps
Sodium	23-2	46690169	47590618	47061028	47113938	0.96	cps
Strontium	86-1	184074	187746	185923	185914	0.99	cps
Strontium	88-1	1672683	1688072	1660178	1673644	0.83	cps
Sulfur	34-1	7533888	7533547	7406167	7491201	0.98	cps
Terbium	159-1	20728528	21187189	20896471	20937396	1.11	cps
Terbium	159-2	6251593	6204603	6115550	6190582	1.12	cps
Thallium	203-1	1243	1177	1427	1282	10.10	cps
Thallium	205-1	2637	2940	3247	2941	10.37	cps
Tin	118-1	4744	4584	4624	4651	1.79	cps
Titanium	47-1	5127233	5130039	5158978	5138750	0.34	cps
Uranium	238-1	840	940	1000	927	8.72	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : ICSA005 Instrumnet Name : P8
Client Sample ID : ICSA005 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:40:53 DataFile Name : 015ICSA.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	350	410	380	380	7.89	cps
Ytterbium	172-1	110	113	163	129	23.18	cps
Ytterbium	172-2	57	33	37	42	29.89	cps
Ytterbium	176-1	1843	1870	1967	1893	3.43	cps
Ytterbium	176-2	337	297	387	340	13.26	cps
Yttrium	89-1	27253494	27165486	26846611	27088530	0.79	cps
Yttrium	89-2	1740354	1705133	1737976	1727821	1.14	cps
Zinc	66-2	5808	5958	5898	5888	1.28	cps
Zirconium	90-1	1343	1337	1207	1296	5.95	cps
Zirconium	91-1	240	283	240	254	9.83	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : ICSAB005 Instrumnet Name : P8
Client Sample ID : ICSAB005 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:44:07 DataFile Name : 016ICSB.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	5994950	6050514	6012247	6019237	0.47	cps
Antimony	121-1	367934	374470	374622	372342	1.03	cps
Arsenic	75-2	6395	6562	6505	6487	1.31	cps
Barium	135-1	90305	89677	91640	90541	1.11	cps
Barium	137-1	156772	160371	158749	158631	1.14	cps
Beryllium	9-1	114145	113164	114488	113932	0.60	cps
Bismuth	209-1	12210319	12504837	12347249	12354135	1.19	cps
Bismuth	209-2	4532101	4527686	4588239	4549342	0.74	cps
Bromine	81-1	5584	5708	5871	5721	2.51	cps
Cadmium	108-1	11505	11682	11738	11642	1.05	cps
Cadmium	106-1	14424	14978	14755	14719	1.89	cps
Cadmium	111-1	90717	91055	91173	90982	0.26	cps
Calcium	43-1	5475465	5503974	5520666	5500035	0.42	cps
Calcium	44-1	88541132	89041159	87942689	88508326	0.62	cps
Carbon	12-1	43781208	44817178	45099161	44565849	1.56	cps
Carbon	12-2	305040	304895	306385	305440	0.27	cps
Chlorine	35-1	83283905	88733885	91018935	87678909	4.53	cps
Chlorine	35-2	352440	351440	356498	353459	0.76	cps
Chromium	52-2	134774	133849	135467	134697	0.60	cps
Cobalt	59-2	135601	137711	137267	136860	0.81	cps
Copper	63-2	142732	141295	143382	142469	0.75	cps
Dysprosium	156-1	77	83	120	93	25.00	cps
Dysprosium	156-2	13	40	30	28	48.51	cps
Erbium	164-1	180	190	160	177	8.65	cps
Erbium	164-2	50	57	37	48	21.31	cps
Gadolinium	160-1	167	217	147	177	20.41	cps
Gadolinium	160-2	37	37	47	40	14.43	cps
Holmium	165-1	20356436	20579217	20516423	20484025	0.56	cps
Holmium	165-2	6491588	6517134	6543364	6517362	0.40	cps
Indium	115-1	15409433	15753983	15620919	15594778	1.11	cps
Indium	115-2	1328865	1327857	1344972	1333898	0.72	cps
Iron	54-2	13914743	13765982	13776885	13819203	0.60	cps
Iron	56-2	251448883	252795470	252437643	252227332	0.28	cps
Iron	57-2	6407290	6374912	6324172	6368792	0.66	cps
Krypton	83-1	280	260	327	289	11.84	cps
Lead	206-1	348955	343612	349910	347492	0.98	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : ICSAB005 Instrumnet Name : P8
Client Sample ID : ICSAB005 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:44:07 DataFile Name : 016ICSB.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	285791	286962	288338	287030	0.44	cps
Lead	208-1	1334026	1336093	1348284	1339468	0.58	cps
Lithium	6-1	7211124	7338204	7134991	7228106	1.42	cps
Magnesium	24-2	22420017	22370046	22652528	22480864	0.67	cps
Manganese	55-2	36995	37305	36797	37032	0.69	cps
Molybdenum	94-1	9434009	9597543	9730634	9587396	1.55	cps
Molybdenum	95-1	17156441	17325048	17299038	17260175	0.53	cps
Molybdenum	96-1	18380093	18484234	18416512	18426946	0.29	cps
Molybdenum	97-1	10543509	10707248	10605771	10618843	0.78	cps
Molybdenum	98-1	27766905	27654863	27940059	27787276	0.52	cps
Neodymium	150-1	70	77	93	80	15.02	cps
Neodymium	150-2	20	20	10	17	34.64	cps
Nickel	60-2	44707	45024	45339	45023	0.70	cps
Phosphorus	31-2	307469	307618	309604	308230	0.39	cps
Potassium	39-2	12509024	12453781	12448709	12470505	0.27	cps
Rhodium	103-1	13875978	13887588	13857661	13873742	0.11	cps
Rhodium	103-2	5051135	5024338	5086858	5054110	0.62	cps
Scandium	45-1	10653580	10619027	10499199	10590602	0.77	cps
Scandium	45-2	189806	188260	189894	189320	0.49	cps
Selenium	82-1	5178	5318	5081	5192	2.29	cps
Selenium	77-2	80	97	90	89	9.44	cps
Selenium	78-2	300	370	327	332	10.63	cps
Silicon	28-1	884565	895282	897462	892436	0.77	cps
Silver	107-1	399566	406427	410427	405473	1.35	cps
Silver	109-1	384188	386227	391978	387464	1.04	cps
Sodium	23-2	47256234	47515498	47878381	47550038	0.66	cps
Strontium	86-1	185800	189458	188823	188027	1.04	cps
Strontium	88-1	1672451	1705470	1709037	1695653	1.19	cps
Sulfur	34-1	7556280	7595385	7600287	7583984	0.32	cps
Terbium	159-1	20806117	21019011	21155095	20993407	0.84	cps
Terbium	159-2	6166700	6246593	6249932	6221075	0.76	cps
Thallium	203-1	343738	351025	345870	346878	1.08	cps
Thallium	205-1	830755	825084	839998	831946	0.90	cps
Tin	118-1	5304	5318	5344	5322	0.38	cps
Titanium	47-1	5323245	5302710	5308434	5311463	0.20	cps
Uranium	238-1	977	973	857	936	7.31	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : ICSAB005 Instrumnet Name : P8
Client Sample ID : ICSAB005 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:44:07 DataFile Name : 016ICSB.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	52001	51981	52322	52101	0.37	cps
Ytterbium	172-1	113	113	123	117	4.95	cps
Ytterbium	172-2	37	47	70	51	33.46	cps
Ytterbium	176-1	1900	2077	1997	1991	4.44	cps
Ytterbium	176-2	253	270	300	274	8.62	cps
Yttrium	89-1	27140585	27803564	27560562	27501570	1.22	cps
Yttrium	89-2	1754140	1715352	1728686	1732726	1.14	cps
Zinc	66-2	16436	16400	16740	16525	1.13	cps
Zirconium	90-1	1207	1223	1237	1222	1.23	cps
Zirconium	91-1	290	277	373	313	16.72	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : CCV021 Instrumnet Name : P8
Client Sample ID : CCV021 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:47:19 DataFile Name : 017CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	3169562	3158720	3180349	3169544	0.34	cps
Antimony	121-1	8126634	8256109	8224264	8202336	0.82	cps
Arsenic	75-2	145802	147048	145903	146251	0.47	cps
Barium	135-1	10236305	10212767	10137241	10195438	0.51	cps
Barium	137-1	17604237	17519349	17311526	17478371	0.86	cps
Beryllium	9-1	2576460	2577644	2639292	2597798	1.38	cps
Bismuth	209-1	11869031	11730148	11919163	11839447	0.83	cps
Bismuth	209-2	4387952	4352452	4342162	4360855	0.55	cps
Bromine	81-1	4527	4387	4731	4549	3.80	cps
Cadmium	108-1	161075	161590	158476	160381	1.04	cps
Cadmium	106-1	229534	231582	229930	230349	0.47	cps
Cadmium	111-1	2018720	2055838	2007825	2027461	1.24	cps
Calcium	43-1	13148554	13299301	12898214	13115356	1.54	cps
Calcium	44-1	212923063	212363910	210857070	212048015	0.50	cps
Carbon	12-1	4363138	4466232	4590958	4473443	2.55	cps
Carbon	12-2	37261	37377	37217	37285	0.22	cps
Chlorine	35-1	5495177	5120077	4779648	5131634	6.97	cps
Chlorine	35-2	15842	14467	14117	14809	6.16	cps
Chromium	52-2	1642859	1676960	1662017	1660612	1.03	cps
Cobalt	59-2	3068121	3057566	3121376	3082354	1.11	cps
Copper	63-2	23192514	22883593	23089720	23055275	0.68	cps
Dysprosium	156-1	437	490	517	481	8.47	cps
Dysprosium	156-2	77	107	157	113	35.66	cps
Erbium	164-1	443	473	417	444	6.38	cps
Erbium	164-2	167	120	127	138	18.32	cps
Gadolinium	160-1	427	383	377	396	6.86	cps
Gadolinium	160-2	107	133	147	129	15.80	cps
Holmium	165-1	20411443	20154997	20036150	20200863	0.95	cps
Holmium	165-2	6471011	6410590	6391546	6424382	0.65	cps
Indium	115-1	15192598	14994091	15205673	15130787	0.78	cps
Indium	115-2	1291644	1281693	1287447	1286928	0.39	cps
Iron	54-2	16391840	16426451	16491236	16436509	0.31	cps
Iron	56-2	300779982	301480556	299584496	300615011	0.32	cps
Iron	57-2	7612747	7563826	7610558	7595710	0.36	cps
Krypton	83-1	290	327	300	306	6.20	cps
Lead	206-1	32927185	33278853	33232704	33146247	0.58	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : CCV021 Instrumnet Name : P8
Client Sample ID : CCV021 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:47:19 DataFile Name : 017CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	28030996	28589648	27969801	28196815	1.21	cps
Lead	208-1	130245371	132158258	129869079	130757569	0.94	cps
Lithium	6-1	6989319	6866152	7044700	6966724	1.31	cps
Magnesium	24-2	56922591	56457906	56650966	56677154	0.41	cps
Manganese	55-2	6443935	6526941	6528521	6499799	0.74	cps
Molybdenum	94-1	28434377	28922759	28848190	28735108	0.92	cps
Molybdenum	95-1	41195852	42220186	41233199	41549745	1.40	cps
Molybdenum	96-1	44650248	45498248	45404404	45184300	1.03	cps
Molybdenum	97-1	25333663	25504583	25677749	25505332	0.67	cps
Molybdenum	98-1	66889372	67446061	66874679	67070037	0.49	cps
Neodymium	150-1	817	783	803	801	2.09	cps
Neodymium	150-2	57	37	60	51	24.69	cps
Nickel	60-2	811957	800282	803952	805397	0.74	cps
Phosphorus	31-2	30171	29710	29440	29774	1.24	cps
Potassium	39-2	15510131	15240570	15348176	15366292	0.88	cps
Rhodium	103-1	13239419	13303216	13345204	13295946	0.40	cps
Rhodium	103-2	4993708	4819310	4900803	4904607	1.78	cps
Scandium	45-1	10331161	10345687	10185383	10287410	0.86	cps
Scandium	45-2	187276	185830	188191	187099	0.64	cps
Selenium	82-1	119667	119070	119079	119272	0.29	cps
Selenium	77-2	1823	1773	1797	1798	1.39	cps
Selenium	78-2	5988	6055	6011	6018	0.56	cps
Silicon	28-1	6806029	6890147	6910203	6868793	0.80	cps
Silver	107-1	10030707	9965765	9920619	9972364	0.55	cps
Silver	109-1	9406048	9317862	9330627	9351512	0.51	cps
Sodium	23-2	114315252	114155588	112389225	113620022	0.94	cps
Strontium	86-1	2667082	2651919	2649561	2656187	0.36	cps
Strontium	88-1	22710773	23095957	23018726	22941819	0.89	cps
Sulfur	34-1	1399744	1409353	1387830	1398975	0.77	cps
Terbium	159-1	20607077	20287746	20418101	20437641	0.79	cps
Terbium	159-2	6157029	6132648	6072720	6120799	0.71	cps
Thallium	203-1	7962308	8077453	8059295	8033019	0.77	cps
Thallium	205-1	19095211	19296247	19574710	19322056	1.25	cps
Tin	118-1	6609116	6517790	6579674	6568860	0.71	cps
Titanium	47-1	12310534	12209617	12080906	12200352	0.94	cps
Uranium	238-1	26703386	26821364	26478126	26667625	0.65	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : CCV021 Instrumnet Name : P8
Client Sample ID : CCV021 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:47:19 DataFile Name : 017CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	1265108	1263519	1293057	1273895	1.30	cps
Ytterbium	172-1	573	537	510	540	5.89	cps
Ytterbium	172-2	200	227	197	208	7.91	cps
Ytterbium	176-1	41866	41843	41050	41586	1.12	cps
Ytterbium	176-2	12946	12850	13163	12986	1.24	cps
Yttrium	89-1	26391441	26406276	26292052	26363257	0.24	cps
Yttrium	89-2	1734264	1703945	1702916	1713708	1.04	cps
Zinc	66-2	2538499	2479406	2459708	2492538	1.65	cps
Zirconium	90-1	14341067	14629049	14514052	14494723	1.00	cps
Zirconium	91-1	3265633	3336903	3199049	3267195	2.11	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : CCB021 Instrumnet Name : P8
Client Sample ID : CCB021 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:50:25 DataFile Name : 018CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	223	157	170	183	19.24	cps
Antimony	121-1	2004	1880	1870	1918	3.87	cps
Arsenic	75-2	7	20	10	12	56.76	cps
Barium	135-1	483	563	353	467	22.71	cps
Barium	137-1	887	757	817	820	7.93	cps
Beryllium	9-1	1172	1077	1002	1084	7.86	cps
Bismuth	209-1	13602345	13556552	13597166	13585354	0.18	cps
Bismuth	209-2	5091491	5063225	5023979	5059565	0.67	cps
Bromine	81-1	4037	4191	4337	4188	3.58	cps
Cadmium	108-1	50	50	20	40	43.30	cps
Cadmium	106-1	8169	8109	8086	8121	0.53	cps
Cadmium	111-1	5870	5781	5739	5797	1.15	cps
Calcium	43-1	817	720	680	739	9.51	cps
Calcium	44-1	33540	32851	32206	32866	2.03	cps
Carbon	12-1	3791397	3747498	3721546	3753480	0.94	cps
Carbon	12-2	24524	24170	24223	24306	0.78	cps
Chlorine	35-1	2413313	2255758	2144908	2271327	5.94	cps
Chlorine	35-2	7245	7169	6592	7002	5.10	cps
Chromium	52-2	757	727	730	738	2.23	cps
Cobalt	59-2	140	200	140	160	21.65	cps
Copper	63-2	3470	3277	3444	3397	3.09	cps
Dysprosium	156-1	17	27	43	29	46.62	cps
Dysprosium	156-2	0	10	7	6	91.64	cps
Erbium	164-1	113	73	117	101	23.85	cps
Erbium	164-2	37	20	47	34	39.11	cps
Gadolinium	160-1	110	117	150	126	17.07	cps
Gadolinium	160-2	43	23	7	24	75.09	cps
Holmium	165-1	21356763	21051391	20935684	21114612	1.03	cps
Holmium	165-2	6774222	6739237	6656540	6723333	0.90	cps
Indium	115-1	16981189	16889178	16766820	16879062	0.64	cps
Indium	115-2	1428029	1431880	1436139	1432016	0.28	cps
Iron	54-2	807	907	877	863	5.94	cps
Iron	56-2	14808	15192	13066	14355	7.89	cps
Iron	57-2	400	383	330	371	9.85	cps
Krypton	83-1	297	287	250	278	8.85	cps
Lead	206-1	5074	4771	4891	4912	3.11	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : CCB021 Instrumnet Name : P8
Client Sample ID : CCB021 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:50:25 DataFile Name : 018CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	4721	4204	4071	4332	7.93	cps
Lead	208-1	20426	19239	18735	19467	4.46	cps
Lithium	6-1	7817610	7793227	7890989	7833942	0.65	cps
Magnesium	24-2	2584	2517	2524	2541	1.45	cps
Manganese	55-2	340	290	287	306	9.78	cps
Molybdenum	94-1	1603	1300	1400	1435	10.78	cps
Molybdenum	95-1	1960	1630	1413	1668	16.51	cps
Molybdenum	96-1	2017	1690	1667	1791	10.93	cps
Molybdenum	97-1	1253	1070	780	1034	23.07	cps
Molybdenum	98-1	3230	2544	2304	2693	17.87	cps
Neodymium	150-1	30	27	27	28	6.92	cps
Neodymium	150-2	3	0	0	1	173.21	cps
Nickel	60-2	893	943	853	897	5.03	cps
Phosphorus	31-2	83	87	97	89	7.81	cps
Potassium	39-2	12913	13009	12302	12741	3.01	cps
Rhodium	103-1	15562460	15619567	15483628	15555218	0.44	cps
Rhodium	103-2	5666831	5643167	5677840	5662612	0.31	cps
Scandium	45-1	11127491	11072012	11074482	11091328	0.28	cps
Scandium	45-2	194925	196014	196136	195692	0.34	cps
Selenium	82-1	-3	-30	60	9	519.86	cps
Selenium	77-2	7	0	0	2	173.21	cps
Selenium	78-2	0	3	0	1	173.21	cps
Silicon	28-1	552574	550187	546972	549911	0.51	cps
Silver	107-1	1277	1277	1160	1238	5.44	cps
Silver	109-1	1150	997	1020	1056	7.83	cps
Sodium	23-2	78985	77524	77685	78065	1.03	cps
Strontium	86-1	657	673	710	680	4.01	cps
Strontium	88-1	1417	1540	1457	1471	4.28	cps
Sulfur	34-1	737293	733184	737981	736153	0.35	cps
Terbium	159-1	21390943	21324923	21431668	21382511	0.25	cps
Terbium	159-2	6463678	6407162	6523040	6464627	0.90	cps
Thallium	203-1	1653	1460	1340	1485	10.65	cps
Thallium	205-1	3737	3564	3394	3565	4.82	cps
Tin	118-1	3737	3701	3667	3702	0.95	cps
Titanium	47-1	1057	880	817	918	13.55	cps
Uranium	238-1	690	543	450	561	21.56	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : CCB021 Instrumnet Name : P8
Client Sample ID : CCB021 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 13:50:25 DataFile Name : 018CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	37	20	13	23	51.52	cps
Ytterbium	172-1	150	130	93	124	23.10	cps
Ytterbium	172-2	20	57	33	37	50.62	cps
Ytterbium	176-1	1997	2077	2074	2049	2.21	cps
Ytterbium	176-2	287	407	327	340	17.97	cps
Yttrium	89-1	28736425	28773251	28317320	28608999	0.89	cps
Yttrium	89-2	1801139	1820729	1780353	1800740	1.12	cps
Zinc	66-2	333	357	283	324	11.55	cps
Zirconium	90-1	1477	1507	1443	1476	2.15	cps
Zirconium	91-1	327	313	327	322	2.39	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-01DLX2 Instrumnet Name : P8
Client Sample ID : YE8C9 Dilution Factor : 2
Date & Time Acquired : 2025-02-11 13:53:45 DataFile Name : 019AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	786934	787424	783325	785894	0.28	cps
Antimony	121-1	5301	5378	5481	5387	1.68	cps
Arsenic	75-2	8683	8599	8933	8738	1.99	cps
Barium	135-1	513366	514803	513554	513908	0.15	cps
Barium	137-1	897467	896559	900540	898189	0.23	cps
Beryllium	9-1	3550	3658	3683	3630	1.94	cps
Bismuth	209-1	13875169	13260921	13231291	13455793	2.70	cps
Bismuth	209-2	4459130	4193644	4973991	4542255	8.73	cps
Bromine	81-1	6472	6602	6878	6651	3.12	cps
Cadmium	108-1	103	130	137	123	14.30	cps
Cadmium	106-1	8283	7696	8072	8017	3.71	cps
Cadmium	111-1	6165	5793	6095	6018	3.29	cps
Calcium	43-1	163865	165986	165236	165029	0.65	cps
Calcium	44-1	2751631	2782490	2752449	2762190	0.64	cps
Carbon	12-1	3755278	3775489	3849006	3793257	1.30	cps
Carbon	12-2	24971	25546	25329	25282	1.15	cps
Chlorine	35-1	85416520	92934442	95239615	91196859	5.63	cps
Chlorine	35-2	377007	383788	382155	380984	0.93	cps
Chromium	52-2	96530	96872	95899	96433	0.51	cps
Cobalt	59-2	103369	102361	101096	102275	1.11	cps
Copper	63-2	268500	267913	270868	269094	0.58	cps
Dysprosium	156-1	38940	38896	40070	39302	1.69	cps
Dysprosium	156-2	13450	13300	12910	13220	2.11	cps
Erbium	164-1	37245	37192	36644	37027	0.90	cps
Erbium	164-2	11418	11715	11218	11451	2.18	cps
Gadolinium	160-1	40016	40358	41524	40633	1.95	cps
Gadolinium	160-2	15499	15746	16353	15866	2.77	cps
Holmium	165-1	21747282	21212370	20677916	21212523	2.52	cps
Holmium	165-2	5978041	5520067	6732015	6076708	10.07	cps
Indium	115-1	16805816	16664419	16428827	16633020	1.14	cps
Indium	115-2	1264795	1167322	1412911	1281676	9.65	cps
Iron	54-2	4640873	4596754	4572174	4603267	0.76	cps
Iron	56-2	85661732	84754087	83996755	84804192	0.98	cps
Iron	57-2	2187726	2178328	2143375	2169810	1.08	cps
Krypton	83-1	390	377	380	382	1.82	cps
Lead	206-1	191067	194698	191358	192374	1.05	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-01DLX2 Instrumnet Name : P8
Client Sample ID : YE8C9 Dilution Factor : 2
Date & Time Acquired : 2025-02-11 13:53:45 DataFile Name : 019AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	154180	154894	154844	154639	0.26	cps
Lead	208-1	730424	739031	732268	733908	0.62	cps
Lithium	6-1	7916853	7900556	7764911	7860774	1.06	cps
Magnesium	24-2	1407189	1428229	1420957	1418792	0.75	cps
Manganese	55-2	353384	349390	345747	349507	1.09	cps
Molybdenum	94-1	6935	7165	7185	7095	1.96	cps
Molybdenum	95-1	4081	4134	4201	4138	1.45	cps
Molybdenum	96-1	5491	5451	5612	5518	1.51	cps
Molybdenum	97-1	2604	2484	2710	2599	4.36	cps
Molybdenum	98-1	6928	6642	6598	6723	2.67	cps
Neodymium	150-1	41604	42086	41918	41869	0.58	cps
Neodymium	150-2	9133	9290	8800	9074	2.76	cps
Nickel	60-2	74108	73907	73411	73809	0.49	cps
Phosphorus	31-2	1120	1087	1143	1117	2.55	cps
Potassium	39-2	53656	57768	56040	55821	3.70	cps
Rhodium	103-1	15403632	15227061	15063333	15231342	1.12	cps
Rhodium	103-2	5021695	4595390	5563967	5060350	9.59	cps
Scandium	45-1	11171588	11095183	11021964	11096245	0.67	cps
Scandium	45-2	177590	160780	197014	178462	10.16	cps
Selenium	82-1	37	143	77	86	62.98	cps
Selenium	77-2	37	60	53	50	24.03	cps
Selenium	78-2	47	37	40	41	12.39	cps
Silicon	28-1	63034696	62269712	60978356	62094255	1.67	cps
Silver	107-1	4924	4451	4337	4571	6.81	cps
Silver	109-1	4504	4314	3984	4267	6.17	cps
Sodium	23-2	94404	95194	94978	94859	0.43	cps
Strontium	86-1	150285	152145	150981	151137	0.62	cps
Strontium	88-1	1313137	1315579	1334872	1321196	0.90	cps
Sulfur	34-1	713149	706556	697304	705670	1.13	cps
Terbium	159-1	21587427	21770691	21210543	21522887	1.33	cps
Terbium	159-2	5638770	5275075	6353839	5755895	9.54	cps
Thallium	203-1	3157	3000	2937	3032	3.74	cps
Thallium	205-1	7609	7492	7269	7457	2.32	cps
Tin	118-1	4224	3971	3774	3990	5.66	cps
Titanium	47-1	21026	45709	23942	30226	44.62	cps
Uranium	238-1	13694	13681	14278	13884	2.46	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-01DLX2 Instrumnet Name : P8
Client Sample ID : YE8C9 Dilution Factor : 2
Date & Time Acquired : 2025-02-11 13:53:45 DataFile Name : 019AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	89477	90071	88518	89355	0.88	cps
Ytterbium	172-1	11061	10978	10845	10961	1.00	cps
Ytterbium	172-2	4154	4131	4177	4154	0.56	cps
Ytterbium	176-1	8890	9150	8643	8894	2.85	cps
Ytterbium	176-2	2894	2854	2794	2847	1.77	cps
Yttrium	89-1	29300030	28633977	28299642	28744550	1.77	cps
Yttrium	89-2	1653632	1521640	1814806	1663359	8.83	cps
Zinc	66-2	39121	38923	38937	38994	0.28	cps
Zirconium	90-1	14230	16986	13248	14821	13.07	cps
Zirconium	91-1	3421	3167	3214	3267	4.13	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-02DLX2 Instrumnet Name : P8
Client Sample ID : YE8C9D Dilution Factor : 2
Date & Time Acquired : 2025-02-11 13:57:03 DataFile Name : 020AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	787887	793634	794972	792164	0.48	cps
Antimony	121-1	5011	5421	4978	5137	4.81	cps
Arsenic	75-2	8706	8519	8773	8666	1.52	cps
Barium	135-1	516247	517418	513193	515619	0.42	cps
Barium	137-1	903102	904596	907246	904981	0.23	cps
Beryllium	9-1	3561	3521	3472	3518	1.26	cps
Bismuth	209-1	13341991	13263071	13500420	13368494	0.90	cps
Bismuth	209-2	4999734	5089382	5005155	5031424	1.00	cps
Bromine	81-1	6548	6832	6438	6606	3.07	cps
Cadmium	108-1	130	113	127	123	7.15	cps
Cadmium	106-1	8223	8099	8389	8237	1.77	cps
Cadmium	111-1	6175	5967	6274	6138	2.55	cps
Calcium	43-1	165759	164600	166305	165555	0.53	cps
Calcium	44-1	2809852	2824930	2755439	2796740	1.31	cps
Carbon	12-1	3764537	3800606	3767791	3777644	0.53	cps
Carbon	12-2	24798	25242	24387	24809	1.72	cps
Chlorine	35-1	92171485	98496822	101397535	97355281	4.85	cps
Chlorine	35-2	388465	390742	394418	391208	0.77	cps
Chromium	52-2	95956	96882	97127	96655	0.64	cps
Cobalt	59-2	102291	102656	101965	102304	0.34	cps
Copper	63-2	272221	272737	273649	272869	0.27	cps
Dysprosium	156-1	39725	39946	39261	39644	0.88	cps
Dysprosium	156-2	13157	13093	13393	13214	1.20	cps
Erbium	164-1	36941	37834	37570	37448	1.22	cps
Erbium	164-2	11322	11615	11542	11493	1.33	cps
Gadolinium	160-1	41401	40421	41775	41199	1.70	cps
Gadolinium	160-2	16126	16033	15976	16045	0.47	cps
Holmium	165-1	21192754	20869823	21286825	21116467	1.04	cps
Holmium	165-2	6702922	6701684	6707447	6704018	0.05	cps
Indium	115-1	16970463	16555997	16425235	16650565	1.71	cps
Indium	115-2	1418939	1422114	1426504	1422519	0.27	cps
Iron	54-2	4758906	4681142	4714342	4718130	0.83	cps
Iron	56-2	85727040	85295132	86172644	85731605	0.51	cps
Iron	57-2	2153417	2189400	2184073	2175630	0.89	cps
Krypton	83-1	347	320	333	333	4.00	cps
Lead	206-1	191020	190852	191727	191200	0.24	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-02DLX2 Instrumnet Name : P8
Client Sample ID : YE8C9D Dilution Factor : 2
Date & Time Acquired : 2025-02-11 13:57:03 DataFile Name : 020AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	155231	155657	154311	155066	0.44	cps
Lead	208-1	732461	734613	731401	732825	0.22	cps
Lithium	6-1	8068817	7925836	7936917	7977190	1.00	cps
Magnesium	24-2	1413547	1415420	1408880	1412615	0.24	cps
Manganese	55-2	350951	349835	351926	350904	0.30	cps
Molybdenum	94-1	7143	7216	7092	7150	0.87	cps
Molybdenum	95-1	4121	4141	3921	4061	3.00	cps
Molybdenum	96-1	6168	5401	5518	5696	7.26	cps
Molybdenum	97-1	2540	2377	2567	2495	4.12	cps
Molybdenum	98-1	6465	6418	6815	6566	3.30	cps
Neodymium	150-1	42674	42975	42808	42819	0.35	cps
Neodymium	150-2	9180	9217	9046	9148	0.98	cps
Nickel	60-2	73636	74051	74185	73957	0.39	cps
Phosphorus	31-2	1017	1027	1013	1019	0.68	cps
Potassium	39-2	55365	53325	53503	54064	2.09	cps
Rhodium	103-1	15652344	15521410	15464494	15546083	0.62	cps
Rhodium	103-2	5668511	5587718	5595542	5617257	0.79	cps
Scandium	45-1	11434626	11338657	11225776	11333020	0.92	cps
Scandium	45-2	200289	196235	201218	199247	1.33	cps
Selenium	82-1	133	113	113	120	9.62	cps
Selenium	77-2	83	63	53	67	22.91	cps
Selenium	78-2	70	50	40	53	28.64	cps
Silicon	28-1	61625207	62732102	64373659	62910323	2.20	cps
Silver	107-1	3891	4077	4241	4070	4.30	cps
Silver	109-1	3307	3694	3844	3615	7.66	cps
Sodium	23-2	88727	89616	89549	89298	0.55	cps
Strontium	86-1	152357	153306	152864	152842	0.31	cps
Strontium	88-1	1323988	1328012	1332306	1328102	0.31	cps
Sulfur	34-1	685435	682378	678770	682195	0.49	cps
Terbium	159-1	21689531	21520336	21356510	21522126	0.77	cps
Terbium	159-2	6413932	6484973	6479243	6459383	0.61	cps
Thallium	203-1	2620	2814	3087	2840	8.26	cps
Thallium	205-1	6185	6535	7436	6718	9.60	cps
Tin	118-1	3847	3767	3937	3851	2.21	cps
Titanium	47-1	30264	27117	29291	28891	5.58	cps
Uranium	238-1	13738	14022	13637	13799	1.45	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-02DLX2 Instrumnet Name : P8
Client Sample ID : YE8C9D Dilution Factor : 2
Date & Time Acquired : 2025-02-11 13:57:03 DataFile Name : 020AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	90581	90282	89149	90004	0.84	cps
Ytterbium	172-1	11041	10347	10908	10766	3.42	cps
Ytterbium	172-2	3994	4127	3984	4035	1.98	cps
Ytterbium	176-1	8369	8786	8256	8471	3.30	cps
Ytterbium	176-2	2807	2760	2930	2833	3.10	cps
Yttrium	89-1	29025150	29242992	29117978	29128707	0.38	cps
Yttrium	89-2	1855512	1812381	1867467	1845120	1.57	cps
Zinc	66-2	38536	39388	39010	38978	1.10	cps
Zirconium	90-1	12179	13409	13277	12955	5.21	cps
Zirconium	91-1	14334	2964	3010	6769	96.78	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-01LDLX10 Instrumnet Name : P8
Client Sample ID : YE8C9L Dilution Factor : 10
Date & Time Acquired : 2025-02-11 14:00:19 DataFile Name : 021AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	158094	157917	156052	157354	0.72	cps
Antimony	121-1	1077	1250	1027	1118	10.49	cps
Arsenic	75-2	1737	1743	1727	1736	0.48	cps
Barium	135-1	103189	103324	103888	103467	0.36	cps
Barium	137-1	179713	180557	180221	180164	0.24	cps
Beryllium	9-1	1075	1171	1074	1107	5.05	cps
Bismuth	209-1	13344141	13335717	13371543	13350467	0.14	cps
Bismuth	209-2	5027330	4999458	4986469	5004419	0.42	cps
Bromine	81-1	5111	4868	4928	4969	2.55	cps
Cadmium	108-1	30	57	40	42	31.91	cps
Cadmium	106-1	7886	7315	7852	7684	4.17	cps
Cadmium	111-1	5676	5221	5569	5488	4.33	cps
Calcium	43-1	33542	33730	33883	33718	0.51	cps
Calcium	44-1	566526	566097	565576	566066	0.08	cps
Carbon	12-1	3401329	3372305	3410232	3394622	0.58	cps
Carbon	12-2	22294	22017	22247	22186	0.67	cps
Chlorine	35-1	23640878	24282560	24549559	24157665	1.93	cps
Chlorine	35-2	89940	90937	90430	90436	0.55	cps
Chromium	52-2	19470	19757	19971	19733	1.27	cps
Cobalt	59-2	20738	20665	20241	20548	1.31	cps
Copper	63-2	56682	56867	56244	56598	0.56	cps
Dysprosium	156-1	7832	8103	8056	7997	1.81	cps
Dysprosium	156-2	2544	2527	2537	2536	0.33	cps
Erbium	164-1	7922	7499	7646	7689	2.80	cps
Erbium	164-2	2344	2267	2350	2320	2.00	cps
Gadolinium	160-1	8333	8456	8196	8328	1.56	cps
Gadolinium	160-2	3320	2990	3154	3155	5.23	cps
Holmium	165-1	21130163	21033594	21145315	21103024	0.29	cps
Holmium	165-2	6744963	6721301	6754975	6740413	0.26	cps
Indium	115-1	16624632	16656734	16653562	16644976	0.11	cps
Indium	115-2	1420798	1433569	1432512	1428960	0.50	cps
Iron	54-2	922965	914357	927957	921760	0.75	cps
Iron	56-2	17298692	17169068	16934226	17133995	1.08	cps
Iron	57-2	425449	419573	420960	421994	0.73	cps
Krypton	83-1	347	300	313	320	7.51	cps
Lead	206-1	41296	41406	41787	41496	0.62	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-01LDLX10 Instrumnet Name : P8
Client Sample ID : YE8C9L Dilution Factor : 10
Date & Time Acquired : 2025-02-11 14:00:19 DataFile Name : 021AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	34436	33824	34593	34284	1.18	cps
Lead	208-1	161422	159033	162024	160826	0.98	cps
Lithium	6-1	8112199	7784870	7781458	7892842	2.41	cps
Magnesium	24-2	277438	280587	277735	278587	0.62	cps
Manganese	55-2	70402	72228	69880	70837	1.74	cps
Molybdenum	94-1	1743	1914	2027	1895	7.53	cps
Molybdenum	95-1	973	947	927	949	2.47	cps
Molybdenum	96-1	1393	1203	1270	1289	7.48	cps
Molybdenum	97-1	727	590	613	643	11.36	cps
Molybdenum	98-1	1640	1483	1610	1578	5.27	cps
Neodymium	150-1	8463	8123	8276	8287	2.06	cps
Neodymium	150-2	1787	1790	1823	1800	1.13	cps
Nickel	60-2	15402	15345	15475	15407	0.42	cps
Phosphorus	31-2	283	293	287	288	1.77	cps
Potassium	39-2	19213	18929	19321	19154	1.06	cps
Rhodium	103-1	15608584	15555282	15515258	15559708	0.30	cps
Rhodium	103-2	5648208	5651796	5566470	5622158	0.86	cps
Scandium	45-1	11277031	11154690	11189552	11207091	0.56	cps
Scandium	45-2	193460	194045	195111	194205	0.43	cps
Selenium	82-1	-3	0	3	0	0.00	cps
Selenium	77-2	7	7	7	7	0.00	cps
Selenium	78-2	20	7	17	14	48.02	cps
Silicon	28-1	15108019	12971658	12964946	13681541	9.03	cps
Silver	107-1	860	983	963	936	7.08	cps
Silver	109-1	850	813	883	849	4.13	cps
Sodium	23-2	68466	68333	67070	67956	1.13	cps
Strontium	86-1	30884	31415	31024	31108	0.88	cps
Strontium	88-1	266199	268715	268934	267950	0.57	cps
Sulfur	34-1	660483	656784	652411	656559	0.62	cps
Terbium	159-1	21170761	21242148	21264433	21225781	0.23	cps
Terbium	159-2	6384698	6380939	6412947	6392861	0.27	cps
Thallium	203-1	1180	1093	1073	1116	5.08	cps
Thallium	205-1	2680	2674	2440	2598	5.26	cps
Tin	118-1	2827	2620	2607	2685	4.60	cps
Titanium	47-1	4184	4935	5757	4959	15.87	cps
Uranium	238-1	2600	2660	2747	2669	2.76	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-01LDLX10 Instrumnet Name : P8
Client Sample ID : YE8C9L Dilution Factor : 10
Date & Time Acquired : 2025-02-11 14:00:19 DataFile Name : 021AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	17631	17644	18419	17898	2.52	cps
Ytterbium	172-1	2350	2194	2350	2298	3.94	cps
Ytterbium	172-2	827	923	843	864	5.98	cps
Ytterbium	176-1	3337	3337	3364	3346	0.46	cps
Ytterbium	176-2	770	820	817	802	3.48	cps
Yttrium	89-1	28809450	28723675	28621494	28718207	0.33	cps
Yttrium	89-2	1809980	1831621	1840499	1827367	0.86	cps
Zinc	66-2	8076	8242	7992	8103	1.57	cps
Zirconium	90-1	3280	3127	3357	3255	3.60	cps
Zirconium	91-1	1277	847	760	961	28.81	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-03DLX2 Instrumnet Name : P8
Client Sample ID : YE8C9S Dilution Factor : 2
Date & Time Acquired : 2025-02-11 14:03:40 DataFile Name : 022AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	797069	797291	790296	794885	0.50	cps
Antimony	121-1	183308	183810	183605	183574	0.14	cps
Arsenic	75-2	9677	9610	9510	9599	0.87	cps
Barium	135-1	1334111	1319852	1333404	1329122	0.60	cps
Barium	137-1	2405793	2347400	2413247	2388813	1.51	cps
Beryllium	9-1	31781	32166	32006	31984	0.60	cps
Bismuth	209-1	13211981	13356044	13213564	13260530	0.62	cps
Bismuth	209-2	4981856	5106176	5018677	5035570	1.27	cps
Bromine	81-1	6328	6175	6275	6259	1.24	cps
Cadmium	108-1	1657	1747	1613	1672	4.07	cps
Cadmium	106-1	9914	9710	9560	9728	1.82	cps
Cadmium	111-1	28014	27188	27453	27552	1.53	cps
Calcium	43-1	190726	190399	190862	190662	0.12	cps
Calcium	44-1	3061299	3059579	3020440	3047106	0.76	cps
Carbon	12-1	3552608	3489118	3554285	3532004	1.05	cps
Carbon	12-2	23208	22748	23101	23019	1.05	cps
Chlorine	35-1	98296909	105543818	106494978	103445235	4.33	cps
Chlorine	35-2	413648	409340	412045	411678	0.53	cps
Chromium	52-2	163735	164416	165323	164491	0.48	cps
Cobalt	59-2	434104	428666	432375	431715	0.64	cps
Copper	63-2	396899	398399	391909	395736	0.86	cps
Dysprosium	156-1	38639	38184	38943	38589	0.99	cps
Dysprosium	156-2	13013	13484	13103	13200	1.89	cps
Erbium	164-1	36884	37108	37115	37036	0.35	cps
Erbium	164-2	11258	11512	11548	11439	1.38	cps
Gadolinium	160-1	40632	40144	40585	40453	0.67	cps
Gadolinium	160-2	15916	15215	16063	15732	2.88	cps
Holmium	165-1	21287211	21277550	21040518	21201759	0.66	cps
Holmium	165-2	6716757	6807545	6765005	6763102	0.67	cps
Indium	115-1	16660537	16584356	16495135	16580009	0.50	cps
Indium	115-2	1416180	1461441	1431521	1436381	1.60	cps
Iron	54-2	4669113	4567689	4580795	4605866	1.20	cps
Iron	56-2	83977314	84314410	84127760	84139828	0.20	cps
Iron	57-2	2137003	2144325	2099925	2127084	1.12	cps
Krypton	83-1	317	417	323	352	15.87	cps
Lead	206-1	215290	217253	216902	216481	0.48	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-03DLX2 Instrumnet Name : P8
Client Sample ID : YE8C9S Dilution Factor : 2
Date & Time Acquired : 2025-02-11 14:03:40 DataFile Name : 022AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	176817	177127	176811	176918	0.10	cps
Lead	208-1	829800	835145	830012	831652	0.36	cps
Lithium	6-1	8126044	7823808	7982229	7977360	1.90	cps
Magnesium	24-2	1415658	1394885	1379623	1396722	1.30	cps
Manganese	55-2	414044	411582	413682	413103	0.32	cps
Molybdenum	94-1	266772	266921	269245	267646	0.52	cps
Molybdenum	95-1	462600	464936	465465	464334	0.33	cps
Molybdenum	96-1	492747	500256	498231	497078	0.78	cps
Molybdenum	97-1	287678	288080	289234	288331	0.28	cps
Molybdenum	98-1	746046	755861	747095	749667	0.72	cps
Neodymium	150-1	41424	41655	42153	41744	0.89	cps
Neodymium	150-2	8693	9063	8986	8914	2.19	cps
Nickel	60-2	165041	164741	164987	164923	0.10	cps
Phosphorus	31-2	1077	1143	1167	1129	4.14	cps
Potassium	39-2	53054	53038	51622	52572	1.56	cps
Rhodium	103-1	15706758	15715586	15271326	15564557	1.63	cps
Rhodium	103-2	5548197	5693487	5588961	5610215	1.34	cps
Scandium	45-1	11277452	11423889	11295612	11332318	0.70	cps
Scandium	45-2	198222	202026	198590	199613	1.05	cps
Selenium	82-1	627	567	600	598	5.03	cps
Selenium	77-2	63	67	73	68	7.51	cps
Selenium	78-2	63	53	70	62	13.48	cps
Silicon	28-1	65918487	61456946	61035739	62803724	4.31	cps
Silver	107-1	108510	109158	109419	109029	0.43	cps
Silver	109-1	103518	103119	105336	103991	1.14	cps
Sodium	23-2	84898	85029	85676	85201	0.49	cps
Strontium	86-1	5875990	5854101	5748111	5826067	1.17	cps
Strontium	88-1	50649066	50575569	50381243	50535293	0.27	cps
Sulfur	34-1	643974	647530	639202	643569	0.65	cps
Terbium	159-1	21510672	21514100	21172030	21398934	0.92	cps
Terbium	159-2	6339971	6636163	6450137	6475424	2.31	cps
Thallium	203-1	86338	88643	88247	87743	1.41	cps
Thallium	205-1	210506	213957	211916	212126	0.82	cps
Tin	118-1	3661	3377	3434	3490	4.30	cps
Titanium	47-1	21618	38599	48649	36289	37.65	cps
Uranium	238-1	13547	14045	13671	13754	1.88	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-03DLX2 Instrumnet Name : P8
Client Sample ID : YE8C9S Dilution Factor : 2
Date & Time Acquired : 2025-02-11 14:03:40 DataFile Name : 022AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	219071	218766	220072	219303	0.31	cps
Ytterbium	172-1	10741	11075	10751	10856	1.75	cps
Ytterbium	172-2	4074	4021	3934	4010	1.76	cps
Ytterbium	176-1	8640	8469	8496	8535	1.07	cps
Ytterbium	176-2	3034	2977	2830	2947	3.56	cps
Yttrium	89-1	29245657	29121253	28759335	29042082	0.87	cps
Yttrium	89-2	1797706	1886557	1838348	1840870	2.42	cps
Zinc	66-2	65862	65202	65574	65546	0.50	cps
Zirconium	90-1	13424	13010	13737	13390	2.73	cps
Zirconium	91-1	3120	3067	4670	3619	25.15	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : PB166570BL Instrumnet Name : P8
Client Sample ID : PBW570 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:15:28 DataFile Name : 023CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	90	117	83	97	18.25	cps
Antimony	121-1	180	133	137	150	17.35	cps
Arsenic	75-2	10	17	7	11	45.82	cps
Barium	135-1	217	227	177	207	12.80	cps
Barium	137-1	257	250	283	263	6.70	cps
Beryllium	9-1	390	401	380	390	2.72	cps
Bismuth	209-1	13670634	13653845	13398026	13574169	1.13	cps
Bismuth	209-2	5187091	5095157	5064013	5115420	1.25	cps
Bromine	81-1	4551	4274	4357	4394	3.23	cps
Cadmium	108-1	33	33	33	33	0.00	cps
Cadmium	106-1	8122	7632	8259	8005	4.12	cps
Cadmium	111-1	5679	5366	5785	5610	3.88	cps
Calcium	43-1	367	463	427	419	11.65	cps
Calcium	44-1	25630	26645	25937	26070	2.00	cps
Carbon	12-1	3605543	3592543	3579270	3592452	0.37	cps
Carbon	12-2	23839	23886	24090	23938	0.56	cps
Chlorine	35-1	1338505	1291060	1256272	1295279	3.19	cps
Chlorine	35-2	4357	4277	4241	4292	1.39	cps
Chromium	52-2	730	787	717	744	4.99	cps
Cobalt	59-2	170	140	170	160	10.83	cps
Copper	63-2	2744	2640	2657	2680	2.07	cps
Dysprosium	156-1	50	40	40	43	13.32	cps
Dysprosium	156-2	0	7	10	6	91.64	cps
Erbium	164-1	90	107	127	108	17.03	cps
Erbium	164-2	20	37	33	30	29.40	cps
Gadolinium	160-1	100	113	110	108	6.44	cps
Gadolinium	160-2	23	20	23	22	8.65	cps
Holmium	165-1	20787672	20804899	21211138	20934570	1.14	cps
Holmium	165-2	6718900	6698914	6811720	6743178	0.89	cps
Indium	115-1	16835605	16913182	16632480	16793756	0.86	cps
Indium	115-2	1447892	1446196	1447121	1447070	0.06	cps
Iron	54-2	513	523	503	513	1.95	cps
Iron	56-2	8062	8066	7946	8025	0.85	cps
Iron	57-2	163	257	247	222	23.06	cps
Krypton	83-1	297	307	240	281	12.79	cps
Lead	206-1	2837	2717	2700	2751	2.71	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : PB166570BL Instrumnet Name : P8
Client Sample ID : PBW570 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:15:28 DataFile Name : 023CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	2294	2274	2110	2226	4.52	cps
Lead	208-1	10685	10912	10402	10666	2.40	cps
Lithium	6-1	7863288	8008021	7889288	7920199	0.97	cps
Magnesium	24-2	2780	2687	2794	2754	2.11	cps
Manganese	55-2	160	170	173	168	4.14	cps
Molybdenum	94-1	433	457	460	450	3.23	cps
Molybdenum	95-1	187	207	167	187	10.71	cps
Molybdenum	96-1	230	313	190	244	25.74	cps
Molybdenum	97-1	130	93	103	109	17.41	cps
Molybdenum	98-1	263	223	260	249	8.92	cps
Neodymium	150-1	17	20	17	18	10.81	cps
Neodymium	150-2	0	3	7	3	100.05	cps
Nickel	60-2	873	867	863	868	0.59	cps
Phosphorus	31-2	103	67	53	74	34.78	cps
Potassium	39-2	10407	10250	10531	10396	1.35	cps
Rhodium	103-1	15457629	15488946	15541379	15495985	0.27	cps
Rhodium	103-2	5692630	5711550	5675582	5693254	0.32	cps
Scandium	45-1	10925194	11284694	11300287	11170058	1.90	cps
Scandium	45-2	194836	196880	198941	196885	1.04	cps
Selenium	82-1	7	-37	63	11	451.11	cps
Selenium	77-2	0	0	0	0	0.00	cps
Selenium	78-2	17	0	17	11	86.60	cps
Silicon	28-1	529888	532878	532114	531627	0.29	cps
Silver	107-1	223	227	293	248	15.94	cps
Silver	109-1	133	170	120	141	18.35	cps
Sodium	23-2	61556	60301	60963	60940	1.03	cps
Strontium	86-1	613	557	613	594	5.50	cps
Strontium	88-1	1180	1227	1390	1266	8.71	cps
Sulfur	34-1	689294	691328	702327	694316	1.01	cps
Terbium	159-1	21541026	21814210	21328514	21561250	1.13	cps
Terbium	159-2	6474682	6448495	6461725	6461634	0.20	cps
Thallium	203-1	780	743	707	743	4.93	cps
Thallium	205-1	1820	1787	1757	1788	1.77	cps
Tin	118-1	3130	3120	3107	3119	0.38	cps
Titanium	47-1	523	583	487	531	9.19	cps
Uranium	238-1	53	23	20	32	56.98	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : PB166570BL Instrumnet Name : P8
Client Sample ID : PBW570 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:15:28 DataFile Name : 023CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	7	17	10	11	45.82	cps
Ytterbium	172-1	127	117	123	122	4.17	cps
Ytterbium	172-2	40	47	40	42	9.12	cps
Ytterbium	176-1	1900	1920	2160	1994	7.26	cps
Ytterbium	176-2	373	260	303	312	18.32	cps
Yttrium	89-1	28024745	28083446	28461445	28189879	0.84	cps
Yttrium	89-2	1869135	1819412	1844663	1844404	1.35	cps
Zinc	66-2	277	207	200	228	18.65	cps
Zirconium	90-1	900	993	1083	992	9.24	cps
Zirconium	91-1	267	200	200	222	17.32	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : PB166570BS Instrumnet Name : P8
Client Sample ID : LCS570 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:18:49 DataFile Name : 024LCSE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	2977	2987	2840	2935	2.79	cps
Antimony	121-1	75366	75648	74682	75232	0.66	cps
Arsenic	75-2	747	730	723	733	1.64	cps
Barium	135-1	87701	88611	88033	88115	0.52	cps
Barium	137-1	150924	152317	152849	152030	0.65	cps
Beryllium	9-1	13068	12988	12857	12971	0.82	cps
Bismuth	209-1	13613427	13381374	13318186	13437662	1.16	cps
Bismuth	209-2	5172859	5024255	5066970	5088028	1.50	cps
Bromine	81-1	4427	4431	4507	4455	1.02	cps
Cadmium	108-1	933	797	850	860	8.01	cps
Cadmium	106-1	9193	9080	9583	9285	2.85	cps
Cadmium	111-1	14995	15028	15487	15170	1.81	cps
Calcium	43-1	59516	60366	59526	59802	0.82	cps
Calcium	44-1	971725	978430	977811	975989	0.38	cps
Carbon	12-1	3703673	3698098	3626819	3676197	1.17	cps
Carbon	12-2	24404	24554	24253	24404	0.62	cps
Chlorine	35-1	2471597	2683294	2719997	2624963	5.11	cps
Chlorine	35-2	10731	10360	10807	10633	2.25	cps
Chromium	52-2	15101	15475	15095	15224	1.43	cps
Cobalt	59-2	14524	14821	14411	14585	1.45	cps
Copper	63-2	27427	27417	27083	27309	0.72	cps
Dysprosium	156-1	53	60	20	44	48.22	cps
Dysprosium	156-2	13	10	3	9	57.30	cps
Erbium	164-1	127	100	127	118	13.07	cps
Erbium	164-2	27	27	33	29	13.31	cps
Gadolinium	160-1	83	97	123	101	20.15	cps
Gadolinium	160-2	17	0	13	10	88.20	cps
Holmium	165-1	21165149	21009167	20945289	21039869	0.54	cps
Holmium	165-2	6931859	6694296	6726824	6784326	1.90	cps
Indium	115-1	17125488	16701437	16724763	16850563	1.41	cps
Indium	115-2	1468491	1452864	1437365	1452907	1.07	cps
Iron	54-2	61029	61883	61582	61498	0.70	cps
Iron	56-2	1123086	1129415	1122975	1125159	0.33	cps
Iron	57-2	28629	27988	28135	28250	1.19	cps
Krypton	83-1	287	307	260	284	8.23	cps
Lead	206-1	31699	32798	32370	32289	1.72	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : PB166570BS Instrumnet Name : P8
Client Sample ID : LCS570 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:18:49 DataFile Name : 024LCSE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	27937	27736	28835	28170	2.08	cps
Lead	208-1	127395	129665	129344	128801	0.95	cps
Lithium	6-1	7987905	7860001	7945430	7931112	0.82	cps
Magnesium	24-2	249722	250900	251155	250592	0.31	cps
Manganese	55-2	3110	3117	3124	3117	0.21	cps
Molybdenum	94-1	133130	133298	133321	133250	0.08	cps
Molybdenum	95-1	194910	196405	194471	195262	0.52	cps
Molybdenum	96-1	211036	212659	210606	211434	0.51	cps
Molybdenum	97-1	120998	122092	120817	121302	0.57	cps
Molybdenum	98-1	310432	313712	312357	312167	0.53	cps
Neodymium	150-1	17	37	17	23	49.48	cps
Neodymium	150-2	0	7	0	2	173.21	cps
Nickel	60-2	4828	4724	4951	4834	2.35	cps
Phosphorus	31-2	250	280	220	250	12.00	cps
Potassium	39-2	145232	145737	146396	145788	0.40	cps
Rhodium	103-1	15702396	15603961	15735173	15680510	0.44	cps
Rhodium	103-2	5721737	5712566	5611819	5682041	1.07	cps
Scandium	45-1	11308777	11119028	11199264	11209023	0.85	cps
Scandium	45-2	203148	200705	197666	200506	1.37	cps
Selenium	82-1	2724	2840	2790	2785	2.10	cps
Selenium	77-2	43	27	30	33	26.45	cps
Selenium	78-2	167	173	150	163	7.36	cps
Silicon	28-1	858676	863063	862674	861471	0.28	cps
Silver	107-1	45505	46455	46605	46188	1.29	cps
Silver	109-1	43636	43980	43967	43861	0.44	cps
Sodium	23-2	563172	563198	570580	565650	0.75	cps
Strontium	86-1	21453	22044	21847	21781	1.38	cps
Strontium	88-1	186315	187855	187589	187253	0.44	cps
Sulfur	34-1	715949	717189	721598	718245	0.41	cps
Terbium	159-1	21719594	21247044	21540591	21502409	1.11	cps
Terbium	159-2	6613024	6455192	6464170	6510795	1.36	cps
Thallium	203-1	36508	36622	37096	36742	0.85	cps
Thallium	205-1	91164	88540	90301	90002	1.49	cps
Tin	118-1	149724	151703	152553	151327	0.96	cps
Titanium	47-1	7319	7175	7255	7250	0.99	cps
Uranium	238-1	110374	112207	111189	111257	0.83	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : PB166570BS Instrumnet Name : P8
Client Sample ID : LCS570 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:18:49 DataFile Name : 024LCSE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	29133	28909	28852	28965	0.51	cps
Ytterbium	172-1	137	133	127	132	3.85	cps
Ytterbium	172-2	57	33	60	50	29.06	cps
Ytterbium	176-1	2197	2374	2244	2271	4.03	cps
Ytterbium	176-2	380	340	380	367	6.30	cps
Yttrium	89-1	29247055	28613975	28421315	28760782	1.50	cps
Yttrium	89-2	1868754	1810392	1796589	1825245	2.10	cps
Zinc	66-2	6271	6118	5888	6092	3.17	cps
Zirconium	90-1	63483	65061	63058	63867	1.65	cps
Zirconium	91-1	14010	14261	13990	14087	1.07	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1135-09 Instrumnet Name : P8
Client Sample ID : YE8H7 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:22:18 DataFile Name : 025AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	187	177	153	172	9.93	cps
Antimony	121-1	160	167	147	158	6.45	cps
Arsenic	75-2	7	10	17	11	45.82	cps
Barium	135-1	310	397	397	368	13.61	cps
Barium	137-1	503	447	540	497	9.47	cps
Beryllium	9-1	374	365	394	377	3.90	cps
Bismuth	209-1	14183191	13980729	13698153	13954024	1.75	cps
Bismuth	209-2	5278453	5217033	5201244	5232243	0.78	cps
Bromine	81-1	4121	4154	4254	4176	1.66	cps
Cadmium	108-1	37	17	30	28	36.66	cps
Cadmium	106-1	8226	8326	8496	8349	1.64	cps
Cadmium	111-1	5776	5863	5973	5871	1.68	cps
Calcium	43-1	787	797	720	768	5.43	cps
Calcium	44-1	31381	31241	31401	31341	0.28	cps
Carbon	12-1	4231120	4687188	4994199	4637502	8.28	cps
Carbon	12-2	34654	35610	36037	35434	2.00	cps
Chlorine	35-1	870118	834926	807230	837425	3.76	cps
Chlorine	35-2	2744	2904	2734	2794	3.42	cps
Chromium	52-2	1520	1607	1537	1555	2.96	cps
Cobalt	59-2	63	103	110	92	27.37	cps
Copper	63-2	2800	2874	2774	2816	1.84	cps
Dysprosium	156-1	40	23	53	39	38.65	cps
Dysprosium	156-2	7	3	0	3	100.05	cps
Erbium	164-1	103	117	130	117	11.43	cps
Erbium	164-2	37	27	17	27	37.50	cps
Gadolinium	160-1	150	90	110	117	26.19	cps
Gadolinium	160-2	23	17	20	20	16.65	cps
Holmium	165-1	21936966	21759127	21289691	21661928	1.54	cps
Holmium	165-2	6811110	6915124	6808524	6844919	0.89	cps
Indium	115-1	17424113	17382266	17230353	17345577	0.59	cps
Indium	115-2	1496153	1485910	1463527	1481863	1.13	cps
Iron	54-2	623	640	780	681	12.63	cps
Iron	56-2	10334	10020	10240	10198	1.58	cps
Iron	57-2	260	223	197	227	14.03	cps
Krypton	83-1	267	337	273	292	13.22	cps
Lead	206-1	4571	3924	3534	4010	13.06	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1135-09 Instrumnet Name : P8
Client Sample ID : YE8H7 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:22:18 DataFile Name : 025AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	3857	3234	3000	3364	13.17	cps
Lead	208-1	17855	14963	13936	15585	13.04	cps
Lithium	6-1	8273041	7982010	8195113	8150055	1.85	cps
Magnesium	24-2	2980	3074	3330	3128	5.80	cps
Manganese	55-2	687	643	807	712	11.88	cps
Molybdenum	94-1	307	443	403	384	18.28	cps
Molybdenum	95-1	137	150	143	143	4.65	cps
Molybdenum	96-1	227	220	220	222	1.73	cps
Molybdenum	97-1	120	100	117	112	9.55	cps
Molybdenum	98-1	230	220	157	202	19.66	cps
Neodymium	150-1	17	3	43	21	96.48	cps
Neodymium	150-2	0	3	3	2	86.60	cps
Nickel	60-2	843	857	897	866	3.21	cps
Phosphorus	31-2	113	107	113	111	3.46	cps
Potassium	39-2	10260	10931	10521	10571	3.20	cps
Rhodium	103-1	16255988	16127922	15905401	16096437	1.10	cps
Rhodium	103-2	5740710	5808568	5741725	5763668	0.67	cps
Scandium	45-1	11452886	11489165	11429967	11457339	0.26	cps
Scandium	45-2	201861	206424	201915	203400	1.29	cps
Selenium	82-1	-10	-90	57	-14	-508.44	cps
Selenium	77-2	3	0	0	1	173.21	cps
Selenium	78-2	3	7	0	3	100.05	cps
Silicon	28-1	687300	696304	704874	696159	1.26	cps
Silver	107-1	417	500	447	454	9.29	cps
Silver	109-1	317	257	233	269	15.99	cps
Sodium	23-2	69786	69967	69846	69866	0.13	cps
Strontium	86-1	760	927	1023	903	14.75	cps
Strontium	88-1	4474	4297	4151	4307	3.76	cps
Sulfur	34-1	723574	723250	725958	724261	0.20	cps
Terbium	159-1	22512836	22125871	22043115	22227274	1.13	cps
Terbium	159-2	6712862	6576837	6556106	6615268	1.29	cps
Thallium	203-1	927	837	663	809	16.55	cps
Thallium	205-1	2197	1800	1590	1862	16.55	cps
Tin	118-1	3190	3277	3320	3263	2.03	cps
Titanium	47-1	340	330	290	320	8.27	cps
Uranium	238-1	27	47	63	46	40.29	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1135-09 Instrumnet Name : P8
Client Sample ID : YE8H7 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:22:18 DataFile Name : 025AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	13	20	10	14	35.26	cps
Ytterbium	172-1	130	123	123	126	3.06	cps
Ytterbium	172-2	50	37	47	44	15.61	cps
Ytterbium	176-1	2087	2134	1930	2050	5.20	cps
Ytterbium	176-2	277	320	327	308	8.82	cps
Yttrium	89-1	30242858	29585202	29054671	29627577	2.01	cps
Yttrium	89-2	1854619	1841045	1811419	1835695	1.20	cps
Zinc	66-2	247	243	253	248	2.06	cps
Zirconium	90-1	1110	1043	907	1020	10.16	cps
Zirconium	91-1	223	183	177	194	12.98	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : PB166317BL Instrumnet Name : P8
Client Sample ID : PBW317 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:25:40 DataFile Name : 026CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	83	93	113	97	15.81	cps
Antimony	121-1	130	97	117	114	14.66	cps
Arsenic	75-2	13	3	0	6	124.93	cps
Barium	135-1	157	250	203	203	22.95	cps
Barium	137-1	323	260	287	290	10.96	cps
Beryllium	9-1	354	344	357	352	2.02	cps
Bismuth	209-1	13607290	13346546	13285507	13413114	1.27	cps
Bismuth	209-2	5077988	5004051	5070557	5050866	0.81	cps
Bromine	81-1	4304	4204	4257	4255	1.18	cps
Cadmium	108-1	40	30	20	30	33.33	cps
Cadmium	106-1	8339	7592	7772	7901	4.93	cps
Cadmium	111-1	5829	5330	5450	5537	4.71	cps
Calcium	43-1	477	390	393	420	11.69	cps
Calcium	44-1	25422	25272	25356	25350	0.30	cps
Carbon	12-1	4081839	3981313	3922052	3995068	2.02	cps
Carbon	12-2	25515	25452	24951	25306	1.22	cps
Chlorine	35-1	671964	655612	643981	657186	2.14	cps
Chlorine	35-2	2224	2160	2177	2187	1.50	cps
Chromium	52-2	787	710	693	730	6.82	cps
Cobalt	59-2	160	120	137	139	14.47	cps
Copper	63-2	2624	2657	2804	2695	3.55	cps
Dysprosium	156-1	57	50	60	56	9.17	cps
Dysprosium	156-2	0	3	3	2	86.60	cps
Erbium	164-1	80	127	77	94	29.60	cps
Erbium	164-2	40	37	43	40	8.32	cps
Gadolinium	160-1	127	110	113	117	7.56	cps
Gadolinium	160-2	20	17	30	22	31.22	cps
Holmium	165-1	20849235	20700029	20724787	20758017	0.39	cps
Holmium	165-2	6630606	6721061	6661490	6671052	0.69	cps
Indium	115-1	16716741	16976759	16474929	16722810	1.50	cps
Indium	115-2	1424596	1423002	1428478	1425359	0.20	cps
Iron	54-2	470	463	487	473	2.54	cps
Iron	56-2	6998	6675	7169	6947	3.61	cps
Iron	57-2	117	173	190	160	24.03	cps
Krypton	83-1	297	307	327	310	4.93	cps
Lead	206-1	2177	2224	2014	2138	5.16	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : PB166317BL Instrumnet Name : P8
Client Sample ID : PBW317 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:25:40 DataFile Name : 026CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	1767	1773	1777	1772	0.29	cps
Lead	208-1	8098	8171	8034	8101	0.84	cps
Lithium	6-1	8023640	7851060	7912857	7929185	1.10	cps
Magnesium	24-2	2884	3190	3000	3025	5.12	cps
Manganese	55-2	143	113	110	122	15.02	cps
Molybdenum	94-1	477	430	410	439	7.80	cps
Molybdenum	95-1	177	153	193	174	11.52	cps
Molybdenum	96-1	270	213	247	243	11.70	cps
Molybdenum	97-1	117	110	140	122	12.89	cps
Molybdenum	98-1	257	297	257	270	8.55	cps
Neodymium	150-1	23	10	10	14	53.28	cps
Neodymium	150-2	3	0	3	2	86.60	cps
Nickel	60-2	790	813	733	779	5.28	cps
Phosphorus	31-2	90	80	73	81	10.34	cps
Potassium	39-2	9660	10471	10127	10086	4.03	cps
Rhodium	103-1	15528866	15403719	15276956	15403181	0.82	cps
Rhodium	103-2	5654509	5636026	5547759	5612765	1.02	cps
Scandium	45-1	10958050	11031114	10947004	10978723	0.42	cps
Scandium	45-2	195596	193825	195548	194990	0.52	cps
Selenium	82-1	-70	-10	-67	-49	-68.97	cps
Selenium	77-2	0	7	0	2	173.21	cps
Selenium	78-2	13	13	7	11	34.61	cps
Silicon	28-1	548179	548170	548837	548395	0.07	cps
Silver	107-1	283	310	247	280	11.36	cps
Silver	109-1	130	157	153	147	9.91	cps
Sodium	23-2	58778	59179	59012	58990	0.34	cps
Strontium	86-1	620	590	650	620	4.84	cps
Strontium	88-1	1403	1267	1283	1318	5.66	cps
Sulfur	34-1	734063	734969	734366	734466	0.06	cps
Terbium	159-1	21220697	21266101	21382802	21289867	0.39	cps
Terbium	159-2	6390756	6477218	6413859	6427278	0.70	cps
Thallium	203-1	603	477	550	543	11.70	cps
Thallium	205-1	1253	1080	1127	1153	7.78	cps
Tin	118-1	3154	3097	3060	3104	1.52	cps
Titanium	47-1	577	517	520	538	6.27	cps
Uranium	238-1	43	57	63	54	18.71	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : PB166317BL Instrumnet Name : P8
Client Sample ID : PBW317 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:25:40 DataFile Name : 026CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	17	23	13	18	28.64	cps
Ytterbium	172-1	110	110	107	109	1.77	cps
Ytterbium	172-2	33	37	47	39	17.85	cps
Ytterbium	176-1	1920	2007	2057	1995	3.47	cps
Ytterbium	176-2	263	337	290	297	12.51	cps
Yttrium	89-1	28525805	27944335	28395065	28288402	1.08	cps
Yttrium	89-2	1802541	1830850	1777917	1803769	1.47	cps
Zinc	66-2	267	247	197	237	15.23	cps
Zirconium	90-1	1020	1023	997	1013	1.43	cps
Zirconium	91-1	190	153	157	167	12.16	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : PB166317BS Instrumnet Name : P8
Client Sample ID : LCS317 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:29:05 DataFile Name : 027LCSE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	2990	2997	3124	3037	2.47	cps
Antimony	121-1	73164	73570	74753	73829	1.12	cps
Arsenic	75-2	703	707	623	678	6.96	cps
Barium	135-1	85568	86799	87302	86556	1.03	cps
Barium	137-1	152251	152195	151750	152065	0.18	cps
Beryllium	9-1	12958	12820	12892	12890	0.53	cps
Bismuth	209-1	13328545	13389641	13465396	13394528	0.51	cps
Bismuth	209-2	5030484	4988085	5151591	5056720	1.68	cps
Bromine	81-1	4231	4291	4121	4214	2.05	cps
Cadmium	108-1	830	883	877	863	3.37	cps
Cadmium	106-1	8930	8623	8706	8753	1.81	cps
Cadmium	111-1	14836	14788	14687	14770	0.51	cps
Calcium	43-1	59395	58585	59218	59066	0.72	cps
Calcium	44-1	960994	973542	956067	963534	0.94	cps
Carbon	12-1	4138020	4170033	4118816	4142290	0.62	cps
Carbon	12-2	27532	27332	27469	27445	0.37	cps
Chlorine	35-1	2134645	2312818	2396382	2281281	5.86	cps
Chlorine	35-2	9713	9613	9393	9573	1.71	cps
Chromium	52-2	14848	14875	14885	14869	0.13	cps
Cobalt	59-2	14838	14641	14724	14734	0.67	cps
Copper	63-2	26926	26698	26859	26828	0.44	cps
Dysprosium	156-1	77	53	43	58	29.61	cps
Dysprosium	156-2	7	7	7	7	0.00	cps
Erbium	164-1	130	80	87	99	27.45	cps
Erbium	164-2	33	30	47	37	24.06	cps
Gadolinium	160-1	123	157	97	126	23.94	cps
Gadolinium	160-2	23	7	27	19	56.72	cps
Holmium	165-1	21165296	20695327	20723832	20861485	1.26	cps
Holmium	165-2	6700262	6684173	6711955	6698797	0.21	cps
Indium	115-1	16722258	16569250	16367012	16552840	1.08	cps
Indium	115-2	1433909	1425313	1460030	1439751	1.26	cps
Iron	54-2	61010	61241	62285	61512	1.10	cps
Iron	56-2	1113026	1104063	1120262	1112450	0.73	cps
Iron	57-2	27363	27804	27500	27556	0.82	cps
Krypton	83-1	283	280	320	294	7.54	cps
Lead	206-1	31335	31388	31595	31439	0.44	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : PB166317BS Instrumnet Name : P8
Client Sample ID : LCS317 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:29:05 DataFile Name : 027LCSE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	27800	27292	27215	27436	1.16	cps
Lead	208-1	125494	125344	124861	125233	0.26	cps
Lithium	6-1	8014041	7922727	7968687	7968485	0.57	cps
Magnesium	24-2	249348	247223	248463	248345	0.43	cps
Manganese	55-2	3010	3147	3080	3079	2.22	cps
Molybdenum	94-1	73405	73067	72927	73133	0.34	cps
Molybdenum	95-1	88784	87755	90177	88905	1.37	cps
Molybdenum	96-1	100315	100171	100097	100194	0.11	cps
Molybdenum	97-1	55085	55855	55099	55346	0.80	cps
Molybdenum	98-1	142735	143718	146215	144222	1.24	cps
Neodymium	150-1	40	20	23	28	38.58	cps
Neodymium	150-2	0	10	3	4	114.60	cps
Nickel	60-2	4831	4671	4814	4772	1.84	cps
Phosphorus	31-2	213	230	193	212	8.65	cps
Potassium	39-2	146340	143916	144569	144942	0.87	cps
Rhodium	103-1	15840646	15491205	15403042	15578298	1.49	cps
Rhodium	103-2	5717431	5611434	5580524	5636463	1.27	cps
Scandium	45-1	11139448	11013459	10915987	11022965	1.02	cps
Scandium	45-2	199251	197548	200035	198945	0.64	cps
Selenium	82-1	2730	2570	2920	2740	6.40	cps
Selenium	77-2	40	27	43	37	24.04	cps
Selenium	78-2	153	163	187	168	10.19	cps
Silicon	28-1	861258	867916	861934	863703	0.42	cps
Silver	107-1	44364	46441	46154	45653	2.46	cps
Silver	109-1	42997	44060	43498	43519	1.22	cps
Sodium	23-2	562873	560366	562423	561887	0.24	cps
Strontium	86-1	11715	12582	12032	12110	3.62	cps
Strontium	88-1	101418	100471	103104	101664	1.31	cps
Sulfur	34-1	728769	730980	730542	730097	0.16	cps
Terbium	159-1	21422802	21596873	21166609	21395428	1.01	cps
Terbium	159-2	6461743	6438002	6506708	6468817	0.54	cps
Thallium	203-1	36916	36357	36702	36658	0.77	cps
Thallium	205-1	87466	87338	87794	87533	0.27	cps
Tin	118-1	148309	150870	149667	149616	0.86	cps
Titanium	47-1	7349	6998	7219	7189	2.46	cps
Uranium	238-1	109229	110345	110417	109997	0.61	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : PB166317BS Instrumnet Name : P8
Client Sample ID : LCS317 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:29:05 DataFile Name : 027LCSE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	29247	28438	28385	28690	1.68	cps
Ytterbium	172-1	137	97	150	128	21.72	cps
Ytterbium	172-2	43	27	53	41	32.76	cps
Ytterbium	176-1	2107	2107	2014	2076	2.60	cps
Ytterbium	176-2	360	360	343	354	2.72	cps
Yttrium	89-1	28717067	28257405	27930601	28301691	1.40	cps
Yttrium	89-2	1800960	1795141	1798872	1798324	0.16	cps
Zinc	66-2	5875	5908	5895	5892	0.29	cps
Zirconium	90-1	62275	62945	62656	62625	0.54	cps
Zirconium	91-1	13947	13837	13620	13801	1.21	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1186-01 Instrumnet Name : P8
Client Sample ID : ME2948 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:32:25 DataFile Name : 028AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	633	660	617	637	3.43	cps
Antimony	121-1	903	930	933	922	1.78	cps
Arsenic	75-2	127	137	97	120	17.35	cps
Barium	135-1	75804	75274	74966	75348	0.56	cps
Barium	137-1	129997	133766	133326	132363	1.56	cps
Beryllium	9-1	320	350	314	328	5.91	cps
Bismuth	209-1	12360129	12159719	12388507	12302785	1.01	cps
Bismuth	209-2	4583568	4728328	4545485	4619127	2.09	cps
Bromine	81-1	97093	109353	119480	108642	10.32	cps
Cadmium	108-1	37	27	33	32	15.80	cps
Cadmium	106-1	7836	7982	8199	8006	2.28	cps
Cadmium	111-1	5519	5602	5753	5625	2.10	cps
Calcium	43-1	19410115	19496383	19674969	19527156	0.69	cps
Calcium	44-1	311745822	314473229	315117255	313778769	0.57	cps
Carbon	12-1	6199554	7069999	7530544	6933366	9.75	cps
Carbon	12-2	55039	56003	57006	56016	1.76	cps
Chlorine	35-1	2997063	3260922	3342415	3200133	5.64	cps
Chlorine	35-2	12916	13126	12799	12947	1.28	cps
Chromium	52-2	2337	2434	2013	2261	9.73	cps
Cobalt	59-2	310	327	340	326	4.62	cps
Copper	63-2	3204	3200	3110	3172	1.67	cps
Dysprosium	156-1	310	287	260	286	8.76	cps
Dysprosium	156-2	73	110	107	97	20.98	cps
Erbium	164-1	397	347	330	358	9.70	cps
Erbium	164-2	97	143	100	113	22.97	cps
Gadolinium	160-1	340	433	310	361	17.81	cps
Gadolinium	160-2	103	83	113	100	15.28	cps
Holmium	165-1	20402756	20299464	20834414	20512211	1.38	cps
Holmium	165-2	6471565	6596433	6540870	6536289	0.96	cps
Indium	115-1	15717160	16056269	15887126	15886852	1.07	cps
Indium	115-2	1341717	1366135	1338539	1348797	1.12	cps
Iron	54-2	112323	111709	112559	112197	0.39	cps
Iron	56-2	2138185	2137755	2093256	2123066	1.22	cps
Iron	57-2	51379	51098	51994	51490	0.89	cps
Krypton	83-1	280	260	233	258	9.08	cps
Lead	206-1	3084	3084	2950	3039	2.53	cps

LB Number :	LB134674	Operator :	Jaswal
Lab Sample ID :	Q1186-01	Instrumnet Name :	P8
Client Sample ID :	ME2948	Dilution Factor :	1
Date & Time Acquired :	2025-02-11 14:32:25	DataFile Name :	028AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	2604	2644	2530	2593	2.22	cps
Lead	208-1	12252	12129	11932	12105	1.33	cps
Lithium	6-1	7521436	7514477	7612071	7549328	0.72	cps
Magnesium	24-2	27104995	27107735	27218039	27143589	0.24	cps
Manganese	55-2	40190	39602	40577	40123	1.22	cps
Molybdenum	94-1	3080	2997	2934	3004	2.45	cps
Molybdenum	95-1	3107	3397	3317	3274	4.58	cps
Molybdenum	96-1	3594	3560	3620	3592	0.84	cps
Molybdenum	97-1	2097	1977	2170	2081	4.69	cps
Molybdenum	98-1	4958	5274	5261	5164	3.47	cps
Neodymium	150-1	297	350	340	329	8.62	cps
Neodymium	150-2	90	70	57	72	23.23	cps
Nickel	60-2	1207	1157	1350	1238	8.11	cps
Phosphorus	31-2	117	103	73	98	22.70	cps
Potassium	39-2	465288	462418	460109	462605	0.56	cps
Rhodium	103-1	13979210	14231310	14297942	14169488	1.19	cps
Rhodium	103-2	5113480	5203076	5125784	5147447	0.94	cps
Scandium	45-1	10645510	10490357	10699503	10611790	1.02	cps
Scandium	45-2	192303	193427	189742	191824	0.98	cps
Selenium	82-1	83	153	250	162	51.59	cps
Selenium	77-2	0	3	0	1	173.21	cps
Selenium	78-2	7	3	20	10	88.20	cps
Silicon	28-1	71972664	72404511	72013361	72130178	0.33	cps
Silver	107-1	643	597	647	629	4.45	cps
Silver	109-1	340	353	287	327	10.80	cps
Sodium	23-2	14386268	14553925	14495396	14478530	0.59	cps
Strontium	86-1	74586052	75384509	75645964	75205508	0.73	cps
Strontium	88-1	655917177	661681910	663061857	660220315	0.57	cps
Sulfur	34-1	29137359	29050366	29303048	29163591	0.44	cps
Terbium	159-1	21058938	20520894	21024976	20868270	1.44	cps
Terbium	159-2	6212668	6300243	6226391	6246434	0.75	cps
Thallium	203-1	533	530	480	514	5.81	cps
Thallium	205-1	1360	1177	1233	1257	7.47	cps
Tin	118-1	16113	15549	15639	15767	1.92	cps
Titanium	47-1	1750	1890	1977	1872	6.11	cps
Uranium	238-1	11555	11188	11228	11324	1.78	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1186-01 Instrumnet Name : P8
Client Sample ID : ME2948 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:32:25 DataFile Name : 028AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	260	317	310	296	10.48	cps
Ytterbium	172-1	210	200	273	228	17.46	cps
Ytterbium	172-2	117	67	97	93	26.96	cps
Ytterbium	176-1	1970	1960	1997	1976	0.96	cps
Ytterbium	176-2	267	327	283	292	10.60	cps
Yttrium	89-1	26897138	27614868	27359726	27290577	1.33	cps
Yttrium	89-2	1759827	1741395	1721147	1740789	1.11	cps
Zinc	66-2	1223	1187	1070	1160	6.90	cps
Zirconium	90-1	3370	3314	3157	3280	3.37	cps
Zirconium	91-1	740	713	710	721	2.28	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-01DL2X4 Instrumnet Name : P8
Client Sample ID : YE8C9 Dilution Factor : 4
Date & Time Acquired : 2025-02-11 14:35:40 DataFile Name : 029AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	395065	398170	400313	397849	0.66	cps
Antimony	121-1	2347	2590	2310	2416	6.30	cps
Arsenic	75-2	4221	4114	4261	4198	1.81	cps
Barium	135-1	263062	263693	262978	263244	0.15	cps
Barium	137-1	454804	458529	455521	456285	0.43	cps
Beryllium	9-1	1957	1850	1796	1868	4.40	cps
Bismuth	209-1	13591849	13756356	13351588	13566598	1.50	cps
Bismuth	209-2	4808751	4506674	4991613	4769013	5.14	cps
Bromine	81-1	24461	21984	19864	22103	10.41	cps
Cadmium	108-1	67	83	57	69	19.55	cps
Cadmium	106-1	8446	8463	8273	8394	1.25	cps
Cadmium	111-1	6106	6049	6024	6060	0.69	cps
Calcium	43-1	85405	85140	84831	85125	0.34	cps
Calcium	44-1	1463817	1430183	1428002	1440667	1.39	cps
Carbon	12-1	4138647	3945005	3858322	3980658	3.61	cps
Carbon	12-2	24958	24310	24641	24636	1.31	cps
Chlorine	35-1	43411424	47142591	48764209	46439408	5.91	cps
Chlorine	35-2	188403	192053	191233	190563	1.00	cps
Chromium	52-2	49020	48990	49683	49231	0.79	cps
Cobalt	59-2	50533	50804	51135	50824	0.59	cps
Copper	63-2	135819	137888	136441	136716	0.78	cps
Dysprosium	156-1	19965	20496	20235	20232	1.31	cps
Dysprosium	156-2	6665	6895	6712	6757	1.80	cps
Erbium	164-1	19154	19000	19010	19055	0.45	cps
Erbium	164-2	5921	5905	5711	5846	2.00	cps
Gadolinium	160-1	20753	20886	21404	21014	1.64	cps
Gadolinium	160-2	8083	8056	7982	8040	0.64	cps
Holmium	165-1	21647737	21639675	20968669	21418694	1.82	cps
Holmium	165-2	6596876	6022539	6760849	6460088	6.00	cps
Indium	115-1	16910936	16796880	16312313	16673376	1.91	cps
Indium	115-2	1371047	1256496	1425605	1351049	6.39	cps
Iron	54-2	2347148	2367068	2359350	2357855	0.43	cps
Iron	56-2	42723857	43277743	42406692	42802764	1.03	cps
Iron	57-2	1052842	1063091	1064368	1060100	0.60	cps
Krypton	83-1	373	303	303	327	12.37	cps
Lead	206-1	102138	103271	102945	102785	0.57	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-01DL2X4 Instrumnet Name : P8
Client Sample ID : YE8C9 Dilution Factor : 4
Date & Time Acquired : 2025-02-11 14:35:40 DataFile Name : 029AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	83577	83973	82806	83452	0.71	cps
Lead	208-1	391785	393009	392610	392468	0.16	cps
Lithium	6-1	8020270	7999226	7939024	7986173	0.53	cps
Magnesium	24-2	692569	705208	699792	699190	0.91	cps
Manganese	55-2	176496	176864	175741	176367	0.32	cps
Molybdenum	94-1	3681	5527	12868	7359	66.04	cps
Molybdenum	95-1	1967	2127	1970	2021	4.52	cps
Molybdenum	96-1	2907	2750	2854	2837	2.81	cps
Molybdenum	97-1	1307	1310	1390	1336	3.53	cps
Molybdenum	98-1	3394	3204	3180	3259	3.59	cps
Neodymium	150-1	21674	20876	21260	21270	1.88	cps
Neodymium	150-2	4444	4601	4411	4485	2.26	cps
Nickel	60-2	37285	37322	37423	37343	0.19	cps
Phosphorus	31-2	667	567	573	602	9.28	cps
Potassium	39-2	31866	32356	35998	33407	6.76	cps
Rhodium	103-1	15602222	15689223	15218615	15503353	1.62	cps
Rhodium	103-2	5408489	4920301	5581312	5303367	6.46	cps
Scandium	45-1	11243628	11158606	10809521	11070585	2.08	cps
Scandium	45-2	189677	173770	195272	186240	5.99	cps
Selenium	82-1	-40	-20	17	-14	-198.99	cps
Selenium	77-2	37	13	27	26	45.82	cps
Selenium	78-2	30	10	20	20	50.00	cps
Silicon	28-1	31199160	30942935	31770877	31304324	1.35	cps
Silver	107-1	2444	2220	2270	2311	5.07	cps
Silver	109-1	2057	2027	2090	2058	1.54	cps
Sodium	23-2	75286	76204	76127	75872	0.67	cps
Strontium	86-1	80627	79732	79314	79891	0.84	cps
Strontium	88-1	696377	696572	697194	696714	0.06	cps
Sulfur	34-1	776270	756753	749033	760686	1.85	cps
Terbium	159-1	21873285	21788643	21273626	21645185	1.50	cps
Terbium	159-2	6287415	5763645	6470174	6173744	5.94	cps
Thallium	203-1	1443	1283	1317	1348	6.26	cps
Thallium	205-1	3020	3254	2964	3079	4.99	cps
Tin	118-1	3480	3107	3057	3215	7.20	cps
Titanium	47-1	12329	16533	16015	14959	15.33	cps
Uranium	238-1	6892	7059	6855	6935	1.56	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-01DL2X4 Instrumnet Name : P8
Client Sample ID : YE8C9 Dilution Factor : 4
Date & Time Acquired : 2025-02-11 14:35:40 DataFile Name : 029AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	44529	45415	45412	45119	1.13	cps
Ytterbium	172-1	5855	5368	5715	5646	4.44	cps
Ytterbium	172-2	1980	2000	1997	1992	0.54	cps
Ytterbium	176-1	5418	5618	5274	5437	3.17	cps
Ytterbium	176-2	1563	1523	1567	1551	1.55	cps
Yttrium	89-1	28890959	28890585	27947100	28576214	1.91	cps
Yttrium	89-2	1760044	1580711	1812029	1717594	7.07	cps
Zinc	66-2	19667	19978	18989	19545	2.59	cps
Zirconium	90-1	7909	7052	8273	7745	8.10	cps
Zirconium	91-1	1477	1647	1714	1612	7.57	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-02DL2X4 Instrumnet Name : P8
Client Sample ID : YE8C9D Dilution Factor : 4
Date & Time Acquired : 2025-02-11 14:38:57 DataFile Name : 030AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	394539	399122	400812	398158	0.82	cps
Antimony	121-1	2364	2347	2337	2349	0.57	cps
Arsenic	75-2	4127	4264	3967	4120	3.61	cps
Barium	135-1	261469	260329	260701	260833	0.22	cps
Barium	137-1	451253	452919	453575	452582	0.26	cps
Beryllium	9-1	1910	1785	1816	1837	3.54	cps
Bismuth	209-1	13976918	13761008	13669871	13802599	1.14	cps
Bismuth	209-2	5035629	4999301	4814980	4949970	2.39	cps
Bromine	81-1	10190	9917	9630	9912	2.83	cps
Cadmium	108-1	73	53	67	64	15.80	cps
Cadmium	106-1	8406	8469	8663	8513	1.57	cps
Cadmium	111-1	6093	6121	6307	6174	1.89	cps
Calcium	43-1	83976	84640	83675	84097	0.59	cps
Calcium	44-1	1424936	1466966	1411174	1434359	2.03	cps
Carbon	12-1	3623350	3648599	3554004	3608651	1.36	cps
Carbon	12-2	23439	23636	23672	23582	0.53	cps
Chlorine	35-1	45358834	48994671	49941608	48098371	5.03	cps
Chlorine	35-2	194606	194228	196559	195131	0.64	cps
Chromium	52-2	48602	48930	48452	48661	0.50	cps
Cobalt	59-2	51048	51429	51322	51266	0.38	cps
Copper	63-2	137602	137942	137946	137830	0.14	cps
Dysprosium	156-1	20025	19641	19918	19862	1.00	cps
Dysprosium	156-2	6622	6622	6625	6623	0.03	cps
Erbium	164-1	19090	18700	18757	18849	1.12	cps
Erbium	164-2	6035	5688	5668	5797	3.56	cps
Gadolinium	160-1	20526	20756	20669	20651	0.56	cps
Gadolinium	160-2	7492	7922	7772	7729	2.83	cps
Holmium	165-1	21964920	21607424	21604339	21725561	0.95	cps
Holmium	165-2	6807637	6715741	6489741	6671040	2.45	cps
Indium	115-1	17547066	17010786	16858082	17138645	2.11	cps
Indium	115-2	1436699	1428601	1380214	1415171	2.16	cps
Iron	54-2	2402154	2371806	2353217	2375726	1.04	cps
Iron	56-2	42985084	42512439	42661534	42719686	0.57	cps
Iron	57-2	1044522	1056428	1060472	1053807	0.79	cps
Krypton	83-1	253	350	290	298	16.39	cps
Lead	206-1	97873	96943	98743	97853	0.92	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-02DL2X4 Instrumnet Name : P8
Client Sample ID : YE8C9D Dilution Factor : 4
Date & Time Acquired : 2025-02-11 14:38:57 DataFile Name : 030AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	79230	79327	79579	79379	0.23	cps
Lead	208-1	372206	371999	375374	373193	0.51	cps
Lithium	6-1	7998777	7854681	7918074	7923844	0.91	cps
Magnesium	24-2	692091	694592	699443	695375	0.54	cps
Manganese	55-2	175459	176210	175477	175715	0.24	cps
Molybdenum	94-1	3494	3577	3494	3522	1.37	cps
Molybdenum	95-1	1953	1963	1930	1949	0.88	cps
Molybdenum	96-1	4233	2650	2674	3186	28.47	cps
Molybdenum	97-1	1217	1170	1277	1221	4.38	cps
Molybdenum	98-1	2944	3284	2984	3070	6.05	cps
Neodymium	150-1	21137	21077	21210	21141	0.32	cps
Neodymium	150-2	4601	4381	4574	4519	2.66	cps
Nickel	60-2	37175	36684	37971	37277	1.74	cps
Phosphorus	31-2	527	633	627	596	10.03	cps
Potassium	39-2	31264	31494	32050	31603	1.28	cps
Rhodium	103-1	15952254	15704533	15706497	15787762	0.90	cps
Rhodium	103-2	5750914	5692490	5401751	5615052	3.33	cps
Scandium	45-1	11468688	11401294	11414894	11428292	0.31	cps
Scandium	45-2	200489	197519	188192	195400	3.28	cps
Selenium	82-1	153	0	140	98	86.87	cps
Selenium	77-2	37	20	17	24	43.83	cps
Selenium	78-2	33	23	43	33	30.00	cps
Silicon	28-1	33876204	31873827	31496015	32415348	3.95	cps
Silver	107-1	1863	1917	2150	1977	7.71	cps
Silver	109-1	1750	1743	1727	1740	0.69	cps
Sodium	23-2	73926	73611	74254	73930	0.44	cps
Strontium	86-1	76280	77791	77835	77302	1.15	cps
Strontium	88-1	673699	683407	675406	677504	0.76	cps
Sulfur	34-1	745852	744975	739757	743528	0.44	cps
Terbium	159-1	22477586	21752451	21711200	21980412	1.96	cps
Terbium	159-2	6504189	6473149	6241538	6406292	2.24	cps
Thallium	203-1	1113	1147	1240	1167	5.63	cps
Thallium	205-1	2660	2687	2884	2744	4.45	cps
Tin	118-1	3170	2974	2937	3027	4.15	cps
Titanium	47-1	10564	15330	11098	12331	21.17	cps
Uranium	238-1	6455	6945	18343	10581	63.57	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-02DL2X4 Instrumnet Name : P8
Client Sample ID : YE8C9D Dilution Factor : 4
Date & Time Acquired : 2025-02-11 14:38:57 DataFile Name : 030AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	45238	45332	45820	45463	0.69	cps
Ytterbium	172-1	5445	5495	5528	5489	0.76	cps
Ytterbium	172-2	2077	2180	1950	2069	5.57	cps
Ytterbium	176-1	5338	5431	5274	5348	1.47	cps
Ytterbium	176-2	1537	1490	1477	1501	2.10	cps
Yttrium	89-1	29887880	29202368	29353155	29481134	1.22	cps
Yttrium	89-2	1823742	1816380	1739654	1793258	2.60	cps
Zinc	66-2	19751	19540	19564	19618	0.59	cps
Zirconium	90-1	17892	6815	7260	10656	58.85	cps
Zirconium	91-1	1707	1507	1393	1536	10.33	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-01LDL2X20 Instrumnet Name : P8
Client Sample ID : YE8C9L Dilution Factor : 20
Date & Time Acquired : 2025-02-11 14:42:16 DataFile Name : 031AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	79372	79797	79312	79494	0.33	cps
Antimony	121-1	560	530	570	553	3.76	cps
Arsenic	75-2	870	900	910	893	2.33	cps
Barium	135-1	51307	52612	51932	51950	1.26	cps
Barium	137-1	91732	89977	89151	90287	1.46	cps
Beryllium	9-1	669	696	627	664	5.21	cps
Bismuth	209-1	13593550	13690941	13583113	13622535	0.44	cps
Bismuth	209-2	5115381	5093541	5117037	5108653	0.26	cps
Bromine	81-1	6665	7095	6472	6744	4.73	cps
Cadmium	108-1	30	33	37	33	10.01	cps
Cadmium	106-1	8309	8042	8219	8190	1.66	cps
Cadmium	111-1	5855	5673	5764	5764	1.58	cps
Calcium	43-1	17381	17327	17568	17425	0.72	cps
Calcium	44-1	299052	301504	300249	300268	0.41	cps
Carbon	12-1	3287613	3269004	3232258	3262959	0.86	cps
Carbon	12-2	20928	20721	21332	20994	1.48	cps
Chlorine	35-1	11481970	12038564	12089150	11869895	2.84	cps
Chlorine	35-2	46060	44920	45063	45348	1.37	cps
Chromium	52-2	10421	10544	10624	10530	0.97	cps
Cobalt	59-2	10167	10004	10327	10166	1.59	cps
Copper	63-2	29758	29397	30019	29725	1.05	cps
Dysprosium	156-1	3944	4084	3884	3971	2.59	cps
Dysprosium	156-2	1390	1250	1307	1316	5.35	cps
Erbium	164-1	3687	3834	3671	3731	2.41	cps
Erbium	164-2	1130	1053	1197	1127	6.37	cps
Gadolinium	160-1	4197	4291	4344	4277	1.74	cps
Gadolinium	160-2	1690	1693	1640	1675	1.78	cps
Holmium	165-1	21417206	21035082	21238486	21230258	0.90	cps
Holmium	165-2	6768829	6681310	6783680	6744606	0.82	cps
Indium	115-1	17108891	16553209	17235246	16965782	2.14	cps
Indium	115-2	1434679	1433619	1432150	1433483	0.09	cps
Iron	54-2	467067	461735	467522	465441	0.69	cps
Iron	56-2	8680549	8612885	8612136	8635190	0.45	cps
Iron	57-2	212648	213565	210980	212398	0.62	cps
Krypton	83-1	367	290	287	314	14.39	cps
Lead	206-1	20864	21414	21040	21106	1.33	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-01LDL2X20 Instrumnet Name : P8
Client Sample ID : YE8C9L Dilution Factor : 20
Date & Time Acquired : 2025-02-11 14:42:16 DataFile Name : 031AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	17285	17222	17238	17248	0.19	cps
Lead	208-1	81257	81660	81674	81531	0.29	cps
Lithium	6-1	7911939	7881317	7897384	7896880	0.19	cps
Magnesium	24-2	142192	141970	141337	141833	0.31	cps
Manganese	55-2	34993	35514	35491	35333	0.83	cps
Molybdenum	94-1	1230	987	907	1041	16.18	cps
Molybdenum	95-1	560	530	540	543	2.81	cps
Molybdenum	96-1	753	770	750	758	1.41	cps
Molybdenum	97-1	283	297	323	301	6.76	cps
Molybdenum	98-1	780	840	763	794	5.08	cps
Neodymium	150-1	4254	4431	4301	4329	2.12	cps
Neodymium	150-2	990	930	970	963	3.17	cps
Nickel	60-2	8346	8119	8189	8218	1.41	cps
Phosphorus	31-2	223	160	167	183	18.98	cps
Potassium	39-2	14504	14751	14694	14650	0.88	cps
Rhodium	103-1	15639084	15929314	15735520	15767973	0.94	cps
Rhodium	103-2	5670506	5713143	5587403	5657017	1.13	cps
Scandium	45-1	11233172	11072895	11252287	11186118	0.88	cps
Scandium	45-2	196443	200153	196953	197850	1.02	cps
Selenium	82-1	-63	17	-23	-23	-171.40	cps
Selenium	77-2	7	7	0	4	86.60	cps
Selenium	78-2	13	3	7	8	65.47	cps
Silicon	28-1	6725510	6715857	6789004	6743457	0.59	cps
Silver	107-1	617	577	503	566	10.16	cps
Silver	109-1	370	380	427	392	7.71	cps
Sodium	23-2	64371	64177	63635	64061	0.60	cps
Strontium	86-1	15892	15796	16186	15958	1.27	cps
Strontium	88-1	138943	138058	136022	137674	1.09	cps
Sulfur	34-1	746827	741995	739402	742741	0.51	cps
Terbium	159-1	21443601	21492268	21838417	21591429	1.00	cps
Terbium	159-2	6421779	6436396	6451816	6436663	0.23	cps
Thallium	203-1	640	533	557	577	9.72	cps
Thallium	205-1	1467	1347	1470	1428	4.92	cps
Tin	118-1	2387	2537	2520	2481	3.31	cps
Titanium	47-1	2467	4286	2717	3157	31.24	cps
Uranium	238-1	1437	1333	1443	1405	4.39	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-01LDL2X20 Instrumnet Name : P8
Client Sample ID : YE8C9L Dilution Factor : 20
Date & Time Acquired : 2025-02-11 14:42:16 DataFile Name : 031AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	9043	9043	8719	8935	2.09	cps
Ytterbium	172-1	1327	1270	1187	1261	5.58	cps
Ytterbium	172-2	483	453	430	456	5.87	cps
Ytterbium	176-1	2827	2594	2620	2680	4.77	cps
Ytterbium	176-2	550	647	547	581	9.77	cps
Yttrium	89-1	28941980	29236530	28515229	28897913	1.25	cps
Yttrium	89-2	1836518	1805555	1814570	1818881	0.88	cps
Zinc	66-2	4134	4307	4351	4264	2.69	cps
Zirconium	90-1	2017	2187	2397	2200	8.66	cps
Zirconium	91-1	500	400	527	476	14.04	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-03DL2X4 Instrumnet Name : P8
Client Sample ID : YE8C9S Dilution Factor : 4
Date & Time Acquired : 2025-02-11 14:45:35 DataFile Name : 032AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	399248	399250	396295	398264	0.43	cps
Antimony	121-1	92121	93198	93175	92831	0.66	cps
Arsenic	75-2	5088	4928	5171	5062	2.44	cps
Barium	135-1	659197	676656	676357	670737	1.49	cps
Barium	137-1	1174499	1174928	1157640	1169022	0.84	cps
Beryllium	9-1	16233	16352	16084	16223	0.83	cps
Bismuth	209-1	13477260	13373927	13372090	13407759	0.45	cps
Bismuth	209-2	5096951	5003598	5068133	5056227	0.95	cps
Bromine	81-1	6555	6855	7045	6818	3.62	cps
Cadmium	108-1	903	987	893	928	5.52	cps
Cadmium	106-1	9507	9423	9403	9444	0.58	cps
Cadmium	111-1	17424	17270	16768	17154	2.00	cps
Calcium	43-1	95642	97037	96659	96446	0.75	cps
Calcium	44-1	1535606	1590562	1540406	1555525	1.96	cps
Carbon	12-1	3540591	3481082	3491315	3504329	0.91	cps
Carbon	12-2	23112	23051	22975	23046	0.30	cps
Chlorine	35-1	44460349	49241239	50759261	48153617	6.83	cps
Chlorine	35-2	196621	198561	198729	197970	0.59	cps
Chromium	52-2	83026	83888	83120	83344	0.57	cps
Cobalt	59-2	217191	217202	217936	217443	0.20	cps
Copper	63-2	200739	200907	200566	200737	0.08	cps
Dysprosium	156-1	19247	19130	19688	19355	1.52	cps
Dysprosium	156-2	6578	6588	6482	6549	0.90	cps
Erbium	164-1	18590	18894	18506	18663	1.09	cps
Erbium	164-2	5525	5454	5765	5581	2.91	cps
Gadolinium	160-1	20139	20252	21170	20520	2.76	cps
Gadolinium	160-2	7956	7846	7609	7803	2.27	cps
Holmium	165-1	21701512	20860310	21360731	21307518	1.99	cps
Holmium	165-2	6756859	6825851	6768275	6783661	0.55	cps
Indium	115-1	16735828	17117051	16559487	16804122	1.70	cps
Indium	115-2	1429242	1413529	1430059	1424277	0.65	cps
Iron	54-2	2381679	2362409	2347861	2363983	0.72	cps
Iron	56-2	42338631	43151463	42247184	42579092	1.17	cps
Iron	57-2	1040689	1043374	1039757	1041273	0.18	cps
Krypton	83-1	347	330	297	324	7.85	cps
Lead	206-1	110022	110960	109188	110057	0.81	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-03DL2X4 Instrumnet Name : P8
Client Sample ID : YE8C9S Dilution Factor : 4
Date & Time Acquired : 2025-02-11 14:45:35 DataFile Name : 032AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	89644	89000	89278	89307	0.36	cps
Lead	208-1	420428	421848	420576	420951	0.19	cps
Lithium	6-1	7922571	8074597	7956145	7984438	1.00	cps
Magnesium	24-2	688272	692805	689107	690061	0.35	cps
Manganese	55-2	206758	209365	206042	207388	0.84	cps
Molybdenum	94-1	135428	135132	137084	135881	0.77	cps
Molybdenum	95-1	232443	232211	236468	233707	1.02	cps
Molybdenum	96-1	250356	254301	250412	251690	0.90	cps
Molybdenum	97-1	146739	145739	146571	146350	0.37	cps
Molybdenum	98-1	376165	377509	379402	377692	0.43	cps
Neodymium	150-1	20796	20703	20890	20796	0.45	cps
Neodymium	150-2	4711	4361	4611	4561	3.95	cps
Nickel	60-2	83915	83784	84529	84076	0.47	cps
Phosphorus	31-2	583	590	530	568	5.79	cps
Potassium	39-2	31194	31321	31070	31195	0.40	cps
Rhodium	103-1	15686615	15630359	15672034	15663003	0.19	cps
Rhodium	103-2	5594299	5609929	5605256	5603161	0.14	cps
Scandium	45-1	11281676	11276946	11266322	11274981	0.07	cps
Scandium	45-2	200150	197529	197000	198227	0.85	cps
Selenium	82-1	227	293	250	257	13.18	cps
Selenium	77-2	33	33	30	32	5.97	cps
Selenium	78-2	63	20	40	41	52.75	cps
Silicon	28-1	31644432	31431149	31150558	31408713	0.79	cps
Silver	107-1	54423	54939	55588	54983	1.06	cps
Silver	109-1	51429	52232	52643	52101	1.19	cps
Sodium	23-2	72385	72522	72060	72322	0.33	cps
Strontium	86-1	2974241	2975719	2979157	2976373	0.08	cps
Strontium	88-1	25694321	25378242	25536191	25536252	0.62	cps
Sulfur	34-1	743850	737159	729548	736852	0.97	cps
Terbium	159-1	21567101	21622441	21530871	21573471	0.21	cps
Terbium	159-2	6449663	6378566	6517109	6448446	1.07	cps
Thallium	203-1	44031	44138	43523	43897	0.75	cps
Thallium	205-1	106039	106265	106462	106255	0.20	cps
Tin	118-1	3044	3007	2864	2971	3.20	cps
Titanium	47-1	13076	11844	11086	12002	8.37	cps
Uranium	238-1	6745	6635	6762	6714	1.03	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : Q1159-03DL2X4 Instrumnet Name : P8
Client Sample ID : YE8C9S Dilution Factor : 4
Date & Time Acquired : 2025-02-11 14:45:35 DataFile Name : 032AREF.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	110324	110925	111483	110911	0.52	cps
Ytterbium	172-1	5224	5354	5471	5350	2.31	cps
Ytterbium	172-2	2077	1927	2067	2024	4.15	cps
Ytterbium	176-1	5351	5224	5411	5329	1.79	cps
Ytterbium	176-2	1523	1577	1647	1582	3.91	cps
Yttrium	89-1	29683754	29052123	29216939	29317605	1.12	cps
Yttrium	89-2	1874750	1807609	1803031	1828463	2.20	cps
Zinc	66-2	33105	33075	32962	33047	0.23	cps
Zirconium	90-1	9641	7263	6405	7769	21.58	cps
Zirconium	91-1	2452	1657	1543	1884	26.28	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : CCV022 Instrumnet Name : P8
Client Sample ID : CCV022 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:50:23 DataFile Name : 033CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	3166656	3219302	3114011	3166656	1.66	cps
Antimony	121-1	8221670	8147852	8174881	8181468	0.46	cps
Arsenic	75-2	145397	144569	145717	145228	0.41	cps
Barium	135-1	9937428	9940063	9997160	9958217	0.34	cps
Barium	137-1	17203885	17037235	17326678	17189266	0.85	cps
Beryllium	9-1	2654376	2638062	2629882	2640773	0.47	cps
Bismuth	209-1	11681266	11866256	11665631	11737718	0.95	cps
Bismuth	209-2	4363116	4439537	4297792	4366815	1.62	cps
Bromine	81-1	4537	4661	4708	4635	1.90	cps
Cadmium	108-1	159518	159413	159338	159423	0.06	cps
Cadmium	106-1	229326	230084	231381	230264	0.45	cps
Cadmium	111-1	1998080	1980420	1949594	1976031	1.24	cps
Calcium	43-1	13159966	13127658	13087303	13124976	0.28	cps
Calcium	44-1	213926557	212261650	213758237	213315481	0.43	cps
Carbon	12-1	4181315	4287135	4447376	4305276	3.11	cps
Carbon	12-2	35727	36181	35927	35945	0.63	cps
Chlorine	35-1	2206331	2077000	2001406	2094912	4.95	cps
Chlorine	35-2	6278	6281	6098	6219	1.69	cps
Chromium	52-2	1653403	1686750	1646399	1662184	1.30	cps
Cobalt	59-2	3036918	3110755	3045672	3064448	1.32	cps
Copper	63-2	22715015	22991670	23074881	22927189	0.82	cps
Dysprosium	156-1	483	447	587	506	14.36	cps
Dysprosium	156-2	73	103	113	97	21.54	cps
Erbium	164-1	437	477	523	479	9.06	cps
Erbium	164-2	163	140	113	139	18.01	cps
Gadolinium	160-1	373	367	467	402	13.90	cps
Gadolinium	160-2	103	97	133	111	17.58	cps
Holmium	165-1	20402021	20415796	19953943	20257253	1.30	cps
Holmium	165-2	6456381	6399193	6303048	6386208	1.21	cps
Indium	115-1	15336206	14961998	14961161	15086455	1.43	cps
Indium	115-2	1272665	1272647	1264877	1270063	0.35	cps
Iron	54-2	16711327	16511395	16316438	16513053	1.20	cps
Iron	56-2	300885262	298991562	301480322	300452382	0.43	cps
Iron	57-2	7455539	7512635	7623014	7530396	1.13	cps
Krypton	83-1	390	310	330	343	12.13	cps
Lead	206-1	32241649	32935878	32613916	32597148	1.07	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : CCV022 Instrumnet Name : P8
Client Sample ID : CCV022 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:50:23 DataFile Name : 033CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	27558165	28544111	28066145	28056141	1.76	cps
Lead	208-1	127315260	130303960	130067121	129228780	1.29	cps
Lithium	6-1	7198645	7053892	6949829	7067455	1.77	cps
Magnesium	24-2	57109904	57753171	56791991	57218355	0.86	cps
Manganese	55-2	6525140	6456763	6498567	6493490	0.53	cps
Molybdenum	94-1	28787649	28763611	28791932	28781064	0.05	cps
Molybdenum	95-1	41298139	42046620	42165827	41836862	1.12	cps
Molybdenum	96-1	46244529	45528104	45893288	45888640	0.78	cps
Molybdenum	97-1	26031702	25624298	26106169	25920723	1.00	cps
Molybdenum	98-1	67525102	67129362	67958714	67537726	0.61	cps
Neodymium	150-1	767	727	793	762	4.40	cps
Neodymium	150-2	57	77	50	61	22.71	cps
Nickel	60-2	795131	798610	800029	797923	0.32	cps
Phosphorus	31-2	29627	29614	28895	29379	1.42	cps
Potassium	39-2	15470017	15285944	15112951	15289637	1.17	cps
Rhodium	103-1	13755072	13393762	13307405	13485413	1.76	cps
Rhodium	103-2	4992220	4898282	4830914	4907139	1.65	cps
Scandium	45-1	10743417	10510735	10438920	10564357	1.51	cps
Scandium	45-2	188139	186078	184643	186286	0.94	cps
Selenium	82-1	119786	117190	119178	118718	1.14	cps
Selenium	77-2	1777	1817	1827	1807	1.46	cps
Selenium	78-2	6185	6241	5831	6086	3.65	cps
Silicon	28-1	6932328	6887328	6952648	6924101	0.48	cps
Silver	107-1	9740600	9861581	9825098	9809093	0.63	cps
Silver	109-1	9345849	9303077	9392373	9347100	0.48	cps
Sodium	23-2	113351995	113559018	114219115	113710043	0.40	cps
Strontium	86-1	2699948	2680982	2684403	2688445	0.38	cps
Strontium	88-1	23195249	23215707	23366180	23259045	0.40	cps
Sulfur	34-1	1461045	1420926	1417068	1433013	1.70	cps
Terbium	159-1	20524883	20552897	20215143	20430974	0.92	cps
Terbium	159-2	6089691	6109674	6000073	6066479	0.96	cps
Thallium	203-1	7965201	8030308	8043287	8012932	0.52	cps
Thallium	205-1	18796496	19019243	19088752	18968163	0.80	cps
Tin	118-1	6496653	6479065	6621326	6532348	1.19	cps
Titanium	47-1	12430604	12281259	12330399	12347421	0.62	cps
Uranium	238-1	25741353	26063259	26611831	26138814	1.68	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : CCV022 Instrumnet Name : P8
Client Sample ID : CCV022 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:50:23 DataFile Name : 033CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	1259175	1254168	1250385	1254576	0.35	cps
Ytterbium	172-1	497	567	473	512	9.48	cps
Ytterbium	172-2	143	200	200	181	18.06	cps
Ytterbium	176-1	41552	40696	41284	41177	1.06	cps
Ytterbium	176-2	12483	12583	12389	12485	0.78	cps
Yttrium	89-1	26981940	26683772	26824623	26830112	0.56	cps
Yttrium	89-2	1735726	1712686	1680361	1709591	1.63	cps
Zinc	66-2	2484632	2468297	2485788	2479572	0.39	cps
Zirconium	90-1	14458267	14582043	14669493	14569934	0.73	cps
Zirconium	91-1	3226371	3328566	3327612	3294183	1.78	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : CCB022 Instrumnet Name : P8
Client Sample ID : CCB022 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:53:49 DataFile Name : 034CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	133	143	153	143	6.98	cps
Antimony	121-1	1943	1767	1633	1781	8.73	cps
Arsenic	75-2	7	7	13	9	43.25	cps
Barium	135-1	513	437	380	443	15.09	cps
Barium	137-1	987	750	740	826	16.91	cps
Beryllium	9-1	969	892	879	913	5.31	cps
Bismuth	209-1	13465650	13521157	13574240	13520349	0.40	cps
Bismuth	209-2	4917479	5017297	4941901	4958892	1.05	cps
Bromine	81-1	4254	4204	4354	4271	1.79	cps
Cadmium	108-1	43	30	60	44	33.82	cps
Cadmium	106-1	7542	8176	7649	7789	4.36	cps
Cadmium	111-1	5339	5832	5449	5540	4.67	cps
Calcium	43-1	690	680	630	667	4.82	cps
Calcium	44-1	32173	29754	29347	30425	5.02	cps
Carbon	12-1	3727055	3700262	3695860	3707726	0.46	cps
Carbon	12-2	23923	23452	23896	23757	1.11	cps
Chlorine	35-1	1188619	1153433	1080475	1140842	4.84	cps
Chlorine	35-2	3697	3807	3524	3676	3.89	cps
Chromium	52-2	700	873	820	798	11.13	cps
Cobalt	59-2	160	163	213	179	16.70	cps
Copper	63-2	3194	3360	3350	3302	2.83	cps
Dysprosium	156-1	30	57	37	41	33.76	cps
Dysprosium	156-2	3	0	3	2	86.60	cps
Erbium	164-1	90	100	103	98	7.10	cps
Erbium	164-2	30	20	23	24	20.83	cps
Gadolinium	160-1	103	137	83	108	25.00	cps
Gadolinium	160-2	27	17	23	22	22.91	cps
Holmium	165-1	20753346	20800179	21049186	20867570	0.76	cps
Holmium	165-2	6634503	6665388	6623623	6641171	0.33	cps
Indium	115-1	16616756	16574874	16616445	16602692	0.15	cps
Indium	115-2	1420781	1423616	1401077	1415158	0.87	cps
Iron	54-2	733	733	710	726	1.86	cps
Iron	56-2	11481	11765	10931	11392	3.72	cps
Iron	57-2	260	287	297	281	6.74	cps
Krypton	83-1	240	267	240	249	6.19	cps
Lead	206-1	4087	3817	3957	3954	3.42	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : CCB022 Instrumnet Name : P8
Client Sample ID : CCB022 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:53:49 DataFile Name : 034CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	207-1	3507	3214	3220	3314	5.06	cps
Lead	208-1	16334	15170	14587	15364	5.79	cps
Lithium	6-1	7785162	7739798	7829087	7784682	0.57	cps
Magnesium	24-2	2530	2387	2567	2495	3.81	cps
Manganese	55-2	300	303	330	311	5.29	cps
Molybdenum	94-1	1437	1037	1057	1177	19.16	cps
Molybdenum	95-1	1643	1110	1237	1330	20.95	cps
Molybdenum	96-1	1850	1487	1217	1518	20.94	cps
Molybdenum	97-1	953	790	643	796	19.49	cps
Molybdenum	98-1	2707	2064	1657	2142	24.72	cps
Neodymium	150-1	17	20	20	19	10.18	cps
Neodymium	150-2	0	3	7	3	100.05	cps
Nickel	60-2	1260	1307	1447	1338	7.26	cps
Phosphorus	31-2	80	77	90	82	8.44	cps
Potassium	39-2	11134	11365	10931	11143	1.95	cps
Rhodium	103-1	15716142	15441275	15807817	15655078	1.22	cps
Rhodium	103-2	5591971	5701259	5524253	5605828	1.59	cps
Scandium	45-1	11047170	10958330	10971621	10992374	0.44	cps
Scandium	45-2	194278	195551	193100	194309	0.63	cps
Selenium	82-1	30	-17	77	30	155.53	cps
Selenium	77-2	0	3	0	1	173.21	cps
Selenium	78-2	0	10	23	11	105.35	cps
Silicon	28-1	546624	539718	540953	542432	0.68	cps
Silver	107-1	1023	937	753	904	15.24	cps
Silver	109-1	987	873	670	843	19.03	cps
Sodium	23-2	69310	69746	68335	69130	1.04	cps
Strontium	86-1	650	627	623	633	2.29	cps
Strontium	88-1	1967	1793	1817	1859	5.06	cps
Sulfur	34-1	728143	723170	724050	725121	0.37	cps
Terbium	159-1	21242294	21554426	21619358	21472026	0.94	cps
Terbium	159-2	6363219	6345993	6358420	6355877	0.14	cps
Thallium	203-1	1050	953	960	988	5.47	cps
Thallium	205-1	2427	2264	2064	2251	8.08	cps
Tin	118-1	3387	3470	3524	3460	1.99	cps
Titanium	47-1	883	837	800	840	4.97	cps
Uranium	238-1	827	517	430	591	35.28	cps

LB Number : LB134674 Operator : Jaswal
Lab Sample ID : CCB022 Instrumnet Name : P8
Client Sample ID : CCB022 Dilution Factor : 1
Date & Time Acquired : 2025-02-11 14:53:49 DataFile Name : 034CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Vanadium	51-2	13	30	20	21	39.75	cps
Ytterbium	172-1	93	83	123	100	20.82	cps
Ytterbium	172-2	60	37	47	48	24.50	cps
Ytterbium	176-1	1910	1954	1923	1929	1.15	cps
Ytterbium	176-2	280	237	297	271	11.42	cps
Yttrium	89-1	28370380	28226698	28540556	28379211	0.55	cps
Yttrium	89-2	1811326	1814681	1779172	1801726	1.09	cps
Zinc	66-2	330	273	287	297	9.99	cps
Zirconium	90-1	1393	1420	1223	1346	7.93	cps
Zirconium	91-1	300	343	197	280	26.91	cps