

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134953

Review By	Sarabjit Jaswal	Review On	3/8/2025 7:35:47 AM
Supervise By	Kareem Khairalla	Supervise On	3/8/2025 7:59:43 AM
STD. NAME	STD REF.#		
ICAL Standard	MP84636,MP84657,MP84655,MP84654,MP84653,MP84652,MP84651		
ICV Standard	MP84658		
CCV Standard	MP84661		
ICSA Standard	MP84659,MP84721		
CRI Standard	MP84810,MP84811		
LCS Standard			
Chk Standard	MP84649,MP84650		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	03/07/25 12:18		Kareem	OK
2	S1	S01	CAL2	03/07/25 12:22		Kareem	OK
3	S2	S02	CAL3	03/07/25 12:26		Kareem	OK
4	S3	S03	CAL4	03/07/25 12:30		Kareem	OK
5	S4	S04	CAL5	03/07/25 12:34		Kareem	OK
6	S5	S05	CAL6	03/07/25 12:38		Kareem	OK
7	S6	S06	CAL7	03/07/25 12:43		Kareem	OK
8	ICV003	ICV003	ICV	03/07/25 12:52	ICV fail Mg,K - Inst. paused for calibration review	Kareem	OK
9	ICB003	ICB003	ICB	03/07/25 12:56		Kareem	OK
10	ICSA003	ICSA003	ICSA	03/07/25 13:01		Kareem	OK
11	ICSAB003	ICSAB003	ICSAB	03/07/25 13:05		Kareem	OK
12	CCV006	CCV006	CCV	03/07/25 13:09		Kareem	OK
13	CCB006	CCB006	CCB	03/07/25 13:13		Kareem	OK
14	Q1106-03A	YE8G6A	PS	03/07/25 13:17	PS for Al - 0.0295ml (mp84810), Mn - 0.006 ml (mp84811) - 10 ml sample	Kareem	OK
15	Q1160-01A	YE8C9A	PS	03/07/25 13:21	PS for Al - 0.034ml (m5808) - 10 ml sample	Kareem	OK
16	CCV007	CCV007	CCV	03/07/25 13:28		Kareem	OK
17	CCB007	CCB007	CCB	03/07/25 13:32		Kareem	OK