

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011

Lab Code: ACE Case No.: 51955 MA No. : SDG No.: YE8D0

Initial Calibration Verification Source : MP84709

Continuing Calibration Verification Source : MP84711

Run Batch: LB134860 Analytical Method: CVAA

Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV024				ID: CCV045				ID: CCV046		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury	4.0	4.0	100		5.0	5.3	106		5.4	107	

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51955 MA No. : SDG No.: YE8D0
Initial Calibration Verification Source : MP84709
Continuing Calibration Verification Source : MP84711
Run Batch: LB134860 Analytical Method: CVAA
Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV047				ID: CCV048		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury					5.0	5.2	105		5.1	102	