

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51900 MA No. : SDG No.: ME2946
 Initial Calibration Verification Source : MP84213
 Continuing Calibration Verification Source : MP84216
 Run Batch: LB134438 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV003				ID: CCV021				ID: CCV022		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	2500	2500	99	1	400000	400000	101	1	400000	99	2
Calcium	10000	9500	95	0	400000	400000	101	1	400000	100	1
Iron	10000	10000	100	0	400000	410000	102	0	400000	101	1
Magnesium	6000	6000	99	1	400000	400000	100	0	390000	98	1
Potassium	9900	9600	97	1	150000	160000	106	0	150000	103	1
Sodium	10000	10000	102	3	400000	380000	96	1	390000	99	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51900 MA No. : SDG No.: ME2946
 Initial Calibration Verification Source : MP84213
 Continuing Calibration Verification Source : MP84216
 Run Batch: LB134438 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV023				ID: CCV024		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	400000	101	1	400000	100	0
Calcium					400000	410000	101	0	410000	101	1
Iron					400000	410000	102	0	410000	102	0
Magnesium					400000	400000	100	0	400000	99	0
Potassium					150000	150000	103	0	150000	103	0
Sodium					400000	380000	94	0	390000	97	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51900 MA No. : SDG No.: ME2946
 Initial Calibration Verification Source : MP84213
 Continuing Calibration Verification Source : MP84216
 Run Batch: LB134460 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV004				ID: CCV025				ID: CCV026		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	2500	2500	98	1	400000	410000	102	0	390000	99	1
Calcium	10000	10000	101	0	400000	420000	104	0	400000	101	0
Iron	10000	9600	96	0	400000	410000	104	0	400000	101	0
Magnesium	6000	6000	101	0	400000	410000	104	0	400000	101	0
Potassium	9900	9600	97	1	150000	160000	108	0	150000	103	0
Sodium	10000	10000	103	3	400000	430000	107	0	400000	100	0