

## FORM 2 - IN

## INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
 Lab Code: ACE Case No.: 51900 MA No. :                      SDG No.: ME2947  
 Initial Calibration Verification Source : MP84213  
 Continuing Calibration Verification Source : MP84216  
 Run Batch: LB134460 Analytical Method: ICP-AES  
 Concentration Units: µg/L

|          | Initial Calibration Verification |       |     |      | Continuing Calibration Verification |       |     |      |            |     |      |
|----------|----------------------------------|-------|-----|------|-------------------------------------|-------|-----|------|------------|-----|------|
|          | ID: ICV004                       |       |     |      | ID: CCV025                          |       |     |      | ID: CCV026 |     |      |
| Analyte  | True                             | Found | %R  | %RSD | True                                | Found | %R  | %RSD | Found      | %R  | %RSD |
| Arsenic  | 1000                             | 1000  | 102 | 0    | 5000                                | 5100  | 102 | 0    | 5000       | 100 | 0    |
| Barium   | 520                              | 560   | 108 | 0    | 10000                               | 11000 | 108 | 2    | 10000      | 102 | 1    |
| Cadmium  | 510                              | 490   | 96  | 0    | 2500                                | 2600  | 105 | 0    | 2600       | 103 | 0    |
| Chromium | 520                              | 550   | 105 | 0    | 15500                               | 16000 | 106 | 0    | 16000      | 103 | 0    |
| Lead     | 1000                             | 980   | 98  | 0    | 25000                               | 26000 | 102 | 0    | 25000      | 100 | 0    |
| Selenium | 1000                             | 990   | 99  | 1    | 5000                                | 5200  | 105 | 0    | 5000       | 100 | 0    |
| Silver   | 250                              | 230   | 92  | 0    | 1250                                | 1400  | 109 | 0    | 1300       | 104 | 0    |

## FORM 2 - IN

## INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
 Lab Code: ACE Case No.: 51900 MA No. :                      SDG No.: ME2947  
 Initial Calibration Verification Source : MP84213  
 Continuing Calibration Verification Source : MP84216  
 Run Batch: LB134460 Analytical Method: ICP-AES  
 Concentration Units: µg/L

|          | Initial Calibration Verification |       |    |      | Continuing Calibration Verification |       |     |      |       |    |      |
|----------|----------------------------------|-------|----|------|-------------------------------------|-------|-----|------|-------|----|------|
|          | ID:                              |       |    |      | ID: CCV027                          |       |     |      | ID:   |    |      |
| Analyte  | True                             | Found | %R | %RSD | True                                | Found | %R  | %RSD | Found | %R | %RSD |
| Arsenic  |                                  |       |    |      | 5000                                | 4900  | 98  | 0    |       |    |      |
| Barium   |                                  |       |    |      | 10000                               | 9900  | 99  | 1    |       |    |      |
| Cadmium  |                                  |       |    |      | 2500                                | 2600  | 104 | 0    |       |    |      |
| Chromium |                                  |       |    |      | 15500                               | 16000 | 104 | 0    |       |    |      |
| Lead     |                                  |       |    |      | 25000                               | 25000 | 100 | 0    |       |    |      |
| Selenium |                                  |       |    |      | 5000                                | 4900  | 98  | 1    |       |    |      |
| Silver   |                                  |       |    |      | 1250                                | 1300  | 103 | 0    |       |    |      |