

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51900 MA No.: _____ SDG No.: ME2947
Instrument ID: CV1 Start Date: 01/29/2025
Analytical Method: CVAA Run Batch: LB134465
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	0	0.0010	0	0.2	0.18	-12	2.5	2.6	4

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51900 MA No.: _____ SDG No.: ME2947
Instrument ID: CV1 Start Date: 01/29/2025
Analytical Method: CVAA Run Batch: LB134465
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	5	4.9	-3	7.5	7.6	1	10	10	0