


LB134465

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000
B= 2.1947e-004 slope
C= -1.6084e-002 y-intercept
Rho= 0.9998183
Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/vl)
0.00	0.000	0.001	0.001	78	0.000	78					
0.05	0.050										
0.20	0.200	0.177	-0.023	881	0.0 %	881					-24
2.50	2.500	2.593	0.093	11887	0.0 %	11887					4
5.00	5.000	4.875	-0.125	22286	0.0 %	22286					-3
7.50	7.500	7.569	0.069	34560	0.0 %	34560					1
10.0	10.000	9.985	-0.015	45571	0.0 %	45571					0

LB134465

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	78	-		0 SFAM01.1	PPB	1/29/2025 13:06	Std
	0.2 S01	881	-		0.2 SFAM01.1	PPB	1/29/2025 13:08	Std
	2.5 S02	11887	-		2.5 SFAM01.1	PPB	1/29/2025 13:11	Std
	5 S03	22286	-		5 SFAM01.1	PPB	1/29/2025 13:13	Std
	7.5 S04	34560	-		7.5 SFAM01.1	PPB	1/29/2025 13:15	Std
	10 S05	45571	-		10 SFAM01.1	PPB	1/29/2025 13:20	Std
ICV073	ICV073	16611	3.6295 -		SFAM01.1	PPB	1/29/2025 13:23	SMPL
ICB073	ICB073	3	-0.0154 -		SFAM01.1	PPB	1/29/2025 13:27	SMPL
CCV082	CCV082	22137	4.8423 -		SFAM01.1	PPB	1/29/2025 13:29	SMPL
CCB082	CCB082	-131	-0.0448 -		SFAM01.1	PPB	1/29/2025 13:31	SMPL
PB166350BL	PBW350	107	0.0074 -		SFAM01.1	PPB	1/29/2025 13:33	SMPL
PB166350TB	LEB235	73	-0.0001 -		SFAM01.1	PPB	1/29/2025 13:36	SMPL
Q1160-01	YE8C9	402	0.0721 -		SFAM01.1	PPB	1/29/2025 13:43	SMPL
Q1160-02	YE8C9D	261	0.0412 -		SFAM01.1	PPB	1/29/2025 13:45	SMPL
Q1160-03	YE8C9S	5459	1.182 -		SFAM01.1	PPB	1/29/2025 13:47	SMPL
Q1160-04	YE8D0	262	0.0414 -		SFAM01.1	PPB	1/29/2025 13:49	SMPL
Q1160-05	YE8D3	515	0.0969 -		SFAM01.1	PPB	1/29/2025 13:52	SMPL
PB166351BL	PBW351	-25	-0.0216 -		SFAM01.1	PPB	1/29/2025 13:54	SMPL
PB166236TB	LEB236	102	0.0063 -		SFAM01.1	PPB	1/29/2025 13:56	SMPL
Q1164-01	ME2946	56	-0.0038 -		SFAM01.1	PPB	1/29/2025 13:59	SMPL
Q1164-02	ME2947	5	-0.015 -		SFAM01.1	PPB	1/29/2025 14:01	SMPL
Q1164-03	ME2947D	29	-0.0097 -		SFAM01.1	PPB	1/29/2025 14:03	SMPL
Q1164-04	ME2947S	5392	1.1673 -		SFAM01.1	PPB	1/29/2025 14:05	SMPL
CCV083	CCV083	21225	4.6421 -		SFAM01.1	PPB	1/29/2025 14:08	SMPL
CCB083	CCB083	-134	-0.0455 -		SFAM01.1	PPB	1/29/2025 14:10	SMPL