

Continuing Calibration Report for SequenceID : FC012925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068144.D

Aliphatic C16-C21	7809500.000	60.000	9.782	13.135	130158.333	135019.289	3.600
Aliphatic C21-C28	9915878.000	80.000	13.136	16.787	123948.475	127207.526	2.562
Aliphatic C28-C40	13340245.000	120.000	16.788	21.601	111168.708	103638.860	-7.265
Aliphatic EPH	44490047.000	360.000	3.130	21.601	123583.464	123700.655	0.095
Aliphatic C9-C12	8081349.000	60.000	3.130	6.397	134689.150	138488.174	2.743
Aliphatic C12-C16	5343075.000	40.000	6.398	9.781	133576.875	137713.071	3.003