

Continuing Calibration Report for SequenceID : FC012925AL

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FC068150.D

|                   |              |         |        |        |            |            |         |
|-------------------|--------------|---------|--------|--------|------------|------------|---------|
| Aliphatic C16-C21 | 7728270.000  | 60.000  | 9.782  | 13.135 | 128804.500 | 135019.289 | 4.603   |
| Aliphatic C21-C28 | 10013282.000 | 80.000  | 13.136 | 16.787 | 125166.025 | 127207.526 | 1.605   |
| Aliphatic C28-C40 | 13682609.000 | 120.000 | 16.788 | 21.601 | 114021.742 | 103638.860 | -10.018 |
| Aliphatic EPH     | 44752286.000 | 360.000 | 3.130  | 21.601 | 124311.906 | 123700.655 | -0.494  |
| Aliphatic C9-C12  | 8029040.000  | 60.000  | 3.130  | 6.397  | 133817.333 | 138488.174 | 3.373   |
| Aliphatic C12-C16 | 5299085.000  | 40.000  | 6.398  | 9.781  | 132477.125 | 137713.071 | 3.802   |