

Continuing Calibration Report for SequenceID : FC012925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068154.D

Aliphatic C16-C21	7061427.000	60.000	9.782	13.135	117690.450	135019.289	12.834
Aliphatic C21-C28	9204213.000	80.000	13.136	16.787	115052.663	127207.526	9.555
Aliphatic C28-C40	12737259.000	120.000	16.788	21.601	106143.825	103638.860	-2.417
Aliphatic EPH	41163293.000	360.000	3.130	21.601	114342.481	123700.655	7.565
Aliphatic C9-C12	7343761.000	60.000	3.130	6.397	122396.017	138488.174	11.620
Aliphatic C12-C16	4816633.000	40.000	6.398	9.781	120415.825	137713.071	12.560