

Continuing Calibration Report for SequenceID : FD012825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048978.D

Aliphatic C9-C12	10308094.000	60.000	3.006	6.243	171801.567	162866.213	-5.486
Aliphatic C12-C16	7083829.000	40.000	6.244	9.613	177095.725	168872.221	-4.870
Aliphatic C16-C21	10835661.000	60.000	9.614	12.956	180594.350	172902.817	-4.448
Aliphatic C21-C28	13896998.000	80.000	12.957	16.595	173712.475	168565.180	-3.054
Aliphatic C28-C40	15703035.000	120.000	16.596	21.296	130858.625	131483.064	0.475
Aliphatic EPH	57827617.000	360.000	3.006	21.296	160632.269	156011.702	-2.962