

Continuing Calibration Report for SequenceID : FD012825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048984.D

Aliphatic C9-C12	10266688.000	60.000	3.006	6.243	171111.467	162866.213	-5.063
Aliphatic C12-C16	7023899.000	40.000	6.244	9.613	175597.475	168872.221	-3.982
Aliphatic C16-C21	10750118.000	60.000	9.614	12.956	179168.633	172902.817	-3.624
Aliphatic C21-C28	13887715.000	80.000	12.957	16.595	173596.438	168565.180	-2.985
Aliphatic C28-C40	16038558.000	120.000	16.596	21.296	133654.650	131483.064	-1.652
Aliphatic EPH	57966978.000	360.000	3.006	21.296	161019.383	156011.702	-3.210