

Continuing Calibration Report for SequenceID : FD012925AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048987.D

Aromatic C12-C16	10439333.000	60.000	5.799	8.402	173988.883	162211.796	-7.260
Aromatic C16-C21	12869048.000	80.000	8.403	12.668	160863.100	154648.237	-4.019
Aromatic C10-C12	6753847.000	40.000	4.079	5.798	168846.175	155230.631	-8.771
Aromatic C21-C36	27753858.000	180.000	12.669	18.079	154188.100	142502.922	-8.200
Aromatic EPH	57816086.000	360.000	4.079	18.079	160600.239	149900.883	-7.138