

Continuing Calibration Report for SequenceID : FD012925AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048993.D

Aromatic C12-C16	10300061.000	60.000	5.799	8.402	171667.683	162211.796	-5.829
Aromatic C16-C21	13123257.000	80.000	8.403	12.668	164040.713	154648.237	-6.073
Aromatic C10-C12	6742873.000	40.000	4.079	5.798	168571.825	155230.631	-8.594
Aromatic C21-C36	27403692.000	180.000	12.669	18.079	152242.733	142502.922	-6.835
Aromatic EPH	57569883.000	360.000	4.079	18.079	159916.342	149900.883	-6.681