

Continuing Calibration Report for SequenceID : FD012925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049004.D

Aliphatic C9-C12	8395803.000	60.000	3.006	6.244	139930.050	162866.213	14.083
Aliphatic C12-C16	5786810.000	40.000	6.245	9.613	144670.250	168872.221	14.332
Aliphatic C16-C21	8946693.000	60.000	9.614	12.956	149111.550	172902.817	13.760
Aliphatic C21-C28	12133447.000	80.000	12.957	16.598	151668.088	168565.180	10.024
Aliphatic C28-C40	16257368.000	120.000	16.599	21.297	135478.067	131483.064	-3.038
Aliphatic EPH	51520121.000	360.000	3.006	21.297	143111.447	156011.702	8.269