

Continuing Calibration Report for SequenceID : FE012825AL

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FE052115.D

|                   |              |         |        |        |            |            |        |
|-------------------|--------------|---------|--------|--------|------------|------------|--------|
| Aliphatic C9-C12  | 6588347.000  | 60.000  | 3.579  | 7.133  | 109805.783 | 109652.183 | -0.140 |
| Aliphatic C12-C16 | 4934096.000  | 40.000  | 7.134  | 10.532 | 123352.400 | 121214.595 | -1.764 |
| Aliphatic C16-C21 | 7628513.000  | 60.000  | 10.533 | 13.875 | 127141.883 | 131187.839 | 3.084  |
| Aliphatic C21-C28 | 11181127.000 | 80.000  | 13.876 | 17.518 | 139764.088 | 132059.255 | -5.834 |
| Aliphatic C28-C40 | 13848671.000 | 120.000 | 17.519 | 22.539 | 115405.592 | 107292.260 | -7.562 |
| Aliphatic EPH     | 44180754.000 | 360.000 | 3.579  | 22.539 | 122724.317 | 118718.880 | -3.374 |