

Continuing Calibration Report for SequenceID : FF012825AL

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FF015349.D

|                   |              |         |        |        |            |            |        |
|-------------------|--------------|---------|--------|--------|------------|------------|--------|
| Aliphatic C9-C12  | 8099560.000  | 60.000  | 3.214  | 6.832  | 134992.667 | 133293.048 | -1.275 |
| Aliphatic C12-C16 | 5645587.000  | 40.000  | 6.833  | 10.265 | 141139.675 | 138403.375 | -1.977 |
| Aliphatic C16-C21 | 8603557.000  | 60.000  | 10.266 | 13.633 | 143392.617 | 144457.893 | 0.737  |
| Aliphatic C21-C28 | 10585659.000 | 80.000  | 13.634 | 17.299 | 132320.738 | 142415.321 | 7.088  |
| Aliphatic C28-C40 | 12383805.000 | 120.000 | 17.300 | 22.207 | 103198.375 | 104686.549 | 1.422  |
| Aliphatic EPH     | 45318168.000 | 360.000 | 3.214  | 22.207 | 125883.800 | 128213.342 | 1.817  |