

Continuing Calibration Report for SequenceID : FF012925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015362.D

Aliphatic C9-C12	7932558.000	60.000	3.214	6.831	132209.300	133293.048	0.813
Aliphatic C12-C16	5483070.000	40.000	6.832	10.266	137076.750	138403.375	0.959
Aliphatic C16-C21	8285632.000	60.000	10.267	13.632	138093.867	144457.893	4.405
Aliphatic C21-C28	10217470.000	80.000	13.633	17.299	127718.375	142415.321	10.320
Aliphatic C28-C40	12239635.000	120.000	17.300	22.207	101996.958	104686.549	2.569
Aliphatic EPH	44158365.000	360.000	3.214	22.207	122662.125	128213.342	4.330