

Continuing Calibration Report for SequenceID : FF012925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015368.D

Aliphatic C9-C12	7741018.000	60.000	3.214	6.831	129016.967	133293.048	3.208
Aliphatic C12-C16	5351070.000	40.000	6.832	10.266	133776.750	138403.375	3.343
Aliphatic C16-C21	8082395.000	60.000	10.267	13.632	134706.583	144457.893	6.750
Aliphatic C21-C28	9950892.000	80.000	13.633	17.299	124386.150	142415.321	12.660
Aliphatic C28-C40	11901552.000	120.000	17.300	22.207	99179.600	104686.549	5.260
Aliphatic EPH	43026927.000	360.000	3.214	22.207	119519.242	128213.342	6.781