

Continuing Calibration Report for SequenceID : FF012925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF015372.D

Aliphatic C9-C12	7763964.000	60.000	3.214	6.831	129399.400	133293.048	2.921
Aliphatic C12-C16	5375013.000	40.000	6.832	10.266	134375.325	138403.375	2.910
Aliphatic C16-C21	8110710.000	60.000	10.267	13.632	135178.500	144457.893	6.424
Aliphatic C21-C28	9946350.000	80.000	13.633	17.299	124329.375	142415.321	12.699
Aliphatic C28-C40	11862963.000	120.000	17.300	22.207	98858.025	104686.549	5.568
Aliphatic EPH	43059000.000	360.000	3.214	22.207	119608.333	128213.342	6.711