

Continuing Calibration Report for SequenceID : FG012925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG015209.D

Aliphatic C9-C12	8782552.000	60.000	3.182	6.819	146375.867	132888.697	-10.149
Aliphatic C12-C16	6210764.000	40.000	6.820	10.271	155269.100	141209.859	-9.956
Aliphatic C16-C21	9473662.000	60.000	10.272	13.651	157894.367	144518.745	-9.255
Aliphatic C21-C28	11979969.000	80.000	13.652	17.325	149749.613	136707.720	-9.540
Aliphatic C28-C40	14618457.000	120.000	17.326	22.263	121820.475	105590.394	-15.371
Aliphatic EPH	51065404.000	360.000	3.182	22.263	141848.344	127500.849	-11.253