

Continuing Calibration Report for SequenceID : FG012925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG015215.D

Aliphatic C9-C12	7233792.000	60.000	3.182	6.819	120563.200	132888.697	9.275
Aliphatic C12-C16	5121769.000	40.000	6.820	10.271	128044.225	141209.859	9.323
Aliphatic C16-C21	7773028.000	60.000	10.272	13.651	129550.467	144518.745	10.357
Aliphatic C21-C28	9804635.000	80.000	13.652	17.325	122557.938	136707.720	10.350
Aliphatic C28-C40	12097784.000	120.000	17.326	22.263	100814.867	105590.394	4.523
Aliphatic EPH	42031008.000	360.000	3.182	22.263	116752.800	127500.849	8.430