

RETENTION TIMES OF INITIAL CALIBRATION

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

Instrument ID: ECD_S Calibration Date(s): 01/14/2025 01/14/2025

Calibration Times: 10:31 12:07

GC Column: RTX-CLP ID: 0.32 (mm)

LAB FILE ID:	RT 200 = <u>PS028901.D</u>	RT 500 = <u>PS028902.D</u>
	RT 750 = <u>PS028903.D</u>	RT 1000 = <u>PS028904.D</u>
		RT 1500 = <u>PS028905.D</u>

COMPOUND	RT 200	RT 500	RT 750	RT 1000	RT 1500	MEAN RT	RT WINDOW	
							FROM	TO
2,4,5-T	9.48	9.48	9.48	9.48	9.48	9.48	9.38	9.58
2,4,5-TP(Silvex)	9.19	9.19	9.19	9.19	9.19	9.19	9.09	9.29
2,4-D	8.32	8.32	8.32	8.32	8.32	8.32	8.22	8.42
2,4-DB	10.05	10.05	10.05	10.05	10.05	10.05	9.95	10.15
2,4-DCAA	7.20	7.20	7.20	7.20	7.20	7.20	7.10	7.30
DICAMBA	7.38	7.38	7.38	7.38	7.38	7.38	7.28	7.48
DICHLORPROP	8.09	8.09	8.09	8.09	8.09	8.09	7.99	8.19
Dinoseb	11.26	11.25	11.25	11.26	11.26	11.25	11.15	11.35



Contract:	<u>CHEM02</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1168</u>	SAS No.:	<u>Q1168</u>	SDG NO.:	<u>Q1168</u>
Instrument ID:	<u>ECD_S</u>	Calibration Date(s):		<u>01/14/2025</u>	<u>01/14/2025</u>		
		Calibration Times:		<u>10:31</u>	<u>12:07</u>		

LAB FILE ID:	RT 200 =	<u>PS028901.D</u>	RT 500 =	<u>PS028902.D</u>
RT 750 =	PS028903.D	RT 1000 =	PS028904.D	RT 1500 =
				PS028905.D

[illegible]