

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 01/14/2025 01/14/2025

Client Sample No.: CCAL01 Date Analyzed: 01/28/2025

Lab Sample No.: HSTDCCC750 Data File : PS028966.D Time Analyzed: 14:53

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.480	9.379	9.579	724.520	712.500	1.7
2,4,5-TP(Silvex)	9.190	9.089	9.289	725.730	712.500	1.9
2,4-D	8.317	8.216	8.416	711.310	705.000	0.9
2,4-DB	10.051	9.950	10.150	722.430	712.500	1.4
2,4-DCAA	7.199	7.097	7.297	743.590	750.000	-0.9
DICAMBA	7.384	7.283	7.483	732.580	705.000	3.9
DICHLORPROP	8.088	7.987	8.187	703.420	705.000	-0.2
Dinoseb	11.255	11.153	11.353	707.640	705.000	0.4

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/14/2025 01/14/2025

Client Sample No.: CCAL01 Date Analyzed: 01/28/2025

Lab Sample No.: HSTDCCC750 Data File : PS028966.D Time Analyzed: 14:53

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.221	10.130	10.330	708.000	712.500	-0.6
2,4,5-TP(Silvex)	9.804	9.713	9.913	718.520	712.500	0.8
2,4-D	8.905	8.813	9.013	670.590	705.000	-4.9
2,4-DB	10.786	10.694	10.894	677.380	712.500	-4.9
2,4-DCAA	7.670	7.577	7.777	697.730	750.000	-7.0
DICAMBA	7.866	7.774	7.974	696.230	705.000	-1.2
DICHLORPROP	8.578	8.486	8.686	675.480	705.000	-4.2
Dinoseb	11.164	11.072	11.272	697.490	705.000	-1.1