

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 01/14/2025 01/14/2025

Client Sample No.: CCAL02 Date Analyzed: 01/28/2025

Lab Sample No.: HSTDCCC750 Data File : PS028972.D Time Analyzed: 17:53

| COMPOUND         | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|------------------|--------|-----------|--------|--------------------|-------------------|------|
|                  |        | FROM      | TO     |                    |                   |      |
| 2,4,5-T          | 9.478  | 9.379     | 9.579  | 707.640            | 712.500           | -0.7 |
| 2,4,5-TP(Silvex) | 9.188  | 9.089     | 9.289  | 707.190            | 712.500           | -0.7 |
| 2,4-D            | 8.315  | 8.216     | 8.416  | 692.600            | 705.000           | -1.8 |
| 2,4-DB           | 10.049 | 9.950     | 10.150 | 707.090            | 712.500           | -0.8 |
| 2,4-DCAA         | 7.196  | 7.097     | 7.297  | 725.420            | 750.000           | -3.3 |
| DICAMBA          | 7.381  | 7.283     | 7.483  | 711.860            | 705.000           | 1.0  |
| DICHLORPROP      | 8.085  | 7.987     | 8.187  | 691.580            | 705.000           | -1.9 |
| Dinoseb          | 11.253 | 11.153    | 11.353 | 684.050            | 705.000           | -3.0 |

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/14/2025 01/14/2025

Client Sample No.: CCAL02 Date Analyzed: 01/28/2025

Lab Sample No.: HSTDCCC750 Data File : PS028972.D Time Analyzed: 17:53

| COMPOUND         | RT     | RT WINDOW<br>FROM TO |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|------------------|--------|----------------------|--------|--------------------|-------------------|------|
| 2,4,5-T          | 10.228 | 10.130               | 10.330 | 708.320            | 712.500           | -0.6 |
| 2,4,5-TP(Silvex) | 9.811  | 9.713                | 9.913  | 717.040            | 712.500           | 0.6  |
| 2,4-D            | 8.911  | 8.813                | 9.013  | 666.130            | 705.000           | -5.5 |
| 2,4-DB           | 10.793 | 10.694               | 10.894 | 678.870            | 712.500           | -4.7 |
| 2,4-DCAA         | 7.676  | 7.577                | 7.777  | 695.960            | 750.000           | -7.2 |
| DICAMBA          | 7.872  | 7.774                | 7.974  | 696.050            | 705.000           | -1.3 |
| DICHLORPROP      | 8.585  | 8.486                | 8.686  | 683.050            | 705.000           | -3.1 |
| Dinoseb          | 11.170 | 11.072               | 11.272 | 684.630            | 705.000           | -2.9 |