

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 01/14/2025 01/14/2025

Client Sample No.: CCAL01 Date Analyzed: 02/04/2025

Lab Sample No.: HSTDCCC750 Data File : PS029063.D Time Analyzed: 14:56

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.471	9.379	9.579	745.170	712.500	4.6
2,4,5-TP(Silvex)	9.179	9.089	9.289	752.750	712.500	5.6
2,4-D	8.308	8.216	8.416	726.990	705.000	3.1
2,4-DB	10.041	9.950	10.150	705.580	712.500	-1.0
2,4-DCAA	7.191	7.097	7.297	775.840	750.000	3.4
DICAMBA	7.376	7.283	7.483	753.640	705.000	6.9
DICHLORPROP	8.079	7.987	8.187	730.940	705.000	3.7
Dinoseb	11.242	11.153	11.353	720.120	705.000	2.1

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/14/2025 01/14/2025

Client Sample No.: CCAL01 Date Analyzed: 02/04/2025

Lab Sample No.: HSTDCCC750 Data File : PS029063.D Time Analyzed: 14:56

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.217	10.130	10.330	715.850	712.500	0.5
2,4,5-TP(Silvex)	9.800	9.713	9.913	730.140	712.500	2.5
2,4-D	8.902	8.813	9.013	670.940	705.000	-4.8
2,4-DB	10.781	10.694	10.894	667.200	712.500	-6.4
2,4-DCAA	7.667	7.577	7.777	693.970	750.000	-7.5
DICAMBA	7.864	7.774	7.974	699.710	705.000	-0.8
DICHLORPROP	8.575	8.486	8.686	681.900	705.000	-3.3
Dinoseb	11.158	11.072	11.272	691.690	705.000	-1.9