

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 01/14/2025 01/14/2025

Client Sample No.: CCAL02 Date Analyzed: 02/04/2025

Lab Sample No.: HSTDCCC750 Data File : PS029069.D Time Analyzed: 17:56

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.469	9.379	9.579	737.960	712.500	3.6
2,4,5-TP(Silvex)	9.177	9.089	9.289	747.520	712.500	4.9
2,4-D	8.307	8.216	8.416	726.240	705.000	3.0
2,4-DB	10.039	9.950	10.150	696.420	712.500	-2.3
2,4-DCAA	7.190	7.097	7.297	769.430	750.000	2.6
DICAMBA	7.374	7.283	7.483	748.450	705.000	6.2
DICHLORPROP	8.078	7.987	8.187	722.320	705.000	2.5
Dinoseb	11.239	11.153	11.353	712.120	705.000	1.0

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/14/2025 01/14/2025

Client Sample No.: CCAL02 Date Analyzed: 02/04/2025

Lab Sample No.: HSTDCCC750 Data File : PS029069.D Time Analyzed: 17:56

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.214	10.130	10.330	717.140	712.500	0.7
2,4,5-TP(Silvex)	9.798	9.713	9.913	732.770	712.500	2.8
2,4-D	8.900	8.813	9.013	674.710	705.000	-4.3
2,4-DB	10.778	10.694	10.894	668.270	712.500	-6.2
2,4-DCAA	7.666	7.577	7.777	698.760	750.000	-6.8
DICAMBA	7.861	7.774	7.974	703.680	705.000	-0.2
DICHLORPROP	8.573	8.486	8.686	688.510	705.000	-2.3
Dinoseb	11.155	11.072	11.272	688.920	705.000	-2.3