

RETENTION TIMES OF INITIAL CALIBRATION

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

Instrument ID: ECD_R Calibration Date(s): 02/10/2025 02/10/2025

Calibration Times: 11:30 14:12

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 = <u>PR070233.D</u>	RT 750 = <u>PR070234.D</u>
RT 500 = <u>PR070235.D</u>	RT 250 = <u>PR070236.D</u>	RT 050 = <u>PR070237.D</u>

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW	
								FROM	TO
Aroclor-1016-1	(1)	4.84	4.85	4.85	4.84	4.84	4.84	4.74	4.94
Aroclor-1016-2	(2)	4.87	4.87	4.87	4.87	4.87	4.87	4.77	4.97
Aroclor-1016-3	(3)	4.93	4.93	4.93	4.93	4.93	4.93	4.83	5.03
Aroclor-1016-4	(4)	5.03	5.03	5.03	5.03	5.03	5.03	4.93	5.13
Aroclor-1016-5	(5)	5.35	5.35	5.35	5.35	5.35	5.35	5.25	5.45
Aroclor-1260-1	(1)	6.56	6.56	6.56	6.56	6.56	6.56	6.46	6.66
Aroclor-1260-2	(2)	6.84	6.84	6.84	6.84	6.84	6.84	6.74	6.94
Aroclor-1260-3	(3)	7.23	7.23	7.23	7.23	7.23	7.23	7.13	7.33
Aroclor-1260-4	(4)	7.48	7.48	7.48	7.48	7.48	7.48	7.38	7.58
Aroclor-1260-5	(5)	7.82	7.82	7.82	7.82	7.82	7.82	7.72	7.92
Decachlorobiphenyl		9.51	9.51	9.51	9.51	9.51	9.51	9.41	9.61
Tetrachloro-m-xylene		3.62	3.63	3.63	3.62	3.62	3.62	3.52	3.72

Contract:	<u>CHEM02</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1168</u>	SAS No.:	<u>Q1168</u>	SDG NO.:	<u>Q1168</u>
Instrument ID:	<u>ECD_R</u>	Calibration Date(s):		<u>02/10/2025</u>	<u>02/10/2025</u>		
		Calibration Times:		<u>11:30</u>	<u>14:12</u>		

LAB FILE ID:		RT 1000 =	<u>PR070233.D</u>	RT 750 =	<u>PR070234.D</u>
RT 500 =	PR070235.D	RT 250 =	PR070236.D	RT 050 =	PR070237.D

[illegible]