

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

Continuing Calib Date: 01/29/2025 Initial Calibration Date(s): 01/28/2025 01/28/2025

Continuing Calib Time: 20:25 Initial Calibration Time(s): 09:37 17:30

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.69	5.69	5.59	5.79	0.00
Aroclor-1016-2	(2)	5.71	5.71	5.61	5.81	0.00
Aroclor-1016-3	(3)	5.78	5.78	5.68	5.88	0.00
Aroclor-1016-4	(4)	5.87	5.87	5.77	5.97	0.00
Aroclor-1016-5	(5)	6.17	6.17	6.07	6.27	0.00
Aroclor-1260-1	(1)	7.29	7.29	7.19	7.39	0.00
Aroclor-1260-2	(2)	7.54	7.54	7.44	7.64	0.00
Aroclor-1260-3	(3)	7.90	7.90	7.80	8.00	0.00
Aroclor-1260-4	(4)	8.12	8.13	8.03	8.23	0.01
Aroclor-1260-5	(5)	8.45	8.45	8.35	8.55	0.00
Tetrachloro-m-xylene		4.54	4.54	4.44	4.64	0.00
Decachlorobiphenyl		10.28	10.28	10.18	10.38	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

Continuing Calib Date: 01/29/2025 Initial Calibration Date(s): 01/28/2025 01/28/2025

Continuing Calib Time: 20:25 Initial Calibration Time(s): 09:37 17:30

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.94	4.93	4.83	5.03	0.00
Aroclor-1016-2	(2)	4.95	4.95	4.85	5.05	0.00
Aroclor-1016-3	(3)	5.13	5.13	5.03	5.23	0.00
Aroclor-1016-4	(4)	5.17	5.17	5.07	5.27	0.00
Aroclor-1016-5	(5)	5.39	5.39	5.29	5.49	0.00
Aroclor-1260-1	(1)	6.43	6.43	6.33	6.53	0.00
Aroclor-1260-2	(2)	6.62	6.62	6.52	6.72	0.01
Aroclor-1260-3	(3)	6.77	6.77	6.67	6.87	0.00
Aroclor-1260-4	(4)	7.24	7.24	7.14	7.34	0.00
Aroclor-1260-5	(5)	7.48	7.48	7.38	7.58	0.00
Tetrachloro-m-xylene		3.84	3.84	3.74	3.94	0.00
Decachlorobiphenyl		8.91	8.91	8.81	9.01	0.00