

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

Continuing Calib Date: 01/29/2025 Initial Calibration Date(s): 01/28/2025 01/28/2025

Continuing Calib Time: 12:45 Initial Calibration Time(s): 11:30 14:20

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.50	4.50	4.40	4.60	0.00
Aroclor-1016-2	(2)	4.52	4.52	4.42	4.62	0.00
Aroclor-1016-3	(3)	4.58	4.58	4.48	4.68	0.01
Aroclor-1016-4	(4)	4.67	4.67	4.57	4.77	0.01
Aroclor-1016-5	(5)	4.95	4.95	4.85	5.05	0.00
Aroclor-1260-1	(1)	6.04	6.04	5.94	6.14	0.00
Aroclor-1260-2	(2)	6.30	6.30	6.20	6.40	0.00
Aroclor-1260-3	(3)	6.65	6.65	6.55	6.75	0.00
Aroclor-1260-4	(4)	6.87	6.86	6.76	6.96	0.00
Aroclor-1260-5	(5)	7.17	7.17	7.07	7.27	0.00
Tetrachloro-m-xylene		3.42	3.42	3.32	3.52	0.01
Decachlorobiphenyl		8.53	8.53	8.43	8.63	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

Continuing Calib Date: 01/29/2025 Initial Calibration Date(s): 01/28/2025 01/28/2025

Continuing Calib Time: 12:45 Initial Calibration Time(s): 11:30 14:20

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	3.79	3.79	3.69	3.89	0.00
Aroclor-1016-2	(2)	3.81	3.81	3.71	3.91	0.00
Aroclor-1016-3	(3)	3.97	3.97	3.87	4.07	0.00
Aroclor-1016-4	(4)	4.02	4.02	3.92	4.12	0.01
Aroclor-1016-5	(5)	4.21	4.21	4.11	4.31	0.00
Aroclor-1260-1	(1)	5.21	5.21	5.11	5.31	0.01
Aroclor-1260-2	(2)	5.39	5.40	5.30	5.50	0.01
Aroclor-1260-3	(3)	5.54	5.54	5.44	5.64	0.00
Aroclor-1260-4	(4)	6.00	6.00	5.90	6.10	0.00
Aroclor-1260-5	(5)	6.24	6.25	6.15	6.35	0.01
Tetrachloro-m-xylene		2.79	2.79	2.69	2.89	0.00
Decachlorobiphenyl		7.56	7.56	7.46	7.66	0.00