

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

Continuing Calib Date: 02/10/2025 Initial Calibration Date(s): 02/10/2025 02/10/2025

Continuing Calib Time: 22:30 Initial Calibration Time(s): 11:30 14:12

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.84	4.85	4.75	4.95	0.01
Aroclor-1016-2	(2)	4.87	4.87	4.77	4.97	0.01
Aroclor-1016-3	(3)	4.93	4.93	4.83	5.03	0.00
Aroclor-1016-4	(4)	5.03	5.03	4.93	5.13	0.00
Aroclor-1016-5	(5)	5.35	5.35	5.25	5.45	0.00
Aroclor-1260-1	(1)	6.56	6.56	6.46	6.66	0.00
Aroclor-1260-2	(2)	6.84	6.84	6.74	6.94	0.00
Aroclor-1260-3	(3)	7.23	7.23	7.13	7.33	0.00
Aroclor-1260-4	(4)	7.48	7.48	7.38	7.58	0.01
Aroclor-1260-5	(5)	7.82	7.82	7.72	7.92	0.00
Tetrachloro-m-xylene		3.62	3.63	3.53	3.73	0.01
Decachlorobiphenyl		9.51	9.51	9.41	9.61	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

Continuing Calib Date: 02/10/2025 Initial Calibration Date(s): 02/10/2025 02/10/2025

Continuing Calib Time: 22:30 Initial Calibration Time(s): 11:30 14:12

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.07	4.07	3.97	4.17	0.00
Aroclor-1016-2	(2)	4.09	4.09	3.99	4.19	0.00
Aroclor-1016-3	(3)	4.27	4.27	4.17	4.37	0.00
Aroclor-1016-4	(4)	4.32	4.32	4.22	4.42	0.00
Aroclor-1016-5	(5)	4.54	4.54	4.44	4.64	0.00
Aroclor-1260-1	(1)	5.64	5.64	5.54	5.74	0.00
Aroclor-1260-2	(2)	5.84	5.84	5.74	5.94	0.00
Aroclor-1260-3	(3)	6.00	6.00	5.90	6.10	0.00
Aroclor-1260-4	(4)	6.51	6.51	6.41	6.61	0.00
Aroclor-1260-5	(5)	6.77	6.77	6.67	6.87	0.00
Tetrachloro-m-xylene		2.96	2.96	2.86	3.06	0.01
Decachlorobiphenyl		8.22	8.23	8.13	8.33	0.01