

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL01 Date Analyzed: 01/28/2025

Lab Sample No.: AR1660CCC500 Data File : PO109138.D Time Analyzed: 09:37

| COMPOUND             | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|----------------------|-------|--------------------|-------------------|------|
| Aroclor-1016-1       | 4.799 | 4.695                | 4.895 | 570.830            | 500.000           | 14.2 |
| Aroclor-1016-2       | 4.819 | 4.714                | 4.914 | 569.510            | 500.000           | 13.9 |
| Aroclor-1016-3       | 4.875 | 4.770                | 4.970 | 568.430            | 500.000           | 13.7 |
| Aroclor-1016-4       | 4.996 | 4.891                | 5.091 | 572.080            | 500.000           | 14.4 |
| Aroclor-1016-5       | 5.254 | 5.149                | 5.349 | 563.660            | 500.000           | 12.7 |
| Aroclor-1260-1       | 6.297 | 6.190                | 6.390 | 552.750            | 500.000           | 10.6 |
| Aroclor-1260-2       | 6.486 | 6.379                | 6.579 | 545.500            | 500.000           | 9.1  |
| Aroclor-1260-3       | 6.855 | 6.747                | 6.947 | 548.430            | 500.000           | 9.7  |
| Aroclor-1260-4       | 7.116 | 7.008                | 7.208 | 550.860            | 500.000           | 10.2 |
| Aroclor-1260-5       | 7.357 | 7.249                | 7.449 | 551.770            | 500.000           | 10.4 |
| Decachlorobiphenyl   | 8.770 | 8.658                | 8.858 | 52.240             | 50.000            | 4.5  |
| Tetrachloro-m-xylene | 3.703 | 3.600                | 3.800 | 57.430             | 50.000            | 14.9 |

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL01 Date Analyzed: 01/28/2025

Lab Sample No.: AR1660CCC500 Data File : PO109138.D Time Analyzed: 09:37

| COMPOUND             | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D  |
|----------------------|-------|----------------------|-------|--------------------|-------------------|-----|
| Aroclor-1016-1       | 4.782 | 4.703                | 4.903 | 519.300            | 500.000           | 3.9 |
| Aroclor-1016-2       | 4.802 | 4.722                | 4.922 | 546.040            | 500.000           | 9.2 |
| Aroclor-1016-3       | 4.978 | 4.898                | 5.098 | 527.410            | 500.000           | 5.5 |
| Aroclor-1016-4       | 5.020 | 4.940                | 5.140 | 523.300            | 500.000           | 4.7 |
| Aroclor-1016-5       | 5.234 | 5.153                | 5.353 | 522.630            | 500.000           | 4.5 |
| Aroclor-1260-1       | 6.268 | 6.186                | 6.386 | 514.480            | 500.000           | 2.9 |
| Aroclor-1260-2       | 6.456 | 6.373                | 6.573 | 510.360            | 500.000           | 2.1 |
| Aroclor-1260-3       | 6.609 | 6.527                | 6.727 | 510.190            | 500.000           | 2.0 |
| Aroclor-1260-4       | 7.082 | 6.999                | 7.199 | 511.780            | 500.000           | 2.4 |
| Aroclor-1260-5       | 7.322 | 7.239                | 7.439 | 520.330            | 500.000           | 4.1 |
| Decachlorobiphenyl   | 8.718 | 8.633                | 8.833 | 50.380             | 50.000            | 0.8 |
| Tetrachloro-m-xylene | 3.697 | 3.618                | 3.818 | 54.790             | 50.000            | 9.6 |