

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL02 Date Analyzed: 01/28/2025

Lab Sample No.: AR1660CCC500 Data File : PO109149.D Time Analyzed: 12:58

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.793	4.695	4.895	527.460	500.000	5.5
Aroclor-1016-2	4.813	4.714	4.914	529.030	500.000	5.8
Aroclor-1016-3	4.869	4.770	4.970	526.570	500.000	5.3
Aroclor-1016-4	4.990	4.891	5.091	528.720	500.000	5.7
Aroclor-1016-5	5.248	5.149	5.349	532.680	500.000	6.5
Aroclor-1260-1	6.291	6.190	6.390	516.140	500.000	3.2
Aroclor-1260-2	6.480	6.379	6.579	510.570	500.000	2.1
Aroclor-1260-3	6.848	6.747	6.947	513.960	500.000	2.8
Aroclor-1260-4	7.109	7.008	7.208	515.730	500.000	3.1
Aroclor-1260-5	7.351	7.249	7.449	518.010	500.000	3.6
Decachlorobiphenyl	8.761	8.658	8.858	49.930	50.000	-0.1
Tetrachloro-m-xylene	3.698	3.600	3.800	52.740	50.000	5.5

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL02 Date Analyzed: 01/28/2025

Lab Sample No.: AR1660CCC500 Data File : PO109149.D Time Analyzed: 12:58

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.782	4.703	4.903	491.340	500.000	-1.7
Aroclor-1016-2	4.801	4.722	4.922	531.170	500.000	6.2
Aroclor-1016-3	4.976	4.898	5.098	516.220	500.000	3.2
Aroclor-1016-4	5.018	4.940	5.140	506.980	500.000	1.4
Aroclor-1016-5	5.231	5.153	5.353	520.940	500.000	4.2
Aroclor-1260-1	6.266	6.186	6.386	502.030	500.000	0.4
Aroclor-1260-2	6.453	6.373	6.573	498.790	500.000	-0.2
Aroclor-1260-3	6.607	6.527	6.727	498.220	500.000	-0.4
Aroclor-1260-4	7.079	6.999	7.199	504.360	500.000	0.9
Aroclor-1260-5	7.319	7.239	7.439	504.050	500.000	0.8
Decachlorobiphenyl	8.715	8.633	8.833	48.370	50.000	-3.3
Tetrachloro-m-xylene	3.697	3.618	3.818	53.000	50.000	6.0