

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL01 Date Analyzed: 01/29/2025

Lab Sample No.: AR1660CCC500 Data File : PP069299.D Time Analyzed: 09:01

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.696	5.592	5.792	528.840	500.000	5.8
Aroclor-1016-2	5.718	5.614	5.814	529.330	500.000	5.9
Aroclor-1016-3	5.781	5.676	5.876	506.270	500.000	1.3
Aroclor-1016-4	5.878	5.774	5.974	537.540	500.000	7.5
Aroclor-1016-5	6.171	6.067	6.267	521.620	500.000	4.3
Aroclor-1260-1	7.292	7.187	7.387	521.270	500.000	4.3
Aroclor-1260-2	7.546	7.441	7.641	499.410	500.000	-0.1
Aroclor-1260-3	7.905	7.800	8.000	522.710	500.000	4.5
Aroclor-1260-4	8.130	8.025	8.225	519.490	500.000	3.9
Aroclor-1260-5	8.452	8.347	8.547	519.490	500.000	3.9
Decachlorobiphenyl	10.283	10.179	10.379	59.660	50.000	19.3
Tetrachloro-m-xylene	4.542	4.438	4.638	54.340	50.000	8.7

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL01 Date Analyzed: 01/29/2025

Lab Sample No.: AR1660CCC500 Data File : PP069299.D Time Analyzed: 09:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.935	4.834	5.034	538.650	500.000	7.7
Aroclor-1016-2	4.954	4.853	5.053	543.470	500.000	8.7
Aroclor-1016-3	5.131	5.030	5.230	548.230	500.000	9.6
Aroclor-1016-4	5.174	5.072	5.272	549.980	500.000	10.0
Aroclor-1016-5	5.389	5.288	5.488	534.010	500.000	6.8
Aroclor-1260-1	6.429	6.326	6.526	529.620	500.000	5.9
Aroclor-1260-2	6.617	6.515	6.715	506.760	500.000	1.4
Aroclor-1260-3	6.772	6.670	6.870	481.540	500.000	-3.7
Aroclor-1260-4	7.245	7.142	7.342	526.010	500.000	5.2
Aroclor-1260-5	7.485	7.384	7.584	520.720	500.000	4.1
Decachlorobiphenyl	8.914	8.811	9.011	50.460	50.000	0.9
Tetrachloro-m-xylene	3.842	3.740	3.940	54.990	50.000	10.0