

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL02 Date Analyzed: 01/29/2025

Lab Sample No.: AR1660CCC500 Data File : PP069310.D Time Analyzed: 12:00

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.692	5.592	5.792	549.820	500.000	10.0
Aroclor-1016-2	5.714	5.614	5.814	535.490	500.000	7.1
Aroclor-1016-3	5.776	5.676	5.876	535.300	500.000	7.1
Aroclor-1016-4	5.874	5.774	5.974	555.870	500.000	11.2
Aroclor-1016-5	6.168	6.067	6.267	551.460	500.000	10.3
Aroclor-1260-1	7.288	7.187	7.387	542.550	500.000	8.5
Aroclor-1260-2	7.542	7.441	7.641	515.830	500.000	3.2
Aroclor-1260-3	7.900	7.800	8.000	528.310	500.000	5.7
Aroclor-1260-4	8.126	8.025	8.225	524.500	500.000	4.9
Aroclor-1260-5	8.447	8.347	8.547	526.480	500.000	5.3
Decachlorobiphenyl	10.280	10.179	10.379	54.880	50.000	9.8
Tetrachloro-m-xylene	4.537	4.438	4.638	54.510	50.000	9.0

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL02 Date Analyzed: 01/29/2025

Lab Sample No.: AR1660CCC500 Data File : PP069310.D Time Analyzed: 12:00

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.934	4.834	5.034	549.650	500.000	9.9
Aroclor-1016-2	4.953	4.853	5.053	558.040	500.000	11.6
Aroclor-1016-3	5.131	5.030	5.230	548.900	500.000	9.8
Aroclor-1016-4	5.172	5.072	5.272	550.040	500.000	10.0
Aroclor-1016-5	5.388	5.288	5.488	551.910	500.000	10.4
Aroclor-1260-1	6.428	6.326	6.526	548.940	500.000	9.8
Aroclor-1260-2	6.616	6.515	6.715	528.230	500.000	5.6
Aroclor-1260-3	6.770	6.670	6.870	502.650	500.000	0.5
Aroclor-1260-4	7.243	7.142	7.342	547.900	500.000	9.6
Aroclor-1260-5	7.484	7.384	7.584	537.920	500.000	7.6
Decachlorobiphenyl	8.912	8.811	9.011	53.510	50.000	7.0
Tetrachloro-m-xylene	3.840	3.740	3.940	55.380	50.000	10.8