

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL01 Date Analyzed: 01/28/2025

Lab Sample No.: AR1660CCC500 Data File : PQ069808.D Time Analyzed: 17:39

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.500	4.399	4.599	576.250	500.000	15.3
Aroclor-1016-2	4.518	4.418	4.618	554.930	500.000	11.0
Aroclor-1016-3	4.576	4.476	4.676	548.260	500.000	9.7
Aroclor-1016-4	4.666	4.566	4.766	560.610	500.000	12.1
Aroclor-1016-5	4.951	4.850	5.050	567.090	500.000	13.4
Aroclor-1260-1	6.043	5.942	6.142	557.420	500.000	11.5
Aroclor-1260-2	6.298	6.198	6.398	564.640	500.000	12.9
Aroclor-1260-3	6.650	6.549	6.749	564.770	500.000	13.0
Aroclor-1260-4	6.865	6.763	6.963	556.900	500.000	11.4
Aroclor-1260-5	7.171	7.070	7.270	568.070	500.000	13.6
Decachlorobiphenyl	8.529	8.425	8.625	56.320	50.000	12.6
Tetrachloro-m-xylene	3.416	3.315	3.515	55.550	50.000	11.1

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL01 Date Analyzed: 01/28/2025

Lab Sample No.: AR1660CCC500 Data File : PQ069808.D Time Analyzed: 17:39

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	3.791	3.691	3.891	517.060	500.000	3.4
Aroclor-1016-2	3.806	3.707	3.907	498.820	500.000	-0.2
Aroclor-1016-3	3.967	3.868	4.068	500.110	500.000	0.0
Aroclor-1016-4	4.016	3.916	4.116	487.150	500.000	-2.6
Aroclor-1016-5	4.211	4.112	4.312	500.800	500.000	0.2
Aroclor-1260-1	5.207	5.108	5.308	502.000	500.000	0.4
Aroclor-1260-2	5.396	5.297	5.497	514.510	500.000	2.9
Aroclor-1260-3	5.538	5.439	5.639	506.700	500.000	1.3
Aroclor-1260-4	5.999	5.901	6.101	494.260	500.000	-1.1
Aroclor-1260-5	6.244	6.146	6.346	494.070	500.000	-1.2
Decachlorobiphenyl	7.561	7.461	7.661	50.110	50.000	0.2
Tetrachloro-m-xylene	2.787	2.687	2.887	48.230	50.000	-3.5