

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL02 Date Analyzed: 01/28/2025

Lab Sample No.: AR1660CCC500 Data File : PQ069819.D Time Analyzed: 20:28

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.499	4.399	4.599	596.060	500.000	19.2
Aroclor-1016-2	4.518	4.418	4.618	583.610	500.000	16.7
Aroclor-1016-3	4.576	4.476	4.676	569.220	500.000	13.8
Aroclor-1016-4	4.666	4.566	4.766	586.280	500.000	17.3
Aroclor-1016-5	4.951	4.850	5.050	589.240	500.000	17.8
Aroclor-1260-1	6.042	5.942	6.142	577.230	500.000	15.4
Aroclor-1260-2	6.298	6.198	6.398	590.460	500.000	18.1
Aroclor-1260-3	6.650	6.549	6.749	583.880	500.000	16.8
Aroclor-1260-4	6.863	6.763	6.963	581.670	500.000	16.3
Aroclor-1260-5	7.171	7.070	7.270	597.470	500.000	19.5
Decachlorobiphenyl	8.526	8.425	8.625	58.510	50.000	17.0
Tetrachloro-m-xylene	3.416	3.315	3.515	57.030	50.000	14.1

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL02 Date Analyzed: 01/28/2025

Lab Sample No.: AR1660CCC500 Data File : PQ069819.D Time Analyzed: 20:28

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	3.790	3.691	3.891	518.070	500.000	3.6
Aroclor-1016-2	3.805	3.707	3.907	496.760	500.000	-0.6
Aroclor-1016-3	3.966	3.868	4.068	496.230	500.000	-0.8
Aroclor-1016-4	4.015	3.916	4.116	480.040	500.000	-4.0
Aroclor-1016-5	4.210	4.112	4.312	491.320	500.000	-1.7
Aroclor-1260-1	5.206	5.108	5.308	496.080	500.000	-0.8
Aroclor-1260-2	5.396	5.297	5.497	508.410	500.000	1.7
Aroclor-1260-3	5.538	5.439	5.639	497.980	500.000	-0.4
Aroclor-1260-4	6.000	5.901	6.101	496.160	500.000	-0.8
Aroclor-1260-5	6.244	6.146	6.346	516.880	500.000	3.4
Decachlorobiphenyl	7.560	7.461	7.661	50.970	50.000	1.9
Tetrachloro-m-xylene	2.786	2.687	2.887	49.650	50.000	-0.7