

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL01 Date Analyzed: 01/29/2025

Lab Sample No.: AR1660CCC500 Data File : PQ069863.D Time Analyzed: 15:52

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.501	4.399	4.599	567.820	500.000	13.6
Aroclor-1016-2	4.520	4.418	4.618	542.870	500.000	8.6
Aroclor-1016-3	4.578	4.476	4.676	535.700	500.000	7.1
Aroclor-1016-4	4.667	4.566	4.766	552.880	500.000	10.6
Aroclor-1016-5	4.952	4.850	5.050	564.460	500.000	12.9
Aroclor-1260-1	6.044	5.942	6.142	531.810	500.000	6.4
Aroclor-1260-2	6.301	6.198	6.398	555.130	500.000	11.0
Aroclor-1260-3	6.652	6.549	6.749	550.180	500.000	10.0
Aroclor-1260-4	6.867	6.763	6.963	543.960	500.000	8.8
Aroclor-1260-5	7.174	7.070	7.270	552.380	500.000	10.5
Decachlorobiphenyl	8.532	8.425	8.625	55.040	50.000	10.1
Tetrachloro-m-xylene	3.417	3.315	3.515	53.130	50.000	6.3

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL01 Date Analyzed: 01/29/2025

Lab Sample No.: AR1660CCC500 Data File : PQ069863.D Time Analyzed: 15:52

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	3.789	3.691	3.891	474.700	500.000	-5.1
Aroclor-1016-2	3.805	3.707	3.907	439.340	500.000	-12.1
Aroclor-1016-3	3.965	3.868	4.068	454.000	500.000	-9.2
Aroclor-1016-4	4.014	3.916	4.116	436.890	500.000	-12.6
Aroclor-1016-5	4.210	4.112	4.312	457.660	500.000	-8.5
Aroclor-1260-1	5.206	5.108	5.308	442.820	500.000	-11.4
Aroclor-1260-2	5.395	5.297	5.497	450.880	500.000	-9.8
Aroclor-1260-3	5.537	5.439	5.639	437.430	500.000	-12.5
Aroclor-1260-4	5.999	5.901	6.101	431.230	500.000	-13.8
Aroclor-1260-5	6.244	6.146	6.346	447.220	500.000	-10.6
Decachlorobiphenyl	7.562	7.461	7.661	45.580	50.000	-8.8
Tetrachloro-m-xylene	2.786	2.687	2.887	41.670	50.000	-16.7