

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL02 Date Analyzed: 01/29/2025

Lab Sample No.: AR1660CCC500 Data File : PQ069873.D Time Analyzed: 18:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.498	4.399	4.599	558.890	500.000	11.8
Aroclor-1016-2	4.517	4.418	4.618	536.040	500.000	7.2
Aroclor-1016-3	4.575	4.476	4.676	525.830	500.000	5.2
Aroclor-1016-4	4.665	4.566	4.766	543.440	500.000	8.7
Aroclor-1016-5	4.950	4.850	5.050	553.970	500.000	10.8
Aroclor-1260-1	6.042	5.942	6.142	524.000	500.000	4.8
Aroclor-1260-2	6.298	6.198	6.398	566.690	500.000	13.3
Aroclor-1260-3	6.650	6.549	6.749	549.740	500.000	9.9
Aroclor-1260-4	6.864	6.763	6.963	547.230	500.000	9.4
Aroclor-1260-5	7.170	7.070	7.270	553.840	500.000	10.8
Decachlorobiphenyl	8.527	8.425	8.625	54.870	50.000	9.7
Tetrachloro-m-xylene	3.415	3.315	3.515	52.180	50.000	4.4

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL02 Date Analyzed: 01/29/2025

Lab Sample No.: AR1660CCC500 Data File : PQ069873.D Time Analyzed: 18:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.789	3.691	3.891	471.850	500.000	-5.6
Aroclor-1016-2	3.805	3.707	3.907	440.880	500.000	-11.8
Aroclor-1016-3	3.965	3.868	4.068	453.380	500.000	-9.3
Aroclor-1016-4	4.014	3.916	4.116	435.250	500.000	-13.0
Aroclor-1016-5	4.209	4.112	4.312	454.920	500.000	-9.0
Aroclor-1260-1	5.205	5.108	5.308	437.230	500.000	-12.6
Aroclor-1260-2	5.395	5.297	5.497	443.520	500.000	-11.3
Aroclor-1260-3	5.536	5.439	5.639	432.040	500.000	-13.6
Aroclor-1260-4	5.998	5.901	6.101	433.870	500.000	-13.2
Aroclor-1260-5	6.242	6.146	6.346	434.880	500.000	-13.0
Decachlorobiphenyl	7.558	7.461	7.661	44.560	50.000	-10.9
Tetrachloro-m-xylene	2.785	2.687	2.887	41.100	50.000	-17.8